



Medtronic

Cardiac Rhythm Disease Management

Product Performance Report

Important Patient Management Information for Physicians

2013

Second Edition – Issue 69

This report is available online at
www.medtronic.com/CRDMProductPerformance

Our Commitment to Quality

Medtronic was founded in 1949 and has grown to become a global leader in medical technology. Seeing what a difference medical technology could make in the lives of patients inspired our founder to develop the Medtronic Mission, which remains unchanged today.

The third tenet of the mission is all about quality:

“To strive without reserve for the greatest possible reliability and quality in our products, to be the unsurpassed standard of comparison, and to be recognized as a company of dedication, honesty, integrity, and service.”

Regardless of function, all CRDM employees play a role in product quality. Whether designing new therapies, sourcing components, manufacturing products, hiring talented people, assigning financial resources to project teams, or serving in one of the hundreds of other roles, every employee has an influence on product quality.

Product performance information is received from many sources through various channels. Medtronic monitors information from many sources from Research and Development through Manufacturing and Field Performance Vigilance.

When a device is returned to Medtronic, laboratory technicians and engineers assess overall device function. Analysis of returned product is performed according to written procedures. These procedures determine the minimum analysis required. The analysis required varies depending on the type of device, age of the device, the associated information received with the device, actual experience with models of similar design, and other factors. Additional analysis is performed as necessary to investigate a performance concern from a customer, or to collect specific reliability data.

When a malfunction is identified, failure analysis is performed to provide the detailed information necessary to investigate possible causes and actions. Medtronic CRDM maintains in-house expertise and performs its failure analysis using facilities it owns and supports. This capability permits detailed failure analysis.

Analysis results are compared to original manufacturing records and design intent. Clinical observations are added to laboratory findings to help determine root cause. Each event is then compared to other events. If a pattern is detected, actions are taken to identify a common root cause, assess patient risk and an appropriate course of action.

Medtronic instituted the industry’s first product performance reports in 1983 by publishing data on our chronic lead studies. Pacemakers and other devices followed as our performance reporting has constantly evolved based on customer needs and feedback. One thing has been a constant. It is our sincere commitment to communicate clearly, offering timely and appropriate product performance data and reliability information. This has always been and will continue to be our goal.



Tim Samsel
Vice President, Quality and Regulatory
Medtronic Cardiac Rhythm Disease Management
Medtronic, Inc.

Contact Information

We invite our customers to use these telephone numbers to call with suggestions, inquiries, or specific problems related to our products.

US Technical Services Department

Phone: 1 (800) 723-4636 (Tachy)
1 (800) 505-4636 (Brady)
Fax: 1 (800) 824-2362
www.medtronic.com/corporate/contact.jsp

International Technical Centers

Europe (Heerlen NL) +31-45-566-8844
Japan (Tokyo) +81-3-5753-4116

For questions related to this CRDM Product Performance Report, please call US Technical Services at the number above, or write to:

Timothy Smith
Medtronic, Inc.
8200 Coral Sea Street NE MVN61
Mounds View, MN 55112 USA

Email: tim.smith@medtronic.com

For questions related to returning explanted product or returning product that shows signs of malfunction, please contact:

Outside the United States:

Your Medtronic representative or international technical center at the number above.

Within the United States:

Your Medtronic representative or

CRDM Returned Product Analysis Laboratory

Phone: 1 (800) 328-2518, ext. 44800

Email: crdm.returnedproduct@medtronic.com

Editorial Staff

Independent Physician Quality Panel

Steven J. Compton, MD, Anchorage, AK
John P. DiMarco, MD, PhD, Charlottesville, VA
Kevin Hackett, MD, Columbus, OH
R. Hardwin Mead, MD, Palo Alto, CA
N.A. Mark Estes, MD, Boston, MA
Andrew Krahn, MD, Vancouver, BC

Editor

Tim Samsel, *Vice President, CRDM Quality and Regulatory*

Trademarks of Medtronic, Inc.

Adapta®	InSync II Marquis™	Prodigy
Advisa®	InSync III Marquis™	Protecta®
Advisa DR MRI™	InSync Maximo®	Quick Look™
AT500®	InSync II	Relia™
Attain®	Protect™	Revo MRI®
Attain Ability®	InSync III	Secura®
Attain StarFix®	Protect™	SelectSecure®
CapSure®	InSync Sentry®	Sensia®
CapSure Sense®	Intrinsic®	Sensing Assurance
CapSureFix®	Jewel®	Sigma®
CapSureFix Novus™	Kappa®	Spectraflex
Capture	Legend®	Sprint®
Management®	Marquis®	Sprint Fidelis®
CareLink®	Maximo®	Sprint Quattro®
Concerto®	Medtronic	Sprint Quattro
Consulta®	CareAlert®	Secure®
EnPulse®	Medtronic	Sprint Quattro
EnRhythm®	CareLink®	Secure S®
EnRhythm MRI™	Micro Jewel	SureFix®
Ensura MRI™	Micro Minix	Syncra™
EnTrust®	Minix	Target Tip®
EnTrust MRI™	Minuet	Tenax
GEM®	MVP®	Thera®-i
InSync®	Onyx®	Transvene
InSync ICD®	Patient Alert™	Versa®
InSync Marquis™	Preva	Virtuoso®
	Prevail®	

CRDM Product Performance Report

2013 Second Edition
Issue 69

Date cutoff for this edition is August 9, 2013 for devices and July 31, 2013 for PSR leads data

This report is available online at www.medtronic.com/CRDM/ProductPerformance

Introduction 1
Method for Estimating CRT, ICD, and IPG Device Performance 6

CRT **Cardiac Resynchronization Therapy** 10
CRT Survival Summary 17
CRT Reference Chart 20

ICD **Implantable Cardioverter Defibrillators** 22
ICD Survival Summary 35
ICD Reference Chart 39

IPG **Implantable Pulse Generators** 41
IPG Survival Summary 63
IPG Reference Chart 70

Leads

Method for Estimating Lead Performance 73

Left-Heart Leads 78
Lead Survival Summary 82
US Returned Product Analysis Summary 82
US Reports of Acute Lead Observations Reference Chart 83

Defibrillation Leads 84
Lead Survival Summary 92
US Returned Product Analysis Summary 93
US Reports of Acute Lead Observations Reference Chart 94

Pacing Leads 95
Lead Survival Summary 115
US Returned Product Analysis Summary 118
US Reports of Acute Lead Observations Reference Chart 120

Epi/Myocardial Pacing Leads 121
Lead Survival Summary 123
US Returned Product Analysis Summary 124
US Reports of Acute Lead Observations Reference Chart 124

VDD Single Pass Pacing Leads 125
Lead Survival Summary 126
US Returned Product Analysis Summary 126
US Reports of Acute Lead Observations Reference Chart 126

ICD and CRT-D Charge Time Performance 127

Advisories 135

(in order of communication date, from most recent to oldest)

Performance Notes 146

Dual Chamber Pacemakers with Measurement Lock-up ERI
Kappa 600, 700, 800, 900, EnPulse, Adapta, Versa, Sensia, Relia,
and Vitatron Models E50A1, E60A1, and G70A1 146
Clinical Management of VCM near Elective Replacement 147
General Follow-Up and Replacement of ICD Leads 148
Clinical Management of High-Voltage Lead System Oversensing 149
Tests and Observations for Clinical Assessment of Chronic Pacing Leads 150

Introduction

All product performance reports are not created equal. For 30 years, Medtronic has monitored performance via both returned product analysis and multicenter clinical studies.

This Product Performance Report (PPR) presents device survival estimates, advisory summaries, performance notes, and other information pertinent to assessing the performance of Medtronic implantable pulse generators (IPGs), implantable cardioverter defibrillators (ICDs), cardiac resynchronization therapy (CRT) devices, and implantable pacing and defibrillation leads.

This Product Performance Report has been prepared in accordance with International Standard ISO 5841-2:2000(E).

The survival estimates provided in this report are considered to be representative of worldwide performance.

Survival Estimates

Medtronic, like other companies, monitors CRT, ICD, and IPG device performance using returned product analysis. We also monitor CRT, ICD, and IPG device performance using an active multicenter clinical study.

Returned product analysis is a passive approach to assessing product performance. This approach provides a suitable measure of product performance only when a significant number of explanted products are returned to the manufacturer. Returned product analysis provides a measure of hardware performance, but not necessarily the total clinical performance (e.g., the incidence of complications such as infection, erosion, muscle stimulation, etc. are not estimated).

The survival estimates provided in this report for CRT, ICD, and IPG devices are based on returned product analysis. This approach is suitable because a significant number of explanted generators are returned for analysis.

Lead performance is monitored differently. In contrast to CRT, ICD, and IPG devices, a very small percentage of leads are returned to the manufacturer due to the difficulty of explanting them. For leads, an active clinical study provides more accurate survival estimates compared to estimates based solely on returned product analysis.

Survival estimates for leads are based on clinical observations recorded via Medtronic CRDM's System Longevity Study. This multicenter clinical study is designed to record clinical observations representative of the total clinical experience. Therefore, the lead survival estimates include both lead hardware failure and lead-related medical complications, and do not differentiate a lead hardware failure from other clinical events such as exit block, perforation, dislodgement, or concurrent pulse generator failure.

The actuarial life table method is applied to the data collected for CRT, ICD, and IPG devices and leads to provide the survival estimates included in this report. A general introduction to understanding this method of survival analysis is given later in this introduction.

ICD Charge Times

Since May 2000, Medtronic has provided important information on charge time performance of ICDs. The information provided in this report shows how ICD charge time can vary during the time a device is implanted. The information is presented in graphical format showing charge time as a function of implant time. The data for charge times are collected from devices enrolled in the Product Surveillance Registry.

Advisory Summaries

This Product Performance Report includes summaries of all advisories applicable to the performance of the products included in the report. An advisory is added to the report when any product affected by the advisory remains in service and at risk of experiencing the behavior described in the advisory. The advisory will remain in the report until Medtronic estimates no product affected by the advisory remains active, or the risk of experiencing the behavior described in the advisory has passed.

For most advisories, the products subject to the advisory retain essentially the same survival probability as the products of the same model(s) not affected by the advisory. For those advisories where the survival probabilities of the affected and non-affected populations do differ significantly, Medtronic will provide separate survival data for each population. The separate survival data will remain in the report until Medtronic estimates no affected product remains in active service.

Performance Notes

This report concludes with a number of Performance Notes developed by Medtronic to provide additional product performance information relevant to follow-up practice and patient management.

continued

Medtronic urges all physicians to return explanted products and to notify Medtronic when a product is no longer in use, regardless of reason for explant or removal from use.

How You Can Help

Medtronic urges all physicians to return explanted products and to notify Medtronic when a product is no longer in use, regardless of the reason for explant or removal from use. The procedures for returning products vary by geographic location.

Mailer kits with prepaid US postage are available for use within the United States to send CRTs, ICDs, IPGs, and leads to Medtronic's CRDM Returned Product Analysis Lab. These mailers are sized to accommodate the devices and leads from a single patient or clinical event and are designed to meet US postal regulations for mailing biohazard materials.

If the product being returned is located outside the United States, please contact your local Medtronic representative for instructions.

Medtronic also requests the return of explanted products from non-clinical sources, such as funeral homes, and will assume responsibility for storage and disposal of the product once received.

Mailer kits can be obtained by contacting the Returned Product Lab. For information on how to contact the Lab, refer to the Contact Information page of this report.

We continually strive to improve this CRDM Product Performance Report. In keeping with this philosophy, we ask for your suggestions on the content and format of this report, as well as any information you have regarding the performance of Medtronic products. For information on how to comment on this report, see the Contact Information page of this report.

Overview of Survival Analysis

Medtronic uses the Cutler-Ederer actuarial life table method to estimate the length of time over which devices and leads will perform within performance limits established by Medtronic. This probability to perform within performance limits over time is called the *survival probability*.

Devices and leads are followed until an *event* occurs where the device or lead ceases to operate within performance limits. The length of time from implant to the event is recorded for individual devices and leads in the *population sample*. The population sample for

CRT, ICD, and IPG devices is made up of patients whose devices are registered as implanted in the United States. For leads, the population sample is the patients enrolled in our multicenter, international prospective Product Surveillance Registry.

For IPGs and ICDs, the events can be normal battery depletion or a device malfunction. For leads, the events are complications as defined in the study protocol.

The actuarial life table method allows Medtronic to account for devices and leads removed from service for reasons unrelated to performance and for device and leads still in service. Devices and leads removed for reasons unrelated to performance or are still in service are said to be suspended. Examples of devices and leads removed from service for reasons unrelated to performance include:

- removed to upgrade the device or lead
- no longer in service due to the death of the patient for reasons unrelated to the device or leads
- implanted in patients who are lost to follow-up

For each suspension, the device or lead has performed within performance limits for a period of time, after which its performance is unknown.

An Example

The following example describes the survival analysis method used to establish the survival probability estimates for Medtronic CRDM devices and leads. The example is intended to provide an overview of the analysis process. The definitions of malfunctions and complications, and other details specific to calculating device and lead survival estimates, are provided in the articles *Method for Estimating CRT, ICD, and IPG Device Performance (page 6)* and *Method for Estimating Lead Performance (page 74)*.

continued

This simple example describes the survival analysis method used to establish the survival probability estimates for Medtronic CRDM devices and leads.

Figure 1

Implant times for devices of 16 patients. Gray bars with a yellow X indicate devices removed from service due to an event. Blue bars indicate suspended devices.




Figure 1 illustrates 16 patients who have implanted devices. The first patient’s device (serial number BH004134) operated within performance limits for 32 months. At that time an event occurred. The fourth patient’s device (serial number BH008926) did not have an event but is suspended, perhaps because it was still in service at the time of the analysis. This patient had 66 months of implant experience. In this example, Figure 1 shows that seven of the 16 devices suffered events, and nine are suspended.

The first step in the life table method is to divide the implant time into intervals of a specific length. This example will use 12-month intervals.

The number of devices entered, suspended, and removed due to an event are counted and summarized, as shown in Table 1. For the first two intervals, all 16 devices survived and none were removed. In the interval (24-36 months), device BH004134 was removed due to an event. Therefore the table entries show that 16 entered the interval, none were suspended, and one was removed due to an event.

For the interval from 36-48 months, only 15 devices entered the interval and one was removed for an event. The remaining intervals are examined and the data entered in columns A, B, and C in like manner. The rest of the columns are filled in using calculations on the data in columns A, B, and C.

The *Effective Sample Size (D)* is the number of devices with full opportunity to experience a qualifying event in the interval. This is computed by subtracting one half the number suspended in the interval from the number that entered the interval. This calculation more accurately reflects the number of devices that could have experienced a qualifying event than simply using the number that entered the interval. Using the number of devices that enter an interval overestimates the sample size because the suspended devices do not complete the interval. Ignoring the suspended devices underestimates the sample size because suspended devices are not credited with their full service time. Using one half the number of suspended devices effectively splits the difference.

The next column in the table is the *Proportion with Event (E)*. This is the proportion of devices that had an event in the interval. It is calculated by dividing the *Number of Events (C)* by the *Effective Sample Size (D)*. The number can be interpreted as the estimated rate at which events occur in the time interval.

The *Interval Survival Probability (F)* is the estimate of probability of surviving to the end of the interval assuming the device was working at the beginning of the interval. It is calculated as 1 minus the *Proportion with Event (E)*. This number can be interpreted as the estimated rate at which events **do not** occur in the time interval.

continued

The **Cumulative Survival Probabilities (G)** from the last column of the life table can be plotted versus time intervals in the first column to give a survival curve. Figure 2 shows the survival curve for the data shown in Table 1.

Table 1 Life Table for Figure 1

	A	B	C	D	E	F	G
Interval in Months	Number Entered	Number Suspended	Number of Events	Effective Sample Size	Proportion with Event	Interval Survival Probability	Cumulative Survival Probability
0	16	0	0	16	0.000	1.000	1.000
0-12	16	0	0	16	0.000	1.000	1.000
12-24	16	0	0	16	0.000	1.000	1.000
24-36	16	0	1	16	0.063	0.938	0.938
36-48	15	0	1	15	0.067	0.933	0.875
48-60	14	0	1	14	0.071	0.929	0.813
60-72	13	1	2	12.5	0.160	0.840	0.683
72-84	10	2	0	9	0.000	1.000	0.683
84-96	8	3	0	6.5	0.000	1.000	0.683
96-108	5	2	2	4	0.500	0.500	0.341
108-120	1	0	0	1	0.000	1.000	0.341
120-132	1	1	0	0.5	0.000	1.000	0.341

Definitions:


A	B	C	D	E	F	G
Number Entered	Number Suspended	Number of Events	Effective Sample Size	Proportion with Event	Interval Survival Probability	Cumulative Survival Probability
Number of devices active at the start of the interval	Number of devices removed from service for reasons other than an event	Number of units removed from service due to an event	Number of units with full opportunity to experience a qualifying event in the interval. Computed by subtracting one half the Number Suspended from the Number Entered.	Proportion of devices that had an event in the interval. Computed by dividing the Number of Events by the Effective Sample Size.	The probability of surviving to the end of the interval, assuming the device was working at the beginning of the interval. Computed as 1 minus the Proportion With Event.	The overall probability of surviving to the end of the interval. Computed by multiplying the Interval Survival Probability by the previous interval's Cumulative Survival Probability.

Cumulative Survival Probability (G) is the estimate of the unconditional probability of surviving to the end of the interval. It is computed by multiplying the *Interval Survival Probability (F)* by the previous interval's *Cumulative Survival Probability*. The probability of surviving to 132 months in the example is estimated for the table to be 0.341, or 34.1%.

The *Cumulative Survival Probabilities (G)* of the life table can be plotted versus time intervals in the first column to give a survival curve. Figure 2 shows the survival curve for the data in Table 1.

continued

Figure 2 Survival Curve for Data Given in Table 1



Confidence Intervals

Since survival curves are based on a sample of the device and lead population, they are only estimates of survival. The larger the effective sample size, the more confident the estimate. A confidence interval can be calculated to assess the confidence in an estimate. In the Product Performance Report, Medtronic provides a 95% confidence interval. This can be interpreted as meaning that 95% of the time, the true survival of the device will fall somewhere in the interval.

Survival Curves in the Product Performance Report

Since the survival estimate can become very imprecise with small effective sample sizes, Medtronic truncates the survival curve when the effective sample size is less than 100 for CRTs, ICDs, and IPGs, and when the number entered is less than 50 for leads. The survival charts in the Product Performance Report show the effective sample size for each year interval where Medtronic has experience. When the effective sample size reaches 100 for CRTs, ICDs, and IPGs or when the number entered reaches 50 for leads, the next data point is added to the survival curve.

Although the report provides tabular data in one-year intervals, the curves are actually computed and plotted using 1-month intervals (for CRT, ICD, and IPG devices) or 3-month intervals (for leads).

A number of references are available for additional information on survival analysis using the Cutler-Ederer life table method.¹

¹ Lee, Elisa T.(2003) Statistical Methods for Survival Data Analysis – 3rd Edition (Wiley Series in Probability and Statistics).

Method for Estimating CRT, ICD, and IPG Device Performance

Medtronic urges all physicians to return explanted products and to notify Medtronic when a product is no longer in use, regardless of reason for explant or removal from use.

The performance of CRT, ICD, and IPG devices is expressed in terms of device survival estimates, where “survival” refers to the function of the device, not the survival of the patient. These survival estimates are intended to illustrate the probability that a device will survive for a given number of years with neither malfunction nor battery depletion.

The survival estimates are determined from the analysis of Medtronic CRDM’s United States device registration data and US returned product analysis data. These data are presented graphically and numerically.

Because this analysis is based on returned product analysis, the performance data does not reflect any device-related medical complications such as erosion, infection, muscle stimulation, or muscle inhibition.

Categorization of Depleted and Malfunctioning Devices for Survival Analysis

For survival estimation, every device returned to Medtronic CRDM and analyzed in the CRDM Returned Product Analysis laboratory is assigned to one of three categories. The device 1) is functioning normally, 2) has reached normal battery depletion, or 3) has malfunctioned. This categorization is combined with data from our device registry for the total number of implants and the implant durations to create the survival curves presented on the following pages.

Definition of Malfunction

Medtronic CRDM considers a device as having malfunctioned whenever the analysis shows that any parameter was outside the performance limits established by Medtronic while implanted and in service. To be considered a malfunction or battery depletion, the device must have been returned to Medtronic and analyzed.

Devices damaged after explant, damaged due to failure to heed warnings or contraindications in the labeling, or damaged due to interaction with other implanted devices (including leads) are not considered device malfunctions.

A device subject to a safety advisory is not considered to have malfunctioned unless it has been returned to Medtronic CRDM and found, through analysis, to actually have performed outside the performance limits established by Medtronic.

Not all malfunctions expose the patient to a loss of pacing or defibrillation therapy. Some malfunctions included in the following survival estimates may not have been detected at all by the physician or the patient. These malfunctions, however, are included in the survival estimates and provide important feedback to our product development organization.

To provide insight into the nature of malfunctions, each malfunction is categorized as Malfunction with Compromised Therapy Function or Malfunction without Compromised Therapy Function.

For this report, Normal Battery Depletion, Malfunction with Compromised Therapy Function, and Malfunction without Compromised Therapy Function are defined as follows:

Normal Battery Depletion – The condition when:

- (a) a device is returned with no associated complaint and the device has reached its elective replacement indicator(s) with implant time that meets or exceeds the nominal (50 percentile) predicted longevity at default (labeled) settings, or
- (b) a device is returned and the device has reached its elective replacement indicator(s) with implant time exceeding 80% of the expected longevity calculated using the available device setting information.

Medtronic CRDM establishes expected longevity by statistically characterizing the power consumed by the device and the power available from the device battery. This characterization is applied to a number of parameter configurations to derive a statistical mean longevity value and standard deviation for each parameter configuration. The statistical mean value minus three standard deviations is used as the expected longevity for determining if a battery depleted normally.

continued

The Standard Actuarial Method is used to estimate IPG and ICD survival. This product performance report has been prepared in accordance with International Standard ISO 5841-2:2000(E).

For reference purposes, the following pages include estimated longevity for each model. The actual longevity achieved for any device while implanted will depend on the actual programmed parameters and patient factors, and may differ significantly from these estimates.

Malfunction with Compromised Therapy Function

The condition when a device is found to have malfunctioned in a manner that compromised pacing or defibrillation therapy (including complete loss or partial degradation), while implanted and in service, as confirmed by returned product analysis.

Examples: Sudden loss of battery voltage; accelerated current drain such that low battery was not detected before loss of therapy; sudden malfunction during defibrillation therapy resulting in aborted delivery of therapy, intermittent malfunction where therapy is compromised while in the malfunction state.

Malfunction without Compromised Therapy Function

The condition when a device is found to have malfunctioned in a manner that *did not* compromise pacing or defibrillation therapy, while implanted and in service, as confirmed by returned product analysis.

Examples: Error affecting diagnostic functions, telemetry function, data storage; malfunction of a component that causes battery to lose power quickly enough to cause premature battery depletion, but slowly enough that the condition is detected through normal follow-up before therapy is lost; mechanical problems with connector header that do not affect therapy.

Expanded Malfunction Detail

The malfunctions are further divided into categories that identify the subject area of the malfunction. The malfunctions are divided into the following subject areas:

Electrical Component – Findings linked to electrical components such as integrated circuits, resistors, capacitors, diodes, etc.

Electrical Interconnect – Findings linked to the connections between electrical components such as wires, solder joints, wire bonds, etc.

Battery – Findings linked to the battery and its components

Software/Firmware – Findings linked to software or firmware function

Possible Early Battery Depletion – Findings where the actual reported implant time is less than 80% of the expected longevity calculated using the available device setting information with no device malfunction observed. There may not be sufficient device setting information to determine conclusively if battery depletion was normal or premature in the absence of a specific root cause finding. However, returned devices meeting the above criteria are conservatively classified as Possible Early Battery Depletion malfunctions.

Other – Findings related to other components such as insulators, grommets, setscrews, and packaging, and findings where analysis is inconclusive

Returned Product Analysis Process

Analysis of returned product is performed according to written procedures. These procedures determine the minimum analysis required. The analysis required varies depending on the type of device, age of the device, the associated information received with the device, actual experience with models of similar design, and other factors. Additional analysis is performed as necessary to investigate a performance concern from a customer, or to collect specific reliability data.

When a device is returned with a performance concern from a customer, the general analysis process includes a preliminary analysis of the device in its as-received condition, followed by an automated functional test using test equipment equivalent to the equipment used in manufacturing.

When a malfunction is identified, failure analysis is performed to provide the detailed information necessary to investigate possible causes and actions. Medtronic CRDM maintains in-house expertise and performs its failure analysis using facilities it owns and supports. This capability permits detailed failure analysis.

continued

Medtronic CRDM adjusts all-cause survival estimates to account for underreporting. While this lowers our all-cause survival estimates, we feel it gives a more accurate perspective on real performance.

Statistical Methods for Survival Analysis

Of the several different statistical methods available for survival analysis, the Standard Actuarial Method, with suspensions assumed distributed evenly within the intervals (Cutler-Ederer Method), is used to determine estimates of IPG and ICD survival. This method is commonly used by medical researchers and clinicians.

Implant times are calculated from the implant date to the earlier of the explant date or the cutoff date of the report. From this data an estimate of the probability of device survival is calculated at each monthly interval.

On the following pages, each graph includes a survival curve where events include malfunctions and normal battery depletions. This survival curve is a good representation of the probability a device will survive a period of time without malfunction and without battery depletion. For example, if a device survival probability is 95% after 5 years of service, then the device has a 5% chance of being removed due to battery depletion or malfunction in the first 5 years following implant.

In addition, a second curve is included to show survival excluding normal battery depletion. This curve is a good representation of the probability for a device to survive without malfunction. This curve includes only malfunctions as events and excludes normal battery depletion.

Since the survival estimate can become very imprecise with small effective sample sizes, Medtronic truncates the survival curve when the effective sample size is less than 100 for CRT, ICD, and IPG devices. The survival charts in the Product Performance Report show the effective sample size for each year interval where we have experience. When the effective sample size reaches 100, the next data point is added to the survival curve.

Although the report provides tabular data in one-year intervals, the curves are actually computed and plotted using one-month intervals.

The data in the tables are rounded to the nearest tenth of one percent. Occasionally, a graph may show 100% survival, but have one or more malfunctions or battery depletions. This occurs because, even with the malfunctions or battery depletions, the data rounds to 100%.

The survival curves are statistical estimates. As performance experience accumulates, the estimation improves. Confidence intervals are provided as a way to indicate the degree of certainty of the estimates.

Greenwood's formula is used to calculate corresponding 95% confidence intervals for the standard errors, and the complementary log-log method is used to produce the confidence bounds.

Sample Size and How the Population and Population Samples Are Defined

The population sample from which the survival estimates are derived is comprised of the devices registered as implanted in the United States as of the report cutoff date. The number of registered implants, as well as an estimate of the number that remain in active service, is listed for each model. To be included in the population, the device must have been registered with Medtronic's registration system and implanted for at least one day.

This sample based on US implants is considered to be representative of the worldwide population, and therefore the survival estimates shown in this report should be representative of the performance worldwide of these models.

A CRT, ICD, or IPG model or model family will be included in this report when it has accumulated at least 10,000 implant months and will remain in the report as long as at least 500 devices remain active.

Methods Used to Adjust for Underreporting of Malfunction and Battery Depletion

The tables on the following pages show the actual number of malfunctions and battery depletions recorded by the analysis lab for US registered devices. Since not all devices are returned to Medtronic CRDM for analysis, these numbers underestimate the true number of malfunctions and battery depletions. To more accurately estimate the all-cause device survival probabilities, the number of malfunctions and battery depletions used to plot each interval of the all-cause survival curves is adjusted (multiplied) by a factor that is based on an estimate of the magnitude of underreporting. The magnitude of underreporting is estimated by analyzing experience in Medtronic's Device And Registrant Tracking (DART) system.

The DART system is an important element of Medtronic's Quality System. The DART system is designed to meet or exceed the US FDA's device tracking requirements set forth by the Safe Medical Devices Act. In the United States, over 98% of Medtronic's CRT, ICD, and IPG implants become registered in the DART system.

continued

Because pacemakers do not cure the patient's underlying health problem, when a pacemaker stops functioning (due to either normal battery replacement or malfunction) it is replaced with a new pacemaker. Therefore, the replacement recorded in the DART system is a good indication that the previous pacemaker experienced either battery depletion or malfunction. The fraction of replaced devices that are subsequently returned can be used to estimate the correction factor for the under reporting of the combination of battery depletion and malfunction.

Note that devices of patients who have expired do not factor into the calculation of the correction. It is possible some proportion of these device experienced battery depletion or malfunction. Since these are not counted into the correction factor based on the return rate of replaced devices, a correction factor based only on the return rate of replaced devices may still underestimate the true rate of battery depletion and malfunction. However, devices that are replaced because the patient is receiving a system upgrade or are removed because the patient no longer needs it (e.g., due to heart transplant) do contribute to the calculation of the correction factor and therefore impart an opposite bias.

Also note that this method of calculating the correction factor cannot distinguish between devices that are removed due to malfunction and those due to normal battery depletion. It might seem intuitive that devices that unexpectedly malfunction should be much more likely to be returned to the manufacturer than a device with ordinary normal battery depletion. But this has not been conclusively demonstrated. Therefore, this method only provides a correction factor reflecting the combination of battery depletion and malfunction.

No adjustment for underreporting is applied to the malfunction-free survival curve because a method for estimating malfunction-only underreporting has not been developed.

Adjustments to Registered Implants to Compensate for Unreported Devices Removed from Service


Devices are at times removed from service for reasons other than device malfunction or battery depletion. Examples are devices removed from service due to non-device related patient mortality and devices removed due to changes in the patient's medical condition. Because an accurate estimate of device survival depends on an accurate estimate of the number of devices in service, it is important not to overstate the number of devices in service.

To ensure the number of devices in service is not overstated, Medtronic addresses this underreporting in two ways. Regular updates obtained from the Social Security Administration about deceased persons is used to update Medtronic's DART data about patients who have died but whose deaths had not been reported to Medtronic. In addition, the patient mortality rate derived from our DART system is monitored and compared to published mortality rates for comparable patient populations. If, during calculation of the survival curves, the patient mortality indicated by the data in DART is significantly different from published rates, an adjustment is applied to correct the difference.

7299 InSync Sentry

Product Characteristics

US Market Release	Apr-05	Malfunctions (US)	178	NBD Code	VVE-DDDR
Registered US Implants	31,200	Therapy Function Not Compromised	168	Serial Number Prefix	PRK
Estimated Active US Implants	2,100	Electrical Component	18	Max Delivered Energy	35 J
Normal Battery Depletions (US)	9,674	Possible Early Battery Depletion	147	Estimated Longevity	See page 20
Advisories	None	Software Malfunction	2		
		Other	1		
		Therapy Function Compromised	10		
		Electrical Component	10		




Years After Implant	Excluding Normal Battery Depletion					Including Normal Battery Depletion	
	1 yr	2 yr	3 yr	4 yr	5 yr	at 61 mo	
%	100.0	99.9	99.7	99.2	99.0	98.9	
%	99.7	98.1	91.7	71.9	13.2	5.0	
#	27,200	23,800	19,000	11,900	1,700	870	
Effective Sample Size							

7304 InSync Maximo

Product Characteristics

US Market Release	Apr-05	Malfunctions (US)	111	NBD Code	VVE-DDDR
Registered US Implants	19,000	Therapy Function Not Compromised	106	Serial Number Prefix	PRL
Estimated Active US Implants	1,800	Electrical Component	14	Max Delivered Energy	35 J
Normal Battery Depletions (US)	5,309	Battery Malfunction	1	Estimated Longevity	See page 20
Advisories	None	Possible Early Battery Depletion	90		
		Other	1		
		Therapy Function Compromised	5		
		Electrical Component	4		
		Other	1		




	Years After Implant					
	1 yr	2 yr	3 yr	4 yr	5 yr	at 62 mo
%	100.0	99.9	99.6	99.2	99.1	99.1
%	99.7	98.1	92.4	74.5	20.6	4.5
#	16,900	14,800	11,900	7,600	1,200	350
Effective Sample Size						

8040 InSync

Product Characteristics

US Market Release	Aug-01	Malfunctions (US)	34
Registered US Implants	15,400	Therapy Function Not Compromised	24
Estimated Active US Implants	1,300	Electrical Component	4
Normal Battery Depletions (US)	1,386	Electrical Interconnect	16
Advisories	None	Possible Early Battery Depletion	3
		Other	1
		Therapy Function Compromised	10
		Electrical Interconnect	10

NBG Code	DDDR
Serial Number Prefix	PIN
Estimated Longevity	See page 21




8042 InSync III

Product Characteristics

US Market Release	Feb-03	Malfunctions (US)	23
Registered US Implants	39,500	Therapy Function Not Compromised	13
Estimated Active US Implants	14,100	Electrical Component	2
Normal Battery Depletions (US)	2,138	Electrical Interconnect	3
Advisories	None	Possible Early Battery Depletion	1
		Other	7
		Therapy Function Compromised	10
		Electrical Interconnect	10


NBG Code	DDDR
Serial Number Prefix	PKF
Estimated Longevity	See page 21



C154DWK, C164AWK, C174AWK Concerto

Product Characteristics

	(N)	(A)		(N)	(A)		
US Market Release	May-06	May-06	Malfunctions (US)	1,377	1,298	NBD Code	DDE-DDDR
Registered US Implants	81,500	3,500	Therapy Function Not Compromised	1,337	1,284	Serial Number Prefix	PVU, PVT, PVR
Estimated Active US Implants	22,600	200	Electrical Component	678	1,280	Max Delivered Energy	35 J
Normal Battery Depletions (US)	16,614	267	Electrical Interconnect	2	0	Estimated Longevity	See page 20
Advisories: See page 139 – 2009 Potential Reduced Device Longevity			Software/Firmware	3	0		
			Possible Early Battery Depletion	651	4	(N) = Non-advisory population	
			Other	3	0	(A) = Advisory population	
			Therapy Function Compromised	40	14		
			Electrical Component	38	12		
			Electrical Interconnect	2	1		
			Other	0	1		



Years After Implant — Excluding Normal Battery Depletion - - - Including Normal Battery Depletion


	Non-Adv	1 yr	2 yr	3 yr	4 yr	5 yr	at 64 mo
%	—	100.0	99.8	99.5	98.4	96.8	96.8
%	- - -	99.8	98.3	93.4	79.3	41.5	18.1
#		72,800	64,100	53,900	38,000	9,000	1,900
Effective Sample Size							

	Advisory	1 yr	2 yr	3 yr	at 43 mo
%	—	99.8	99.4	79.1	51.2
%	- - -	99.7	97.5	59.9	8.0
#		3,100	2,700	1,500	200
Effective Sample Size					

D224TRK, D234TRK, D204TRM, D214TRM Consulta CRT-D

Product Characteristics


US Market Release	Sep-08	Malfunctions (US)	440	NBD Code	DDE-DDDR
Registered US Implants	66,700	Therapy Function Not Compromised	424	Serial Number Prefix	PUC, PUD, PYZ, PZA
Estimated Active US Implants	48,000	Electrical Component	24	Max Delivered Energy	35 J
Normal Battery Depletions (US)	2,844	Electrical Interconnect	1	Estimated Longevity	See page 20
Advisories	None	Possible Early Battery Depletion	393		
		Software Malfunction	5		
		Other	1		
		Therapy Function Compromised	16		
		Electrical Component	16		



D274TRK, D294TRK Concerto II CRT-D

Product Characteristics


US Market Release	Aug-09	Malfunctions (US)	110	NBD Code	DDE-DDDR
Registered US Implants	30,200	Therapy Function Not Compromised	108	Serial Number Prefix	PZV, PZU
Estimated Active US Implants	22,900	Electrical Component	11	Max Delivered Energy	35 J
Normal Battery Depletions (US)	784	Possible Early Battery Depletion	96	Estimated Longevity	See page 20
Advisories	None	Software/Firmware	1		
		Therapy Function Compromised	2		
		Electrical Component	2		



D264TRM, D284TRK Maximo II CRT-D

Product Characteristics


US Market Release	Sep-08	Malfunctions (US)	112	NBD Code	VVE-DDDR
Registered US Implants	14,900	Therapy Function Not Compromised	109	Serial Number Prefix	PZP, PZO
Estimated Active US Implants	10,300	Electrical Component	3	Max Delivered Energy	35 J
Normal Battery Depletions (US)	799	Possible Early Battery Depletion	106	Estimated Longevity	See page 20
Advisories	None	Therapy Function Compromised	3		
		Electrical Component	3		



D314TRG, D314TRM, D334TRG, D334TRM, D354TRG, D354TRM, D364TRG, D364TRM Protecta/Protecta XT DR

Product Characteristics

US Market Release	Mar-11	Malfunctions (US)	16	NBD Code	DDE-DDDR, VVE-DDDR
Registered US Implants	57,000	Therapy Function Not Compromised	13	Serial Number Prefix	PFS, PSE, PSF, PSI, PSL, PSO, PTB, PTE, PXI, PXJ
Estimated Active US Implants	53,700	Electrical Component	8	Max Delivered Energy	35J
Normal Battery Depletions (US)	67	Possible Early Battery Depletion	5	Estimated Longevity	See page 20
Advisories	None	Therapy Function Compromised	3		
		Electrical Component	3		



C2TR01 Syncra CRT-P, C3TR01, C4TR01 Consulta CRT-P


Product Characteristics

US Market Release	Mar-11
Registered US Implants	15,300
Estimated Active US Implants	14,000
Normal Battery Depletions (US)	2

Malfunctions (US)	0
Therapy Function Not Compromised	0
Therapy Function Compromised	0

NBG Code	OAE-DDDR, OOE-DDDR
Serial Number Prefix	PVX, PZI, PZX
Max Delivered Energy	NA
Estimated Longevity	See page 21

Advisories: [See page 135](#) 2013 Potential Loss of Device Hermeticity



Device Survival Summary (95% Confidence Interval)

The following table shows CRT device survival estimates with 95% confidence intervals. Estimates are shown both with and without normal battery depletions included.

Model Number	Family	US Market Release	Registered US Implants	Estimated Active US Implants	Normal Battery Depletions (US)	Malfunctions (US)		Device Survival Probability (%)											
						Therapy Function Compromised	Therapy Function Not Compromised	Years After Implant											
						1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	10 yr	12 yr				
7299	InSync Sentry	Apr-05	31,200	2,100	9,674	10 + 168 = 178	178	100.0 +0.0/-0.0	99.9 +0.0/-0.1	99.7 +0.1/-0.1	99.2 +0.1/-0.1	99.0 +0.1/-0.2	98.9 +0.2/-0.2 at 61 mo						
						5 + 106 = 111	111	99.7 +0.1/-0.1	98.1 +0.2/-0.2	91.7 +0.4/-0.4	71.9 +0.6/-0.7	13.2 +0.7/-0.7	5.0 +0.5/-0.5 at 61 mo						
7304	InSync Maximo	Apr-05	19,000	1,800	5,309	5 + 106 = 111	111	100.0 +0.0/-0.0	99.9 +0.0/-0.1	99.6 +0.1/-0.1	99.2 +0.2/-0.2	99.1 +0.2/-0.2	99.1 +0.2/-0.2 at 62 mo						
						10 + 24 = 34	34	99.7 +0.1/-0.1	98.1 +0.2/-0.2	92.4 +0.4/-0.5	74.5 +0.8/-0.8	20.6 +1.0/-1.0	4.5 +0.8/-0.7 at 62 mo						
8040	InSync	Aug-01	15,400	1,300	1,386	10 + 24 = 34	34	100.0 +0.0/-0.0	100.0 +0.0/-0.0	99.9 +0.0/-0.1	99.8 +0.1/-0.1	99.7 +0.1/-0.2	99.6 +0.1/-0.2	99.6 +0.1/-0.2	99.6 +0.1/-0.2	99.4 +0.2/-0.3 at 124 mo			
						10 + 13 = 23	23	99.8 +0.1/-0.1	99.5 +0.1/-0.1	98.0 +0.3/-0.3	95.7 +0.4/-0.5	90.8 +0.7/-0.7	83.5 +0.9/-1.0	73.1 +1.2/-1.3	58.5 +1.6/-1.6	20.6 +1.9/-1.9	10.3 +1.9/-1.7 at 124 mo		
8042	InSync III	Feb-03	39,500	14,100	2,138	10 + 13 = 23	23	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.1	99.9 +0.1/-0.1 at 104 mo			
						40 + 1,337 = 1,377	1,377	99.9 +0.0/-0.0	99.8 +0.0/-0.1	99.2 +0.1/-0.1	97.7 +0.2/-0.2	95.1 +0.3/-0.3	89.7 +0.5/-0.5	77.2 +0.9/-0.9	50.8 +1.5/-1.5	15.4 +2.3/-2.2 at 104 mo			
C154DVK, C164AWK, C174AWK (Non- advisory population)	Concerto	May-06	81,500	22,600	16,614	40 + 1,337 = 1,377	1,377	100.0 +0.0/-0.0	99.8 +0.0/-0.0	99.5 +0.1/-0.1	98.4 +0.1/-0.1	96.8 +0.2/-0.2	96.8 +0.2/-0.2 at 64 mo						
						99.8 +0.0/-0.0	98.3 +0.1/-0.1	93.4 +0.2/-0.2	79.2 +0.3/-0.3	41.5 +0.6/-0.6	18.1 +0.7/-0.7 at 64 mo								

continued

Device Survival Summary continued

		Malfunctions (US)				Device Survival Probability (%)													
Model Number	Family	US Market Release	Registered US Implants	Estimated Active US Implants	Normal Battery Depletions (US)	Malfunctions (US)		Years After Implant											
						Therapy Function Compromised	Therapy Function Not Compromised	1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	10 yr			
C154DWK, C164AWK, C174AWK (Advisory population)	Concerto	May-06	3,500	200	267	14	1,284	=	1,298	99.8 +0.1/-0.2	99.4 +0.2/-0.3	79.1 +1.6/-1.7	51.2 +2.2/-2.2 at 43 mo						
Advisories: See page 139–2009 Potential Reduced Device Longevity										99.7 +0.1/-0.3	97.5 +0.5/-0.6	59.9 +1.9/-2.0	8.0 +1.5/-1.3 at 45 mo						
D224TRK, D234TRK, D204TRM, D214TRM	Consulta CRT-D/ CRT-D	Sep-08	66,700	48,000	2,845	16	424	=	440	100.0 +0.0/-0.0	99.7 +0.0/-0.0	99.2 +0.1/-0.1	98.5 +0.2/-0.2 at 53 mo	98.3 +0.2/-0.2 at 53 mo					
D274TRK, D294TRK	Concerto II CRT-D	Aug-09	30,200	22,900	784	2	108	=	110	99.8 +0.0/-0.0	98.5 +0.1/-0.1	92.9 +0.3/-0.3	77.58 +0.7/-0.7 at 53 mo	61.2 +1.8/-1.9 at 53 mo					
D264TRM, D284TRK	Maximo II CRT-D	Sep-08	14,900	10,300	799	3	109	=	112	100.0 +0.0/-0.0	99.7 +0.1/-0.1	99.2 +0.2/-0.2	98.3 +0.3/-0.4 at 54 mo	97.9 +0.4/-0.5 at 54 mo					
D314TRG, D314TRM, D334TRG, D334TRM, D354TRG, D354TRM, D364TRG, D364TRM	Protecta/Protecta XT CRT-D	Mar-11	57,000	53,700	68	3	13	=	16	100.0 +0.0/-0.0	99.9 +0.0/-0.1	99.6 +0.3/-1.9 at 28 mo							
										99.9 +0.0/-0.0	99.1 +0.2/-0.2	97.9 +0.9/-1.5 at 28 mo							

Device Survival Summary continued

Model Number	Family	US Market Release	US Implants Registered	Estimated Active US Implants	Normal Battery Depletions (US)	Malfunctions (US)			Device Survival Probability (%)									
						Therapy Function Compromised	Therapy Function Not Compromised	Total	Years After Implant									
									1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	10 yr	
C2TR01, C3TR01, C4TR01	Syncra CRT-P Consulta CRT-P	Mar-11	15,200	14,000	2	0	0	0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0 at 29 mo							
	Advisories; See page 139 – 2013 Potential Loss of Device Hermeticity								100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0 at 29 mo							

continued

Reference Chart

The longevity estimates provided are mean values calculated for the parameters given. The actual longevity achieved for any device while implanted will depend on the actual programmed parameters and patient factors, and may differ significantly from these estimates.

Model Number	Family	Volume/ Mass*	Delivered Energy	Estimated Longevity					Elective Replacement (ERI)**		End of Life (EOL) Battery Voltage
				Charging Frequency**	100% Pacing#	50% Pacing#	15% Pacing#	100% Sensing	Battery Voltage	Charge Time	
7299	InSync Sentry	40 cc 78 g	35 J	Monthly	3.3	3.8	4.1	4.3	≤ 2.62 V	> 16 second charge time	3 months after ERI
				Quarterly	4.5	5.3	6.2	6.6			
				Semiannual	5.0	6.0	7.1	7.7			
7304	InSync Maximo	40 cc 78 g	35 J	Monthly	3.3	3.8	4.1	4.3	≤ 2.62 V	> 16 second charge time	3 months after ERI
				Quarterly	4.5	5.3	6.2	6.6			
				Semiannual	5.0	6.0	7.1	7.7			

Model Number	Family	Volume/ Mass*	Delivered Energy	Estimated Longevity					Recommended Replacement (RRT)**		End of Service (EOS)
				Charging Frequency**	100% Pacing#	50% Pacing#	15% Pacing#	100% Sensing	Battery Voltage	Charge Time	
C154DWK, C164AWK, C174AWK	Concerto	38 cc 68 g	35 J	Monthly	3.8	4.3	4.8	5.0	≤ 2.62 V	—	3 month after RRT or > 16-second charge time
				Quarterly	5.5	6.8	8.0	8.8			
				Semiannual	6.3	8.0	9.8	10.9			
D224TRK, D234TRK, D204TRM, D214TRM	Consulta CRT-D CRT-D M4	38 cc/ 68 g	35 J	Monthly	3.2	3.8	4.4	4.7	≤ 2.63 V	—	3 month after RRT or > 16-second charge time
				Quarterly	4.4	5.5	6.8	7.5			
				Semiannual	4.8	6.2	7.9	9.0			
D274TRK D294TRK	Concerto II	38 cc/ 68 g	35 J	Monthly	3.2	3.8	4.4	4.7	≤ 2.63 V	—	3 month after RRT or > 16-second charge time
				Quarterly	4.4	5.5	6.8	7.5			
				Semiannual	4.8	6.2	7.9	9.0			
D284TRK, D264TRM	Maximo II CRT-D CRT-D M4	38 cc/ 68 g	35 J	Monthly	3.2	3.8	4.4	4.7	≤ 2.63 V	—	3 month after RRT or > 16-second charge time
				Quarterly	4.4	5.5	6.8	7.5			
				Semiannual	4.8	6.2	7.9	9.0			
D314TRG, D354TRG, D314TRM, D354TRM	Protecta XT CRT-D CRT-D M4	38 cc/ 68 g	35 J	Monthly	3.2	3.8	4.4	4.7	≤ 2.63 V	—	3 month after RRT
				Quarterly	4.4	5.5	6.8	7.5			
				Semiannual	4.8	6.2	7.9	9.0			
D334TRG, D364TRG, D334TRM, D364TRM	Protecta CRT-D CRT-D M4	38 cc/ 68 g	35 J	Monthly	3.2	3.8	4.4	4.7	≤ 2.63 V	—	3 month after RRT
				Quarterly	4.4	5.5	6.8	7.5			
				Semiannual	4.8	6.2	7.9	9.0			

* Volume and mass differ by connector style.

** A full charge is a full energy therapeutic shock or capacitor reformation.

*** The minimum time between ERI and EOL (or RRT and EOS) is 3 months (100% pacing, two charges per month).

‡ Pacing mode is DDD for CRT models. Parameter settings; lower rate at 60 ppm, sensing rate at 70 bpm, (A, RV, LV) 3.0 V amplitude, 0.4 ms pulse width, and 510-ohm pace load per applicable channel.

Reference Chart

The longevity estimates provided are mean values calculated for the parameters given. The actual longevity achieved for any device while implanted will depend on the actual programmed parameters and patient factors, and may differ significantly from these estimates.

Model Number	Family	Estimated Longevity		
		Amplitude Setting	500 Lead Ω	1,000 Lead Ω
8040	InSync	Low 2.5 V (A, RV)	11.9	13.7
		Nominal 3.5 V (A, RV)	8.9	11.4
		High 5.0 V (A, RV)	6.6	9.1
8042	InSync III	Low 2.5 V (A, RV, LV)	8.3	9.9
		Nominal 3.5 V (A, RV, LV)	5.9	7.8
		High 5.0 V (A, RV, LV)	4.1	6.0
C2TR01	Syncra CRT-P	Low 2.5 V (A, RV)	8.7	10.7
		Normal 3.5 V (A, RV)	6.0	8.2
		High 5.0 V (A, RV)	3.3	5.1
C3TR01 C4TR01	Consulta CRT-P	Low 2.5 V (A, RV)	8.7	10.7
		Normal 3.5 V (A, RV)	6.0	8.2
		High 5.0 V (A, RV)	3.3	5.1

7230Cx, B, E Marquis VR


Product Characteristics

US Market Release	Dec-02
Registered US Implants	19,400
Estimated Active US Implants	2,770
Normal Battery Depletions (US)	2,753

Advisories: [See page 145](#) – 2005 Potential Premature Battery Depletion Due to Battery Short

Malfunctions (US)	61
Therapy Function Not Compromised	32
Electrical Component	15
Software/Firmware	1
Battery Malfunction (<i>1 malfunction due to advisory</i>)	1
Possible Early Battery Depletion	14
Other	1
Therapy Function Compromised	29
Battery Malfunction (<i>19 malfunctions due to advisory</i>)	20
Electrical Component	9

NBD Code	VVE-VVIR
Serial Number Prefix	PKD, PLW, PLY
Max Delivered Energy	30 J
Estimated Longevity	See page 39




	1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	9 yr	at 112 mo
% ——— Excluding Normal Battery Depletion	99.9	99.9	99.9	99.8	99.7	99.6	99.5	99.4	99.4	99.4
% - - - - - Including Normal Battery Depletion	99.7	99.4	99.1	98.7	98.1	94.6	86.8	75.0	40.2	12.4
#	17,300	13,600	11,300	10,200	9,100	7,900	6,600	5,000	1,500	250
Effective Sample Size										

7231Cx GEM III VR

Product Characteristics


US Market Release	Dec-00	Malfunctions (US)	37	NBD Code	VVE-VVIR
Registered US Implants	17,500	Therapy Function Not Compromised	27	Serial Number Prefix	PJL
Estimated Active US Implants	1,800	Electrical Component	22	Max Delivered Energy	30 J
Normal Battery Depletions (US)	3,684	Battery Malfunction	1	Estimated Longevity	See page 39
		Possible Early Battery Depletion	4		
		Therapy Function Compromised	10		
		Electrical Component	8		
		Electrical Interconnect	1		
		Battery Malfunction	1		



7232B, Cx, E Maximo VR

Product Characteristics


US Market Release	Oct-03	Malfunctions (US)	73	NBD Code	VVE-VVIR
Registered US Implants	44,300	Therapy Function Not Compromised	57	Serial Number Prefix	PRN, PVF, PVG
Estimated Active US Implants	17,100	Electrical Component	28	Max Delivered Energy	35 J
Normal Battery Depletions (US)	3,737	Possible Early Battery Depletion	23	Estimated Longevity	See page 39
Advisories: See page 145 – 2005 Potential Premature Battery Depletion Due to Battery Short		Other	6		
		Therapy Function Compromised	16		
		Electrical Component	13		
		Electrical Interconnect	1		
		Possible Early Battery Depletion	1		
		Other	1		



7274 Marquis DR

Product Characteristics

US Market Release	Mar-02	Malfunctions (US)	196	NBD Code	VVE-DDDR
Registered US Implants	48,400	Therapy Function Not Compromised	89	Serial Number Prefix	PKC
Estimated Active US Implants	2,300	Battery Malfunction (73 malfunctions due to advisory)	6	Max Delivered Energy	30 J
Normal Battery Depletions (US)	8,904	Electrical Component	31	Estimated Longevity	See page 39
Advisories: See page 145 – 2005 Potential Premature Battery Depletion Due to Battery Short		Possible Early Battery Depletion	51		
		Other	1		
		Therapy Function Compromised	107		
		Battery Malfunction (3 malfunctions due to advisory)	80		
		Electrical Component	27		




	1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	at 90 mo
% ———	100.0	99.9	99.8	99.6	99.4	99.3	99.2	99.2
% - - - - -	99.8	99.5	98.6	97.3	92.1	72.5	34.2	1.1
#	43,000	34,600	26,600	22,500	18,500	12,100	4,100	400
Effective Sample Size								



7278 Maximo DR

Product Characteristics

US Market Release	Oct-03	Malfunctions (US)	70	NBD Code	VVE-DDDR
Registered US Implants	37,700	Therapy Function Not Compromised	60	Serial Number Prefix	PRM
Estimated Active US Implants	5,800	Electrical Component	22	Max Delivered Energy	35 J
Normal Battery Depletions (US)	9,134	Possible Early Battery Depletion	34	Estimated Longevity	See page 39
Advisories: See page 145 – 2005 Potential Premature Battery Depletion Due to Battery Short		Other	4		
		Therapy Function Compromised	10		
		Electrical Component	9		
		Possible Early Battery Depletion	1		




Years After Implant	1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	at 88 mo
%	100.0	99.9	99.9	99.8	99.8	99.7	99.7	99.7
%	99.9	99.6	99.1	97.7	90.2	67.0	28.5	3.2
#	34,000	30,400	27,200	23,800	19,000	11,200	3,100	470
Effective Sample Size								

7288 Intrinsic

Product Characteristics

US Market Release	Jun-04	Malfunctions (US)	71	NBD Code	VVE-DDDR
Registered US Implants	30,700	Therapy Function Not Compromised	64	Serial Number Prefix	PUB
Estimated Active US Implants	3,900	Electrical Component	27	Max Delivered Energy	35 J
Normal Battery Depletions (US)	9,100	Battery Malfunction	2	Estimated Longevity	See page 39
Advisories	None	Possible Early Battery Depletion	33		
		Software Malfunction	1		
		Other	1		
		Therapy Function Compromised	7		
		Electrical Component	5		
		Other	2		




Years After Implant	Excluding Normal Battery Depletion							Including Normal Battery Depletion	
	1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	at 86 mo	
%	100.0	99.9	99.9	99.8	99.8	99.7	99.7	99.7	
%	99.9	99.6	99.2	98.2	90.7	70.4	22.1	7.8	
#	28,700	26,400	23,800	21,000	17,500	11,500	2,600	970	
Effective Sample Size									



D154ATG, D154DRG EnTrust

Product Characteristics


US Market Release	Jun-05	Malfunctions (US)	122	NBD Code	DDE-DDDR, VVE-DDDR
Registered US Implants	28,200	Therapy Function Not Compromised	108	Serial Number Prefix	PNR
Estimated Active US Implants	7,200	Electrical Component (9 malfunctions due to advisory)	29	Max Delivered Energy	35 J
Normal Battery Depletions (US)	5,787	Electrical Interconnect	1	Estimated Longevity	See page 40
Advisories: See page 136 – 2012 Potential Rapid Battery Depletion		Software Malfunction	3		
		Possible Early Battery Depletion	74		
		Other	1		
		Therapy Function Compromised	14		
		Electrical Component (2 malfunctions due to advisory)	14		



D154AWG, D164AWG Virtuoso DR

Product Characteristics

	(N)	(A)		(N)	(A)		
US Market Release	May-06	May-06	Malfunctions (US)	1,219	1,874	NBD Code	DDE-DDDR
Registered US Implants	72,700	4,100	Therapy Function Not Compromised	1,190	1,861	Serial Number Prefix	PVV, PUL
Estimated Active US Implants	40,200	300	Electrical Component	1,056	1,860	Max Delivered Energy	35 J
Normal Battery Depletions (US)	4,514	117	Electrical Interconnect	2	0	Estimated Longevity	See page 40
Advisories: See page 139 – 2009 Potential Reduced Device Longevity			Possible Early Battery Depletion	127	0		
			Software Malfunction	1	0		
			Other	4	1		
			Therapy Function Compromised	29	13		
			Electrical Component	26	13		
			Possible Early Battery Depletion	1	0		
			Other	2	0		



Years After Implant — Excluding Normal Battery Depletion — Including Normal Battery Depletion

Non-Adv	1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	at 77 mo
%	100.0	99.9	99.9	99.4	97.3	97.2	97.2
%	99.9	99.7	99.4	98.0	89.5	70.6	46.9
#	67,700	62,400	56,900	49,400	30,500	7,700	690
Effective Sample Size							


Advisory	1 yr	2 yr	3 yr	4 yr	at 49 mo
%	100.0	99.9	90.5	49.5	46.0
%	99.9	99.6	84.1	13.9	6.5
#	3,800	3,500	2,800	400	250
Effective Sample Size					



D154VRC EnTrust VR

Product Characteristics


US Market Release	Jun-05	Malfunctions (US)	99	NBD Code	VVE-VVIR
Registered US Implants	14,500	Therapy Function Not Compromised	80	Serial Number Prefix	PNT
Estimated Active US Implants	6,900	Battery Malfunction	2	Max Delivered Energy	35 J
Normal Battery Depletions (US)	678	Electrical Component	44	Estimated Longevity	See page 40
Advisories	None	Possible Early Battery Depletion	24		
		Other	10		
		Therapy Function Compromised	19		
		Electrical Component	18		
		Other	1		



D154VWC, D164VWC Virtuoso VR

Product Characteristics


US Market Release	May-06	Malfunctions (US)	306	NBD Code	VVE-VVIR
Registered US Implants	33,100	Therapy Function Not Compromised	292	Serial Number Prefix	PUN, PUP
Estimated Active US Implants	20,800	Electrical Component (<i>4 malfunctions due to advisory</i>)	273	Max Delivered Energy	35 J
Normal Battery Depletions (US)	432	Electrical Interconnect	1	Estimated Longevity	See page 40
Advisories: See page 139 – 2009 Potential Reduced Device Longevity		Possible Early Battery Depletion	14		
		Other	4		
		Therapy Function Compromised	14		
		Electrical Component	14		



D224DRG, D234DRG, D204DRM, D214DRM Secura DR/DR

Product Characteristics


US Market Release	Sep-08	Malfunctions (US)	63	NBD Code	DDE-DDDR
Registered US Implants	50,800	Therapy Function Not Compromised	49	Serial Number Prefix	PUF, PUG, PZC, PZD
Estimated Active US Implants	42,400	Electrical Component	15	Max Delivered Energy	35 J
Normal Battery Depletions (US)	207	Possible Early Battery Depletion	23	Estimated Longevity	See page 40
Advisories	None	Software Malfunction	9		
		Other	2		
		Therapy Function Compromised	14		
		Electrical Component	12		
		Possible Early Battery Depletion	1		
		Software Malfunction	1		



D224VRC, D234VRC, D204VRM, D214VRM Secura VR/VR

Product Characteristics


US Market Release	Sep-08	Malfunctions (US)	15	NBD Code	VVE-VVIR
Registered US Implants	20,600	Therapy Function Not Compromised	10	Serial Number Prefix	PUK, PUX, PZF, PZG
Estimated Active US Implants	17,300	Electrical Component	2	Max Delivered Energy	35 J
Normal Battery Depletions (US)	38	Possible Early Battery Depletion	5	Estimated Longevity	See page 40
Advisories	None	Software Malfunction	2		
		Other	1		
		Therapy Function Compromised	5		
		Electrical Component	4		
		Software/Firmware	1		



D274DRG, D294DRG Virtuoso II DR

Product Characteristics


US Market Release	Aug-09	Malfunctions (US)	3	NBD Code	DDE-DDDR
Registered US Implants	22,200	Therapy Function Not Compromised	2	Serial Number Prefix	PZS, PZT
Estimated Active US Implants	18,800	Battery Malfunction	1	Max Delivered Energy	35 J
Normal Battery Depletions (US)	38	Possible Early Battery Depletion	1	Estimated Longevity	See page 40
Advisories	None	Therapy Function Compromised	1		
		Electrical Component	1		



D274VRC, D294VRC Virtuoso II VR

Product Characteristics


US Market Release	Aug-09	Malfunctions (US)	3	NBD Code	VVE-VVIR
Registered US Implants	9,100	Therapy Function Not Compromised	3	Serial Number Prefix	PZQ, PZR
Estimated Active US Implants	7,800	Electrical Component	1	Max Delivered Energy	35 J
Normal Battery Depletions (US)	11	Possible Early Battery Depletion	1	Estimated Longevity	See page 40
Advisories	None	Software Malfunction	1		
		Therapy Function Compromised	0		



D264DRM, D284DRG Maximo II DR/DR

Product Characteristics


US Market Release	Sep-08	Malfunctions (US)	20	NBD Code	VVE-DDDR
Registered US Implants	19,700	Therapy Function Not Compromised	15	Serial Number Prefix	PUJ, PZM
Estimated Active US Implants	16,300	Electrical Component	4	Max Delivered Energy	35 J
Normal Battery Depletions (US)	84	Possible Early Battery Depletion	11	Estimated Longevity	See page 40
Advisories	None	Therapy Function Compromised	5		
		Electrical Component	5		



D264VRM, D284VRC Maximo II VR/VR

Product Characteristics


US Market Release	Sep-08	Malfunctions (US)	9	NBD Code	VVE-VVIR
Registered US Implants	12,700	Therapy Function Not Compromised	6	Serial Number Prefix	PUZ, PZN
Estimated Active US Implants	10,700	Electrical Component	3	Max Delivered Energy	35 J
Normal Battery Depletions (US)	27	Possible Early Battery Depletion	1	Estimated Longevity	See page 40
Advisories	None	Software Malfunction	2		
		Therapy Function Compromised	3		
		Electrical Component	2		
		Software Malfunction	1		



D314DRG, D314DRM, D334DRG, D334DRM, D354DRG, D354DRM, D364DRG, D364DRM Protecta/Protecta XT DR

Product Characteristics


US Market Release	Mar-11	Malfunctions (US)	18	NBD Code	DDE-DDDR, VVE-DDDR
Registered US Implants	52,300	Therapy Function Not Compromised	13	Serial Number Prefix	PSC, PSD, PSJ, PSK, PSM, PSP, PTC, PTF, PXF, PXP
Estimated Active US Implants	50,000	Electrical Component	13	Max Delivered Energy	35 J
Normal Battery Depletions (US)	26	Therapy Function Compromised	5	Estimated Longevity	See page 40
Advisories	None	Electrical Component	5		



D314VRG, D314VRM, D334VRG, D334VRM, D354VRG, D354VRM, D364VRG, D364VRM Protecta/Protecta XT VR

Product Characteristics

US Market Release	Mar-11	Malfunctions (US)	5	NBD Code	VVE-VVIR
Registered US Implants	17,600	Therapy Function Not Compromised	3	Serial Number Prefix	PSA, PSG, PSH, PSN, PSX, PTD, PTG, P XK, PXO
Estimated Active US Implants	16,800	Electrical Component	2	Max Delivered Energy	35 J
Normal Battery Depletions (US)	8	Other	1	Estimated Longevity	See page 40
Advisories	None	Therapy Function Compromised	2		
		Electrical Component	2		



Device Survival Summary (95% Confidence Interval)

The following table shows ICD device survival estimates with 95% confidence intervals. Estimates are shown both with and without normal battery depletions included.

Model Number	Family	US Market Release	Registered US Implantants	Estimated Active US Implantants	Normal Battery Depletions (US)	Malfunctions (US)		Device Survival Probability (%)									
						Therapy Function Compromised	Therapy Function Not Compromised	Years After Implant									
								1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	10 yr	
7230Cx B, E	Marquis VR	Dec-02	19,400	2,700	2,770	29 + 32 = 61	Total	99.9 +0.0/-0.0	99.9 +0.0/-0.1	99.9 +0.0/-0.1	99.8 +0.1/-0.1	99.7 +0.1/-0.1	99.6 +0.1/-0.1	99.5 +0.1/-0.2	99.4 +0.1/-0.2	99.4 +0.1/-0.2	99.4 +0.2/-0.2 at 112 mo
	Advisories: See page 145 – 2005 Potential Premature Battery Depletion Due to Battery Short					(19) + (1) = (20) (advisory-related subset)		99.7 +0.1/-0.1	99.4 +0.1/-0.1	99.1 +0.1/-0.2	98.7 +0.2/-0.2	98.1 +0.2/-0.3	94.6 +0.4/-0.5	86.8 +0.7/-0.7	75.0 +1.0/-1.0	12.4 +1.7/-1.6 at 112 mo	
7231Cx	GEM III VR	Dec-00	17,500	1,800	3,684	10 + 27 = 37	Total	99.9 +0.0/-0.0	99.9 +0.0/-0.1	99.8 +0.1/-0.1	99.8 +0.1/-0.1	99.7 +0.1/-0.1	99.7 +0.1/-0.1	99.7 +0.1/-0.1	99.7 +0.1/-0.1	99.7 +0.1/-0.1	99.7 +0.1/-0.1 at 107 mo
7232Cx B, E	Maximo VR	Oct-03	44,300	17,200	3,737	16 + 57 = 73	Total	100.0 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.8 +0.0/-0.0	99.8 +0.0/-0.0	99.8 +0.0/-0.1	99.8 +0.1/-0.1	99.8 +0.1/-0.1	99.8 +0.1/-0.1 at 102 mo
	Advisories: See page 145 – 2005 Potential Premature Battery Depletion Due to Battery Short					(0) + (0) = (0) (advisory-related subset)		99.9 +0.0/-0.0	99.7 +0.0/-0.1	99.4 +0.1/-0.1	99.1 +0.1/-0.1	97.7 +0.2/-0.2	91.5 +0.3/-0.4	83.1 +0.5/-0.5	71.5 +0.7/-0.7	44.8 +2.5/-2.5 at 102 mo	
7274	Marquis DR	Mar-02	48,400	2,800	8,904	107 + 89 = 196	Total	100.0 +0.0/-0.0	99.9 +0.0/-0.0	99.8 +0.0/-0.1	99.6 +0.1/-0.1	99.4 +0.1/-0.1	99.3 +0.1/-0.1	99.2 +0.1/-0.1	99.2 +0.1/-0.1 at 90 mo	99.2 +0.1/-0.1 at 90 mo	
	Advisories: See page 145 – 2005 Potential Premature Battery Depletion Due to Battery Short					(73) + (3) = (76) (advisory-related subset)		99.8 +0.0/-0.1	99.5 +0.1/-0.1	98.6 +0.1/-0.1	97.3 +0.2/-0.2	92.1 +0.3/-0.4	72.5 +0.6/-0.7	34.2 +0.8/-0.8	1.1 +0.3/-0.3 at 90 mo		
7278	Maximo DR	Oct-03	37,600	5,800	9,134	10 + 60 = 70	Total	100.0 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.8 +0.0/-0.1	99.8 +0.0/-0.1	99.7 +0.1/-0.1	99.7 +0.1/-0.1	99.7 +0.1/-0.1 at 88 mo	99.7 +0.1/-0.1 at 88 mo	
	Advisories: See page 145 – 2005 Potential Premature Battery Depletion Due to Battery Short					(0) + (0) = (0) (advisory-related subset)		99.9 +0.0/-0.0	99.6 +0.1/-0.1	99.1 +0.1/-0.1	97.7 +0.2/-0.2	90.2 +0.4/-0.4	67.0 +0.7/-0.7	28.5 +0.8/-0.8	3.2 +0.6/-0.5 at 88 mo		
7288	Intrinsic	Jun-04	30,700	3,900	9,100	7 + 64 = 71	Total	100.0 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.8 +0.0/-0.1	99.8 +0.1/-0.1	99.7 +0.1/-0.1	99.7 +0.1/-0.1	99.7 +0.1/-0.1 at 86 mo	99.7 +0.1/-0.1 at 86 mo	
								99.9 +0.0/-0.1	99.6 +0.1/-0.1	99.2 +0.1/-0.1	98.2 +0.2/-0.2	90.7 +0.4/-0.4	70.4 +0.7/-0.7	22.1 +0.8/-0.8	7.8 +0.6/-0.6 at 86 mo		

continued

Model Number	Family	US Market Release	Registered US Implants	Estimated Active US Implants	Normal Battery Depletions	Malfunctions		Device Survival Probability (%)										
						Therapy Function Compromised	Therapy Function Not Compromised	Years After Implant										
								1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	10 yr		
D154ATG, D154DRG	EnTrust	Jun-05	28,200	7,200	5,787	14 + 108 = 122	122	100.0 +0.0/-0.0	99.9 +0.0/-0.0	99.8 +0.0/-0.1	99.7 +0.1/-0.1	99.4 +0.1/-0.1	99.4 +0.1/-0.1	99.4 +0.1/-0.1	99.4 +0.1/-0.1	99.4 +0.1/-0.1	99.4 +0.1/-0.1	
	Advisories: See page 136 – 2012 Potential Rapid Battery Depletion					(2) + (9) = (11) (advisory-related subset)	(11)	99.8 +0.0/-0.1	99.6 +0.1/-0.1	99.1 +0.1/-0.1	97.9 +0.2/-0.2	90.5 +0.4/-0.4	69.6 +0.7/-0.7	32.8 +1.0/-1.0	21.4 +1.2/-1.2 at 86 mo			
D154AWG D164AWG (Non-advisory population)	Virtuoso DR	May-06	72,700	40,200	4,514	29 + 1,190 = 1,219	1,219	100.0 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.4 +0.1/-0.1	97.3 +0.1/-0.2	97.2 +0.2/-0.2	97.2 +0.2/-0.2	97.2 +0.2/-0.2	97.2 +0.2/-0.2	97.2 +0.2/-0.2	97.2 +0.2/-0.2
	Advisories: See page 139 – 2009 Potential Reduced Device Longevity					13 + 1,861 = 1,874	1,874	99.9 +0.0/-0.0	99.7 +0.0/-0.0	99.4 +0.1/-0.1	98.0 +0.1/-0.1	89.5 +0.3/-0.3	70.6 +0.6/-0.6	46.9 +1.8/-1.9 at 77 mo				
D154VRC	Virtuoso DR	May-06	4,100	200	117	13 + 1,861 = 1,874	1,874	100.0 +0.0/-0.0	99.9 +0.1/-0.2	90.5 +1.0/-1.1	49.5 +1.9/-1.9	46.0 +1.9/-1.9 at 49 mo						
	Advisories: See page 139 – 2009 Potential Reduced Device Longevity							99.9 +0.1/-0.1	99.6 +0.2/-0.3	84.1 +1.2/-1.3	13.9 +1.5/-1.4	6.5 +1.2/-1.0 at 49 mo						
D154VRC	EnTrust VR	Jun-05	14,500	6,900	678	19 + 80 = 99	99	100.0 +0.0/-0.0	99.9 +0.0/-0.1	99.8 +0.1/-0.1	99.7 +0.1/-0.1	99.5 +0.1/-0.2	99.2 +0.2/-0.2	98.9 +0.2/-0.3	98.8 +0.2/-0.3	98.8 +0.2/-0.3	98.8 +0.2/-0.3	98.8 +0.2/-0.3
	Advisories: See page 136 – 2012 Potential Rapid Battery Depletion					(9) + (39) = (48) (advisory-related subset)	(48)	99.9 +0.0/-0.1	99.6 +0.1/-0.1	99.4 +0.1/-0.2	98.9 +0.2/-0.2	97.9 +0.3/-0.3	92.3 +0.6/-0.6	85.4 +0.8/-0.9	78.0 +1.8/-1.9 at 92 mo			
D154VWC D164VWC	Virtuoso VR	May-06	33,100	20,800	432	14 + 292 = 306	306	100.0 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.7 +0.1/-0.1	98.6 +0.2/-0.2	98.5 +0.2/-0.2	98.5 +0.2/-0.2	98.5 +0.2/-0.2	98.5 +0.2/-0.2	98.5 +0.2/-0.2	98.5 +0.2/-0.2
	Advisories: See page 139 – 2009 Potential Reduced Device Longevity					(0) + (4) = (4) (advisory-related subset)	(4)	99.9 +0.0/-0.0	99.8 +0.0/-0.1	99.6 +0.1/-0.1	99.1 +0.1/-0.1	96.7 +0.2/-0.3	92.1 +0.5/-0.5	87.4 +1.3/-1.5 at 79 mo				

continued

Device Survival Summary continued

Model Number	Family	US Market Release	Registered US Implants	Estimated Active US Implants	Normal Battery Depletions	Malfunctions		Device Survival Probability (%)								
						Therapy Function Compromised	Therapy Function Not Compromised	Years After Implant								
								1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	
D224DRG, D234DRG, D204DRM, D214DRM	Secura DR	Sep-08	50,800	42,400	207	14	49	63	100.0 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.8 +0.0/-0.1	99.7 +0.1/-0.3 at 58 mo			
									99.8 +0.0/-0.0	99.6 +0.1/-0.1	99.3 +0.1/-0.1	98.4 +0.2/-0.2	97.1 +0.4/-0.5 at 58 mo			
									Excluding Normal Battery Depletion	Including Normal Battery Depletion	Excluding Normal Battery Depletion	Including Normal Battery Depletion	Excluding Normal Battery Depletion	Including Normal Battery Depletion		
D224VRC, D234VRC, D204VRM, D214VRM	Secura VR	Sep-08	20,600	17,300	38	5	10	15	100.0 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.1	99.9 +0.0/-0.1	99.9 +0.1/-0.1 at 58 mo			
									99.9 +0.0/-0.0	99.7 +0.1/-0.1	99.5 +0.1/-0.1	99.4 +0.1/-0.2	99.4 +0.1/-0.2 at 58 mo			
									Excluding Normal Battery Depletion	Including Normal Battery Depletion	Excluding Normal Battery Depletion	Including Normal Battery Depletion	Excluding Normal Battery Depletion	Including Normal Battery Depletion		
D274DRG, D294DRG	Virtuoso II DR	Aug-09	22,200	18,800	38	1	2	3	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0 at 47 mo				
									99.9 +0.0/-0.0	99.9 +0.0/-0.1	99.6 +0.1/-0.1	99.4 +0.2/-0.2 at 47 mo				
									Excluding Normal Battery Depletion	Including Normal Battery Depletion	Excluding Normal Battery Depletion	Including Normal Battery Depletion	Excluding Normal Battery Depletion	Including Normal Battery Depletion		
D274VRC, D294VRC	Virtuoso II VR	Aug-09	9,100	7,800	11	0	3	3	100.0 +0.0/-0.0	100.0 +0.0/-0.0	99.9 +0.0/-0.1	99.9 +0.1/-0.1 at 46 mo				
									99.8 +0.1/-0.1	99.8 +0.1/-0.1	99.7 +0.1/-0.2	99.7 +0.1/-0.2 at 46 mo				
									Excluding Normal Battery Depletion	Including Normal Battery Depletion	Excluding Normal Battery Depletion	Including Normal Battery Depletion	Excluding Normal Battery Depletion	Including Normal Battery Depletion		
D284DRG, D264DRM	Maximo II DR	Sep-08	19,700	16,300	84	5	15	20	100.0 +0.0/-0.0	100.0 +0.0/-0.0	99.9 +0.0/-0.1	99.8 +0.1/-0.1	99.8 +0.1/-0.2 at 58 mo			
									100.0 +0.0/-0.0	99.8 +0.1/-0.1	99.4 +0.1/-0.1	98.5 +0.3/-0.3	96.7 +0.7/-1.0 at 58 mo			
									Excluding Normal Battery Depletion	Including Normal Battery Depletion	Excluding Normal Battery Depletion	Including Normal Battery Depletion	Excluding Normal Battery Depletion	Including Normal Battery Depletion		
D284VRC, D264VRM	Maximo II VR	Sep-08	12,700	10,700	27	3	6	9	100.0 +0.0/-0.0	100.0 +0.0/-0.0	99.9 +0.0/-0.1	99.9 +0.0/-0.1 at 58 mo				
									99.9 +0.0/-0.1	99.7 +0.1/-0.1	99.5 +0.1/-0.2	99.4 +0.2/-0.2	99.3 +0.2/-0.3 at 58 mo			
									Excluding Normal Battery Depletion	Including Normal Battery Depletion	Excluding Normal Battery Depletion	Including Normal Battery Depletion	Excluding Normal Battery Depletion	Including Normal Battery Depletion		

continued

Device Survival Summary continued

Model Number	Family	US Market Release	Registered US Implants	Estimated Active US Implants	Normal Battery Depletions	Malfunctions			Device Survival Probability (%)								
						Therapy Function Compromised	Therapy Function Not Compromised	Total	1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	
D314DRG, D354DRG, D314DRM, D354DRM, D334DRG, D364DRG, D334DRM, D364DRM, Protecta/ Protecta XTDR	Protecta XTDR/ DR	Mar-11	54,700	52,300	29	5	13	=	18	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0 at 28 mo					
D314VRG, D354VRG, D314VRM, D354VRM, D334VRG, D364VRG, D334VRM, D364VRM, Protecta/ Protecta XTVR	Protecta XTVR/ VR	Mar-11	25,300	24,200	8	2	3	=	5	99.9 +0.0/-0.0	99.7 +0.1/-0.1	99.6 +0.1/-0.2 at 28 mo					
										100.0 +0.0/-0.0	99.9 +0.0/-0.1	99.9 +0.0/-0.1 at 28 mo					
										99.9 +0.0/-0.0	99.8 +0.1/-0.1	99.8 +0.1/-0.1 at 28 mo					

continued

Reference Chart

The longevity estimates provided are mean values calculated for the parameters given. The actual longevity achieved for any device while implanted will depend on the actual programmed parameters and patient factors, and may differ significantly from these estimates.

Model Number	Family	Volume/ Mass*	Delivered Energy	Estimated Longevity				Elective Replacement (ERI)***		End of Life (EOL) Battery Voltage	
				Charging Frequency**	100% Pacing#	50% Pacing#	15% Pacing#	100% Sensing	Battery Voltage		Charge Time
7230	Marquis VR	36 cc 75 g	30 J	Monthly	4.9	5.2	5.4	5.5	≤ 2.62 V	> 16-second charge time	3 months after ERI
				Quarterly	7.3	8.0	8.5	8.7			
				Semiannual	8.5	9.3	10.0	10.4			
7231Cx	GEM III VR	39 cc 77 g	30 J	Monthly	4.3	4.7	5.0	5.2	≤ 2.55 V	—	≤ 2.40 V
				Quarterly	6.0	6.8	7.4	7.8			
				Semiannual	6.6	7.5	8.5	8.9			
7232 Cx, E	Maximo VR	39 cc 76 g	35 J	Monthly	4.4	4.7	4.8	4.9	≤ 2.62 V	> 16-second charge time	3 months after ERI
				Quarterly	7.0	7.5	8.0	8.3			
				Semiannual	8.2	9.0	9.7	10.0			
7274	Marquis DR	36 cc 75 g	30 J	Monthly	4.0	4.4	4.8	4.9	≤ 2.62 V	> 16-second charge time	3 months after ERI
				Quarterly	5.6	6.4	7.1	7.5			
				Semiannual	6.2	7.2	8.1	8.6			
7278	Maximo DR	39 cc 77 g	35 J	Monthly	3.7	4.1	4.3	4.5	≤ 2.62 V	> 16-second charge time	3 months after ERI
				Quarterly	5.3	6.1	6.8	7.1			
				Semiannual	6.0	7.0	8.0	8.5			
7288	Intrinsic	38 cc 76 g	35 J	Monthly	3.7	4.1	4.3	4.5	≤ 2.62 V	> 16-second charge time	3 months after ERI
				Quarterly	5.3	6.1	6.8	7.1			
				Semiannual	6.0	7.0	8.0	8.5			

* Volume and mass differ by connector style.

** A full charge is a full energy therapeutic shock or capacitor reformation.

*** The minimum time between ERI and EOL is 3 months (100% pacing, two charges per month).

‡ Pacing mode is VVI for single chamber models and DDD for dual chamber models. Parameter settings; lower rate at 60 ppm, sensing rate at 70 bpm, (A, RV, LV) 3.0 V amplitude, 0.4 ms pulse width, and 510-ohm pace load per applicable channel.



Reference Chart continued

Model Number	Family	Volume/ Mass*	Delivered Energy	Charging Frequency**	Estimated Longevity				Recommended Replacement (RRT)***		End of Service (EOS)
					100% Pacing†	50% Pacing†	15% Pacing†	100% Sensing	Battery Voltage	Charge Time	
D154ATG, D154DRG	EnTrust	35 cc 68 g	35 J	Monthly	3.8	4.2	4.4	4.6	≤ 2.61 V	—	3 months after RRT or > 16-second charge time
				Quarterly	5.5	6.1	6.8	7.0			
				Semiannual	6.1	7.0	7.9	8.3			
D154AWG, D164AWG	Virtuoso DR	37 cc 68 g	35 J	Monthly	4.1	4.5	4.8	5.0	≤ 2.62 V	—	3 months after RRT or > 16-second charge time
				Quarterly	6.3	7.3	8.3	8.8			
				Semiannual	7.3	8.7	10.1	11.0			
D154VRC	EnTrust VR	35 cc 68 g	35 J	Monthly	4.8	5.0	5.2	5.3	≤ 2.61 V	—	3 months after RRT or > 16-second charge time
				Quarterly	7.5	8.3	8.8	9.0			
				Semiannual	9.0	10.0	10.7	11.0			
D154VWC, D164VWC	Virtuoso	37 cc 68 g	35 J	Monthly	4.8	5.1	5.3	5.4	≤ 2.62 V	—	3 months after RRT or > 16-second charge time
				Quarterly	8.1	9.0	9.6	10.0			
				Semiannual	10.0	11.2	12.3	12.9			
D224DRG, D234DRG, D204DRM, D214DRM	Secura DR/ DR M4	37 cc 68 g	35 J	Monthly	3.6	4.1	4.5	4.7	≤ 2.63 V	—	3 months after RRT or > 16-second charge time
				Quarterly	5.1	6.1	7.0	7.5			
				Semiannual	5.7	7.0	8.3	9.0			
D224VRC, D234VRC, D204VRM, D214VRM	Secura VR/ VR M4	37 cc 68 g	35 J	Monthly	4.3	4.7	4.9	5.0	≤ 2.63 V	—	3 months after RRT or > 19-second charge time
				Quarterly	6.7	7.7	8.1	8.4			
				Semiannual	7.8	8.9	9.8	10.3			
D274DRG, D294DRG	Virtuoso II DR	37 cc 68 g	35 J	Monthly	3.6	4.1	4.5	4.7	≤ 2.63 V	—	3 months after RRT or > 16-second charge time
				Quarterly	5.1	6.0	7.0	7.5			
				Semiannual	5.7	7.0	8.3	9.0			
D274VRC, D294VRC	Virtuoso II VR	37 cc 68 g	35 J	Monthly	4.3	4.7	4.9	5.0	≤ 2.63 V	—	3 months after RRT or > 19-second charge time
				Quarterly	6.7	7.5	8.1	8.4			
				Semiannual	7.8	8.9	9.8	10.3			
D284DRG, D264DRM	Maximo II DR/DR M4	37 cc 68 g	35 J	Monthly	3.6	4.1	4.5	4.6	≤ 2.63 V	—	3 months after RRT or > 16-second charge time
				Quarterly	5.1	6.1	7.0	7.5			
				Semiannual	5.7	7.0	8.2	9.0			
D284VRC, D264VRM	Maximo II VR/ VR M4	37 cc 68 g	35 J	Monthly	4.3	4.6	4.9	5.0	≤ 2.63 V	—	3 months after RRT or > 19-second charge time
				Quarterly	6.6	7.4	8.1	8.4			
				Semiannual	7.7	8.8	9.7	10.2			
D314DRG, D354DRG, D314DRM, D354DRM	Protecta XT DR/DR M4	37 cc 68 g	35 J	Monthly	3.6	4.1	4.5	4.7	≤ 2.63 V	—	3 months after RRT
				Quarterly	5.1	6.0	7.0	7.5			
				Semiannual	5.7	7.0	8.3	9.0			
D314VRG, D354VRG, D314VRM, D354VRM	Protecta XT VR/VR M4	37 cc 68 g	35 J	Monthly	4.3	4.7	4.9	5.0	≤ 2.63 V	—	3 months after RRT
				Quarterly	6.7	7.5	8.1	8.4			
				Semiannual	7.8	8.9	9.8	10.3			
D334DRG, D364DRG, D334DRM, D364DRM	Protecta DR/ DR M4	37 cc 68 g	35 J	Monthly	3.6	4.1	4.5	4.7	≤ 2.63 V	—	3 months after RRT
				Quarterly	5.1	6.0	7.0	7.5			
				Semiannual	5.7	7.0	8.3	9.0			
D334VRG, D364VRG, D334VRM, D364VRM	Protecta VR/ VR M4	37 cc 68 g	35 J	Monthly	4.3	4.7	4.9	5.0	≤ 2.63 V	—	3 months after RRT
				Quarterly	6.7	7.5	8.1	8.4			
				Semiannual	7.8	8.9	9.8	10.3			

* Volume and mass differ by connector style.

** A full charge is a full energy therapeutic shock or capacitor reformation.


*** The minimum time between RRT and EOS is 3 months (100% pacing, two charges per month).

† Pacing mode is VVI for single chamber models and DDD for dual chamber models. Parameter settings; lower rate at 60 ppm, sensing rate at 70 bpm, (A, RV, LV) 3.0 V amplitude, 0.4 ms pulse width, and 510-ohm pace load per applicable channel.

Adapta DR ADDR01, ADDR03, ADDR06, ADD01

Product Characteristics


US Market Release	Jul-06	Malfunctions (US)	60	NBG Code	DDD, DDDR
Registered US Implants	361,100	Therapy Function Not Compromised	38	Serial Number Prefix	NWB, NWC, NWD, NWF, PWB, PWC, PWD, PWF
Estimated Active US Implants	295,800	Electrical Component	36		
Normal Battery Depletions (US)	570	Electrical Interconnect	1		
Performance Note: See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI		Other	1	Estimated Longevity	See page 70
		Therapy Function Compromised	22		
		Electrical Component	18		
		Electrical Interconnect	2		
		Other	2		



Adapta DR ADDRL1

Product Characteristics


US Market Release	Jul-06	Malfunctions (US)	9	NBG Code	DDDR
Registered US Implants	81,800	Therapy Function Not Compromised	6	Serial Number Prefix	PWE, NWE
Estimated Active US Implants	73,700	Electrical Component	5	Estimated Longevity	See page 70
Normal Battery Depletions (US)	36	Electrical Interconnect	1		
Performance Note: See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI		Therapy Function Compromised	3		
		Electrical Interconnect	1		
		Other	2		



Adapta DR ADDR1

Product Characteristics


US Market Release	Jul-06	Malfunctions (US)	10	NBG Code	DDDR
Registered US Implants	35,200	Therapy Function Not Compromised	6	Serial Number Prefix	PWA, NWA
Estimated Active US Implants	26,100	Electrical Component	5	Estimated Longevity	See page 70
Normal Battery Depletions (US)	514	Possible Early Battery Depletion	1		
Performance Note: See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI		Therapy Function Compromised	4		
		Electrical Component	2		
		Other	2		



Adapta SR ADSR01, ADSR03, ADSR06

Product Characteristics


US Market Release	Jul-06	Malfunctions (US)	9	NBG Code	SSIR
Registered US Implants	69,000	Therapy Function Not Compromised	5	Serial Number Prefix	NWN, NWM, NWP, PWP, PWM, PWN
Estimated Active US Implants	48,700	Electrical Component	3	Estimated Longevity	See page 70
Normal Battery Depletions (US)	236	Electrical Interconnect	1		
		Possible Early Battery Depletion	1		
Advisories	None	Therapy Function Compromised	4		
		Electrical Component	3		
		Electrical Interconnect	1		



Adapta VDD ADVDD01

Product Characteristics


US Market Release	Jul-06	Malfunctions (US)	0	NBG Code	VDD
Registered US Implants	1,100	Therapy Function Not Compromised	0	Serial Number Prefix	PWG, NWG
Estimated Active US Implants	800	Therapy Function Compromised	0	Estimated Longevity	See page 70
Normal Battery Depletions (US)	2				
Performance Note: See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI					



Advisa DR / DR MRI A2DR01, A3DR01, A4DR01, A5DR01 Ensura MRI EN1DR01

Product Characteristics


US Market Release	Apr-11	Malfunctions (US)	0	NBG Code	OAE - DDDR OOE - DDDR
Registered US Implants	8,500	Therapy Function Not Compromised	0	Serial Number Prefix	PZK, PZJ, PZL, PZW, PVY
Estimated Active US Implants	8,400	Therapy Function Compromised	0	Estimated Longevity	See page 70
Normal Battery Depletions (US)	0				
Performance Note: See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI					



EnPulse DR E1DR01, E1DR03, E1DR06

Product Characteristics


US Market Release	Dec-03	Malfunctions (US)	1	NBG Code	DDDR
Registered US Implants	6,800	Therapy Function Not Compromised	1	Serial Number Prefix	PRA, PRB, PRE
Estimated Active US Implants	1,200	Electrical Component	1	Estimated Longevity	See page 70
Normal Battery Depletions (US)	1,298	Therapy Function Compromised	0		
Performance Note: See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI					



EnPulse DR E1DR21

Product Characteristics


US Market Release	Dec-03	Malfunctions (US)	0	NBG Code	DDDR
Registered US Implants	1,900	Therapy Function Not Compromised	0	Serial Number Prefix	PPT
Estimated Active US Implants	150	Therapy Function Compromised	0	Estimated Longevity	See page 70
Normal Battery Depletions (US)	374				
Performance Note: See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI					



EnPulse 2 DR E2DR01, E2DR03, E2DR06, E2D01, E2D03

Product Characteristics

US Market Release	Feb-04	Malfunctions (US)	26	NBG Code	DDD, DDDR
Registered US Implants	101,000	Therapy Function Not Compromised	20	Serial Number Prefix	PNB, PNC, PNF, PNG, PNH
Estimated Active US Implants	40,300	Electrical Component	17	Estimated Longevity	See page 70
Normal Battery Depletions (US)	8,189	Possible Early Battery Depletion	2		
Performance Note: See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI		Other	1		
		Therapy Function Compromised	6		
		Electrical Component	3		
		Electrical Interconnect	2		
		Battery Malfunction	1		




	Years After Implant									
	1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	at 103 mo	
% ———	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0
% - - - -	99.9	99.9	99.8	99.4	98.7	97.1	91.4	70.9	34.3	
#	94,800	87,700	80,600	73,800	67,200	60,700	48,300	16,800	1,190	
Effective Sample Size										



EnPulse 2 DR E2DR21

Product Characteristics

US Market Release	Feb-04	Malfunctions (US)	1	NBG Code	DDDR
Registered US Implants	12,200	Therapy Function Not Compromised	0	Serial Number Prefix	PMU
Estimated Active US Implants	1,900	Therapy Function Compromised	1	Estimated Longevity	See page 70
Normal Battery Depletions (US)	2,074	Electrical Component	1		
Performance Note: See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI					




	Years After Implant								
	1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	at 85 mo	
%	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0	
%	99.9	99.5	98.9	97.1	91.0	67.7	22.5	14.5	
#	10,900	9,800	8,800	7,700	6,300	3,700	500	210	
Effective Sample Size									

EnPulse 2 DR E2DR31, E2DR33

Product Characteristics

US Market Release	Feb-04	Malfunctions (US)	0	NBG Code	DDDR
Registered US Implants	600	Therapy Function Not Compromised	0	Serial Number Prefix	PNL, PNM
Estimated Active US Implants	400	Therapy Function Compromised	0	Estimated Longevity	See page 70
Normal Battery Depletions (US)	10				
Performance Note: See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI					




	Years After Implant									
	1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	at 104 mo	
%	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0	
%	100.0	100.0	100.0	100.0	100.0	99.7	98.7	98.2	98.2	
#	1,400	1,400	1,300	1,300	1,200	1,100	1,000	400	110	
Effective Sample Size										

EnPulse 2 SR E2SR01, E2SR03, E2SR06

Product Characteristics


US Market Release	Dec-03	Malfunctions (US)	4	NBG Code	SSIR
Registered US Implants	25,500	Therapy Function Not Compromised	3	Serial Number Prefix	PMW, PMY, PNA
Estimated Active US Implants	6,400	Electrical Component	2	Estimated Longevity	See page 70
Normal Battery Depletions (US)	1,754	Possible Early Battery Depletion	1		
Advisories	None	Therapy Function Compromised	1		
		Other	1		



EnPulse 2 VDD E2VDD01

Product Characteristics


US Market Release	Dec-03	Malfunctions (US)	0	NBG Code	VDD
Registered US Implants	600	Therapy Function Not Compromised	0	Serial Number Prefix	PMV
Estimated Active US Implants	200	Therapy Function Compromised	0	Estimated Longevity	See page 70
Normal Battery Depletions (US)	73				
Performance Note: See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI					



EnRhythm DR P1501DR

Product Characteristics


US Market Release	May-05	Malfunctions (US)	7,445	NBG Code	DDDRP
Registered US Implants	110,200	Therapy Function Not Compromised	7,392	Serial Number Prefix	PNP
Estimated Active US Implants	63,600	Battery Malfunction (229 malfunctions due to advisory)	7,307	Estimated Longevity	See page 70
Normal Battery Depletions (US)	1,651	Electrical Component (2 malfunctions due to advisory)	38		
Advisories: See page 137 – 2010 Low Battery Voltage Displayed at Device Interrogation		Possible Early Battery Depletion	45		
		Electrical Interconnect	2		
		Therapy Function Compromised	53		
		Electrical Component	37		
		Electrical Interconnect	4		
		Battery Malfunction	5		
		Possible Early Battery Depletion	2		
		Other	5		



EnRhythm MRI EMDR01

Product Characteristics


US Market Release	N/A	Malfunctions (US)	10	NBG Code	DDDRP
Registered US Implants	100	Therapy Function Not Compromised	10	Serial Number Prefix	PTA
Estimated Active US Implants	80	Battery Malfunction	10	Estimated Longevity	See page 70
Normal Battery Depletions (US)	0	Therapy Function Compromised	0		
Advisories	None				



Kappa 400 DR KDR401, KDR403

Product Characteristics

US Market Release	Jan-98	Malfunctions (US)	27	NBG Code	DDDR
Registered US Implants	46,700	Therapy Function Not Compromised	15	Serial Number Prefix	PER, PET
Estimated Active US Implants	4,000	Electrical Component	10	Estimated Longevity	See page 71
Normal Battery Depletions (US)	7,981	Electrical Interconnect	1		
Advisories	None	Possible Early Battery Depletion	2		
		Other	2		
		Therapy Function Compromised	12		
		Electrical Component	6		
		Electrical Interconnect	6		




Years After Implant	Excluding Normal Battery Depletion		Including Normal Battery Depletion							
	1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	at 107 mo	
%	100.0	100.0	100.0	100.0	99.9	99.9	99.9	99.9	99.9	
%	99.9	99.8	99.7	99.4	98.8	97.1	88.1	58.8	7.8	
#	44,200	41,100	37,900	34,800	31,500	27,700	21,300	9,800	670	
Effective Sample Size										



Kappa 400 SR KSR401, KSR403

Product Characteristics

US Market Release	Feb-98	Malfunctions (US)	5	NBG Code	SSIR
Registered US Implants	15,400	Therapy Function Not Compromised	4	Serial Number Prefix	PEU, PGD
Estimated Active US Implants	1,500	Electrical Component	3	Estimated Longevity	See page 71
Normal Battery Depletions (US)	1,547	Possible Early Battery Depletion	1		
Advisories	None	Therapy Function Compromised	1		
		Electrical Interconnect	1		



Kappa 700 DR KD701, KD703, KD706


Product Characteristics

US Market Release	Jan-99	Malfunctions (US)	0	NBG Code	DDD
Registered US Implants	300	Therapy Function Not Compromised	0	Serial Number Prefix	PHK, PHM, PHL
Estimated Active US Implants	70	Therapy Function Compromised	0	Estimated Longevity	See page 71
Normal Battery Depletions (US)	21				

Advisories: [See page 140](#) – 2009 Potential Separation of Interconnect Wires

Performance Note: [See page 146](#) –


Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI



Kappa 700 DR KDR701, KDR703, KDR706

Product Characteristics

US Market Release	Jan-99	Malfunctions (US)	744	NBG Code	DDD, DDD/RO
Registered US Implants	206,200	Therapy Function Not Compromised	53	Serial Number Prefix	PGT, PGU, PGW, PGY, PHJ
Estimated Active US Implants	22,200	Electrical Component	26		
Normal Battery Depletions (US)	35,745	Electrical Interconnect	19	Estimated Longevity	See page 71
Advisories: See page 140 – 2009 Potential Separation of Interconnect Wires		Battery Malfunction	1		
		Possible Early Battery Depletion	4		
Performance Note: See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI		Other	3		
		Therapy Function Compromised	691		
		Electrical Component	17		
		Electrical Interconnect	673		
		(206 malfunctions due to advisory)			
		Possible Early Battery Depletion	1		




	Years After Implant									
	1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	9 yr	at 109 mo
%	100.0	100.0	99.9	99.9	99.9	99.9	99.7	99.6	99.4	99.4
%	99.9	99.8	99.6	99.1	98.0	95.4	85.9	59.3	12.4	3.2
#	180,500	165,500	151,000	136,600	122,400	107,000	83,000	41,900	3,700	1,600
Effective Sample Size										



Kappa 700 DR KDR721

Product Characteristics


US Market Release	Feb-99	Malfunctions (US)	5	NBG Code	DDD/RO
Registered US Implants	9,800	Therapy Function Not Compromised	1	Serial Number Prefix	PGR
Estimated Active US Implants	700	Electrical Component	1	Estimated Longevity	See page 71
Normal Battery Depletions (US)	1,349	Therapy Function Compromised	4		
		Electrical Interconnect	4		
Advisories: See page 140 – 2009 Potential Separation of Interconnect Wires					
Performance Note: See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI					



Kappa 700 SR KSR701, KSR703, KSR706

Product Characteristics

US Market Release	Jan-99	Malfunctions (US)	28	NBG Code	SSIR
Registered US Implants	55,200	Therapy Function Not Compromised	4	Serial Number Prefix	PHR, PHT, PHU, PHW
Estimated Active US Implants	5,700	Electrical Component	2	Estimated Longevity	See page 71
Normal Battery Depletions (US)	5,272	Electrical Interconnect	1		
Advisories: See page 140 – 2009 Potential Separation of Interconnect Wires		Possible Early Battery Depletion	1		
		Therapy Function Compromised	24		
		Electrical Component	4		
		Electrical Interconnect	20		




Years After Implant	Excluding Normal Battery Depletion										Including Normal Battery Depletion									
	1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	9 yr	at 116 mo	1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	9 yr	at 116 mo
%	100.0	100.0	100.0	100.0	100.0	100.0	99.9	99.9	99.8	99.8	99.9	99.7	99.3	98.5	97.0	93.5	84.2	61.5	29.5	8.5
#	48,300	41,700	35,700	30,600	25,900	21,400	15,800	7,900	1,800	130	48,300	41,700	35,700	30,600	25,900	21,400	15,800	7,900	1,800	130

Effective Sample Size

Kappa 900 DR KDR901, KDR903, KDR906

Product Characteristics

US Market Release	Jan-02	Malfunctions (US)	74	NBG Code	DDD, DDDR
Registered US Implants	125,600	Therapy Function Not Compromised	21	Serial Number Prefix	PKM, PKN, PKP, PLB, PLC, PLD
Estimated Active US Implants	22,900	Electrical Component	16	Estimated Longevity	See page 71
Normal Battery Depletions (US)	21,661	Electrical Interconnect	4		
Advisories	None	Other	1		
Performance Note: See page 146 –		Therapy Function Compromised	53		
Performance note on Dual Chamber Pacemakers with Measurement Lock-up ER		Electrical Component	9		
		Electrical Interconnect	44		




	Years After Implant									
	1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	at 107 mo	
%	100.0	100.0	100.0	100.0	100.0	100.0	99.9	99.9	99.9	
%	99.9	99.9	99.7	99.3	98.4	96.2	87.9	60.4	11.8	
#	117,400	108,100	99,000	90,200	81,500	72,500	57,900	28,000	2,400	
Effective Sample Size										

Kappa 900 SR KSR901, KSR903, KSR906

Product Characteristics


US Market Release	Jan-02	Malfunctions (US)	17	NBG Code	SSIR
Registered US Implants	37,000	Therapy Function Not Compromised	8	Serial Number Prefix	PLF, PLG, PLH
Estimated Active US Implants	5,700	Electrical Component	7	Estimated Longevity	See page 71
Normal Battery Depletions (US)	3,480	Possible Early Battery Depletion	1		
Advisories: See page 140 – 2009 Potential Separation of Interconnect Wires		Therapy Function Compromised	9		
		Electrical Interconnect	9		



Kappa 900 VDD KVDD901

Product Characteristics


US Market Release	Jan-02	Malfunctions (US)	1	NBG Code	VDD
Registered US Implants	700	Therapy Function Not Compromised	1	Serial Number Prefix	PLE
Estimated Active US Implants	100	Other	1	Estimated Longevity	See page 71
Normal Battery Depletions (US)	82	Therapy Function Compromised	0		
Advisories: See page 140 – 2009 Potential Separation of Interconnect Wires					
Performance Note: See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI					



Kappa 920 DR KDR921

Product Characteristics


US Market Release	Jan-02	Malfunctions (US)	4	NBG Code	DDDR
Registered US Implants	16,300	Therapy Function Not Compromised	1	Serial Number Prefix	PKR
Estimated Active US Implants	1,400	Electrical Component	1	Estimated Longevity	See page 71
Normal Battery Depletions (US)	2,823	Therapy Function Compromised	3		
Advisories: See page 140 – 2009 Potential Separation of Interconnect Wires		Electrical Interconnect	3		
Performance Note: See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ER					



Revo MRI SureScan RVDR01

Product Characteristics


US Market Release	Feb-11	Malfunctions (US)	5	NBG Code	DDDRP
Registered US Implants	58,700	Therapy Function Not Compromised	4	Serial Number Prefix	PTN
Estimated Active US Implants	56,600	Electrical Component	4	Estimated Longevity	See page 71
Normal Battery Depletions (US)	2	Therapy Function Compromised	1		
Advisories	None	Electrical Component	1		



Sensia DR SEDR01, SED01

Product Characteristics


US Market Release	Jul-06	Malfunctions (US)	24	NBG Code	DDD, DDDR
Registered US Implants	124,600	Therapy Function Not Compromised	14	Serial Number Prefix	PWL, PWK, NWL, NWK
Estimated Active US Implants	93,600	Electrical Component	12	Estimated Longevity	See page 71
Normal Battery Depletions (US)	250	Electrical Interconnect	1		
Performance Note: See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ER		Other	1		
		Therapy Function Compromised	10		
		Electrical Component	4		
		Electrical Interconnect	1		
		Other	5		



Sensia SR SESR01, SES01

Product Characteristics

US Market Release	Jul-06	Malfunctions (US)	8	NBG Code	SSIR, SSI
Registered US Implants	87,800	Therapy Function Not Compromised	6	Serial Number Prefix	PWR, PWS, NWR, NWS
Estimated Active US Implants	61,200	Electrical Component	5	Estimated Longevity	See page 71
Normal Battery Depletions (US)	198	Other	1		
Advisories	None	Therapy Function Compromised	2		
		Electrical Component	1		
		Electrical Interconnect	1		




Sigma 100 S SS103, SS106

Product Characteristics

US Market Release	Aug-99	Malfunctions (US)	0	NBG Code	SSI
Registered US Implants	800	Therapy Function Not Compromised	0	Serial Number Prefix	PJG, PJH
Estimated Active US Implants	100	Therapy Function Compromised	0	Estimated Longevity	See page 71
Normal Battery Depletions (US)	26				

Advisories: [See page 144](#) – 2005 Potential Separation of Interconnect Wires




Sigma 200 DR SDR203

Product Characteristics

US Market Release	Aug-99	Malfunctions (US)	36	NBG Code	DDD/RO
Registered US Implants	15,900	Therapy Function Not Compromised	10	Serial Number Prefix	PJD
Estimated Active US Implants	3,100	Electrical Component	1	Estimated Longevity	See page 72
Normal Battery Depletions (US)	915	Electrical Interconnect (1 malfunction due to advisory)	9		

Advisories: [See page 144](#) – 2005 Potential Separation of Interconnect Wires; [See also page 140](#) – 2009 Potential Separation of Interconnect Wires

Therapy Function Compromised	26
Electrical Component	2
Electrical Interconnect (20 malfunctions due to advisory)	23
Other	1




Sigma 200 SR SSR203, SS203

Product Characteristics

US Market Release	Sep-99	Malfunctions (US)	14	NBG Code	SSIR, SSI
Registered US Implants	12,100	Therapy Function Not Compromised	0	Serial Number Prefix	PJG
Estimated Active US Implants	1,700	Therapy Function Compromised	14	Estimated Longevity	See page 72
Normal Battery Depletions (US)	434	Electrical Interconnect <i>(13 malfunctions due to advisory)</i>	14		


Advisories: [See page 144](#) – 2005 Potential Separation of Interconnect Wires; [See also page 140](#) – 2009 Potential Separation of Interconnect Wires



Sigma 300 DR SDR303, SDR306

Product Characteristics

US Market Release	Aug-99	Malfunctions (US)	249	NBG Code	DDD/RO, DDD
Registered US Implants	106,900	Therapy Function Not Compromised	53	Serial Number Prefix	PJD, PJE
Estimated Active US Implants	28,400	Electrical Component	7	Estimated Longevity	See page 72
Normal Battery Depletions (US)	4,553	Electrical Interconnect <i>(4 malfunctions due to advisory)</i>	44		
Advisories: See page 144 – 2005 Potential Separation of Interconnect Wires; See also page 140 – 2009 Potential Separation of Interconnect Wires		Possible Early Battery Depletion	1		
		Other	1		
		Therapy Function Compromised	196		
		Electrical Component	7		
		Electrical Interconnect <i>(123 malfunctions due to advisory)</i>	188		
		Other	1		




	1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	9 yr	10 yr	11 yr	12 yr	at 146 mo
% — Excluding Normal Battery Depletion	100.0	100.0	100.0	99.9	99.9	99.9	99.8	99.6	99.5	99.4	99.4	99.4	99.4
% - - - Including Normal Battery Depletion	99.9	99.9	99.8	99.6	99.3	98.7	97.7	95.1	90.9	83.6	69.7	43.5	18.2
# Effective Sample Size	96,600	86,500	77,500	69,100	61,400	54,000	45,200	35,500	26,300	16,600	7,900	1,100	200

Sigma 300 SR SSR303, SSR306

Product Characteristics

US Market Release	Aug-99	Malfunctions (US)	57	NBG Code	SSIR, SSI
Registered US Implants	54,200	Therapy Function Not Compromised	15	Serial Number Prefix	PJG, PJH
Estimated Active US Implants	10,300	Electrical Component	1	Estimated Longevity	See page 72
Normal Battery Depletions (US)	1,420	Electrical Interconnect <i>(3 malfunctions due to advisory)</i>	12		
Advisories: See page 144 – 2005 Potential Separation of Interconnect Wires; See also page 140 – 2009 Potential Separation of Interconnect Wires		Other	2		
		Therapy Function Compromised	42		
		Electrical Component	3		
		Electrical Interconnect <i>(20 malfunctions due to advisory)</i>	39		




Sigma 300 VDD SVDD303

Product Characteristics

US Market Release	Sep-99	Malfunctions (US)	1	NBG Code	VDDD
Registered US Implants	650	Therapy Function Not Compromised	0	Serial Number Prefix	PJD
Estimated Active US Implants	80	Therapy Function Compromised	1	Estimated Longevity	See page 72
Normal Battery Depletions (US)	71	Electrical Interconnect (1 malfunction due to advisory)	1		


Advisories: [See page 144](#) – 2005 Potential Separation of Interconnect Wires



Versa DR VEDR01

Product Characteristics

US Market Release	Jul-06	Malfunctions (US)	13	NBG Code	DDDR
Registered US Implants	97,400	Therapy Function Not Compromised	7	Serial Number Prefix	PWH, NWH
Estimated Active US Implants	73,700	Electrical Component	5	Estimated Longevity	See page 72
Normal Battery Depletions (US)	245	Electrical Interconnect	2		
Performance Note: See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI		Other	0		
		Therapy Function Compromised	6		
		Electrical Component	2		
		Other	4		



Device Survival Summary (95% Confidence Interval)
 The following table shows IPG device survival estimates with 95% confidence intervals. Estimates are shown both with and without normal battery depletions included.

Family	Model Number	US Market Release	Registered US Implants	Estimated Active US Implants	Normal Battery Depletions (US)	Malfunctions (US)		Device Survival Probability (%)													
						Therapy Function Compromised	Therapy Function Not Compromised	Years After Implant													
						1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	10 yr	12 yr	14 yr					
Adapta DR	ADDR01, ADDR03, ADDR06, ADD01 See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI	Jul-06	361,100	295,800	570	22	+ 38	= 60	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0				
						Excluding Normal Battery Depletion	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0
Adapta DR	ADDRL1 See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI	Jul-06	81,800	73,700	36	3	+ 6	= 9	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0				
						Excluding Normal Battery Depletion	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0
Adapta DR	ADDRS1 See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI	Jul-06	35,200	26,100	514	4	+ 6	= 10	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0				
						Excluding Normal Battery Depletion	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0
Adapta SR	ADSR01, ADSR03, ADSR06 See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI	Jul-06	69,000	48,700	236	4	+ 5	= 9	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0				
						Excluding Normal Battery Depletion	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0
Adapta VDD	ADVDD01 See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI	Jul-06	1,100	800	2	0	+ 0	= 0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0				
						Excluding Normal Battery Depletion	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0
Advista DR MRI Ensura	A2DR01, A3DR01, A4DR01, A5DR01, ENDR01 See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI	Apr-11	8,500	8,400	0	0	+ 0	= 0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0				
						Excluding Normal Battery Depletion	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0
						100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0				

continued



Device Survival Summary continued

		Malfunctions		Device Survival Probability (%)																			
Family	Model Number	US Market Release	Registered US Implants	Estimated Active US Implants	Normal Battery Depletions	Therapy Compromised		Therapy Function Not Compromised		Years After Implant													
						Therapy Compromised	Therapy Function Not Compromised	Total	1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	10 yr	12 yr	14 yr				
Kappa 700 DR	KDR701, KDR703, KDR706	Jan-99	206,200	22,200	33,745	691	+ 53 = 744	Excluding Normal Battery Depletion	100.0	100.0	99.9	99.9	99.9	99.9	99.9	99.6	99.4						
	See page 140 – 2009 Potential Separation of Interconnect Wires					(206)	+ (0) = (206)	Including Normal Battery Depletion	99.9	99.8	99.6	99.1	98.0	95.4	85.9	59.3	3.2						
See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI																							
Kappa 700 DR	KDR721	Feb-99	9,800	700	1,349	4	+ 1 = 5	Excluding Normal Battery Depletion	100.0	100.0	100.0	100.0	99.9	99.9	99.9	99.9							
	See page 140 – 2009 Potential Separation of Interconnect Wires					(0)	+ (0) = (0)	Including Normal Battery Depletion	99.9	99.5	98.7	96.6	90.7	68.8	22.2	12.4							
See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI																							
Kappa 700 SR	KSR701, KSR703, KSR706	Jan-99	55,300	5,700	5,273	24	+ 4 = 28	Excluding Normal Battery Depletion	100.0	100.0	100.0	100.0	100.0	100.0	99.9	99.9	99.8						
	See page 140 – 2009 Potential Separation of Interconnect Wires					(0)	+ (0) = (0)	Including Normal Battery Depletion	99.9	99.7	99.3	98.5	97.0	93.5	84.2	61.5	8.5						
See page 146 – Performance note on Dual Chamber Pacemakers with Measurement Lock-up ERI																							
(advisory-related subset)																							
(advisory-related subset)																							

continued

Family	Model Number	US Market Release	Registered US Implants	Estimated Active US Implants	Normal Battery Depletions	Malfunctions		Device Survival Probability (%)																			
						Therapy Function Compromised	Therapy Function Not Compromised	Years After Implant																			
							Total	1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	10 yr	12 yr	14 yr	16 yr								
Kappa 900 DR	KDR901, KDR903, KDR906	Jan-02	125,600	22,900	21,661	53	+ 21 = 74	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	99.9 +0.0/-0.0 at 107 mo	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0		
Kappa 900 SR	KSR901, KSR903, KSR906	Jan-02	37,000	5,700	3,480	9	+ 8 = 17	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	99.9 +0.0/-0.0 at 107 mo	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	
Kappa 900 VDD	KVDD901	Jan-02	650	75	82	0	+ 1 = 1	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	99.8 +0.2/-1.3 at 94 mo	99.8 +0.2/-1.3	99.8 +0.2/-1.3	99.8 +0.2/-1.3	99.8 +0.2/-1.3	99.8 +0.2/-1.3	99.8 +0.2/-1.3	99.8 +0.2/-1.3	99.8 +0.2/-1.3	99.8 +0.2/-1.3	99.8 +0.2/-1.3	99.8 +0.2/-1.3
Kappa 920 DR	KDR921	Jan-02	16,300	1,400	2,823	3	+ 1 = 4	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0
Revo MRI SureScan	RVDR01	Feb-11	58,700	56,600	2	1	+ 4 = 5	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0
Sensia DR	SEDR01, SED01	Jul-06	124,600	93,600	250	10	+ 14 = 24	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0
Sensia SR	SESR01, SES01	Jul-06	87,800	61,200	198	2	+ 6 = 8	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0

continued

Device Survival Summary continued

Family	Model Number	US Market Release	Registered US Implants	Estimated Active US Implants	Normal Battery Depletions	Malfunctions		Total	Device Survival Probability (%)												
						Therapy Function Compromised	Therapy Function Not Compromised		Years After Implant												
									1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	10 yr	12 yr	14 yr	16 yr	
Sigma 300 VDD	SVDD303	Sep-99	650	80	71	1	+ 0	= 1	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	99.8 +0.2/-0.9	99.2 +0.5/-1.2	98.8 +0.7/-1.8	98.8 +0.7/-1.8 at 127 mo			
						(1)	+ 0	= (1)	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	99.7 +0.2/-0.9	99.1 +0.5/-1.1	96.6 +1.2/-1.8	93.2 +1.8/-2.4	64.6 +4.8/-5.3	59.7 +5.3/-5.7 at 127 mo			
Versa DR	VEDR01	Jul-06	97,400	73,700	245	6	+ 7	= 13	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0	100.0 +0.0/-0.0 at 82 mo						
									100.0 +0.0/-0.0	99.9 +0.0/-0.0	99.9 +0.0/-0.0	99.7 +0.0/-0.1	99.2 +0.1/-0.1	98.2 +0.2/-0.2	97.0 +0.4/-0.4 at 82 mo						
							(advisory-related subset)														

Reference Chart

The longevity estimates provided are mean values calculated for the parameters given. The longevity estimates shown here assume a lower rate of 60 ppm, 100% pacing, and pulse width of 0.4 ms unless noted otherwise. The actual longevity achieved for any device while implanted will depend on the actual programmed parameters and patient factors, and may differ significantly from these estimates.

Family	Model Number	Amplitude Setting	Estimated Longevity	
			500 Lead Ω	1000 Lead Ω
Adapta DR	ADDR01, ADDR03, ADDR06, ADD01	Low 2.5 V (A, RV)	7.4	8.2
		Nominal 3.5 V (A, RV)	6.0	7.4
		High 5.0 V (A, RV)	4.5	6.0
Adapta DR	ADDRS1	Low 2.5 V (A, RV)	5.5	6.1
		Nominal 3.5 V (A, RV)	4.3	5.4
		High 5.0 V (A, RV)	3.2	4.4
Adapta DR	ADDRL1	Low 2.5 V (A, RV)	9.0	10.0
		Nominal 3.5 V (A, RV)	7.3	8.9
		High 5.0 V (A, RV)	5.4	7.2
Adapta SR	ADSR01, ADSR03, ADSR06	Low 2.5 V (RV)	7.4	7.9
		Nominal 3.5 V (RV)	6.5	7.5
		High 5.0 V (RV)	5.1	6.3
Adapta VDD	ADVDD01	Low 2.5 V (RV)	6.2	6.5
		Nominal 3.5 V (RV)	5.5	6.2
		High 5.0 V (RV)	4.4	5.4
Advisa DR	A4DR01, A5DR01	Low 2.5 V (A, RV)	5.5	6.1
		Nominal 3.5 V (A, RV)	4.3	5.4
		High 5.0 V (A, RV)	3.2	4.4
Advisa DR MRI+C82	A3DR01	Low 2.5 V (A, RV)	5.4	6.0
		Nominal 3.5 V (A, RV)	4.3	5.4
		High 5.0 V (A, RV)	3.0	4.2
EnPulse DR	E1DR01, E1DR03, E1DR06	Low 2.5 V (A, RV)	7.5	8.5
		Nominal 3.5 V (A, RV)	6.2	7.6
		High 5.0 V (A, RV)	4.4	5.9
EnPulse DR	E1DR21	Low 2.5 V (A, RV)	5.4	6.0
		Nominal 3.5 V (A, RV)	4.3	5.4
		High 5.0 V (A, RV)	3.0	4.2
EnPulse 2 DR	E2DR01, E2DR03, E2DR06, E2D01, E2D03	Low 2.5 V (A, RV)	7.5	8.5
		Nominal 3.5 V (A, RV)	6.2	7.6
		High 5.0 V (A, RV)	4.4	5.9
EnPulse 2 DR	E2DR21	Low 2.5 V (A, RV)	5.4	6.0
		Nominal 3.5 V (A, RV)	4.3	5.4
		High 5.0 V (A, RV)	3.0	4.2
EnPulse 2 DR	E2DR31, E2DR33	Low 2.5 V (A, RV)	9.0	10.1
		Nominal 3.5 V (A, RV)	7.4	9.1
		High 5.0 V (A, RV)	5.2	7.1
EnPulse 2 SR	E2SR01, E2SR03, E2SR06	Low 2.5 V (A, RV)	7.2	7.7
		Nominal 3.5 V (A, RV)	6.3	7.3
		High 5.0 V (A, RV)	4.8	6.1
EnPulse 2 VDD	E2VDD01	Low 2.5 V (RV)	6.1	6.5
		Nominal 3.5 V (RV)	5.5	6.2
		High 5.0 V (RV)	4.3	5.4
EnRhythm DR	P1501DR	Low 2.5 V (A, RV)	9.7	11.2
		Nominal 3.5 V (A, RV)	7.3	9.4
		High 5.0 V (A, RV)	4.9	7.1
EnRhythm MRI	EMDR01	Low 2.5 V (RV)	6.1	6.5
		Nominal 3.5 V (RV)	5.5	6.2
		High 5.0 V (RV)	4.3	5.4

continued

Reference Chart continued

Estimated Longevity

Family	Model Number	Amplitude Setting	500 Lead Ω	1000 Lead Ω
Kappa 400 DR	KDR401, KDR403	Low 2.5 V (A, RV)	7.8	8.5
		Nominal 3.5 V (A, RV)	6.4	7.5
		High 5.0 V (A, RV)	5.1	6.5
Kappa 400 SR	KSR401, KSR403	Low 2.5 V (RV)	7.9	8.4
		Nominal 3.5 V (RV)	6.9	7.7
		High 5.0 V (RV)	5.8	7.0
Kappa 700 D	KD701, KD703, KD706	Low 2.5 V (A, RV)	7.7	8.6
		Nominal 3.5 V (A, RV)	6.3	7.7
		High 5.0 V (A, RV)	4.4	6.0
Kappa 700 DR	KDR701, KDR703, KDR706	Low 2.5 V (A, RV)	7.7	8.6
		Nominal 3.5 V (A, RV)	6.3	7.7
		High 5.0 V (A, RV)	4.4	6.0
Kappa 700 DR	KDR721	Low 2.5 V (A, RV)	5.5	6.1
		Nominal 3.5 V (A, RV)	4.4	5.5
		High 5.0 V (A, RV)	3.0	4.2
Kappa 700 SR	KSR701, KSR703, KSR706	Low 2.5 V (RV)	7.4	7.9
		Nominal 3.5 V (RV)	6.5	7.5
		High 5.0 V (RV)	4.9	6.2
Kappa 900 DR	KDR901, KDR903, KDR906	Low 2.5 V (A, RV)	7.7	8.6
		Nominal 3.5 V (A, RV)	6.3	7.7
		High 5.0 V (A, RV)	4.4	6.0
Kappa 920 DR	KDR921	Low 2.5 V (A, RV)	5.5	6.1
		Nominal 3.5 V (A, RV)	4.4	5.5
		High 5.0 V (A, RV)	3.0	4.3
Kappa 900 SR	KSR901, KSR903, KSR906	Low 2.5 V (RV)	7.3	7.9
		Nominal 3.5 V (RV)	6.4	7.4
		High 5.0 V (RV)	4.9	6.1
Kappa 900 VDD	KVDD901	Low 2.5 V (RV)	6.2	6.6
		Nominal 3.5 V (RV)	5.6	6.3
		High 5.0 V (RV)	4.4	5.4
Revo MRI SureScan	RVDR01	Low 2.5 V (A, RV)	9.7	11.2
		Nominal 3.5 V (A, RV)	7.3	9.4
		High 5.0 V (A, RV)	4.9	7.1
Sensia DR	SED01, SED01	Low 2.5 V (A, RV)	7.4	8.2
		Nominal 3.5 V (A, RV)	6.0	7.4
		High 5.0 V (A, RV)	4.5	6.0
Sensia SR	SES01, SES01	Low 2.5 V (RV)	7.4	7.9
		Nominal 3.5 V (RV)	6.5	7.5
		High 5.0 V (RV)	5.1	6.3
Sigma 100 S	SS103, SS106	Low 2.5 V (RV)	10.1	11.1
		Nominal 3.5 V (RV)	8.2	9.8
		High 5.0 V (RV)	6.4	8.4

continued

Reference Chart continued

Family	Model Number	Amplitude Setting	Estimated Longevity	
			500 Lead Ω	1000 Lead Ω
Sigma 200 DR	SDR203	Low 2.5 V (RV)	10.1	11.1
		Nominal 3.5 V (RV)	8.2	9.8
		High 5.0 V (RV)	6.4	8.4
Sigma 200 SR	SSR203	Low 2.5 V (RV)	10.1	11.1
		Nominal 3.5 V (RV)	8.2	9.8
		High 5.0 V (RV)	6.4	8.4
Sigma 300 DR	SDR303, SDR306	Low 2.5 V (A, RV)	10.1	11.7
		Nominal 3.5 V (A, RV)	7.5	9.6
		High 5.0 V (A, RV)	5.5	7.8
Sigma 300 SR	SSR303, SSR306	Low 2.5 V (RV)	10.1	11.1
		Nominal 3.5 V (RV)	8.2	9.8
		High 5.0 V (RV)	6.4	8.4
Sigma 300 VDD	SVDD303	Low 2.5 V (RV)	8.9	9.7
		Nominal 3.5 V (RV)	7.3	8.6
		High 5.0 V (RV)	5.8	7.4
Versa DR	VEDR01	Low 2.5 V (A, RV)	7.4	8.2
		Nominal 3.5 V (A, RV)	6.0	7.4
		High 5.0 V (A, RV)	4.5	6.0

Method for Estimating Lead Performance

Medtronic CRDM has tracked lead survival for over 30 years with its multicenter, global chronic lead studies.

Leads Performance Analysis

Implanted leads operate in the challenging biochemical environment of the human body and the body's response to foreign objects. Implanted leads are also subject to mechanical stresses associated with heart motion, body motion, and patient anatomy.

In this environment, pacemaker and defibrillation leads cannot be expected to last forever. While IPGs and ICDs have a battery that will deplete after a predictable length of time, a lead's longevity cannot be predicted, nor are there simple indicators that a lead is approaching the end of its service life. Therefore, regular monitoring while implanted, and evaluation of lead integrity upon IPG or ICD replacement, is necessary to determine if a lead may be approaching the end of its service life.

Shortfalls of Using Returned Product and Complaints to Estimate Lead Performance

Leads and lead segments returned to Medtronic are analyzed to determine whether or not they meet performance limits established by Medtronic. Although returned product analyses are valuable for gaining insight into lead failure mechanisms, this data cannot be used by itself for determining the survival probability of leads because only a small fraction of leads are explanted and returned for analysis. Additionally, those leads that are returned cannot be assumed to be statistically representative of the performance of the total population for a given lead model. Partial or total lead extraction can result in significant damage to a lead, making a definitive analysis of a suspected failure, and its cause, impossible.

To account for the under reporting inherent with lead survival analysis based solely on returned product, some manufacturers add reported complaints where adverse product performance is evident but the product itself has not been returned. The improvement to the accuracy of survival estimates depends on the degree to which all complaints are actually communicated to the manufacturer. Since not all complaints are communicated to the manufacturer, adding complaints to the survival analysis does not completely solve the under reporting problem.

Thus, lead survival probabilities are more appropriately determined through a clinical surveillance study that includes active follow-up with the patients. Although Medtronic monitors returned product analysis and complaints, these are not used to determine lead survival estimates.

Medtronic tracks lead survival through its Product Surveillance Registry. The registry is designed to record clinical observations representative of the total clinical experience. Therefore, the lead survival estimates include both lead hardware failure and lead-related medical complications, and do not differentiate a lead hardware failure from other clinical events such as exit block, perforation, dislodgement, or concurrent pulse generator failure. The lead need not be returned to Medtronic.

Product Surveillance Registry (PSR)

Medtronic has been monitoring the performance of its cardiac therapy products with a multicenter study since 1983 and has evaluated the performance of more than 95,000 leads, with data reported from countries around the world. Throughout this time period, Medtronic has continually worked to adapt systems and processes to more effectively monitor product performance following market release. The following summarizes current registry requirements.

Medtronic's global Product Surveillance Registry has a prospective, non-randomized, observational design. A key purpose of the registry is to provide continuing evaluation and periodic reporting of the long-term reliability and performance of Medtronic market-released cardiac therapy products. Product-related adverse events, indicating the status of the product, are collected to measure survival probabilities. The data gathered may also be used to support the design and development of investigational plans for new cardiac therapy products. The registry is designed to continue indefinitely, encompassing new products as they become commercially available.

To ensure a sufficiently large and representative source of data, participating clinical sites must meet pre-specified selection criteria. The number of participants is regularly reviewed to ensure the necessary capacity to meet Medtronic's ongoing prospective post-market surveillance needs is available. Every effort is made to ensure participants are representative of the range of clinical environments in which Medtronic cardiac rhythm products are used. Eligible products for enrollment include Medtronic market-released cardiac rhythm therapy products for which additional information to further characterize product performance following market release is desired. Enrollment may be capped at a product when the number enrolled ensures sufficient precision to effectively characterize product survivability.

continued

The Standard Actuarial Method is used to determine estimates of lead survival.

Enrolled patients are followed in accordance with the standard care practices of their care provider from their implant date until they can no longer be followed (e.g., death, lost to follow-up, etc.). However, to ensure regular patient status assessments are completed, follow-up windows consistent with typical care practices have been established with a minimum annual follow-up requirement. Product-related adverse events, system modifications and changes in patient status (e.g., death and withdrawal from the study) are required to be reported upon occurrence. This active surveillance model ensures a robust dataset for effectively monitoring product performance.

Patients are eligible for enrollment if:

- They are intended to be implanted or are within 30 days post-implant of a Medtronic market-released lead connected to a market-released CRT, ICD, or IPG device, and the lead is used for a pacing, sensing, or defibrillation application, or
- They participated in a qualifying investigational study of a Medtronic cardiac therapy product that is now market-released; complete implant and follow-up data are available; and the data can be appropriately and legally released.

Each site must inform Medtronic whenever a lead complication has occurred or when a patient is no longer participating. Chronic product performance is analyzed as a function of time using the survival analysis method.

Medtronic continually evaluates the quality and integrity of the data through a combination of on-site and centralized monitoring activities.

Lead Complications

The data presented characterizes chronic lead performance by estimating lead-related complication-free survival probabilities. For analysis purposes, the complication criteria, which align with the AdvaMed 'Industry Guidance for Uniform Reporting of Clinical Performance of Cardiac Rhythm Management Pulse Generators and Leads', are defined below. These criteria do not, however, enable a lead integrity or "hardware" failure to be conclusively differentiated from other clinical events such as an undetected lead dislodgement, exit block, or concurrent pulse generator failure manifested as a sensing or capture problem.

¹ During the evolution of PSR, event adjudication was transitioned from a Medtronic technical review committee to an independent event adjudication committee in 2011. Data analyses include adjudication using both methods.

All reported lead-related adverse events are classified by the reporting investigator and are adjudicated by an independent event adjudication committee.¹ A lead-related complication is considered to have occurred if a clinical observation occurs more than 30 days after implant, is adjudicated with at least one of the following event classifications and at least one of the following clinical actions is made. Events with an onset date 30 days or less after the implant are considered procedure-related and therefore not included as chronic lead-related complications.

Event Classifications

- Failure to capture
- Failure to sense/undersensing
- Oversensing
- Abnormal pacing impedance (based on lead model, but normal range is typically 200 - 2,000 ohms)
- Abnormal defibrillation impedance (based on lead model, but normal range is typically 20 - 200 ohms)
- Insulation breach
- Conductor fracture, confirmed electrically, visually or radiographically
- Extracardiac stimulation
- Cardiac perforation
- Lead dislodgement

Clinical Actions

- Lead surgically abandoned/capped
- Lead electrically abandoned
- Lead explanted
- Lead replaced
- Lead conductor taken out of service (polarity reprogrammed to remove defective conductor, e.g., bipolar to unipolar)

continued

- Lead use continued based on medical judgment despite a known clinical performance issue
- Other lead-related surgery performed (e.g., lead mechanical alteration or unsuccessful repositioning)

Note: Successful lead repositioning is not a qualifying action.

Data Analysis Methods

The performance of leads is expressed in terms of lead survival estimates, where “survival” refers to the function of the lead, not the survival of the patient. These survival estimates are intended to illustrate the probability that a lead will survive for a given number of years without a chronic lead-related complication.

Survival times are calculated from the implant date to the earliest of the complication date, out-of-service date (for example, subject leaves the study or the lead is no longer being used) or the last follow-up date. If a lead experiences more than one complication, the first is used to calculate survival time; although all complications associated with a lead are included in the summary tables.

Of the several different statistical methods available for survival analysis, the Standard Actuarial Method, with suspensions assumed distributed across the intervals (Cutler-Ederer Method), is used to determine estimates of lead survival.

On the following pages, each graph includes a survival curve where events include qualifying lead-related complications. This survival estimate is a good representation of the probability a lead will survive a period of time without a lead-related complication. For example, if a survival probability is 95% after 5 years of service, then the lead has a 5% chance of experiencing a lead-related complication in the first 5 years following implant.

Although tabular data is provided in 1-year intervals, the curves are actually computed and plotted using 3-month intervals.

The data in the tables is rounded to the nearest tenth of one percent. Occasionally, a graph may show 100% survival, but have one or more complications. This occurs because even with the complications, the data rounds to 100%.

The survival curves are statistical estimates. As performance experience accumulates, the estimation improves. Confidence intervals are provided as a way to indicate the degree of certainty of the estimates. Greenwood’s formula is used to calculate the standard errors, and the log-log method is used to produce the 2-sided 95% confidence bounds.

Since the survival estimate can become very imprecise with small effective sample sizes, Medtronic truncates the survival curve when the number of leads entering an interval is less than 50 leads. When the number of leads entering an interval reaches 50, the next data point is added to the survival curve. For those lead models that do not have sufficient sample size, a survival curve will not be presented.

Definition of Analysis Dataset

The survival estimates are derived from all device components successfully enrolled as of the data cutoff date. The number of enrollments is listed for each lead model.

This sample is considered to be representative of the worldwide population, and therefore the survival estimates shown should be representative of the performance worldwide of these models.

Criteria for Model Inclusion

Performance information for a model or model family will be published when more than 100 leads have been enrolled and no fewer than 50 leads followed for at least 6 months. Models will continue to be published for at least 20 years as long as Medtronic estimates at least 500 leads remain active in the United States, based on estimated US implants.

Medtronic, at its discretion, may stop providing updated performance information on lead models that received original US market-release approval 20 or more years ago.

Medtronic urges all physicians to return explanted products and to notify Medtronic when a product is no longer in use, regardless of reason for explant or removal from use.

Returned Product Analysis Results

Although the returned product analysis data is not used to generate the survival estimates, the data provides valuable insight into the causes of lead malfunction.

For reporting returned product analysis results, Medtronic CRDM considers a lead as having malfunctioned whenever the analysis shows that any parameter was outside the performance limits established by Medtronic while implanted and in service. To be considered a malfunction for returned product analysis reporting, the lead must have been returned to Medtronic and analyzed.

The results of the analysis is presented in four categories. The lead reporting categories are:

- **Conductor Fracture:** Conductor malfunction with complete or intermittent loss of continuity that could interrupt current flow (e.g., fractured conductors), including those associated with clavicle flex fatigue or crush damage.
- **Insulation Breach:** A malfunction of the insulation allowing inappropriate entry of body fluids or inappropriate current flow between the conductors, or between the conductor and the body. Examples include cuts, tears, depressions, abrasions, and material degradation.
- **Crimps/Welds/Bonds:** Any malfunction in a conductor or lead body associated with a point of connection.
- **Other:** Malfunctions of specific lead mechanical attributes, such as sensors, connectors, seal rings, or malfunction modes not included in the three categories above.

A lead subject to a safety advisory is not considered to have malfunctioned unless it has been returned to Medtronic CRDM and found, through analysis, to actually have performed outside the performance limits established by Medtronic.

For leads designed for either ventricular or atrial use, the numbers listed in the Returned Product Analysis tables include both.

The numbers of malfunctions listed in the Returned Product Analysis tables are the actual numbers confirmed in the returned product analysis from the United States. The numbers of complications listed in the complications

tables are the actual numbers observed in the PSR centers around the world.

US Reports of Acute Lead Observations (Occurring within First Month of Service)

In the first weeks following lead implantation, physiologic responses and lead performance can vary until long-term lead stability is attained. Acute (defined as the first month after implant) lead performance may be subject to a number of factors, including patient-specific anatomy, clinical conditions and/or varying implant conditions/ techniques. After a period of time, the implant and the lead performance stabilizes. It is for this reason that the Product Surveillance Registry results, which are intended to measure long-term performance, do not include complications that occur within the first 30 days after implant.

Information about the clinical experience in the first month of service is included in this report. The source for this information is Medtronic's complaint handling system database. The information is summarized in tables titled "US Reports of Acute Lead Observations." To be included in this summary of observations, a lead must first be successfully implanted and registered in Medtronic's Device and Registrant Tracking system.

Each Event Report received by Medtronic's complaint handling system is assigned one or more Reason for Report codes based on the information received. The Reason for Report codes have been grouped into Acute Lead Observation categories. The categories used for this product performance report are drawn from the "FDA Guidance for Submission of Research and Marketing Applications for Permanent Pacemaker Leads and for Pacemaker Lead Adapter 510(k) Submissions." The categories are:

1. Cardiac Perforation
2. Conductor Fracture
3. Lead Dislodgement
4. Failure to Capture
5. Oversensing
6. Failure to Sense
7. Insulation Breach
8. Impedance Abnormal

9. Extracardiac Stimulation

10. Unspecified

Although multiple observations are possible for any given lead, only one observation is reported per lead. The observation reported is the observation highest on the list. For example, if an Event Report includes observations for both Lead Dislodgement and Failure to Sense, Lead Dislodgement is reported.

The lead event reported to Medtronic may or may not have involved clinical action or product returned to Medtronic. The lead may have remained implanted and in service.

Estimated Number of Implanted and Active Leads in the United States

In addition to providing the number of leads enrolled in the PSR, this report also provides the number of leads registered as implanted and the number remaining active in the United States based on the status recorded in the Medtronic Device and Registrant Tracking system.

Some lead models do not have a survival curve presented in this report. These lead models do not have a survival curve because they have insufficient sample size in the Product Surveillance Registry. Returned Product Analysis results for these models are included here for reference and comparison.

Left-Heart Leads continued

2187 Attain LV


Product Characteristics

US Market Release	Aug-01	Serial Number Prefix	LEY	US Returned Product Analysis	
Registered US Implants	12,000	Type and/or Fixation	Transvenous, Left Ventricular Cardiac Vein, Distal Continuous Curve	Conductor Fracture	1
Estimated Active US Implants	2,700	Polarity	Unipolar	Crimp/Weld/Bond	0
Advisories	None	Steroid	No	Insulation Breach	0
				Other	4

Product Surveillance Registry Results

Qualifying Complications 1 Total

Number of Leads Enrolled in Study	134	Failure to Capture	1
Cumulative Months of Follow-Up	6,763		
Number of Leads Active in Study	11		



4193 Attain OTW


Product Characteristics

US Market Release	May-02	Serial Number Prefix	BAA	US Returned Product Analysis	
Registered US Implants	100,800	Type and/or Fixation	Transvenous, Left Ventricular Cardiac Vein, Distal Double Curve	Conductor Fracture	52
Estimated Active US Implants	34,100	Polarity	Unipolar	Crimp/Weld/Bond	0
Advisories	None	Steroid	Yes	Insulation Breach	8
				Other	48

Product Surveillance Registry Results

Qualifying Complications 39 Total

Number of Leads Enrolled in Study	675	Conductor Fracture	1	Lead Dislodgement	14
Cumulative Months of Follow-Up	31,621	Extra Cardiac Stimulation	8	Unspecified Clinical Failure	3
Number of Leads Active in Study	103	Failure to Capture	13		



Left-Heart Leads

4194 Attain OTW


Product Characteristics

US Market Release	Aug-04	Serial Number Prefix	LFG	US Returned Product Analysis	
Registered US Implants	109,700	Type and/or Fixation	Transvenous, Left Ventricular Cardiac Vein, Distal Double Curve	Conductor Fracture	13
Estimated Active US Implants	65,000	Polarity	Bipolar	Crimp/Weld/Bond	0
Advisories	None	Steroid	Yes	Insulation Breach	58
				Other	9

Product Surveillance Registry Results

Qualifying Complications 36 Total

Number of Leads Enrolled in Study	1,223	Extra Cardiac Stimulation	8	Lead Dislodgement	14
Cumulative Months of Follow-Up	42,016	Failure to Capture	11		
Number of Leads Active in Study	648	Insulation (ESC)	1		
		Insulation (not further defined)	2		



4195 Attain StarFix


Product Characteristics

US Market Release	Aug-08	Serial Number Prefix	AAD	US Returned Product Analysis	
Registered US Implants	15,100	Type and/or Fixation	Transvenous, Left Ventricular Cardiac Vein, Deployable Lobe Fixation	Conductor Fracture	1
Estimated Active US Implants	11,900	Polarity	Unipolar	Crimp/Weld/Bond	0
Advisories	None	Steroid	Yes	Insulation Breach	1
				Other	3

Product Surveillance Registry Results

Qualifying Complications 15 Total

Number of Leads Enrolled in Study	1,279	Conductor Fracture	1	Insulation (not further defined)	1
Cumulative Months of Follow-Up	30,282	Extra Cardiac Stimulation	8	Lead Dislodgement	3
Number of Leads Active in Study	886	Failure to Capture	2		



Left-Heart Leads continued

4196 Attain Ability


Product Characteristics

US Market Release	May-09	Serial Number Prefix	PVI	US Returned Product Analysis	
Registered US Implants	52,000	Type and/or Fixation	Transvenous, Left Ventricular Cardiac Vein, Double Curve	Conductor Fracture	6
Estimated Active US Implants	43,100	Polarity	Bipolar	Crimp/Weld/Bond	0
Advisories	None	Steroid	Yes	Insulation Breach	0
				Other	6

Product Surveillance Registry Results

Qualifying Complications 46 Total

Number of Leads Enrolled in Study	1,878	Conductor Fracture	1	Lead Dislodgement	16
Cumulative Months of Follow-Up	43,029	Extra Cardiac Stimulation	14		
Number of Leads Active in Study	1,104	Failure to Capture	15		



4296 Attain Ability Plus


Product Characteristics

US Market Release	Apr-11	Serial Number Prefix	RRA	US Returned Product Analysis	
Registered US Implants	16,000	Type and/or Fixation	Transvenous, Left Ventricular Cardiac Vein, Distal Double Curve	Conductor Fracture	0
Estimated Active US Implants	15,000	Polarity	Dual Electrodes	Crimp/Weld/Bond	0
Advisories	None	Steroid	Yes	Insulation Breach	0
				Other	2

Product Surveillance Registry Results

Qualifying Complications 3 Total

Number of Leads Enrolled in Study	819	Extracardiac Stimulation	1		
Cumulative Months of Follow-Up	6,360	Lead Dislodgement	2		
Number of Leads Active in Study	679				



Left-Heart Leads continued

4396 Attain Ability Straight


Product Characteristics

		Product Characteristics		US Returned Product Analysis	
US Market Release	Mar-11	Serial Number Prefix	RAE	Conductor Fracture	0
Registered US Implants	3,600	Type and/or Fixation	Transvenous, Left Ventricular Cardiac Vein, Tines	Crimp/Weld/Bond	0
Estimated Active US Implants	3,300	Polarity	Dual Electrodes	Insulation Breach	0
Advisories	None	Steroid	Yes	Other	0

Product Surveillance Registry Results

Qualifying Complications 0 Total

Number of Leads Enrolled in Study	211
Cumulative Months of Follow-Up	2,520
Number of Leads Active in Study	165



Left-Heart Leads continued

Lead Survival Summary (95% Confidence Interval)

Model Number	Family	US Market Release	Leads Enrolled	Leads Active in Study	Qualifying Complications	Cumulative Months of Follow-Up in Study	Device Survival Probability (%)									
							Years After Implant									
							1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	9 yr	10 yr
2187	Attain LV	Aug-01	134	11	1	6,763	99.1 +0.7/-5.1	99.1 +0.7/-5.1	99.1 +0.7/-5.1	99.1 +0.7/-5.1	99.1 +0.7/-5.1 at 54 mo					
4193	Attain OTW	May-02	675	103	39	31,621	95.9 +1.3/-1.9	95.0 +1.5/-2.1	94.3 +1.6/-2.2	94.0 +1.7/-2.3	93.6 +1.8/-2.5	93.1 +2.0/-2.7	93.1 +2.0/-2.7	92.2 +2.4/-3.4	92.2 +2.4/-3.4 at 105 mo	
4194	Attain OTW	Aug-04	1,223	648	36	42,016	98.6 +0.5/-0.9	97.3 +0.8/-1.2	97.2 +0.9/-1.2	96.6 +1.1/-1.5	96.1 +1.3/-2.0	93.6 +2.4/-3.6 at 81 mo	93.6 +2.4/-3.6 at 81 mo			
4195	Attain StarFix	Aug-08	1,279	886	15	30,282	99.2 +0.4/-0.8	98.6 +0.6/-1.0	98.2 +0.8/-1.3	98.2 +0.8/-1.3	98.2 +0.8/-1.3	98.2 +0.8/-1.3 at 63 mo				
4196	Attain Ability	May-09	1,878	1,104	46	43,029	98.0 +0.6/-0.8	97.4 +0.7/-0.9	96.8 +0.8/-1.1	96.8 +0.8/-1.1 at 45 mo						
4296	Attain Ability Plus	Apr-11	819	679	3	6,360	99.4 +0.4/-1.3	99.4 +0.4/-1.3 at 15 mo								
4396	Attain Ability Straight	Mar-11	211	165	0	2,520	100.0	100.0								

Source: Product Surveillance Registry
Data as of July 31, 2013

US Returned Product Analysis Summary

Model Number	Family	US Market Release	Estimated US Implants	Estimated US Active	Conductor Fracture	Insulation Breach	Crimp/Weld/Bond	Other
2187	Attain LV	Aug-01	12,000	2,700	1	0	0	4
4193	Attain OTW	May-02	100,800	34,100	52	8	0	48
4194	Attain OTW	Aug-04	109,700	65,000	13	58	0	9
4195	Attain StarFix	Aug-08	15,100	11,900	1	1	0	3
4196	Attain Ability	May-09	52,000	43,100	6	0	0	6
4296	Attain Ability Plus	Apr-11	16,000	15,000	0	0	0	2
4396	Attain Ability Straight	Mar-11	3,600	3,300	0	0	0	0

Source: Returned Product Analysis
Data as of July 31, 2013

US Reports of Acute Lead Observations

Model Number	Family	Estimated US Implants	Cardiac Perforation	Conductor Fracture	Lead Dislodgement	Failure to Capture	Oversensing	Failure to Sense	Insulation Breach	Impedance Abnormal	Extracardiac Stimulation	Unspecified
2187	Attain LV	12,000	0	0	9	3	0	1	0	0	1	0
4193	Attain OTW	100,800	0	0	45	11	1	0	0	0	16	2
4194	Attain OTW	109,700	2	2	121	36	2	0	0	6	29	5
4195	Attain StarFix	15,100	0	0	23	15	0	0	0	1	22	1
4196	Attain Ability	52,000	1	1	129	35	1	1	1	5	62	3
4296	Attain Ability Plus	16,000	1	0	51	8	0	0	1	0	17	0
4396	Attain Ability Straight	3,600	0	1	16	4	0	0	0	0	5	0

Report Cutoff Date: September 10, 2013

Reference Chart

Model Number	Family	Type	Insulation	Conductor Material	Tip Electrode	Connector Type
2187	Attain LV	Transvenous Cardiac Vein Preformed Body	Polyurethane (55D)	MP35N	Platinized Platinum	IS-1 UNI
4193	Attain OTW	Transvenous Cardiac Vein Preformed Body	Polyurethane (55D)	MP35N	Platinized Platinum	IS-1 UNI
4194	Attain OTW	Transvenous Cardiac Vein Preformed Body	Polyurethane (55D)/ Silicone (4719)	MP35N	Platinum Alloy	IS-1 BI
4195	Attain StarFix	Transvenous Cardiac Vein Deployable Lobes	Polyurethane (55D)	MP35N	Platinum Alloy	IS-1 UNI
4196	Attain Ability	Transvenous Cardiac Vein Preformed Body	Polyurethane (outer) SI-polyimide (inner)	Ag-core- MP35N	Platinum iridium alloy with titanium nitride coating	IS-1 BI
4296	Attain Ability Plus	Transvenous Cardiac Vein Distal Double Curve	Polyurethane (outer) Silicone (inner)	Ag-core- MP35N	Platinum iridium alloy with titanium nitride coating	IS-1 BI
4396	Attain Ability Straight	Transvenous Cardiac Vein Tines	Polyurethane (outer) Silicone (inner)	Ag-core- MP35N	Platinum iridium alloy with titanium nitride coating	IS-1 BI

Defibrillation Leads

6721 Epicardial Patch

Product Characteristics

US Market Release	Mar-94	Serial Number Prefix	TBH, TBG, TBB
Registered US Implants	2,900	Type and/or Fixation	Epi Patch, Epicardial Suture
Estimated Active US Implants	1,100	Polarity	NA
Advisories	None	Steroid	No

US Returned Product Analysis


Conductor Fracture	12
Crimp/Weld/Bond	0
Insulation Breach	1
Other	0

Product Surveillance Registry Results

Number of Leads Enrolled in Study	407
Cumulative Months of Follow-Up	23,401
Number of Leads Active in Study	4

Qualifying Complications 47 Total

Conductor Fracture	21	Insulation (not further defined)	2
Failure to Capture	8	Oversensing	12
Impedance Out of Range	4		



6930 Sprint Fidelis

Product Characteristics

US Market Release	Sep-04	Serial Number Prefix	LFK
Registered US Implants	400	Type and/or Fixation	Transvenous, Right Ventricle, Tines
Estimated Active US Implants	200	Polarity	True Bipolar/One Coil
Advisories:		Steroid	Yes
See page 142 - 2007 Potential Conductor Wire Fracture			


US Returned Product Analysis

Conductor Fracture	3
Crimp/Weld/Bond	0
Insulation Breach	0
Other	0

Product Surveillance Registry Results

Number of Leads Enrolled in Study	4
Cumulative Months of Follow-Up	188
Number of Leads Active in Study	2

Qualifying Complications 0 Total



Defibrillation Leads continued

6931 Sprint Fidelis

Product Characteristics

US Market Release	Sep-04	Serial Number Prefix	LFL
Registered US Implants	8,100	Type and/or Fixation	Transvenous, Right Ventricle, Active Screw-in
Estimated Active US Implants	3,500	Polarity	True Bipolar/One Coil
Advisories		Steroid	Yes
See page 142 – 2007 Potential Conductor Wire Fracture			

US Returned Product Analysis


Conductor Fracture	532
Crimp/Weld/Bond	0
Insulation Breach	1
Other	5

Product Surveillance Registry Results

Number of Leads Enrolled in Study	294
Cumulative Months of Follow-Up	13,526
Number of Leads Active in Study	113

Qualifying Complications 43 Total

Conductor Fracture	24	Lead Dislodgement	2
Failure to Capture	3	Oversensing	5
Failure to Sense	1		
Impedance Out of Range	8		



6932 Sprint

Product Characteristics

US Market Release	Aug-96	Serial Number Prefix	TCA
Registered US Implants	14,900	Type and/or Fixation	Transvenous, Right Ventricle, Tines
Estimated Active US Implants	4,200	Polarity	True Bipolar/One Coil
Advisories	None	Steroid	Yes

US Returned Product Analysis


Conductor Fracture	22
Crimp/Weld/Bond	0
Insulation Breach	24
Other	2

Product Surveillance Registry Results

Number of Leads Enrolled in Study	412
Cumulative Months of Follow-Up	25,811
Number of Leads Active in Study	38

Qualifying Complications 11 Total

Extra Cardiac Stimulation	1	Impedance Out of Range	2
Failure to Capture	2	Oversensing	4
Failure to Sense	2		



Defibrillation Leads continued

6933, 6937, 6937A, SVC/CS

Product Characteristics

US Market Release	Apr-94	Serial Number Prefix	TAT, TBU, TDB
Registered US Implants	11,900	Type and/or Fixation	Transvenous, SVC/CS, Passive
Estimated Active US Implants	2,400	Polarity	One Coil
Advisories	None	Steroid	No

US Returned Product Analysis


Conductor Fracture	127
Crimp/Weld/Bond	0
Insulation Breach	17
Other	1

Product Surveillance Registry Results

Number of Leads Enrolled in Study	966
Cumulative Months of Follow-Up	54,489
Number of Leads Active in Study	11

Qualifying Complications

Conductor Fracture	16	Insulation (not further defined)	2
Extra Cardiac Stimulation	4	Lead Dislodgement	1
Failure to Capture	6	Oversensing	10
Failure to Sense	1	Unspecified Clinical Failure	4
Impedance Out of Range	3		



6935 Sprint Quattro Secure

Product Characteristics

US Market Release	Nov-08	Serial Number Prefix	TAU
Registered US Implants	42,900	Type and/or Fixation	Transvenous, Right Ventricle, Active Screw-in
Estimated Active US Implants	38,800	Polarity	True Bipolar/One Coil
Advisories	None	Steroid	Yes

US Returned Product Analysis


Conductor Fracture	61
Crimp/Weld/Bond	0
Insulation Breach	2
Other	35

Product Surveillance Registry Results

Number of Leads Enrolled in Study	2,026
Cumulative Months of Follow-Up	36,705
Number of Leads Active in Study	1,550

Qualifying Complications

Conductor Fracture	4	Oversensing	3
Failure to Capture	2		
Failure to Sense	1		
Lead Dislodgement	4		



Defibrillation Leads continued

6935M Sprint Quattro Secure


Product Characteristics

US Market Release	Aug-12	Serial Number Prefix	TDL	US Returned Product Analysis	
Registered US Implants	12,700	Type and/or Fixation	Transvenous, Right Ventricle, Active Screw-in	Conductor Fracture	0
Estimated Active US Implants	12,400	Polarity	True Bipolar/One Coil	Crimp/Weld/Bond	0
Advisories	None	Steroid	Yes	Insulation Breach	0
				Other	1

Product Surveillance Registry Results

Qualifying Complications 1 Total

Number of Leads Enrolled in Study	339	Other	1
Cumulative Months of Follow-Up	943		
Number of Leads Active in Study	307		



6942 Sprint


Product Characteristics

US Market Release	Jul-97	Serial Number Prefix	TCB	US Returned Product Analysis	
Registered US Implants	17,700	Type and/or Fixation	Transvenous, Right Ventricle, Tines	Conductor Fracture	15
Estimated Active US Implants	5,300	Polarity	Integrated Bipolar/Two Coils	Crimp/Weld/Bond	1
Advisories	None	Steroid	Yes	Insulation Breach	22
				Other	4

Product Surveillance Registry Results

Qualifying Complications 7 Total

Number of Leads Enrolled in Study	351	Conductor Fracture	1	Oversensing	3
Cumulative Months of Follow-Up	19,347	Failure to Sense	1	Unspecified Clinical Failure	1
Number of Leads Active in Study	25	Lead Dislodgement	1		



Defibrillation Leads continued

6943 Sprint

Product Characteristics

US Market Release	Oct-97	Serial Number Prefix	TCE
Registered US Implants	20,600	Type and/or Fixation	Transvenous, Right Ventricle, Active Screw-in
Estimated Active US Implants	6,200	Polarity	True Bipolar/One Coil
Advisories	None	Steroid	Yes

US Returned Product Analysis

Conductor Fracture	73
Crimp/Weld/Bond	1
Insulation Breach	29
Other	5

Product Surveillance Registry Results


Number of Leads Enrolled in Study	1,311
Cumulative Months of Follow-Up	81,556
Number of Leads Active in Study	220

Qualifying Complications

Conductor Fracture	24
Failure to Capture	8
Failure to Sense	7
Impedance Out of Range	8

94 Total

Insulation (not further defined)	1
Lead Dislodgement	2
Other	1
Oversensing	40
Unspecified Clinical Failure	3



6944 Sprint Quattro

Product Characteristics

US Market Release	Dec-00	Serial Number Prefix	TDC
Registered US Implants	41,900	Type and/or Fixation	Transvenous, Right Ventricle, Tines
Estimated Active US Implants	21,900	Polarity	True Bipolar/Two Coils
Advisories	None	Steroid	Yes

US Returned Product Analysis

Conductor Fracture	106
Crimp/Weld/Bond	1
Insulation Breach	3
Other	4


Product Surveillance Registry Results

Number of Leads Enrolled in Study	520
Cumulative Months of Follow-Up	18,398
Number of Leads Active in Study	270

Qualifying Complications

Conductor Fracture	2
Failure to Sense	1
Impedance Out of Range	1
Oversensing	2

7 Total



Defibrillation Leads continued

6945 Sprint


Product Characteristics

US Market Release	Sep-97	Serial Number Prefix	TDA	<u>US Returned Product Analysis</u>	
Registered US Implants	42,700	Type and/or Fixation	Transvenous, Right Ventricle, Active Screw-in	Conductor Fracture	127
Estimated Active US Implants	12,600	Polarity	Integrated Bipolar/Two Coils	Crimp/Weld/Bond	1
Advisories	None	Steroid	Yes	Insulation Breach	40
				Other	6

Product Surveillance Registry Results

Qualifying Complications 40 Total

Number of Leads Enrolled in Study	1,155	Conductor Fracture	9	Impedance Out of Range	6
Cumulative Months of Follow-Up	68,458	Extra Cardiac Stimulation	1	Oversensing	17
Number of Leads Active in Study	111	Failure to Capture	2	Unspecified Clinical Failure	1
		Failure to Sense	4		



6947M Sprint Quattro Secure


Product Characteristics

US Market Release	Feb-12	Serial Number Prefix	TDK	<u>US Returned Product Analysis</u>	
Registered US Implants	30,200	Type and/or Fixation	Transvenous, Right Ventricle, Active Screw-in	Conductor Fracture	3
Estimated Active US Implants	29,300	Polarity	True Bipolar/Two Coils	Crimp/Weld/Bond	0
Advisories	None	Steroid	Yes	Insulation Breach	0
				Other	4

Product Surveillance Registry Results

Qualifying Complications 2 Total

Number of Leads Enrolled in Study	1,212	Failure to Capture	2
Cumulative Months of Follow-Up	7,132		
Number of Leads Active in Study	1,125		



Leads

Defibrillation Leads continued

6947 Sprint Quattro Secure


Product Characteristics

				US Returned Product Analysis	
US Market Release	Nov-01	Serial Number Prefix	TDG	Conductor Fracture	504
Registered US Implants	362,600	Type and/or Fixation	Transvenous, Right Ventricle, Active Screw-in	Crimp/Weld/Bond	4
Estimated Active US Implants	243,500	Polarity	True Bipolar/Two Coils	Insulation Breach	32
Advisories	None	Steroid	Yes	Other	211

Product Surveillance Registry Results

Qualifying Complications 41 Total

Number of Leads Enrolled in Study	2,718	Conductor Fracture	8	Lead Dislodgement	4
Cumulative Months of Follow-Up	122,934	Failure to Capture	1	Oversensing	13
Number of Leads Active in Study	1,026	Failure to Sense	2	Unspecified Clinical Failure	2
		Impedance Out of Range	8		
		Insulation (not further defined)	3		



6948 Sprint Fidelis


Product Characteristics

				US Returned Product Analysis	
US Market Release	Sep-04	Serial Number Prefix	LFH	Conductor Fracture	154
Registered US Implants	10,400	Type and/or Fixation	Transvenous, Right Ventricle, Tines	Crimp/Weld/Bond	0
Estimated Active US Implants	4,700	Polarity	True Bipolar/Two Coils	Insulation Breach	2
Advisories		Steroid	Yes	Other	2
See page 142 – 2007 Potential Conductor Wire Fracture					

Product Surveillance Registry Results

Qualifying Complications 2 Total

Number of Leads Enrolled in Study	30	Conductor Fracture	2
Cumulative Months of Follow-Up	1,462		
Number of Leads Active in Study	12		



Defibrillation Leads continued

6949 Sprint Fidelis


Product Characteristics

Product Characteristics				US Returned Product Analysis	
US Market Release	Sep-04	Serial Number Prefix	LFI	Conductor Fracture	6,376
Registered US Implants	186,800	Type and/or Fixation	Transvenous, Right Ventricle, Active Screw-in	Crimp/Weld/Bond	3
Estimated Active US Implants	76,100	Polarity	True Bipolar/Two Coils	Insulation Breach	29
Advisories		Steroid	Yes	Other	69
See page 142 – 2007 Potential Conductor Wire Fracture					

Product Surveillance Registry Results

Qualifying Complications 80 Total

Number of Leads Enrolled in Study	796	Conductor Fracture	38	Insulation (not further defined)	2
Cumulative Months of Follow-Up	38,922	Failure to Capture	2	Lead Dislodgement	1
Number of Leads Active in Study	217	Failure to Sense	4	Oversensing	16
		Impedance Out of Range	16	Unspecified Clinical Failure	
				Other	1



6996 Sub-Q Lead


Product Characteristics

Product Characteristics				US Returned Product Analysis	
US Market Release	Jun-01	Serial Number Prefix	TCR	Conductor Fracture	20
Registered US Implants	3,900	Type and/or Fixation	Subcutaneous, Defibrillation, Suture on Anchor Sleeve	Crimp/Weld/Bond	0
Estimated Active US Implants	2,300	Polarity	One Coil	Insulation Breach	0
Advisories	None	Steroid	No	Other	0

Product Surveillance Registry Results

Qualifying Complications 2 Total

Number of Leads Enrolled in Study	43	Conductor Fracture	1
Cumulative Months of Follow-Up	1,190	Impedance Out of Range	1
Number of Leads Active in Study	20		



Defibrillation Leads continued

Lead Survival Summary (95% Confidence Interval)

Model Number	Family	US Market Release	Leads Enrolled	Leads Active in Study	Qualifying Complications	Cumulative Months of Follow-Up in Study	Device Survival Probability (%)														
							Years After Implant														
							1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	10 yr	12 yr	14 yr	16 yr	18 yr		
6721	Epicardial Patch	Mar-94	407	4	47	23,401	96.5 +1.4/-2.4	95.0 +1.8/-2.8	92.7 +2.3/-3.3	91.9 +2.5/-3.5	90.0 +2.9/-4.0	85.1 +3.9/-5.2	83.7 +4.3/-5.5	83.7 +4.3/-5.5	83.7 +4.3/-5.5	83.7 +4.3/-5.5					
6930	Sprint Fidelis	Sep-04	4	2	0	188	100.0 at 0 mo														
	Advisories: See page 142 – 2007 Potential Conductor Wire Fracture																				
6931	Sprint Fidelis	Sep-04	294	113	43	13,526	98.2 +1.0/-2.5	96.2 +1.7/-3.2	93.0 +2.6/-4.1	88.4 +3.6/-5.1	81.9 +5.0/-6.5	78.1 +5.8/-7.5									
	Advisories: See page 142 – 2007 Potential Conductor Wire Fracture																				
6932	Sprint	Aug-96	412	38	11	25,811	99.2 +0.5/-1.7	98.3 +0.9/-2.0	98.3 +0.9/-2.0	98.3 +0.9/-2.0	97.8 +1.2/-2.5	97.8 +1.2/-2.5	96.8 +1.7/-3.7	95.6 +2.4/-5.0	92.6 +3.8/-7.5						
6933, 6937, 6937A	SVC/CS	Apr-94	966	11	47	54,489	98.4 +0.6/-1.1	97.5 +0.9/-1.3	97.2 +0.9/-1.4	96.7 +1.1/-1.5	95.4 +1.4/-1.9	94.9 +1.5/-2.1	93.9 +1.8/-2.5	93.4 +1.9/-2.7	91.2 +2.8/-4.1						
	Advisories: See page 142 – 2007 Potential Conductor Wire Fracture																				
6935	Sprint Quattro Secure	Nov-08	2,026	1,550	14	36,705	99.5 +0.3/-0.5	99.2 +0.4/-0.7	98.6 +0.7/-1.3	98.6 +0.7/-1.3											
	Advisories: See page 142 – 2007 Potential Conductor Wire Fracture																				
6935M	Sprint Quattro Secure	Aug-12	339	307	1	75,241	99.6 +0.3/-2.4														
	Advisories: See page 142 – 2007 Potential Conductor Wire Fracture																				
6942	Sprint	Jul-97	351	25	7	19,347	99.1 +0.6/-1.9	99.1 +0.6/-1.9	98.1 +1.1/-2.7	97.5 +1.4/-3.1	96.8 +1.8/-3.7	96.8 +1.8/-3.7	96.8 +1.8/-3.7	96.8 +1.8/-3.7	96.8 +1.8/-3.7	96.8 +1.8/-3.7					
	Advisories: See page 142 – 2007 Potential Conductor Wire Fracture																				
6943	Sprint	Oct-97	1,311	199	94	83,217	98.6 +0.5/-0.8	97.7 +0.7/-1.0	96.6 +0.9/-1.2	95.7 +1.1/-1.4	93.8 +1.4/-1.8	92.2 +1.6/-2.0	91.0 +1.8/-2.3	90.1 +2.0/-2.5	87.0 +2.7/-3.4	82.3 +4.2/-5.3					
	Advisories: See page 142 – 2007 Potential Conductor Wire Fracture																				
6944	Sprint Quattro	Dec-00	520	270	7	18,398	100.0	100.0	99.5 +0.5/-3.2	97.9 +1.4/-4.4	95.6 +2.6/-6.1	95.6 +2.6/-6.1	95.6 +2.6/-6.1								
	Advisories: See page 142 – 2007 Potential Conductor Wire Fracture																				
6945	Sprint	Sep-97	1,155	111	40	68,458	99.4 +0.3/-0.7	98.7 +0.6/-1.0	98.2 +0.7/-1.1	97.6 +0.8/-1.3	96.8 +1.1/-1.6	96.1 +1.2/-1.8	95.6 +1.4/-2.0	94.5 +1.7/-2.4	92.3 +2.3/-3.3	90.7 +3.0/-4.2					
	Advisories: See page 142 – 2007 Potential Conductor Wire Fracture																				
6947	Sprint Quattro Secure	Nov-01	2,718	1,026	41	122,934	99.5 +0.2/-0.4	99.2 +0.3/-0.4	98.9 +0.4/-0.5	98.5 +0.5/-0.7	98.1 +0.6/-0.9	97.5 +0.8/-1.2	97.0 +1.0/-1.4	96.7 +1.1/-1.6	95.9 +1.4/-2.1						
	Advisories: See page 142 – 2007 Potential Conductor Wire Fracture																				
6947M	Sprint Quattro Secure	Feb-12	1,212	1,125	2	7,132	99.8 +0.2/-0.7														
	Advisories: See page 142 – 2007 Potential Conductor Wire Fracture																				
6948	Sprint Fidelis	Sep-04	30	12	2	1,462	100.0 at 0 mo														
	Advisories: See page 142 – 2007 Potential Conductor Wire Fracture																				
6949	Sprint Fidelis	Sep-04	796	217	80	38,922	98.5 +0.6/-1.2	96.8 +1.1/-1.6	94.1 +1.6/-2.1	92.1 +1.9/-2.5	89.3 +2.4/-3.0	84.9 +3.1/-3.8	81.8 +3.9/-4.8	80.4 +4.5/-5.6							
	Advisories: See page 142 – 2007 Potential Conductor Wire Fracture																				
6996	Sub-Q Lead	Jun-01	43	20	2	1,190	100.0 at 0 mo														

US Returned Product Analysis Summary

Model Number	Family	US Market Release	Estimated US Implants	Estimated US Active	Conductor Fracture	Insulation Breach	Crimp/Weld/Bond	Other
6721	Epicardial Patch	Mar-94	2,900	1,100	12	1	0	0
6930	Sprint Fidelis	Sep-04	400	200	3	0	0	0
6931	Sprint Fidelis	Sep-04	8,100	3,500	532	1	0	5
6932	Sprint	Aug-96	14,900	4,200	22	24	0	2
6933, 6937, 6937A	SVC/CS	Apr-94	11,900	2,400	127	17	0	1
6935	Sprint Quattro Secure	Nov-08	42,900	38,800	61	2	0	35
6935M	Sprint Quattro Secure	Aug-12	12,700	12,400	0	0	0	1
6942	Sprint	Jul-97	17,700	5,300	15	22	1	4
6943	Sprint	Oct-97	20,600	6,200	73	29	1	5
6944	Sprint Quattro	Dec-00	41,900	21,900	106	3	1	4
6945	Sprint	Sep-97	42,700	12,600	127	40	1	6
6947	Sprint Quattro Secure	Nov-01	362,600	243,500	504	36	4	211
6947M	Sprint Quattro Secure	Feb-12	30,200	29,300	3	0	0	4
6948	Sprint Fidelis	Sep-04	10,400	4,700	154	2	0	2
6949	Sprint Fidelis	Sep-04	186,800	76,100	6,376	29	3	69
6996	Sub-Q Lead	Jun-01	3,900	2,300	20	0	0	0

Source: Returned Product Analysis
Data as of July 31, 2013

US Reports of Acute Lead Observations

Model Number	Family	Estimated US Implants	Cardiac Perforation	Conductor Fracture	Lead Dislodgement	Failure to Capture	Oversensing	Failure to Sense	Insulation Breach	Impedance Abnormal	Extracardiac Stimulation	Unspecified
6721	Epicardial Patch	2,900	1	2	0	0	1	1	0	3	0	0
6930	Sprint Fidelis	400	0	0	0	0	0	0	0	0	0	1
6931	Sprint Fidelis	8,100	1	2	1	1	3	1	0	0	0	1
6932	Sprint	14,900	0	0	4	2	0	2	0	1	0	2
6933, 6937, 6937A	SVC/CS	11,900	0	0	1	1	0	0	0	0	0	5
6935	Sprint Quattro Secure	42,900	12	0	27	16	27	5	1	10	0	5
6935M	Sprint Quattro Secure	12,700	1	0	16	13	16	1	0	1	3	0
6942	Sprint	17,700	0	1	1	4	1	0	0	2	0	1
6943	Sprint	20,600	1	0	0	1	1	1	1	2	0	0
6944	Sprint Quattro	41,900	0	2	17	11	11	3	0	8	0	6
6945	Sprint	42,700	1	1	4	6	7	2	2	0	1	2
6947	Sprint Quattro Secure	362,600	24	18	99	66	113	29	4	47	2	22
6947M	Sprint Quattro Secure	30,200	2	1	32	18	11	3	0	2	1	0
6948	Sprint Fidelis	10,400	0	2	7	6	1	0	0	0	0	3
6949	Sprint Fidelis	186,800	10	41	23	32	30	19	6	17	0	25
6996	SubQ	3,900	0	0	1	1	0	0	0	2	0	0

Report Cutoff Date: September 10, 2013

Defibrillation Leads continued

Reference Chart

Model Number	Family	Type	Pin Configuration		Lead Body Diameter	Insulation, Lead Body	Fixation, Steroid
			Pace/Sense	High Voltage			
6721	Epicardial Patch	Epi Patch	—	DF-1	S, M, L	Silicone, Single Lumen	Suture
6930	Sprint Fidelis	Endo RV True Bipolar Sensing	IS-1	DF-1	6.6 Fr	Silicone with Polyurethane Overlay, Multilumen	Passive, Steroid
6931	Sprint Fidelis	Endo RV True Bipolar Sensing	IS-1	DF-1	6.6 Fr	Silicone with Polyurethane Overlay, Multilumen	Active, Steroid
6932	Sprint	Endo RV True Bipolar Sensing	IS-1	DF-1	7.8 Fr	Silicone, Multilumen	Passive, Steroid
6933	SVC/CS	Endo SVC/CS Coil	—	DF-1	7 Fr	Silicone, Single Lumen	Passive
6935	Sprint Quattro Secure	Endo RV True Bipolar Sensing	IS-1	DF-1	8.6 Fr	Silicone with Polyurethane Overlay, Multilumen	Active, Steroid
6935M	Sprint Quattro Secure	Endo RV True Bipolar Sensing	DF-4	DF-4	8.6 Fr	Silicone with Polyurethane Overlay, Multilumen	Active, Steroid
6937	SVC/CS	Endo SVC Coil	—	DF-1	5.5 Fr	Silicone, Single Lumen	Passive
6937A	SVC/CS	Endo SVC Coil	—	DF-1	7.5 Fr	Silicone with Polyurethane Overlay, Single Lumen	Passive
6942	Sprint	Endo RV/SVC Integrated Bipolar Sensing	IS-1	2 DF-1	7.8 Fr	Silicone, Multilumen	Passive, Steroid
6943	Sprint	Endo RV True Bipolar Sensing	IS-1	DF-1	7.8 Fr	Silicone, Multilumen	Active, Steroid
6944	Sprint Quattro	Endo RV/SVC True Bipolar Sensing	IS-1	2 DF-1	8.2 Fr	Silicone with Polyurethane Overlay, Multilumen	Passive, Steroid
6945	Sprint	Endo RV/SVC Integrated Bipolar Sensing	IS-1	2 DF-1	7.8 Fr	Silicone, Multilumen	Active, Steroid
6947	Sprint Quattro Secure	Endo RV/SVC True Bipolar Sensing	IS-1	2 DF-1	8.6 Fr	Silicone with Polyurethane Overlay, Multilumen	Active, Steroid
6947M	Sprint Quattro Secure	Endo RV/SVC True Bipolar Sensing	DF4	DF4	8.6 Fr	Silicone with Polyurethane Overlay, Multilumen	Active, Steroid
6948	Sprint Fidelis	Endo RV/SVC True Bipolar Sensing	IS-1	2 DF-1	6.6 Fr	Silicone with Polyurethane Overlay, Multilumen	Passive, Steroid
6949	Sprint Fidelis	Endo RV/SVC True Bipolar Sensing	IS-1	2 DF-1	6.6 Fr	Silicone with Polyurethane Overlay, Multilumen	Active, Steroid
6996	Sub-Q Lead	SQ Coil	—	DF-1	7.5 Fr	Silicone, Single Lumen	Passive

Pacing Leads

3830 SelectSecure

Product Characteristics

Product Characteristics				US Returned Product Analysis	
US Market Release	Aug-05	Serial Number Prefix	LFF	Conductor Fracture	4
Registered US Implants	21,600	Type and/or Fixation	Transvenous, Atrium or Right Ventricle, Fixed Screw	Crimp/Weld/Bond	0
Estimated Active US Implants	16,400	Polarity	Bipolar	Insulation Breach	14
Advisories	None	Steroid	Yes	Other	3


Atrial Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	778
Cumulative Months of Follow-Up	25,544
Number of Leads Active in Study	493

Qualifying Complications

Qualifying Complications	7	Total	
Cardiac Perforation	1	Failure to Capture	1
Conductor Fracture	1	Failure to Sense	1
Extra Cardiac Stimulation	1	Lead Dislodgement	2




Ventricular Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	482
Cumulative Months of Follow-Up	17,389
Number of Leads Active in Study	283

Qualifying Complications

Qualifying Complications	4	Total
Failure To Capture	1	
Impedance Out of Range	1	
Lead Dislodgement	2	



Pacing Leads continued

4024 CapSure SP


Product Characteristics

US Market Release	Oct-91	Serial Number Prefix	LAJ	US Returned Product Analysis	
Registered US Implants	219,400	Type and/or Fixation	Transvenous, Right Ventricle, Tines	Conductor Fracture	28
Estimated Active US Implants	40,700	Polarity	Bipolar	Crimp/Weld/Bond	0
Advisories	None	Steroid	Yes	Insulation Breach	183
				Other	12

Product Surveillance Registry Results

Qualifying Complications 4 Total

Number of Leads Enrolled in Study	1,215	Failure to Capture	3
Cumulative Months of Follow-Up	50,987	Insulation (not further defined)	1
Number of Leads Active in Study	10		



Pacing Leads continued

4068 CapSureFix

Product Characteristics

Product Characteristics				US Returned Product Analysis	
US Market Release	Mar-96	Serial Number Prefix	LCE	Conductor Fracture	50
Registered US Implants	124,300	Type and/or Fixation	Transvenous, Atrium or Right Ventricle, Active Screw-in	Crimp/Weld/Bond	0
Estimated Active US Implants	29,800	Polarity	Bipolar	Insulation Breach	167
Advisories	None	Steroid	Yes	Other	93


Atrial Placement

Product Surveillance Registry Results

Qualifying Complications

86 Total

Number of Leads Enrolled in Study	2,413	Conductor Fracture	3	Insulation (ESC)	2
Cumulative Months of Follow-Up	135,071	Extra Cardiac Stimulation	3	Insulation (MIO)	2
Number of Leads Active in Study	224	Failure to Capture	23	Insulation (not further defined)	2
		Failure to Sense	16	Lead Dislodgement	8
		Impedance Out of Range	11	Oversensing	13
				Unspecified Clinical Failure	3




Ventricular Placement

Product Surveillance Registry Results

Qualifying Complications

56 Total

Number of Leads Enrolled in Study	1,799	Conductor Fracture	3	Impedance Out of Range	16
Cumulative Months of Follow-Up	96,956	Extra Cardiac Stimulation	2	Insulation (not further defined)	1
Number of Leads Active in Study	128	Failure to Capture	22	Oversensing	7
		Failure to Sense	3	Unspecified Clinical Failure	2



Pacing Leads continued

4074 CapSure Sense

Product Characteristics


US Market Release	Jun-02	Serial Number Prefix	BBD	US Returned Product Analysis	
Registered US Implants	94,100	Type and/or Fixation	Transvenous, Right Ventricle, Tines	Conductor Fracture	3
Estimated Active US Implants	58,800	Polarity	Bipolar	Crimp/Weld/Bond	0
Advisories	None	Steroid	Yes	Insulation Breach	20
				Other	0

Atrial Placement

Product Surveillance Registry Results

Qualifying Complications 2 Total

Number of Leads Enrolled in Study	214	Failure to Sense	1
Cumulative Months of Follow-Up	15,716	Lead Dislodgement	1
Number of Leads Active in Study	132		




Ventricular Placement

Product Surveillance Registry Results

Qualifying Complications 3 Total

Number of Leads Enrolled in Study	988	Failure to Capture	1
Cumulative Months of Follow-Up	35,488	Impedance Out of Range	1
Number of Leads Active in Study	618	Lead Dislodgement	1



Pacing Leads continued

4076 CapSureFix Novus

Product Characteristics

Product Characteristics				US Returned Product Analysis	
US Market Release	Feb-04	Serial Number Prefix	BBL	Conductor Fracture	39
Registered US Implants	452,200	Type and/or Fixation	Transvenous, Atrium or Right Ventricle, Active Screw-in	Crimp/Weld/Bond	1
Estimated Active US Implants	347,600	Polarity	Bipolar	Insulation Breach	35
Advisories	None	Steroid	Yes	Other	18


Atrial Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,663
Cumulative Months of Follow-Up	65,593
Number of Leads Active in Study	894

Qualifying Complications

Qualifying Complications	9	Total
Conductor Fracture	1	Lead Dislodgement 3
Failure to Capture	2	
Failure to Sense	2	
Insulation (not further defined)	1	




Ventricular Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,229
Cumulative Months of Follow-Up	51,284
Number of Leads Active in Study	653

Qualifying Complications

Qualifying Complications	6	Total
Extra Cardiac Stimulation	1	
Failure to Capture	3	
Impedance Out of Range	2	



Leads

Pacing Leads continued

4092 CapSure SP Novus

Product Characteristics


US Market Release	Sep-98	Serial Number Prefix	LEP	US Returned Product Analysis	
Registered US Implants	179,200	Type and/or Fixation	Transvenous, Right Ventricle, Tines	Conductor Fracture	11
Estimated Active US Implants	80,700	Polarity	Bipolar	Crimp/Weld/Bond	0
Advisories	None	Steroid	Yes	Insulation Breach	53
				Other	23

Product Surveillance Registry Results

Qualifying Complications

19 Total

Number of Leads Enrolled in Study	1,147	Conductor Fracture	3	Impedance Out of Range	2
Cumulative Months of Follow-Up	66,685	Extra Cardiac Stimulation	1	Lead Dislodgement	4
Number of Leads Active in Study	261	Failure to Capture	9		



4524 CapSure SP

Product Characteristics


US Market Release	Oct-91	Serial Number Prefix	LAR	US Returned Product Analysis	
Registered US Implants	100,300	Type and/or Fixation	Transvenous, Atrium, J-Shape, Tines	Conductor Fracture	1
Estimated Active US Implants	23,200	Polarity	Bipolar	Crimp/Weld/Bond	0
Advisories	None	Steroid	Yes	Insulation Breach	67
				Other	3

Product Surveillance Registry Results

Qualifying Complications

6 Total

Number of Leads Enrolled in Study	911	Failure to Capture	3
Cumulative Months of Follow-Up	40,951	Failure to Sense	2
Number of Leads Active in Study	29	Lead Dislodgement	1



Pacing Leads continued

4558M Screw-In


Product Characteristics

Product Characteristics				US Returned Product Analysis	
US Market Release	Nov-94	Serial Number Prefix	LDC	Conductor Fracture	1
Registered US Implants	19,700	Type and/or Fixation	Transvenous, Atrium-J, Active Screw-in	Crimp/Weld/Bond	0
Estimated Active US Implants	3,800	Polarity	Bipolar	Insulation Breach	21
Advisories	None	Steroid	No	Other	20

Product Surveillance Registry Results

Qualifying Complications 12 Total

Number of Leads Enrolled in Study	539	Electrical Abandonment	1	Impedance Out of Range	2
Cumulative Months of Follow-Up	23,280	Failure to Capture	3	Insulation (not further defined)	2
Number of Leads Active in Study	5	Failure to Sense	2	Oversensing	2



Pacing Leads continued

4568 CapSureFix


Product Characteristics

US Market Release	Jan-97	Serial Number Prefix	LDD	US Returned Product Analysis	
Registered US Implants	69,500	Type and/or Fixation	Transvenous, Atrium, J-Shape, Screw-in	Conductor Fracture	5
Estimated Active US Implants	21,100	Polarity	Bipolar	Crimp/Weld/Bond	0
Advisories	None	Steroid	Yes	Insulation Breach	86
				Other	52

Product Surveillance Registry Results

Qualifying Complications 36 Total

Number of Leads Enrolled in Study	656	Failure to Capture	18	Lead Dislodgement	9
Cumulative Months of Follow-Up	32,664	Failure to Sense	4	Medical Judgment	1
Number of Leads Active in Study	115	Impedance Out of Range	4		



4574 CapSure Sense


Product Characteristics

US Market Release	Jun-02	Serial Number Prefix	BBE	US Returned Product Analysis	
Registered US Implants	62,900	Type and/or Fixation	Transvenous, Atrium, J-Shape, Tines	Conductor Fracture	7
Estimated Active US Implants	42,200	Polarity	Bipolar	Crimp/Weld/Bond	0
Advisories	None	Steroid	Yes	Insulation Breach	3
				Other	0

Product Surveillance Registry Results

Qualifying Complications 4 Total

Number of Leads Enrolled in Study	520	Failure to Capture	1
Cumulative Months of Follow-Up	7,525	Lead Dislodgement	3
Number of Leads Active in Study	387		



Pacing Leads continued

4592 CapSure SP Novus


Product Characteristics

Product Characteristics				US Returned Product Analysis	
US Market Release	Oct-98	Serial Number Prefix	LER	Conductor Fracture	7
Registered US Implants	86,300	Type and/or Fixation	Transvenous, Atrium, J-Shape, Tines	Crimp/Weld/Bond	0
Estimated Active US Implants	40,700	Polarity	Bipolar	Insulation Breach	18
Advisories	None	Steroid	Yes	Other	1

Product Surveillance Registry Results

Qualifying Complications 7 Total

Number of Leads Enrolled in Study	283	Failure to Capture	4
Cumulative Months of Follow-Up	14,887	Failure to Sense	1
Number of Leads Active in Study	56	Lead Dislodgement	2



Pacing Leads continued

5033 CapSure Z

Product Characteristics


US Market Release	Feb-96	Serial Number Prefix	LDK	US Returned Product Analysis	
Registered US Implants	2,300	Type and/or Fixation	Transvenous, Right Ventricle, Tines	Conductor Fracture	1
Estimated Active US Implants	500	Polarity	Unipolar	Crimp/Weld/Bond	0
Advisories	None	Steroid	Yes	Insulation Breach	0
				Other	3

Product Surveillance Registry Results

Qualifying Complications

31 Total

Number of Leads Enrolled in Study	1,899	Cardiac Perforation	1	Impedance Out of Range	4
Cumulative Months of Follow-Up	101,933	Conductor Fracture	8	Insulation (not further defined)	1
Number of Leads Active in Study	134	Failure to Capture	15	Lead Dislodgement	2



Pacing Leads continued

5034 CapSure Z

Product Characteristics

US Market Release	Feb-96	Serial Number Prefix	LDF	US Returned Product Analysis	
Registered US Implants	55,400	Type and/or Fixation	Transvenous, Right Ventricle, Tines	Conductor Fracture	12
Estimated Active US Implants	12,200	Polarity	Bipolar	Crimp/Weld/Bond	0
Advisories	None	Steroid	Yes	Insulation Breach	15
				Other	7


Atrial Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	386
Cumulative Months of Follow-Up	44,612
Number of Leads Active in Study	111

Qualifying Complications

Conductor Fracture	5	Total
Failure to Capture	1	
Failure to Sense	2	
Impedance Out of Range	1	




Ventricular Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,209
Cumulative Months of Follow-Up	44,680
Number of Leads Active in Study	11

Qualifying Complications

Conductor Fracture	11	Total
Lead Dislodgement	1	
Failure to Capture	7	
Failure to Sense	2	



Pacing Leads continued

5054 CapSure Z Novus

Product Characteristics


US Market Release	Jun-98	Serial Number Prefix	LEH	US Returned Product Analysis	
Registered US Implants	96,600	Type and/or Fixation	Transvenous, Right Ventricle, Tines	Conductor Fracture	10
Estimated Active US Implants	41,100	Polarity	Bipolar	Crimp/Weld/Bond	1
Advisories	None	Steroid	Yes	Insulation Breach	25
				Other	3

Atrial Placement

Product Surveillance Registry Results

Qualifying Complications 2 Total

Number of Leads Enrolled in Study	424	Failure to Capture	1
Cumulative Months of Follow-Up	34,515	Lead Dislodgement	1
Number of Leads Active in Study	115		




Ventricular Placement

Product Surveillance Registry Results

Qualifying Complications 9 Total

Number of Leads Enrolled in Study	967	Failure to Capture	6	Lead Dislodgement	1
Cumulative Months of Follow-Up	39,578	Failure to Sense	1		
Number of Leads Active in Study	84	Impedance Out of Range	1		



Pacing Leads continued

5068 CapSureFix

Product Characteristics

US Market Release	Jan-97	Serial Number Prefix	LDJ	US Returned Product Analysis	
Registered US Implants	102,400	Type and/or Fixation	Transvenous, Atrium or Right Ventricle, Active Screw-in	Conductor Fracture	41
Estimated Active US Implants	28,800	Polarity	Bipolar	Crimp/Weld/Bond	2
Advisories	None	Steroid	Yes	Insulation Breach	56
				Other	83


Atrial Placement

Product Surveillance Registry Results

Qualifying Complications

6 Total

Number of Leads Enrolled in Study	968	Failure to Capture	2	Lead Dislodgement	1
Cumulative Months of Follow-Up	34,293	Impedance Out of Range	1	Oversensing	1
Number of Leads Active in Study	29	Insulation (not further defined)	1		




Ventricular Placement

Product Surveillance Registry Results

Qualifying Complications

8 Total

Number of Leads Enrolled in Study	1,362	Conductor Fracture	1	Lead Dislodgement	1
Cumulative Months of Follow-Up	40,745	Failure to Capture	2	Oversensing	1
Number of Leads Active in Study	56	Impedance Out of Range	1	Unspecified Clinical Failure	1
		Insulation (not further defined)	1		



Pacing Leads continued

5072 SureFix

Product Characteristics


Product Characteristics				US Returned Product Analysis	
US Market Release	Jun-98	Serial Number Prefix	LEM	Conductor Fracture	3
Registered US Implants	10,000	Type and/or Fixation	Transvenous, Atrium or Right Ventricle, Fixed Screw	Crimp/Weld/Bond	0
Estimated Active US Implants	4,100	Polarity	Bipolar	Insulation Breach	6
Advisories	None	Steroid	Yes	Other	0

Product Surveillance Registry Results

Qualifying Complications

3 Total

Number of Leads Enrolled in Study	508	Cardiac Perforation	1
Cumulative Months of Follow-Up	28,227	Failure to Capture	1
Number of Leads Active in Study	50	Failure to Sense	1



Pacing Leads continued

5076 CapSureFix Novus

Product Characteristics


US Market Release	Aug-00	Serial Number Prefix	PJN	US Returned Product Analysis	
Registered US Implants	1,552,400	Type and/or Fixation	Transvenous, Atrium or Right Ventricle, Active Screw-in	Conductor Fracture	435
Estimated Active US Implants	981,100	Polarity	Bipolar	Crimp/Weld/Bond	0
Advisories	None	Steroid	Yes	Insulation Breach	452
				Other	186

Atrial Placement

Product Surveillance Registry Results

Qualifying Complications

			28	Total		
Number of Leads Enrolled in Study	2,744	Cardiac Perforation	1	Impedance Out of Range	4	
Cumulative Months of Follow-Up	142,851	Conductor Fracture	3	Insulation (not further defined)	1	
Number of Leads Active in Study	593	Extra Cardiac Stimulation	2	Lead Dislodgement	5	
		Failure to Capture	6	Oversensing	2	
		Failure to Sense	2	Other	2	




Ventricular Placement

Product Surveillance Registry Results

Qualifying Complications

			14	Total		
Number of Leads Enrolled in Study	1,539	Cardiac Perforation	1	Failure to Sense	1	
Cumulative Months of Follow-Up	69,416	Conductor Fracture	2	Impedance Out of Range	3	
Number of Leads Active in Study	247	Failure to Capture	5	Lead Dislodgement	2	



Pacing Leads continued

5086MRI CapSureFix Novus


Product Characteristics

US Market Release		Feb-11	Serial Number Prefix	LFP	US Returned Product Analysis	
Registered US Implants	128,800		Type and/or Fixation	Transvenous, Atrium or Right Ventricle, Active Screw-in	Conductor Fracture	3
Estimated Active US Implants	124,700		Polarity	Bipolar	Crimp/Weld/Bond	0
Advisories	None		Steroid	Yes	Insulation Breach	5
					Other	8

Product Surveillance Registry Results

Qualifying Complications 2 Total

Number of Leads Enrolled in Study	2,532	Lead Dislodgement	2
Cumulative Months of Follow-Up	25,622		
Number of Leads Active in Study	2,216		




Ventricular Placement

Product Surveillance Registry Results

Qualifying Complications 4 Total

Number of Leads Enrolled in Study	2,518	Failure to Capture	2
Cumulative Months of Follow-Up	25,436	Failure to Sense	1
Number of Leads Active in Study	2,206	Lead Dislodgement	1



Pacing Leads continued

5092 CapSure SP Novus

Product Characteristics


Product Characteristics				US Returned Product Analysis	
US Market Release	Jun-98	Serial Number Prefix	LET	Conductor Fracture	11
Registered US Implants	134,300	Type and/or Fixation	Transvenous, Right Ventricle, Tines	Crimp/Weld/Bond	0
Estimated Active US Implants	61,100	Polarity	Bipolar	Insulation Breach	38
Advisories	None	Steroid	Yes	Other	3

Product Surveillance Registry Results

Qualifying Complications

9 Total

Number of Leads Enrolled in Study	1,172	Extra Cardiac Stimulation	1
Cumulative Months of Follow-Up	48,445	Failure to Capture	2
Number of Leads Active in Study	125	Impedance Out of Range	1
		Lead Dislodgement	5



Pacing Leads continued

5534 CapSure Z

Product Characteristics

US Market Release	Feb-96	Serial Number Prefix	LDG
Registered US Implants	25,800	Type and/or Fixation	Transvenous, Atrium-J, Tines
Estimated Active US Implants	6,700	Polarity	Bipolar
Advisories	None	Steroid	Yes

US Returned Product Analysis


Conductor Fracture	3
Crimp/Weld/Bond	0
Insulation Breach	4
Other	4

Product Surveillance Registry Results

Qualifying Complications

6 Total

Number of Leads Enrolled in Study	264	Failure to Capture	5
Cumulative Months of Follow-Up	13,141	Impedance Out of Range	1
Number of Leads Active in Study	5		



5554 CapSure Z Novus

Product Characteristics

US Market Release	Jun-98	Serial Number Prefix	LEJ
Registered US Implants	62,500	Type and/or Fixation	Transvenous, Atrium-J, Tines
Estimated Active US Implants	29,100	Polarity	Bipolar
Advisories	None	Steroid	Yes

US Returned Product Analysis


Conductor Fracture	9
Crimp/Weld/Bond	0
Insulation Breach	18
Other	1

Product Surveillance Registry Results

Qualifying Complications

5 Total

Number of Leads Enrolled in Study	344	Failure to Capture	2	Oversensing	1
Cumulative Months of Follow-Up	15,883	Impedance Out of Range	1		
Number of Leads Active in Study	33	Lead Dislodgement	1		



Pacing Leads continued

5568 CapSureFix


Product Characteristics

US Market Release	Jan-97	Serial Number Prefix	LDN	US Returned Product Analysis	
Registered US Implants	92,200	Type and/or Fixation	Transvenous, Atrium-J, Active Screw-in	Conductor Fracture	13
Estimated Active US Implants	50,900	Polarity	Bipolar	Crimp/Weld/Bond	0
Advisories	None	Steroid	Yes	Insulation Breach	28
				Other	37

Product Surveillance Registry Results

Qualifying Complications 13 Total

Number of Leads Enrolled in Study	1,053	Conductor Fracture	1	Failure to Sense	2
Cumulative Months of Follow-Up	41,190	Extra Cardiac Stimulation	1	Lead Dislodgement	1
Number of Leads Active in Study	82	Failure to Capture	6	Oversensing	2



5592 CapSure SP Novus


Product Characteristics

US Market Release	Jun-98	Serial Number Prefix	LEU	US Returned Product Analysis	
Registered US Implants	35,000	Type and/or Fixation	Transvenous, Atrium-J, Tines	Conductor Fracture	4
Estimated Active US Implants	19,000	Polarity	Bipolar	Crimp/Weld/Bond	0
Advisories	None	Steroid	Yes	Insulation Breach	4
				Other	0

Product Surveillance Registry Results

Qualifying Complications 5 Total

Number of Leads Enrolled in Study	672	Failure to Capture	3
Cumulative Months of Follow-Up	31,531	Lead Dislodgement	2
Number of Leads Active in Study	106		



Pacing Leads continued

5594 CapSure SP Novus


Product Characteristics

US Market Release	Jun-01	Serial Number Prefix	LFD	US Returned Product Analysis	
Registered US Implants	16,100	Type and/or Fixation	Transvenous, Atrium-J, Tines	Conductor Fracture	5
Estimated Active US Implants	10,500	Polarity	Bipolar	Crimp/Weld/Bond	0
Advisories	None	Steroid	Yes	Insulation Breach	7
				Other	1

Product Surveillance Registry Results

Qualifying Complications 0 Total

Number of Leads Enrolled in Study	20
Cumulative Months of Follow-Up	1,568
Number of Leads Active in Study	9



6940 CapSureFix


Product Characteristics

US Market Release	Oct-98	Serial Number Prefix	TCP	US Returned Product Analysis	
Registered US Implants	25,400	Type and/or Fixation	Transvenous, Atrium-J, Active Screw-in	Conductor Fracture	12
Estimated Active US Implants	7,100	Polarity	Bipolar	Crimp/Weld/Bond	0
Advisories	None	Steroid	Yes	Insulation Breach	18
				Other	12

Product Surveillance Registry Results

Qualifying Complications 13 Total

Number of Leads Enrolled in Study	818	Conductor Fracture	1	Lead Dislodgement	3
Cumulative Months of Follow-Up	44,783	Failure to Capture	1	Oversensing	5
Number of Leads Active in Study	67	Failure to Sense	3		



Lead Survival Summary (95% Confidence Interval)

Model Number	Family	Chamber	US Market Release	Leads Enrolled	Leads Active in Study	Qualifying Complications	Cumulative Months of Follow-Up in Study	Device Survival Probability (%)																							
								Years After Implant																							
								1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	10 yr	12 yr	14 yr	16 yr	18 yr	20 yr										
3830	SelectSecure	Atrial	Aug-05	778	493	7	25,544	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0					
3830	SelectSecure	Vent	Aug-05	482	283	4	17,389	99.3 +0.4/-1.4	99.3 +0.4/-1.4	99.3 +0.4/-1.4	99.3 +0.4/-1.4	99.3 +0.4/-1.4	99.3 +0.4/-1.4	99.3 +0.4/-1.4	99.3 +0.4/-1.4	99.3 +0.4/-1.4	99.3 +0.4/-1.4	99.3 +0.4/-1.4	99.3 +0.4/-1.4	99.3 +0.4/-1.4	99.3 +0.4/-1.4	99.3 +0.4/-1.4	99.3 +0.4/-1.4	99.3 +0.4/-1.4	99.3 +0.4/-1.4	99.3 +0.4/-1.4	99.3 +0.4/-1.4				
4024	CapSure SP	Vent	Oct-91	1,215	10	4	50,987	99.9 +0.1/-0.5	99.8 +0.2/-0.7	99.8 +0.2/-0.7	99.8 +0.2/-0.7	99.8 +0.2/-0.7	99.8 +0.2/-0.7	99.8 +0.2/-0.7	99.8 +0.2/-0.7	99.8 +0.2/-0.7	99.8 +0.2/-0.7	99.8 +0.2/-0.7	99.8 +0.2/-0.7	99.8 +0.2/-0.7	99.8 +0.2/-0.7	99.8 +0.2/-0.7	99.8 +0.2/-0.7	99.8 +0.2/-0.7	99.8 +0.2/-0.7	99.8 +0.2/-0.7	99.8 +0.2/-0.7	99.8 +0.2/-0.7			
4068	CapSureFix	Atrial	Mar-96	2,413	224	86	135,071	98.9 +0.4/-0.5	98.7 +0.4/-0.6	98.2 +0.5/-0.7	98.0 +0.6/-0.8	97.4 +0.7/-0.9	97.2 +0.7/-1.0	96.5 +0.9/-1.2	94.9 +1.3/-1.7	93.1 +1.7/-2.2	90.3 +2.4/-3.2	85.8 +4.1/-5.7													
4068	CapSureFix	Vent	Mar-96	1,799	128	56	96,956	99.3 +0.3/-0.6	98.7 +0.5/-0.7	98.7 +0.5/-0.7	98.2 +0.6/-0.9	97.8 +0.7/-1.0	97.0 +0.9/-1.3	95.6 +1.3/-1.8	94.6 +1.5/-2.1	92.2 +2.1/-2.9	89.4 +3.4/-4.9														
4074	CapSure Sense	Atrial	Jun-02	214	132	2	15,716	99.1 +0.7/-2.8	99.1 +0.7/-2.8	99.1 +0.7/-2.8	99.1 +0.7/-2.8	99.1 +0.7/-2.8	99.1 +0.7/-2.8	99.1 +0.7/-2.8	99.1 +0.7/-2.8	99.1 +0.7/-2.8	99.1 +0.7/-2.8	99.1 +0.7/-2.8	99.1 +0.7/-2.8	99.1 +0.7/-2.8	99.1 +0.7/-2.8	99.1 +0.7/-2.8	99.1 +0.7/-2.8	99.1 +0.7/-2.8	99.1 +0.7/-2.8	99.1 +0.7/-2.8	99.1 +0.7/-2.8	99.1 +0.7/-2.8			
4074	CapSure Sense	Vent	Jun-02	998	618	3	35,488	99.8 +0.2/-0.7	99.6 +0.3/-0.9	99.6 +0.3/-0.9	99.6 +0.3/-0.9	99.6 +0.3/-0.9	99.6 +0.3/-0.9	99.6 +0.3/-0.9	99.6 +0.3/-0.9	99.6 +0.3/-0.9	99.6 +0.3/-0.9	99.6 +0.3/-0.9	99.6 +0.3/-0.9	99.6 +0.3/-0.9	99.6 +0.3/-0.9	99.6 +0.3/-0.9	99.6 +0.3/-0.9	99.6 +0.3/-0.9	99.6 +0.3/-0.9	99.6 +0.3/-0.9	99.6 +0.3/-0.9	99.6 +0.3/-0.9	99.6 +0.3/-0.9		
4076	CapSureFix Novus	Atrial	Feb-04	1,663	894	9	65,593	99.8 +0.2/-0.4	99.6 +0.2/-0.5	99.4 +0.3/-0.6	99.4 +0.3/-0.6	99.4 +0.3/-0.6	99.4 +0.3/-0.6	99.4 +0.3/-0.6	99.4 +0.3/-0.6	99.4 +0.3/-0.6	99.4 +0.3/-0.6	99.4 +0.3/-0.6	99.4 +0.3/-0.6	99.4 +0.3/-0.6	99.4 +0.3/-0.6	99.4 +0.3/-0.6	99.4 +0.3/-0.6	99.4 +0.3/-0.6	99.4 +0.3/-0.6	99.4 +0.3/-0.6	99.4 +0.3/-0.6	99.4 +0.3/-0.6	99.4 +0.3/-0.6	99.4 +0.3/-0.6	
4076	CapSureFix Novus	Vent	Feb-04	1,229	653	6	51,284	99.8 +0.1/-0.5	99.8 +0.1/-0.5	99.7 +0.2/-0.6	99.6 +0.3/-0.8	99.0 +0.6/-1.4	99.0 +0.6/-1.4	99.0 +0.6/-1.4	99.0 +0.6/-1.4	99.0 +0.6/-1.4	99.0 +0.6/-1.4	99.0 +0.6/-1.4	99.0 +0.6/-1.4	99.0 +0.6/-1.4	99.0 +0.6/-1.4	99.0 +0.6/-1.4	99.0 +0.6/-1.4	99.0 +0.6/-1.4	99.0 +0.6/-1.4	99.0 +0.6/-1.4	99.0 +0.6/-1.4	99.0 +0.6/-1.4	99.0 +0.6/-1.4	99.0 +0.6/-1.4	99.0 +0.6/-1.4
4092	CapSure SP Novus	Vent	Sep-98	1,147	261	19	66,685	98.9 +0.5/-0.9	98.8 +0.5/-0.9	98.7 +0.6/-0.9	98.4 +0.6/-1.1	98.1 +0.7/-1.2	97.7 +0.9/-1.3	97.7 +0.9/-1.3	97.7 +0.9/-1.3	97.7 +0.9/-1.3	97.7 +0.9/-1.3	97.7 +0.9/-1.3	97.7 +0.9/-1.3	97.7 +0.9/-1.3	97.7 +0.9/-1.3	97.7 +0.9/-1.3	97.7 +0.9/-1.3	97.7 +0.9/-1.3	97.7 +0.9/-1.3	97.7 +0.9/-1.3	97.7 +0.9/-1.3	97.7 +0.9/-1.3	97.7 +0.9/-1.3	97.7 +0.9/-1.3	97.7 +0.9/-1.3

continued

Pacing Leads continued

Lead Survival Summary continued

Model Number	Family	Chamber	US Market Release	Leads Enrolled	Leads Active in Study	Qualifying Complications	Cumulative Months of Follow-Up in Study	Device Survival Probability (%)																
								Years After Implant																
								1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	10 yr	12 yr	14 yr	16 yr	18 yr	20 yr			
4524	CapSure SP	Atrial	Oct-91	911	29	6	40,951	99.6 +0.2/-0.8	99.3 +0.4/-1.0	99.3 +0.4/-1.0	99.0 +0.5/-1.3	99.0 +0.5/-1.3	99.0 +0.5/-1.3	99.0 +0.5/-1.3	99.0 +0.5/-1.3	99.0 +0.5/-1.3	99.0 +0.5/-1.3 at 132 mo							
4558M	Screw-In	Atrial	Nov-94	539	5	12	23,280	99.3 +0.5/-1.4	99.3 +0.5/-1.4	99.3 +0.5/-1.4	99.3 +0.5/-1.4	99.3 +0.5/-1.4	97.6 +1.5/-4.2	96.5 +2.1/-5.2	96.5 +2.1/-5.2	91.8 +4.3/-8.7 at 108 mo								
4568	CapSureFix	Atrial	Jan-97	656	115	36	32,664	96.8 +1.2/-1.8	96.4 +1.3/-1.9	95.3 +1.5/-2.2	94.7 +1.7/-2.4	94.0 +1.8/-2.6	94.0 +1.8/-2.6	92.9 +2.2/-3.2	92.2 +2.5/-3.5	90.8 +3.3/-5.0 at 111 mo								
4574	CapSure Sense	Atrial	Jun-02	520	387	4	7,525	99.0 +0.7/-1.8	99.0 +0.7/-1.8	99.0 +0.7/-1.8 at 30 mo														
4592	CapSure SP Novus	Atrial	Oct-98	283	56	7	14,887	98.2 +1.1/-3.0	98.2 +1.1/-3.0	98.2 +1.1/-3.0	98.2 +1.1/-3.0	97.4 +1.6/-3.8	97.4 +1.6/-3.8	97.4 +1.6/-3.8	96.0 +2.3/-5.5 at 108 mo									
5033	CapSure Z	Vent	Feb-96	1,899	134	31	101,933	99.7 +0.2/-0.4	99.6 +0.2/-0.5	99.1 +0.4/-0.7	99.0 +0.4/-0.8	98.8 +0.5/-0.9	98.3 +0.7/-1.1	97.7 +0.9/-1.4	96.9 +1.1/-1.7	96.0 +1.4/-2.1	95.6 +1.6/-2.4	93.3 +2.5/-4.0 at 186 mo						
5034	CapSure Z	Atrial	Feb-96	386	111	5	45,612	99.5 +0.4/-1.5	99.5 +0.4/-1.5	99.5 +0.4/-1.5	99.5 +0.4/-1.5	99.5 +0.4/-1.5	99.5 +0.4/-1.5	98.8 +0.7/-1.9	98.8 +0.7/-1.9	98.8 +0.7/-1.9	98.8 +0.7/-1.9	97.8 +1.4/-4.1 at 168 mo						
5034	CapSure Z	Vent	Feb-96	1,209	11	11	44,680	99.7 +0.2/-0.6	99.4 +0.3/-0.8	99.1 +0.5/-1.0	98.8 +0.6/-1.1	98.8 +0.6/-1.1	98.8 +0.6/-1.1	98.1 +1.1/-2.7	96.6 +2.2/-5.6 at 96 mo									
5054	CapSure Z Novus	Atrial	Jun-98	424	115	2	34,515	99.5 +0.4/-1.4	99.5 +0.4/-1.4	99.5 +0.4/-1.4	99.5 +0.4/-1.4	99.5 +0.4/-1.4	99.5 +0.4/-1.4	99.5 +0.4/-1.4	99.5 +0.4/-1.4	99.5 +0.4/-1.4	99.5 +0.4/-1.4 at 126 mo							
5054	CapSure Z Novus	Vent	Jun-98	967	84	9	39,578	99.5 +0.3/-0.8	99.4 +0.4/-0.9	99.4 +0.4/-0.9	99.1 +0.5/-1.2	99.1 +0.5/-1.2	99.1 +0.5/-1.2	97.9 +1.1/-2.3	97.9 +1.1/-2.3	97.9 +1.1/-2.3	97.9 +1.1/-2.3	97.9 +1.1/-2.3	97.9 +1.1/-2.3	97.9 +1.1/-2.3	97.9 +1.1/-2.3	97.9 +1.1/-2.3 at 123 mo		

continued

Lead Survival Summary continued

Model Number	Family	Chamber	US Market Release	Leads Enrolled	Leads Active in Study	Qualifying Complications	Cumulative Months of Follow-Up in Study	Device Survival Probability (%)																					
								Years After Implant																					
								1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	10 yr	12 yr	14 yr	16 yr	18 yr	20 yr								
5068	CapSureFix	Atrial	Jan-97	968	29	6	34,293	99.6 +0.3/-0.9	99.6 +0.3/-0.9	99.3 +0.4/-1.2	99.3 +0.4/-1.2	99.3 +0.4/-1.2	99.3 +0.4/-1.2	99.3 +0.4/-1.2	99.3 +0.4/-1.2	99.3 +0.4/-1.2	99.3 +0.4/-1.2	97.6 +1.5/-4.2 at 126 mo	97.6 +1.5/-4.2 at 126 mo										
5068	CapSureFix	Vent	Jan-97	1,362	56	8	40,745	99.9 +0.1/-0.7	99.7 +0.2/-0.9	99.5 +0.4/-1.2	99.2 +0.5/-1.5	99.2 +0.5/-1.5	98.5 +0.9/-2.6	97.2 +1.6/-3.7	97.2 +1.6/-3.7	97.2 +1.6/-3.7	97.2 +1.6/-3.7	97.2 +1.6/-3.7	97.2 +1.6/-3.7	97.2 +1.6/-3.7	97.2 +1.6/-3.7	97.2 +1.6/-3.7	97.2 +1.6/-3.7	97.2 +1.6/-3.7	97.2 +1.6/-3.7	97.2 +1.6/-3.7			
5072	SureFix	A or V	Jun-98	508	50	3	28,227	99.8 +0.2/-1.4	99.8 +0.2/-1.4	99.4 +0.4/-1.7	99.4 +0.4/-1.7	99.4 +0.4/-1.7	99.4 +0.4/-1.7	99.4 +0.4/-1.7	99.4 +0.4/-1.7	99.4 +0.4/-1.7	99.4 +0.4/-1.7	99.4 +0.4/-1.7	98.5 +1.1/-4.1	98.5 +1.1/-4.1	98.5 +1.1/-4.1	98.5 +1.1/-4.1	98.5 +1.1/-4.1	98.5 +1.1/-4.1	98.5 +1.1/-4.1	98.5 +1.1/-4.1	98.5 +1.1/-4.1		
5076	CapSureFix Novus	Atrial	Aug-00	2,744	593	28	142,851	99.6 +0.2/-0.3	99.6 +0.2/-0.4	99.4 +0.3/-0.5	98.9 +0.4/-0.6	98.7 +0.5/-0.7	98.5 +0.5/-0.8	98.5 +0.5/-0.8	98.3 +0.6/-0.8	98.3 +0.6/-0.8	98.3 +0.6/-0.8	98.3 +0.6/-0.8	97.8 +0.9/-1.6 at 129 mo	97.8 +0.9/-1.6 at 129 mo									
5076	CapSureFix Novus	Vent	Aug-00	1,539	247	14	69,416	99.6 +0.2/-0.5	99.4 +0.3/-0.6	99.3 +0.4/-0.7	99.0 +0.5/-0.9	99.0 +0.5/-0.9	98.6 +0.7/-1.2	98.3 +0.8/-1.4	98.3 +0.8/-1.4	98.3 +0.8/-1.4	98.3 +0.8/-1.4	98.3 +0.8/-1.4	97.7 +1.2/-2.3 at 120 mo	97.7 +1.2/-2.3 at 120 mo									
5086MRI	CapSureFix Novus	Atrial	Feb-11	2,532	2,216	2	25,622	99.9 +0.1/-0.3	99.9 +0.1/-0.3 at 21 mo																				
5086MRI	CapSureFix Novus	Vent	Feb-11	2,518	2,206	4	25,436	99.8 +0.1/-0.3	99.8 +0.1/-0.3 at 21 mo																				
5092	CapSure SP Novus	Vent	Jun-98	1,172	125	9	48,445	99.5 +0.3/-0.7	99.4 +0.3/-0.8	99.2 +0.4/-0.9	99.0 +0.5/-1.1	99.0 +0.5/-1.1	98.6 +0.8/-1.6	98.6 +0.8/-1.6	98.6 +0.8/-1.6	98.6 +0.8/-1.6	98.6 +0.8/-1.6	98.6 +0.8/-1.6	98.6 +0.8/-1.6	98.6 +0.8/-1.6	98.6 +0.8/-1.6	98.6 +0.8/-1.6	98.6 +0.8/-1.6	98.6 +0.8/-1.6	98.6 +0.8/-1.6	98.6 +0.8/-1.6	98.6 +0.8/-1.6		
5534	CapSure Z	Atrial	Feb-96	264	5	6	13,141	98.3 +1.0/-2.7	97.8 +1.3/-3.0	97.8 +1.3/-3.0	97.1 +1.6/-3.5	97.1 +1.6/-3.5	97.1 +1.6/-3.5	97.1 +1.6/-3.5	97.1 +1.6/-3.5	97.1 +1.6/-3.5	97.1 +1.6/-3.5	97.1 +1.6/-3.5	97.1 +1.6/-3.5	97.1 +1.6/-3.5	97.1 +1.6/-3.5	97.1 +1.6/-3.5	97.1 +1.6/-3.5	97.1 +1.6/-3.5	97.1 +1.6/-3.5	97.1 +1.6/-3.5	97.1 +1.6/-3.5	97.1 +1.6/-3.5	
5554	CapSure Z Novus	Atrial	Jun-98	344	33	5	15,883	100.0	99.2 +0.6/-2.4	98.7 +0.9/-2.6	98.2 +1.1/-3.0	98.2 +1.1/-3.0	98.2 +1.1/-3.0	98.2 +1.1/-3.0	98.2 +1.1/-3.0	98.2 +1.1/-3.0	98.2 +1.1/-3.0	98.2 +1.1/-3.0	98.2 +1.1/-3.0	98.2 +1.1/-3.0	98.2 +1.1/-3.0	98.2 +1.1/-3.0	98.2 +1.1/-3.0	98.2 +1.1/-3.0	98.2 +1.1/-3.0	98.2 +1.1/-3.0	98.2 +1.1/-3.0	98.2 +1.1/-3.0	
5568	CapSureFix	Atrial	Jan-97	1,053	82	13	41,190	99.7 +0.2/-0.7	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	99.2 +0.4/-1.0	
5592	CapSure SP Novus	Atrial	Jun-98	672	106	5	31,531	99.7 +0.3/-1.0	99.3 +0.5/-1.2	99.3 +0.5/-1.2	98.9 +0.6/-1.6	98.9 +0.6/-1.6	98.9 +0.6/-1.6	98.9 +0.6/-1.6	98.9 +0.6/-1.6	98.9 +0.6/-1.6	98.9 +0.6/-1.6	98.9 +0.6/-1.6	98.9 +0.6/-1.6	98.9 +0.6/-1.6	98.9 +0.6/-1.6	98.9 +0.6/-1.6	98.9 +0.6/-1.6	98.9 +0.6/-1.6	98.9 +0.6/-1.6	98.9 +0.6/-1.6	98.9 +0.6/-1.6	98.9 +0.6/-1.6	
5594	CapSure SP Novus	Atrial	Jun-01	20	9	0	1,568	100.0 at 0 mo																					
6940	CapSureFix	Atrial	Oct-98	818	67	13	44,783	99.7 +0.2/-0.8	99.6 +0.3/-0.9	98.3 +0.8/-1.5	98.1 +0.9/-1.6	98.1 +0.9/-1.6	98.1 +0.9/-1.6	98.1 +0.9/-1.6	98.1 +0.9/-1.6	98.1 +0.9/-1.6	98.1 +0.9/-1.6	98.1 +0.9/-1.6	98.1 +0.9/-1.6	98.1 +0.9/-1.6	98.1 +0.9/-1.6	98.1 +0.9/-1.6	98.1 +0.9/-1.6	98.1 +0.9/-1.6	98.1 +0.9/-1.6	98.1 +0.9/-1.6	98.1 +0.9/-1.6	98.1 +0.9/-1.6	93.7 +3.5/-7.6 at 141 mo

US Returned Product Analysis Summary

Model Number	Family	US Market Release	Estimated US Implants	Estimated US Active	Conductor Fracture	Insulation Breach	Crimp/Weld/Bond	Other
3830	SelectSecure	Aug-05	21,600	16,400	4	14	0	3
4024	CapSure SP	Oct-91	219,400	40,700	28	183	0	12
4068	CapSureFix	Mar-96	124,300	29,800	50	167	0	93
4074	CapSure Sense	Jun-02	94,100	58,800	3	20	0	0
4076	CapSureFix Novus	Feb-04	452,200	347,600	39	35	1	18
4092	CapSure SP Novus	Sep-98	179,200	80,700	11	53	0	2
4524	CapSure SP	Oct-91	100,300	23,200	1	67	0	3
4558M	Screw-in	Nov-94	19,700	3,800	1	21	0	20
4568	CapSureFix	Jan-97	69,500	21,100	5	86	0	52
4574	CapSure Sense	Jun-02	62,900	42,200	7	3	0	0
4592	CapSure SP Novus	Oct-98	86,300	40,700	7	18	0	1
5033	CapSure Z	Feb-96	2,300	500	1	0	0	3
5034	CapSure Z	Feb-96	55,400	12,200	12	15	0	7
5054	CapSure Z Novus	Jun-98	96,600	41,100	10	25	1	3
5068	CapSureFix	Jan-97	102,400	28,800	41	56	2	83
5072	SureFix	Jun-98	10,000	4,100	3	6	0	0
5076	CapSureFix Novus	Aug-00	1,552,400	981,100	435	452	0	186
5086MRI	CapSureFix Novus MRI	Feb-11	128,800	124,700	3	5	0	8
5092	CapSure SP Novus	Jun-98	134,300	61,100	11	38	0	3
5534	CapSure Z	Feb-96	25,800	6,700	3	4	0	4
5554	CapSure Z Novus	Jun-98	62,500	29,100	9	18	0	1
5568	CapSureFix	Jan-97	92,900	50,900	13	28	0	37
5592	CapSure SP Novus	Jun-98	35,000	19,000	4	4	0	0
5594	CapSure SP Novus	Jun-01	16,100	10,500	5	7	0	1
6940	CapSureFix	Oct-98	25,400	7,100	12	18	0	12

Source: Returned Product Analysis
Data as of July 31, 2013

US Reports of Acute Lead Observations

Model Number	Family	Estimated US Implants	Cardiac Perforation	Conductor Fracture	Lead Dislodgement	Failure to Capture	Oversensing	Failure to Sense	Insulation Breach	Extracardiac Stimulation	Impedance Abnormal	Unspecified
3830	SelectSecure	21,600	6	1	29	14	3	1	1	0	0	2
4024	CapSure SP	219,400	13	11	49	103	1	16	1	2	8	20
4068	CapSureFix	124,300	5	3	31	23	0	5	1	1	2	4
4074	CapSure Sense	94,100	10	1	32	30	0	1	0	0	3	0
4076	CapSureFix Novus	452,200	54	4	134	66	6	17	1	9	12	12
4092	CapSure SP Novus	179,200	1	4	22	29	0	0	1	2	0	2
4524	CapSure SP	100,300	2	2	23	15	0	4	2	1	0	12
4558M	Screw-in	19,700	2	0	2	2	0	1	0	1	1	1
4568	CapSureFix	69,500	3	1	4	6	1	1	0	0	2	1
4574	CapSure Sense	62,900	0	1	39	16	1	8	0	0	0	4
4592	CapSure SP Novus	86,300	0	0	25	8	1	1	1	0	0	2
5033	CapSure Z	2,300	0	0	0	0	0	0	0	0	0	1
5034	CapSure Z	55,400	2	2	14	28	0	3	3	0	0	12
5054	CapSure Z Novus	96,600	1	1	22	22	0	0	1	0	2	9
5068	CapSureFix	102,400	16	4	20	31	1	5	1	0	1	7
5072	SureFix	10,000	0	0	2	2	0	0	0	0	0	0
5076	CapSureFix Novus	1,552,400	167	11	529	241	28	36	8	13	14	31
5086MRI	CapsureFix Novus	128,800	119	2	151	74	20	13	1	10	4	0
5092	CapSure SP Novus	134,300	5	1	51	34	1	6	3	3	0	9
5534	CapSure Z	25,800	0	0	6	3	0	1	0	1	0	4
5554	CapSure Z Novus	62,500	0	1	31	28	0	2	0	0	0	3
5568	CapSureFix	92,200	8	0	33	19	2	2	1	2	2	4
5592	CapSure SP Novus	35,000	1	0	27	4	0	2	0	0	0	1
5594	CapSure SP Novus	16,100	0	0	8	0	0	0	0	0	0	2
6940	CapSureFix	25,400	0	1	6	1	0	0	0	0	1	0

Report Cutoff Date: September 10, 2013

Reference Chart

Model Number	Family	Type	Insulation	Conductor Material	Tip Electrode	Connector Type
3830	SelectSecure	Transvenous V or A Screw-In	Polyurethane/Silicone (55D,4719)	MP35N 5 Filars/ Cable	1.8 mm Helix/Steroid	IS-1 BI
4024	CapSure SP	Transvenous Ventricular Tines	Polyurethane (55D)	MP35N 4/5 Filars	Porous Platinized/ Steroid	IS-1 BI
4068	CapSureFix	Transvenous V or A Screw-In	Polyurethane (80A/55D)	MP35N 4/3 Filars	1.8 mm Helix/Steroid	IS-1 BI
4074	CapSure Sense	Transvenous Ventricular Tines	Polyurethane/Silicone (55D, 4719)	MP35N 5/5 Filars	TiN Coated Platinum Iridium/ Steroid	IS-1 BI
4076	CapSureFix Novus	Transvenous V or A Screw-In	Polyurethane/Silicone (55D, 4719)	MP35N 4/6 Filars	TiN Coated Platinum Alloy/Steroid	IS-1 BI
4092	CapSure SP Novus	Transvenous Ventricular Tines	Polyurethane/Silicone (55D/4719)	MP35N 6/4 Filars	Porous Platinized/ Steroid	IS-1 BI
4524	CapSure SP	Transvenous Atrial-J Tines	Polyurethane (55D)	MP35N 4/5 Filars	Porous Platinized/ Steroid	IS-1 BI
4558M	Screw-In	Transvenous Atrial-J Screw-In	Polyurethane (80A/55D)	MP35N 6/3 Filars	1.8 mm Helix/Steroid	IS-1 BI
4568	CapSureFix	Transvenous Atrial-J Screw-In	Polyurethane (80A/55D)	MP35N 6/3 Filars	1.8 mm Helix/Steroid	IS-1 BI
4574	CapSure Sense	Transvenous Atrial-J Tines	Polyurethane/Silicone (55D,4719)	MP35N 5/5 Filars	TiN Coated Platinum Iridium	IS-1 BI
4592	CapSure SP Novus	Transvenous Atrial-J Tines	Polyurethane/Silicone (55D/4719)	MP35N 6/3 Filars	Porous Platinized/ Steroid	IS-1 BI
5033	CapSure Z	Transvenous Ventricular Tines	Silicone	MP35N 4 Filars	CapSure Z Platinized/Steroid	IS-1 UNI
5034	CapSure Z	Transvenous Ventricular Tines	Silicone	MP35N 4/5 Filars	CapSure Z Platinized/Steroid	IS-1 BI
5054	CapSure Z Novus	Transvenous Ventricular Tines	Silicone (4719)	MP35N 5/5 Filars	CapSure Z Porous/Platinized/ Steroid	IS-1 BI
5068	CapSureFix	Transvenous V or A Screw-In	Silicone	MP35N 4/3 Filars	1.8 mm Helix/Steroid	IS-1 BI
5072	SureFix	Transvenous V or A Screw-In	Silicone	MP35N 4/5 Filars	1.8 mm Helix/Steroid	IS-1 BI
5076	CapSureFix Novus	Transvenous V or A Screw-In	Silicone (4719)	MP35N 4/6 Filars	Porous Platinized/ Steroid	IS-1 BI
5086MRI	CapSureFix Novus	Transvenous A or V Screw-in	Silicone	MP35N	Titanium nitride coated platinum alloy	IS-1 BI
5092	CapSure SP Novus	Transvenous Ventricular Tines	Silicone (4719)	MP35N 5/5 Filars	Porous Platinized/ Steroid	IS-1 BI
5534	CapSure Z	Transvenous Atrial-J Tines	Silicone	MP35N 4/5 Filars	CapSure Z Platinized/Steroid	IS-1 BI
5554	CapSure Z Novus	Transvenous Atrial-J Tines	Silicone (4719)	MP35N 6/5 Filars	CapSure Z Porous Platinized/ Steroid	IS-1 BI
5568	CapSureFix	Transvenous Atrial-J Screw-In	Silicone	MP35N 6/3 Filars	1.8 mm Helix/Steroid	IS-1 BI
5592	CapSure SP Novus	Transvenous Atrial-J Tines	Silicone (4719)	MP35N 6/5 Filars	Porous Platinized/ Steroid	IS-1 BI
5594	CapSure SP Novus	Transvenous Atrial-J Tines	Silicone (4719)	MP35N 6/5 Filars	Platinized Platinum/ Steroid	IS-1 BI
6940	CapSureFix	Transvenous Atrial-J Screw-In	Silicone	MP35N 3/6 Filars	Platinum Alloy	IS-1 BI

Epi/Myocardial Pacing Leads

4965 CapSure Epi

Product Characteristics


Product Characteristics				US Returned Product Analysis	
US Market Release	Sep-96	Serial Number Prefix	LBT	Conductor Fracture	171
Registered US Implants	20,900	Type and/or Fixation	Myocardial, Atrium or Right Ventricle, Suture	Crimp/Weld/Bond	1
Estimated Active US Implants	9,400	Polarity	Unipolar	Insulation Breach	37
Advisories	None	Steroid	Yes	Other	0

Product Surveillance Registry Results

Qualifying Complications

12 Total

Number of Leads Enrolled in Study	219	Conductor Fracture	5	Insulation (not further defined)	1
Cumulative Months of Follow-Up	7,786	Failure to Capture	3	Oversensing	2
Number of Leads Active in Study	24	Failure to Sense	1		



Epi/Myocardial Pacing Leads continued

4968 CapSure Epi


Product Characteristics

US Market Release	Sep-99	Serial Number Prefix	LEN	US Returned Product Analysis	
Registered US Implants	29,000	Type and/or Fixation	Myocardial, Atrium or Right Ventricle, Suture	Conductor Fracture	35
Estimated Active US Implants	17,900	Polarity	Bipolar	Crimp/Weld/Bond	0
Advisories	None	Steroid	Yes	Insulation Breach	17
				Other	1

Product Surveillance Registry Results

Qualifying Complications 59 Total

Number of Leads Enrolled in Study	819	Conductor Fracture	12	Impedance Out of Range	5
Cumulative Months of Follow-Up	41,146	Extra Cardiac Stimulation	2	Insulation (not further defined)	3
Number of Leads Active in Study	312	Failure to Capture	21	Oversensing	12
		Failure to Sense	3	Other	1



5071 Screw-in


Product Characteristics

US Market Release	Dec-92	Serial Number Prefix	LAQ	US Returned Product Analysis	
Registered US Implants	45,200	Type and/or Fixation	Myocardial, Right Ventricle, Fixed Screw-in.	Conductor Fracture	13
Estimated Active US Implants	15,400	Polarity	Unipolar	Crimp/Weld/Bond	0
Advisories	None	Steroid	No	Insulation Breach	2
				Other	0

Product Surveillance Registry Results

Qualifying Complications 15 Total

Number of Leads Enrolled in Study	302	Failure to Capture	12
Cumulative Months of Follow-Up	8,521	Impedance Out of Range	1
Number of Leads Active in Study	63	Oversensing	2



Epi/Myocardial Pacing Leads continued

Lead Survival Summary (95% Confidence Interval)

Model Number	Family	US Market Release	Leads Enrolled	Leads Active in Study	Qualifying Complications	Cumulative Months of Follow-Up in Study	Device Survival Probability (%)											
							Years After Implant											
							1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	10 yr	12 yr	14 yr	16 yr
4965	CapSure Epi	Sep-96	219	24	12	7,786	99.0 +0.8/-3.0	98.2 +1.3/-3.9	97.2 +1.8/-4.7	92.8 +3.6/-7.0	88.1 +5.4/-9.3 at 54 mo							
4968	CapSure Epi	Sep-99	819	312	59	41,146	99.5 +0.3/-0.9	97.67 +0.9/-1.5	95.9 +1.3/-2.0	93.6 +1.8/-2.5	92.5 +2.0/-2.8	90.9 +2.4/-3.2	89.0 +2.9/-3.8	89.0 +2.9/-3.8	84.9 +4.4/-6.0 at 120 mo			
5071	Screw-in	Dec-92	302	63	15	8,521	96.9 +1.6/-3.3	94.0 +2.7/-4.9	91.8 +3.6/-6.1	91.8 +3.6/-6.1	90.1 +4.3/-7.3 at 60 mo							

Epi/Myocardial Pacing Leads continued

US Returned Product Analysis Summary

Model Number	Family	US Market Release	Estimated US Implants	Estimated US Active	Conductor Fracture	Insulation Breach	Crimp/Weld/Bond	Other
4965	CapSure Epi	Sep-96	21,200	9,400	171	37	1	0
4968	CapSure Epi	Sep-99	29,000	17,900	35	17	0	1
5071	Screw-in	Dec-92	45,200	15,400	13	2	0	0

Source: Returned Product Analysis
Data as of July 31, 2013

US Reports of Acute Lead Observations

Model Number	Family	Estimated US Implants	Cardiac Perforation	Conductor Fracture	Lead Dislodgement	Failure to Capture	Oversensing	Failure To Sense	Impedance Abnormal	Extracardiac Stimulation
4965	CapSure Epi	21,200	0	1	0	4	1	4	4	0
4968	CapSure Epi	29,000	0	0	3	14	2	0	2	0
5071	Screw-in	45,200	1	0	0	33	0	2	2	3

Model Number	Family	Estimated US Implants	Insulation Breach	Unspecified
4965	CapSure Epi	20,900	0	3
4968	CapSure Epi	27,800	1	0
5071	Screw-in	44,300	0	1

Report Cutoff Date: Data as of September 10, 2013

Reference Chart

Model Number	Family	Type	Insulation	Conductor Material	Tip Electrode	Connector Type
4965	CapSure Epi	Epicardial Suture V or A	Silicone	MP35N 5 Filars	Porous Platinized/ Steroid	IS-1 UNI
4968	CapSure Epi	Epicardial Suture V or A	Silicone	MP35N 5 Filars	Porous Platinized/ Steroid	IS-1 B1
5071	Screw-in	Myocardial Screw-In Ventricular	Silicone	MP35N Multifilars	2-Turn Helix	IS-1 UNI

VDD Single Pass Pacing Leads

5032 CapSure VDD


Product Characteristics

US Market Release	Mar-96	Serial Number Prefix	LCL, LCN, LCM	US Returned Product Analysis	
Registered US Implants	5,300	Type and/or Fixation	Transvenous, Right Ventricle, Tines		Conductor Fracture 7
Estimated Active US Implants	1,200	Polarity	Quadripolar		Crimp/Weld/Bond 0
Advisories	None	Steroid	Yes		Insulation Breach 6 Other 0

Product Surveillance Registry Results

Qualifying Complications 1 Total

Number of Leads Enrolled in Study	38	Failure to Sense	1
Cumulative Months of Follow-Up	1,683		
Number of Leads Active in Study	0		



5038 CapSure VDD-2


Product Characteristics

US Market Release	Sep-98	Serial Number Prefix	LEE, LEG, or LEF	US Returned Product Analysis	
Registered US Implants	9,200	Type and/or Fixation	Transvenous, Right Ventricle, Tines		Conductor Fracture 4
Estimated Active US Implants	3,600	Polarity	Quadripolar		Crimp/Weld/Bond 0
Advisories	None	Steroid	Yes		Insulation Breach 1 Other 0

Product Surveillance Registry Results

Qualifying Complications 6 Total

Number of Leads Enrolled in Study	558	Conductor Fracture	3
Cumulative Months of Follow-Up	20,842	Failure to Capture	1
Number of Leads Active in Study	41	Failure to Sense	2



VDD Single Pass Pacing Leads

Lead Survival Summary (95% Confidence Interval)

Model Number	Family	US Market Release	Leads Enrolled	Leads Active in Study	Qualifying Complications	Cumulative Months of Follow-Up in Study	Device Survival Probability (%)													
							Years After Implant													
							1 yr	2 yr	3 yr	4 yr	5 yr	6 yr	7 yr	8 yr	9 yr	10 yr				
5032	CapSure VDD	Mar-96	38	0	1	1,683	100.0 at 0 mo													
5038	CapSure VDD-2	Sep-98	558	41	6	20,842	99.8 +0.2/-1.3	99.8 +0.2/-1.3	99.8 +0.2/-1.3	98.8 +0.8/-2.7	98.2 +1.2/-3.2	98.2 +1.2/-3.2	95.2 +2.9/-7.2 at 81 mo							

Source: Product Surveillance Registry
Data as of July 31, 2013

US Returned Product Analysis Summary

Model Number	Family	US Market Release	Estimated US Implants	Estimated US Active	Conductor Fracture	Insulation Breach	Crimp/Weld/Bond	Other
5032	CapSure VDD	Mar-96	5,300	1,200	7	6	0	0
5038	CapSure VDD-2	Sep-98	9,200	3,600	4	1	0	0

Source: Returned Product Analysis
Data as of July 31, 2013

US Reports of Acute Lead Observations

Model Number	Family	Estimated US Implants	Lead Dislodgement	Failure to Capture	Failure to Sense	Extracardiac Stimulation	Unspecified
5032	CapSure VDD	5,300	1	1	1	0	1
5038	CapSure VDD-2	9,200	2	1	1	1	0

Report Cutoff Date: September 10, 2013

Reference Chart

Model Number	Family	Type	Insulation	Conductor Material	Tip Electrode	Connector Type
5032	CapSure VDD	Transvenous V and A Tines	Silicone	MP35N 5/6/1 Filars	Porous Platinized/ Steroid	Atr. IS-1 BI, Vent. IS-1 BI
5038	CapSure VDD-2	Transvenous V and A Tines	Silicone	MP35N	Porous Platinized/ Steroid	Atr. IS-1 BI, Vent. IS-1 BI

ICD and CRT-D Charge Time Performance

Medtronic continues its commitment to providing updated information on charge time performance.

Introduction

Information on charge time performance of Medtronic products is presented in this section of the CRDM Product Performance Report. Medtronic implemented the collection of charge time data on July 1, 1999. The data are collected via our ongoing active clinical study of long-term system performance called the Product Surveillance Registry. The study protocol requests device data be routinely taken and sent to Medtronic at no more than 6-month intervals.

In our analysis performed for this report, only charge times resulting from full energy charges are considered. To ensure consistent reporting across devices, the charge time reported at implant represents the last charge time available from date of implant. When more than one charge time is available in a 6-month interval, a conservative approach has been adopted whereby only the maximum charge time in each 6-month interval is reported. As charge time is directly proportional to the time elapsed since the last capacitor reformation, charges occurring within 15 days of a previous charge are excluded. This precludes the reporting of overly optimistic results.


Data from over 20,000 devices contribute to the charge time data in this report. By tracking and reporting this charge time data, Medtronic is able to ascertain the actual performance of its charging circuitry. The insight gained through this information is applied to Medtronic's ongoing efforts to provide charge times that are short and consistent over the life of the product.

Data Presentation


Charge time data for ICD and CRT-D models are presented using boxplots at 6-month intervals. The shaded box on the plots represents the middle half of the data – the Interquartile Range (IQR). The white line in the middle of each box is the median charge time. The top of the box representing the IQR is the third quartile or the 75th percentile (i.e., 75% of all charge times fall below this line), whereas the bottom of the box represents the first quartile or the 25th percentile. Vertical lines are drawn from the quartiles to the farthest value not more than 1.5 times the interquartile range. Any values more extreme than the vertical lines are considered outliers.

Results

The graph below shows the overall maximums, minimums, and medians for Medtronic ICD and CRT-D products.




Medtronic CRT-D and ICD Charge Time Performance




ICD Charge Times


7230 Marquis VR Charge Time




7231 GEM III VR Charge Time




7232 Maximo VR Charge Time




7271 GEM DR Charge Time



7274 Marquis DR Charge Time




7275 GEM III DR Charge Time




ICD and CRT-D Charge Time Performance continued


7276 GEM III AT Charge Time




7278 Maximo DR Charge Time




7279 InSync III Marquis Charge Time




7285 InSync III Protect Charge Time



7288 Intrinsic Charge Time




7289 InSync II Marquis Charge Time




ICD and CRT-D Charge Time Performance continued


7297 InSync Sentry Charge Time




7298 InSync Sentry Charge Time




7299 InSync Sentry Charge Time




7303 InSync Maximo Charge Time




7304 InSync Maximo Charge Time




C154DWK, C164AWK, C174AWK
Concerto Charge Time




D153ATG/DRG EnTrust Charge Time




D154ATG/DRG EnTrust Charge Time




D154AWG/164 Virtuoso Charge Time



D154VRC/153 EnTrust Charge Time




D154VWC/164 Virtuoso Charge Time




ICD Charge Times

ICD and CRT-D Charge Time Performance continued


D224VRC/234 Secura VR Charge Time




D224TRK/234/204TRM Consulta Charge Time




D224DRG/234/204DRM Secura DR Charge Time




D274DRG Virtuoso II DR Charge Time



D274TRK/294 Concerto II CRT-D Charge Time




D274VRC Virtuoso II VR Charge Time




ICD and CRT-D Charge Time Performance continued


D284DRG Maximo II DR Charge Time




D284TRK Maximo II CRT-D Charge Time




D284VRC Maximo II VR Charge Time




D314DRG/DRM Protecta XT DR Charge Time



D314TRG/TRM Protecta XT CRT-D Charge Time




D314VRG/VRM Protecta XT VR Charge Time




ICD and CRT-D Charge Time Performance continued


D334DRG/DRM Protecta DR Charge Time




D334TRG/TRM Protecta CRT-D Charge Time




D334VRG/VRM Protecta VR Charge Time




D354DRG/DRM Protecta XT DR Charge Time



D354TRG/TRM Protecta XT CRT-D Charge Time



D354VRG/VRM Protecta XT VR Charge Time



Consulta CRT-P and Syncra CRT-P

Original Date of Advisory: June 2013

Potential Loss Of Device Hermeticity

Product

Consulta® CRT-P and Syncra® CRT-P

Advisory

Medtronic has identified an issue with a connector bracket weld on a subset of Consulta CRT-P models and Syncra CRT-P devices manufactured between April 1 and May 13, 2013. This type of connector bracket weld is unique to Consulta and Syncra CRT-P devices and no other Medtronic device models are affected.

An out-of-specification weld could result in a loss of device hermeticity and compromised device functionality. **There have been no reported or confirmed device failures or patient injuries.** Medtronic estimates the rate of out-of-specification welds to be 1-2% in this subset of devices.

Non-implanted devices from this subset have been recalled to Medtronic for re-inspection with additional controls to ensure that the weld meets specification. In June 2013, Medtronic communicated to impacted physicians that up to 779 devices worldwide (43 in the U.S.) may have been implanted from this subset.

Patient Management Recommendations (As of June 2013)

As a result of on-going investigation and consultation with Medtronic's Independent Physician Quality Panel, Medtronic offers the following patient management recommendations:

- Physicians should advise their patients to seek medical attention immediately if they experience a return of symptoms related to bradycardia or heart failure.
- If considering prophylactic device replacement for pacemaker-dependent patients with a device in the identified subset, physicians should carefully assess individual patient circumstances against the known risk of a device replacement.
- Physicians should continue routine follow-up in accordance with standard practice.

Status Update

As of September 10, 2013, 536 of the 779 devices have been returned from field inventory. Medtronic estimates the remaining 243 devices (44 in the U.S.) have been implanted. There have been no reported or confirmed device failures or patient injuries.

Initial Affected Population	Number of Confirmed Advisory Related Events	Estimated Remaining Active Population	Current Malfunction Rate (Confirmed Malfunctions over total population)
Up to 779 Worldwide (44 United States)	0 Worldwide (0 United States)	243 Worldwide (43 United States)	0% Worldwide (0% United States)

EnTrust ICDs

Original Date of Advisory: March 2012

Potential Rapid Battery Depletion

Product

All EnTrust ICDs.

Advisory

A small percentage of EnTrust ICDs may not meet expected longevity or provide at least three months of device operation between the Elective Replacement Indicator (ERI) and End of Life (EOL) due to a more-rapid-than-expected drop in battery voltage. No patient deaths or serious injuries have been reported as a result of this issue.

The reported events have involved a drop in battery voltage from ~3.0 V to ERI (2.61 V) over a time period ranging from approximately one week to six months. All reported events have occurred at least 30 months after implant.

Medtronic has identified the cause of these occurrences to be an internal battery short that develops as the battery capacity is consumed.

Patient Management Recommendations (As of March 2012)

After consultation with Medtronic’s Independent Physician Quality Panel, Medtronic offers the following patient management recommendations:

- Physicians should continue routine follow-up sessions at least every three months in accordance with product labeling.
- Physicians should program the audible patient alerts for “Low Battery Voltage ERI” and “Excessive Charge Time EOL” to ON.
- Physicians should replace devices promptly after they reach ERI if the decline in voltage is more rapid than expected.
- Prophylactic replacement of EnTrust ICDs is not recommended.

Status Update

As of September 10, 2013, there have been 83 confirmed events. No patient deaths have been reported due to this issue. No reports have been made of a failure to deliver high voltage therapy.

Initial Affected Population	Number of Confirmed Advisory Related Events	Estimated Remaining Active Population	Current Malfunction Rate (Confirmed Malfunctions over total population)
69,000 Worldwide (43,200) United States)	83 Worldwide (63) United States)	23,300 Worldwide (14,600) United States)	0.12% Worldwide (0.15%) United States)

EnRhythm Pacemakers

Original Date of Advisory: February 2010

Low Battery Voltage Displayed at Device Interrogation

Product

All EnRhythm pacemakers.

Advisory

Two specific battery issues with EnRhythm pacemakers have been identified and both are addressed by a Medtronic software update.

First Issue

As of February 2010, Medtronic had received 62 reports (out of approximately 110,000 devices worldwide) indicating that the battery voltage at device interrogation was lower than the battery voltage that is tracked by the device to provide data for the elective replacement indicator (ERI) notification. The lower voltage measurement caused confusion and occasionally resulted in unnecessary explants.

Medtronic's investigation has shown that none of these reports resulted in loss of therapy. Importantly, the ERI notification, which uses the nightly battery measurement, is unaffected and accurate. Medtronic has identified the root cause as higher than expected battery resistance.

Medtronic's internal testing has shown that there is no current risk for compromised therapy delivery. If the software update referenced above is not implemented, there will be a potential risk of loss of device functionality in a small percent (less than 0.08% 6 years post-implant) of devices. The software update eliminates this risk.

Second Issue

Through internal accelerated testing, Medtronic has identified a second issue that projects battery voltage could decrease sooner than expected due to a slightly increased rate of lithium depletion. This issue has not been clinically observed and is not expected to occur until approximately 9 years post-implant. If the software update referenced above is not implemented, there may be a potential risk for loss of therapy at or near ERI in a small number of devices. The software will eliminate this issue by changing ERI criteria.

Software Update (As of October 2010)

The battery issues described above and subsequent software update are summarized in the table below. When a device receives the software update, if battery impedance is greater than the new ERI threshold, ERI will be triggered shortly thereafter. Therefore, clinicians may observe an ERI/EOL indicator at the next patient follow-up. When ERI is triggered by battery impedance, additional battery capacity remains and can support device function at ERI parameters for at least one

year. Medtronic is not aware of any reports of loss of therapy due to this issue.

As a reminder, when ERI is triggered, EnRhythm devices revert to VVI pacing at 65 ppm at the programmed output settings. EOL is declared 90 days after ERI or at a battery voltage of 2.69 V, whichever comes sooner.

Battery Issue	Software Update
Battery voltage could decrease sooner than expected due to a slightly increased rate of lithium depletion	Changed ERI battery voltage threshold from 2.59 V to 2.81 V to ensure 90 days of therapy from ERI to EOL
Higher than expected battery impedance	Added a secondary ERI trigger based on battery impedance. This new criteria will identify devices with increased battery impedance before device performance is impacted. If triggered, displayed battery voltage is reset to 2.81 V to ensure alignment with ERI battery voltage threshold.

Updated Performance Information (as of August 2011)

We now have access to battery impedance and ERI performance on more than 5,000 EnRhythm devices that have received the EnRhythm software update. Our modeling based on these data shows that approximately 6-10% of devices will reach ERI within 5 years post-implant. Consistent with our previous communications, we continue to expect average device longevity to be reduced by approximately 10-15%, with the expected average longevity remaining at 8.5 to 10.5 years, depending on device settings.¹

Updated Patient Management Recommendations (as of August 2011)

After consultation with Medtronic's Independent Physician Quality Panel, we recommend:

Performing a device follow-up within 90 days after the software download to identify devices that triggered ERI shortly after the software update. Subsequent follow-up can be performed per standard practice. During programmer interrogation of a device at ERI, there is a slight possibility a transient drop in pacing amplitude could occur. If this is noted, either remove the programmer head or temporarily program to a higher output voltage.

continued

EnRhythm Pacemakers

Original Date of Advisory: February 2010

Low Battery Voltage Displayed at Device Interrogation, continued

If an unanticipated ERI/EOL is declared, it is likely due to battery impedance. In such cases, additional battery capacity remains and can support device function at ERI parameters for at least one year. However, when ERI or EOL (typically 90 days after ERI) declaration is seen, schedule device replacement.

¹ The 8.5 year estimate represents a high use scenario (DDD, 100% pacing in atrium and ventricle with 3.0 V output in both chambers). The 10.5 year estimate represents a typical use scenario for a sinus node dysfunction patient with the MVP function ON (AAI(R) <=> DDD(R), 50% pacing in atrium and 5% pacing in ventricle with 3.0 V output in both chambers). Projections are based on modeling and not actual field returns, due to limited availability of implant experience beyond 6 years. Field performance will continue to be monitored and modeling updated to reflect actual data.

Status Update

As of September 10, 2013, 381 devices out of approximately 146,500 devices worldwide have been confirmed as having exhibited an advisory event related to the original advisory, in which higher than expected battery impedance caused a drop in battery voltage at interrogation.

Initial Affected Population	Number of Confirmed Advisory Related Events	Estimated Remaining Active Population	Current Malfunction Rate (confirmed malfunctions over total population)	The software update eliminates any potential future risk of the two battery issues described above by changing the ERI criteria.
All EnRhythm pacemakers (146,500 Worldwide)	381 Worldwide	84,700 Worldwide	0.25%	

Included in the August 2011 Performance Update was information about the projected percentage of devices that would encounter an early ERI due to unexpected high battery impedance. As of February 25, 2013, percentage of devices that have encountered ERI due to battery impedance has not exceeded the rate of 6-10% within 5 years post-implant as communicated with our August 2011 Performance Update.

Initial Affected Population	Number of Confirmed Advisory Related Events	Estimated Remaining Active Population	Only devices using the updated software can trigger ERI due to impedance.
All EnRhythm pacemakers (146,500 Worldwide)	8,805 Worldwide	84,700 Worldwide	

Concerto CRT-D and Virtuoso ICD

Original Date of Advisory: September 2009

Potential Reduced Device Longevity

Product

A subset of Concerto CRT-D and Virtuoso ICD devices may not meet expected device longevity. Go to www.medtronic.com/CRDMProductPerformance to determine if a specific device is affected.

Advisory

A subset of Concerto CRT-D and Virtuoso ICD devices may not meet expected device longevity due to gradually increasing current drain caused by low voltage capacitor degradation. This issue may present in the affected devices as reaching the Recommended Replacement Time (RRT) earlier than projected. This issue does not compromise device functionality or affect therapy delivery.

Based on information from returned devices, Medtronic expects that affected devices will continue to provide at least 3 months of normal device function between RRT and End of Service (EOS) as described in device labeling.

A total of 8,900 devices worldwide are affected by this advisory. An estimated 6,300 of these devices were active at the time of the original advisory communication.

Concerto and Virtuoso devices in the affected subset were manufactured primarily in 2006 and can be traced to a specific subset of low voltage capacitors.

Patient Management Recommendations

After consultation with Medtronic’s Independent Physician Quality Panel, Medtronic offers the following recommendations for patients with devices in the affected subset:

Physicians should continue routine follow-up sessions at least every 3 months in accordance with product labeling.

Physicians should verify that the Low Battery Voltage RRT alert is programmed to “On-High.” This provides an audible, alternating tone when the device reaches RRT. These devices are shipped with this alert programmed nominally to “On-High.”

Physicians may consider monitoring patients through CareLink. The CareLink home monitor can be used to automatically notify the clinician when the device reaches RRT.

Status Update

As of September 10, 2013, 3,684 devices out of approximately 8,900 devices in this subset worldwide have been confirmed as having exhibited this capacitor degradation. Out of the initial advisory population of 8,900 worldwide, less than 300 remain implanted worldwide.

Initial Affected Population	Number of Confirmed Advisory Related Events	Estimated Remaining Active Population	Current Malfunction Rate (Confirmed Malfunctions over total population)
8,900 Implanted Worldwide (7,000 United States)	3,684 Worldwide (3,168 United States)	< 300 Worldwide (< 300 United States)	42% Worldwide (45% United States)

Kappa 600/700/900 Pacemakers **Sigma 100/200/300 Pacemakers** Original Date of Advisory: May 2009

Potential Separation of Interconnect Wires (2009)

Product

A specific subset of Kappa and Sigma series pacemakers may fail due to separation of interconnect wires from the hybrid circuit. Go to www.medtronic.com/CRDMProductPerformance to determine if a specific device is affected.

Advisory Population

Specific subsets of Kappa and Sigma series pacemakers may fail at a higher than expected rate due to separation of wires that connect the electronic circuit to other pacemaker components (e.g., battery, connector). This may present clinically as loss of rate response, premature battery depletion, loss of telemetry, or no output.

Some patients, whose devices experience a wire separation resulting in a loss of pacing output, will experience a return of bradycardia symptoms (e.g., fainting or lightheadedness). In rare cases involving pacemaker dependent patients, loss of pacing output may result in death or serious injury.

Since 1997, there have been over 1.7 million Kappa and Sigma devices implanted worldwide. At the time of the original advisory communication, an estimated 15,200 Kappa and 6,100 Sigma devices affected by the advisory remained implanted and active. These devices were manufactured primarily between November 2000 and November 2002. Most of these devices have been implanted in patients for five years or longer and may be nearing normal elective replacement time.

There is no provocative testing that can predict which specific devices may fail, and no device programming can mitigate this issue if it occurs.

Patient Management Recommendations

We realize that each patient requires unique clinical consideration and we support your judgment in caring for your patients. After consultation with Medtronic's Independent Physician Quality Panel, Medtronic offers the following recommendations for patients:

- Physicians should advise their patients to seek medical attention immediately if they experience symptoms (e.g., fainting or lightheadedness).
- Physicians should consider device replacement for patients who are both pacemaker dependent and who have been implanted with a device in the affected subsets. Medtronic will offer a supplemental device warranty if the device is not already at elective replacement time.
- Physicians should continue routine follow-up in accordance with standard practice for those patients who are not pacemaker dependent.

Status Update

Advisory Population

Patient management recommendations remain unchanged. As of September 10, 2013, Medtronic has observed 458 Kappa devices and 299 Sigma devices with this failure mechanism from the Kappa and new Sigma device subsets. This represents 0.79% (Kappa) and 2% (Sigma) of the original affected implant population.

Four hundred twenty-one (421) of the Kappa devices (0.72%) and 232 of the Sigma devices (1.56%) were returned with information indicating a problem with the patient's pacing system prior to explant. The remaining 37 Kappa devices (0.06%) and 67 Sigma devices (0.44%) were returned with no information indicating a potential malfunction while implanted or with insufficient information to determine the state of the device at explant. Lacking definite information indicating proper operation until explant, these remaining devices are conservatively categorized as having experienced interconnect wire separation while implanted.

As of May 2009, our modeling predicts failure rates due to this issue of 1.1% (Kappa) and 4.8% (Sigma) over the remaining lifetime of those pacemakers still in service at that time.

Out of the initial advisory population of 58,300 Kappa devices and 14,900 Sigma devices worldwide, less than 200 Kappa devices remain implanted worldwide and 2,000 Sigma devices remain implanted worldwide. Of these, 500 Sigma devices are in the United States.

Continued Vigilance

Included in the advisory communication was information about an additional subset of Kappa devices where we have observed a much lower rate of occurrence of this issue. Less than 200 devices of this subset remain active. We have observed a failure rate of approximately 0.096% in this subset and our May 2009 modeling predicts a failure rate of 0.12% over the remaining device life of those pacemakers still in service at that time. After review with our Independent Physician Quality Panel, we do not recommend any specific actions for this group of devices. We will continue to monitor performance and inform you if any specific patient management recommendations are warranted.

continued

Kappa 600/700/900 Pacemakers

Sigma 100/200/300 Pacemakers

Original Date of Advisory: May 2009

Potential Separation of Interconnect Wires, continued

Initial Affected Population	Number of Confirmed Advisory Related Events	Estimated Remaining Active Population	Current Malfunction Rate (Confirmed and Unconfirmed Malfunctions over total population)	Predicted Malfunction Rate Over the Remaining Life of the Devices Still Implanted Population
Kappa Pacemakers				
58,300 Implanted Worldwide (est.) (17,600 United States)	421 Worldwide (221 United States) with information indicating a clinical presentation. An additional 37 worldwide (25 US) without information indicating a clinical presentation or with insufficient information to determine the state of the device at explant.	< 200 Worldwide (< 100 United States)	0.79% Worldwide 1.40% (United States)	1.1%
Sigma Pacemakers				
14,900 Implanted Worldwide (est.) (3,700 United States)	232 Worldwide (49 United States) with information indicating a clinical presentation. An additional 67 worldwide (17 US) without information indicating a clinical presentation or with insufficient information to determine the state of the device at explant.	2,000 Worldwide (500 United States)	2% Worldwide 1.8% (United States)	4.8%

6930, 6931, 6948, 6949 Sprint Fidelis Defibrillation Leads

Original Date of Advisory: October 2007

Potential Conductor Wire Fracture

Product

All Model 6930, 6931, 6948, and 6949 implantable defibrillation leads

Advisory

There are two primary locations where chronic conductor fractures have occurred on Sprint Fidelis leads: 1) the distal portion of the lead, affecting the anode (ring electrode) and 2) near the anchoring sleeve tie-down, predominantly affecting the cathode (helix tip electrode), and occasionally the high voltage conductor. These two locations account for approximately 90% of the chronic fractures identified in Returned Product Analysis (RPA). The remaining 10% of chronic fractures occurred in the DF-1 connector leg and the proximal portion of the RV coil. High voltage conductor fractures could result in the inability to deliver defibrillation therapy. Anode or cathode conductor fractures (at either location) may present clinically as increased impedance, oversensing, increased interval counts, multiple inappropriate shocks, and/or loss of pacing output.

Patient Management Recommendations (Updated April 2011)

The Lead Integrity Alert (LIA) provides three days advance notice prior to inappropriate therapy to 76% of patients with lead fractures.¹ As a result, we strongly recommend that all Sprint Fidelis patients who have the ability to upgrade to Lead Integrity Alert do so promptly. Also ensure that high voltage lead impedance alerts (maximum of 100 ohms) are programmed. When a lead fracture is suspected or confirmed, immediate patient attention is strongly recommended. Physicians should inform their patients to seek medical attention without delay if they experience unexpected shocks

- **If a Fidelis lead fracture of any type has occurred, we recommend implanting a new high voltage lead with or without extraction of the Fidelis lead.**
- In patients with normal device function and no manifestation of lead fracture, no action is recommended. The risk of prophylactic intervention appears to be greater than serious injury resulting from lead fracture even for pacemaker dependent patients, except in select individual patient circumstances as determined by the physician.
- In the event of a device change-out or upgrade procedure, with no manifestation of lead fracture, consider the patient age and lead model data above, as well as patient life expectancy, co-morbidities, ease of extraction related to implant time, patient preference, etc., for the following options:
 - Leave a properly performing lead intact.
 - Implant a new ICD lead without extraction of the existing lead.

- Carefully consider all factors before prophylactic placement of a pace-sense lead. Data shows an increased risk of high voltage conductor fracture if a pace-sense conductor fracture has previously occurred. This data is available at www.medtronic.com/fidelis
- Individual patient circumstances may warrant extracting and implanting a new ICD lead. If warranted, Medtronic's Independent Physician Quality Panel recommends the lead extraction procedure be performed by a physician with extensive lead extraction experience.²

Status Update

As of September 10, 2013, of the initial implant population of 205,600 in the United States, approximately 88,300 remain implanted. According to Product Surveillance Registry results, lead survival is estimated to be 80.4% (+4.5/-5.6) at 87 months. As the implanted population ages and the sample size increases for each time interval, the accuracy of the estimated survival probability will increase as shown by tighter confidence intervals.

¹ Swerdlow C, Gunderson B, Ousdigian K, et al. Downloadable Algorithm to Reduce Inappropriate Shocks Caused by Fractures of Implantable Cardioverter-Defibrillator Leads, *Circulation*. November 2008;118:2122-2129.

² Wilkoff B, Love C, Byrd C, et al. Transvenous Lead Extraction: Heart Rhythm Society Expert Consensus on Facilities, Training, Indications, and Patient Management. *Heart Rhythm*. July 2009;6:1085-1104.

continued

6930, 6931, 6948, 6949 Sprint Fidelis Defibrillation Leads

Original Date of Advisory: October 2007

Potential Conductor Wire Fracture, continued

Keeping Physicians Informed

The most recent Sprint Fidelis lead performance information, including survival curves, physician letters, and subpopulation data, can be found at www.medtronic.com/fidelis and will be updated semi-annually. Medtronic's website also has a selected list of peer-reviewed publications related to Fidelis lead performance and extraction. Medtronic is committed to answering your questions and keeping you informed. If you have any questions or concerns, please contact your Medtronic Representative or Medtronic Technical Services at 1 (800) 723-4636 (US).

Initial Affected Population	Number of Confirmed Advisory Related Events	Estimated Remaining Active Population	Additional information about the Sprint Fidelis lead is available at: www.medtronic.com/fidelis .
279,500 Worldwide (205,600 United States)	6,330 Worldwide (4,471 United States)	120,000 Worldwide (83,300 United States)	

Sigma Implantable Pulse Generators

Original Date of Advisory: November 2005

Potential Separation of Interconnect Wires (2005)

Product

A specific subset of Sigma series pacemakers may fail due to separation of interconnect wires from the hybrid circuit. Go to www.medtronic.com/CRDMProductPerformance to determine if a specific device is affected.

Advisory

This subset of Sigma series pacemakers that may fail due to separation of interconnect wires from the hybrid circuit may present clinically as loss of rate response, premature battery depletion, intermittent or total loss of telemetry, or no output.

Separation of redundant interconnect wires has been observed on hybrid terminal blocks. Device failure occurs only where both interconnect wires separate from a hybrid terminal block. In October 2005, testing and analysis identified the root cause of these failures and the affected population. Hybrid circuits used in this subset of devices were cleaned during manufacturing with a particular cleaning solvent that could potentially reduce the strength of the interconnect wire bond over time.

No provocative testing can predict which devices may fail.

Patient Management Recommendations

Recommendation for the management of patients who have pacemakers affected by this advisory were changed in May 2009. Current recommendations are:

We realize that each patient requires unique clinical consideration and we support your judgment in caring for your patients. After consultation with Medtronic's Independent Physician Quality Panel, Medtronic offers the following recommendations for patients in the 2005 Sigma advisory:

- Physicians should advise their patients to seek medical attention immediately if they experience symptoms (e.g., fainting or lightheadedness).
- Physicians should consider device replacement for patients who are both pacemaker dependent and who have been implanted with a device in the affected subsets. Medtronic will offer a supplemental device warranty if the device is not already at elective replacement time.
- Physicians should continue routine follow-up in accordance with standard practice for those patients who are not pacemaker dependent.

Status Update

Patient management recommendations remain unchanged. As of September 10, 2013, 806 devices out of approximately 40,000 devices worldwide have been confirmed as having experienced interconnect wire separation.

Four hundred forty-two (442) of the Sigma devices (1%) were returned with information indicating a problem with the patient's pacing system prior to explant. The remaining 364 Sigma devices (0.90%) were returned with no information indicating a potential malfunction while implanted or with insufficient information to determine the state of the device at explant. Lacking definite information indicating proper operation until explant, these remaining devices are conservatively categorized as having experienced interconnect wire separation while implanted.

Our original modeling predicted a failure rate from 0.17% to 0.30% over the remaining lifetime of these pacemakers. However, as of May 2009, updated modeling now predicts a failure rate of 3.9% over the remaining device life of those devices still in service at that time.

Out of the initial advisory population of 40,000 worldwide, approximately 5,900 remain implanted. Approximately 1,400 of these are in the United States.

Initial Affected Population	Number of Confirmed Advisory Related Events	Estimated Remaining Active Population	Current Malfunction Rate (Confirmed and Unconfirmed Malfunctions over total population)	Predicted Malfunction Rate Over the Remaining Life of the Devices Still in Service as of May 2009
40,000 Implanted Worldwide (est.) (9,900) United States)	442 Worldwide (86) United States) with information indicating a clinical presentation. An additional 364 Worldwide (66) US) without information indicating a clinical presentation or with insufficient information to determine the state of the device at explant.	5,900 Worldwide (1,400) United States)	2.0% Worldwide 1.5% (United States)	3.9%

7274 Marquis DR 7278 Maximo DR 7277 InSync Marquis 7279 InSync III Marquis
7230 Marquis VR 7232 Maximo VR 7289 InSync II Marquis 7285 InSync III Protect
 Original Date of Advisory: February 2005

Potential Premature Battery Depletion Due to Battery Short

Product

The specific subset of Marquis family ICD and CRT-D devices having batteries manufactured prior to December 2003 is affected. Devices manufactured with batteries produced after December 2003 are not affected. Go to www.medtronic.com/CRDMProductPerformance to determine if a specific device is affected.

Advisory

Medtronic Marquis family of ICD and CRT-D devices having batteries manufactured prior to December 2003 may experience rapid battery depletion due to a specific internal battery short mechanism. Battery design changes were implemented in December 2003 that eliminate the possibility of this internal shorting mechanism.

Highly accelerated bench testing indicated the rate of this shorting mechanism may increase as the battery is depleted. As of February 2005, the rate of shorting was approximately 1 in 10,000 (0.01%); bench test data indicated the rate may increase to between 0.2% and 1.5% over the second half of device life.

No provocative testing can predict which of these devices will experience this issue. Once a short occurs, battery depletion can take place within a few hours to a few days. After depletion the device ceases to function. It is also possible that as the battery depletes quickly, patients may experience temporary warmth in the area surrounding the ICD.

Patient Management Recommendations

We recommend you consider the following patient management options:

- Conduct quarterly (i.e., every 3 months) follow-up procedures

- Inform patients that should they experience warmth in the area surrounding the ICD to seek follow-up care promptly
- Program Low Battery Voltage ERI Patient Alert to “On-High.” This will result in an audible, alternating tone in the limited circumstances where a battery depletes slowly over a number of days. Data indicates most shorts will occur rapidly and will not be detected by this feature.
- Provide a hand-held magnet to patients to check device status and program the Low Battery Voltage ERI Patient Alert to “On-High.” Device operation may be monitored periodically (e.g., daily) by patients placing the magnet over the device for 1-2 seconds. If the device is functional, a steady tone will sound for approximately 20 seconds. If no tone is heard, follow-up care should be sought promptly.

Status Update

The Marquis Family device performance related to the battery shorting mechanism continues to be within Medtronic’s engineering projections. As of September 10, 2013, 193 Marquis Family devices have been confirmed as having this internal battery shorting mechanism. One hundred fifteen (115) of these devices were returned from the United States.

Out of the initial advisory population of 87,000 worldwide, approximately 1,600 remain implanted. Approximately 1,400 of these are in the United States.

The Patient Management Recommendations set forth in the advisory remain unchanged.

Initial Affected Population	Number of Confirmed Advisory Related Events	Estimated Remaining Active Population	Current Malfunction Rate (Confirmed Malfunctions over total population)	Predicted Malfunction Rate Over the Remaining Life of the Devices Still Implanted Population
87,000 Implanted Worldwide (76,000 United States)	193 Worldwide (115 United States)	1,600 Worldwide (1,400 United States)	0.22% Worldwide (0.15% United States)	Consistent with Medtronic projections, the observed rate of shorting may increase to between 0.2% and 1.5% over the second half of device life.

Performance Notes

Dual Chamber Pacemakers with Measurement Lock-up ERI

Kappa 600, 700, 800, 900, EnPulse, Adapta, Versa, Sensia, Relia, and Vitatron Models E50A1, E60A1, and G70A1

Purpose of this Information

This Performance Note describes a rare measurement lock-up issue that impacts the Medtronic dual chamber pacemakers listed above. If this measurement lock-up occurs, the device will trigger a false Elective Replacement Indicator (ERI). A reset is available to clear this condition and there is no need to explant the device. This issue does not impact battery longevity.

Background

If this rare measurement lock-up occurs in the pacemaker, it causes the device to read a value of zero for battery voltage. After four measurements of zero, the device will trigger ERI and revert to a VVI pacing mode at 65 bpm. There is no loss of ventricular pacing and the output voltage will remain the same.

Programmer Software Reset Method (Adapta, Versa, Sensia, Relia, Vitatron Series E and G)

Programmer software is available which can differentiate a regular ERI and an ERI caused by the measurement lock-up issue. Upon interrogation of a device with the measurement

lock-up ERI, the programmer software recognizes the issue and guides the clinician to clear the ERI (Example 1). Following an ERI reset, the device parameters should be reviewed and reprogrammed to clinician specifications.

Reset Method for Kappa and EnPulse


A service tool continues to be available through Medtronic Technical Services to clear the measurement lock-up issue for Kappa and EnPulse devices.

The issue can be identified using the programmer or via CareLink transmission; the battery voltage measurements and remaining longevity will appear as blank values (Example 2). If this measurement lock-up occurs, contact Medtronic Brady Technical Services at 1-800-505-4636 for assistance.

Example 1 – Programmer Software Detects Measurement Lock-up ERI



Example 2 – Programmer Screens for Measurement Lock-up ERI (Kappa and EnPulse)



Clinical Management of VCM near Elective Replacement

Background

Medtronic Technical Services has received reports of devices going to ERI or end of life (EOL) sooner than expected after a normal follow-up in which the device longevity was projected to be approximately 18 months. It has been noted that these cases typically involve Kappa 700 devices where Ventricular Capture Management set the ventricular lead to high output (5 V, 1 ms), which occurs by device design when a high threshold is measured. It is important for physicians and allied professionals to understand VCM behavior as it relates to longevity so that they can, in turn, understand how this affects management of the device and follow-up visits as VCM equipped IPGs near the end of their expected longevity.

Device Longevity and VCM Behavior

Ventricular Capture Management is a feature that uses evoked response sensing to determine the stimulation threshold needed to capture the ventricular chamber. Proper detection of the evoked response is crucial to the VCM algorithm determining an accurate capture threshold. There are rare conditions, however, during which the VCM algorithm will not be able to measure the evoked response accurately.¹ When this occurs, for safety reasons the VCM algorithm will reprogram the output to 5 V, 1 ms until the subsequent VCM measurement.

If the device has considerable remaining longevity, these occasional excursions to high output do not substantially affect remaining longevity. However, if the device has less than approximately 18 months remaining longevity, there is the possibility that the high output condition caused by the 5 V, 1 ms output will drain the battery and trigger ERI.

When ERI is declared by the device, VCM is disabled and the outputs are left at 5 V, 1 ms until the device is reprogrammed at an in-office follow-up. This increased current drain of a high output condition will speed depletion of the device, possibly resulting in the device getting to the EOL (battery voltage ≤ 2.15 V).

Please note that the following parameter changes occur when the device goes to ERI:

Table: IPG Therapy Parameter Changes at ERI

Parameter	Value
Pacing Mode	VVI
Lower Rate	65 bpm
Single Chamber Hysteresis	OFF
Sleep Function	OFF
Ventricular Capture Management	OFF
Atrial Sensing Assurance	OFF
Ventricular Sensing Assurance	OFF

Kappa 700 is Medtronic's first-generation VCM algorithm, which has a relatively higher incidence of evoked response undersensing compared to subsequent algorithms, resulting in more frequent high output conditions. Therefore, Kappa 700 products are the primary focus of this note. It should be noted that IPGs equipped with the second-generation VCM algorithm (Kappa 900, EnPulse, Adapta/Versa/Sensia, and Relia) have not been observed with evoked response undersensing in the general population, though the items listed in "Follow-Up Considerations" may also be used on these devices.

Follow-Up Considerations

- Estimated longevity in the event the device goes to high output can be determined by the following steps. This allows the clinician to determine follow-up frequency if he or she is concerned the device may go to ERI due to high output.
 - Program the ventricular channel to 5 V, 1 ms
 - Navigate to Data/Battery and Lead Measurements
 - When the message stating "Warning – Old Data" is displayed, select "Yes" to measure battery voltage and lead impedance at the new ventricular outputs
 - An updated remaining longevity estimate will be calculated on the elevated outputs. Note the "Minimum Remaining Longevity." Clinical decisions can be based on this value.
 - Program the Amplitude and Pulse Widths back to their original values before leaving the session
- If the capture trends and lead impedance trends are stable, VCM can be programmed to "Monitor Only" for the remaining device life. This should be considered only if remaining longevity is 18 months or less.
- Follow-up frequency can be increased for those patients who do not have stable capture or lead impedance trends. This can be done via a CareLink Home Monitor, or in-office.

¹ Medtronic, Inc. (2001). Medtronic Kappa 700/600 Series Pacemaker Reference Guide (Chapter 4, p. 27). Can be retrieved from <http://manuals.medtronic.com>.

Performance Notes

General Follow-Up and Replacement of ICD Leads

Implanted leads operate in the challenging biochemical environment of the human body and the body's response to foreign objects. Implanted leads are also subject to mechanical stresses associated with heart motion, body motion, and patient anatomy.

In this environment, pacemaker and defibrillation leads cannot be expected to last forever. Unlike implantable cardioverter defibrillators (ICDs), a lead's longevity cannot be predicted nor are there simple indicators that a lead is approaching the end of its service life. The determination that a lead may be approaching end of service life requires follow-up of the chronically implanted lead and thorough evaluation of lead integrity at ICD replacement.

Follow-Up of Chronically Implanted Leads

The frequency of follow-up for ICD patients will depend on a number of factors including the patient's medical condition, ICD system implant time, hospital/clinic follow-up practice, and Medicare guidelines. In all cases, it is important to assess the functionality of the ICD system and the integrity. For newly implanted leads, it is beneficial to establish a baseline of chronic performance parameters once the lead has stabilized, generally within 6 to 12 months after implant. These performance parameters should include pacing and sensing thresholds and impedance. During routine patient follow-up, these procedures can be used to evaluate lead integrity.

- Measure pacing and sensing threshold and compare to the chronic baseline. Significant increases or decreases may be indicative of lead failure, dislodgement, perforation, exit block, etc.
- Measure pacing impedance where possible and compare to the chronic baseline. Decreases of 30% or more or pacing impedances below 200-250 ohms may be indicative of insulation failure. Sudden and significant increases in pacing impedance may be indicative of conductor fracture.
- High voltage lead circuit impedance should be between 10-75 ohms at system implant. Chronic measurements below 10 and above 200 ohms may be indicative of high voltage lead circuit failure.
- Carefully review ECGs or the nonsustained detection log on Medtronic ICDs for indications of pacing and/or sensing abnormalities such as oversensing, undersensing, and loss of capture
- Elicit and investigate any patient complaints/symptoms that may be suggestive of potential lead failure

Where routine follow-up indicates, additional tools should be used to further evaluate performance. Tools include radiographic data, ICD electrograms, ICD Patient Alert and performance information from the Product Surveillance Registry (PSR).

The final decision on the functional integrity and continued use of an implanted lead must be a matter of medical judgment based on these factors as well as specific patient conditions.

General Criteria for Lead Replacement

The evaluation of a chronically implanted lead is an important part of the decision to continue to use the lead with a new ICD. However, these results alone do not necessarily predict the future integrity of that lead. With the expected longevity of today's ICDs varying between approximately 5 and 10 years, a physician replacing a device should consider a number of factors, including those listed below.

Factors that should be considered in a decision to replace or continue to use include:

- Pacing and sensing thresholds should be evaluated for the potential to maintain acceptable levels
- Pacing impedance should be measured. Bear in mind that pacing impedance below 250 ohms results in excessive battery current drain, which may seriously compromise ICD longevity, regardless of lead integrity.
- The physical appearance of the lead should be examined for insulation cracks, breaches, or other indications of lead wear or degradation
- Medtronic System Longevity Study data should be referenced. Actuarial survival of the lead and the observed lead failure mechanisms are specific factors to consider. Use of a new lead should be considered if failure mechanisms suggest an increased time dependency as suggested in the shape of performance curve for the specific lead model.
- Current publications may provide additional information on the clinical management of leads.¹⁻³ Ultimately, the decision to replace an implanted lead involves medical judgment.

¹ Hauser RG, Cannom D, Hayes DL, et al. Long-term structural failure of coaxial polyurethane implantable cardioverter defibrillator leads. *PACE*. June 2002;25(6):879-882.

² Ellenbogen KA, Wood MA, Shepard RK, et al. Detection and management of an implantable cardioverter defibrillator lead failure: incidence and clinical implications. *J Am Coll Cardiol*. January 1, 2003;41(1):73-80.

³ Hauser RG, Kallinen LM, Almquist AK, Gornick CC, Katsiyannis WT. Early failure of a small-diameter high-voltage implantable cardioverter-defibrillator lead. *Heart Rhythm*. July 2007;4(7):892-896.

Clinical Management of High-Voltage Lead System Oversensing

Appropriate sensing by an ICD system refers to the sensing of cardiac events that may or may not require therapy delivery. ICD systems must sense relatively large QRS complexes while avoiding sensing of smaller T waves, yet continue to sense often small variable amplitude ventricular fibrillation. Thus, ICD systems attempt to dynamically adjust sensing of electrical events and discriminate between them based on detection algorithms and programmed settings.

Inappropriate sensing can occur when an ICD system classifies events of non-cardiac origin as QRS/VF events, or senses and counts T and far-field P waves as ventricular depolarizations. This is often referred to as “oversensing,” and may result in delivery of inappropriate high-voltage therapies. This is due, in part, to the desire to err on the side of delivering lifesaving high voltage therapy rather than withholding

it. Thus, an ICD system that is experiencing oversensing issues will continue to deliver therapeutic shocks as required, but may also subject the patient to unnecessary shocks.

Oversensing can be difficult to manage, in that the precipitating cause of the oversensing can be problematic to isolate. Oversensing can be caused by many factors, including myopotentials/far-field sensing, electromagnetic interference, T wave sensing, connector issues, incomplete or complete conductor fractures, and insulation breaches. While the individual physician must exercise medical judgment in determination of appropriate clinical management of ICD systems, the chart below may assist in the process of causal factor differentiation and possible intervention.

Phenomenon	Causal Factors	Characteristics	Management/Comments
Myopotentials/ Far-field sensing	Diaphragmatic muscle potentials in breathing, wide tip-to-ring (coil on integrated bipolar leads) spacing	Nonphysiological sensed event on EGM, which may confuse detection potentially resulting in false positive shocks	Check R waves for deterioration. Reprogram sensitivity. Try repositioning lead. Consider change-out to true bipolar lead, or if true bipolar lead in use, one with closer tip-to-ring spacing than current lead.
EMI (Electro-Magnetic Interference)	Arc welders, electrical generators, store walk-through security scanners, poorly insulated electrical equipment	Multiple and consecutive short intervals (< 140 ms) independent of underlying sinus beats. Associated with proximity to the EMI source.	Avoid EMI areas. True bipolar leads less susceptible.
T-wave sensing	Drugs, ischemic tissue, exercise, Long QT syndrome, electrolyte imbalance	Sense markers seen on EGM related to T wave. False positive detection.	Check for R wave deterioration and characteristics. If R wave > 3.0 mV, reprogram sensitivity. If R wave < 3.0 mV, reposition/replace lead. Address causal factor (e.g., drugs [if appropriate/medically viable]).
Connector problems	Loose setscrew, cross-threaded setscrew, incomplete lead insertion into header	This is an acute phenomenon seen within 6 months of implant (usually sooner)	Requires invasive check of connections. May be reproducible with pocket manipulation.
Incomplete conductor fracture	One or more filars of a multifilar conductor fracturing while leaving enough filars intact to provide a conduction circuit	Characterized by chaotic oversensing related to motion of the fracture site	Check EGMs and x-rays. Manipulate lead at suspected fracture site if possible as a provocative test. If confirmed, replace lead.
Lead insulation breach	Cuts, tears, metal ion oxidization, abrasion, cold flow, environmental stress cracking	Characterized by cyclical and/or erratic, intermittent, spontaneous oversensing; often post-pace or post-shock can cause false positives	Replace lead. If acute, usually secondary to implant damage/replacement damage. If late, material characteristic.
Oversensing during interrogation with programming head (not wireless telemetry) with complete lead fracture	Interrogation with a programming head in combination with complete lead fracture that creates an open circuit can induce noise on the sensing circuitry inside the ICD can	Nonphysiologic sensed event on EGM. If detection is enabled during interrogation, oversensing may result in inappropriate therapy.	Quickly remove the programming head. CANCEL the interrupted interrogation and manually load the software for the specific device model. Reposition the programmer head over the device and immediately select SUSPEND. Device will resume detection when programming head is removed, or when RESUME is selected. Replace lead.

Technical Services is available at all times to advise clinicians in the troubleshooting and management of Medtronic products. For assistance in the United States, please call 1 (800) 723-4636. In other countries, please contact your local Medtronic representative.

Performance Notes

Tests and Observations for Clinical Assessment of Chronic Pacing Leads

Test/Observation	Possible Insulation Failure	Possible Conductor Failure	Possible Other System Failure	Effect on Test/Observation
Pacing Impedance (Telemetered or Measured Invasively)	Sudden and Significant Decrease	Sudden and Significant Increase	Dislodgement. Perforation. Electrolyte Imbalance. Improper IPG/Lead Connection. . .	Decrease Increase or Decrease Increase or Decrease Increase or Decrease
Pacing Thresholds (Telemetered/Programmed or Measured Invasively)	Sudden and Significant Increase, Especially in Bipolar System	Sudden and Significant Increase	Dislodgement. Exit Block. Infarct at Electrode Site. Perforation. Improper IPG/Lead Connection. . .	Increase Increase Increase Increase Increase
Electrograms (Telemetered or Measured Invasively)	Sudden and Significant Decrease in Amplitudes and/or Slew Rates for P and/or R Waves	Sudden and Significant Decrease or Disappearance of Amplitudes and/or Slew Rates for P and/or R Waves	Dislodgement. Perforation Infarct at Electrode Site. Electrolyte Imbalance. Improper IPG/Lead Connection. . .	Decrease Decrease Decrease Decrease Decrease
Waveform Analysis (Oscillographs of Pacer Artifact from ECG Electrodes)	Sudden Increase in Ratios of Leading-Edge Voltages to Trailing-Edge Voltages (i.e., over 25% increase)	Intermittent or No Pacer Artifacts (Even in Asynchronous Mode)	Improper IPG/Lead Connection	Intermittent or No Pacer Artifacts (Even in Asynchronous Mode)
Radiographs (Post-Implant, Recent, Current)	Not Discernible	Visual Observation of Conductor/Connector/ Electrode Fracture (Sometimes Discernible)	Dislodgement or Perforation. Improper IPG/Lead Connection.	Sometimes Discernible
Visual Inspection (Invasive)	Insulation Breach and/or Degradation, or Ligature Cut-Through	Not Easily Discernible	Connector Defect or Connector Pulled Apart. Improper IPG/Lead Connection.	Sometimes Discernible
Pectoral Muscle Stimulation	Sudden Onset, Especially in Bipolar System		Connector Defect in Bipolar or Unipolar. Hypersensitivity to Unipolar Pulse Generator Can. Anti-Stim Coating or Protection Deficient.	
Phrenic Nerve/ Diaphragmatic Stimulation	Sudden Onset in Bipolar or Unipolar Systems		Perforation or Displacement of Atrial Lead (Phrenic Nerve)	
Pacemaker ECG Stimulus Artifact Size and Morphology Change (May Not Be Possible with Digital ECG)	Sudden Onset and Significant Change, Especially in Bipolar System (Increase in Size)	Sudden Changes, Usually a Decrease in Size	Perforation or Dislodgement. Connector Defect. Improper IPG/Lead Connection.	Sometimes Discernible
Oversensing (Intermittent or Continuous)	Sudden Onset, Especially in Bipolar Systems		Physical Contact between the Electrode(s) on the Lead and that of Another Lead. Inappropriate IPG Parameter Setting. Improper IPG/Lead Connection.	Sometimes Discernible
Undersensing (Intermittent or Continuous)	Sudden Onset in Either Unipolar or Bipolar Systems	Sudden Onset in Either Unipolar or Bipolar Systems	Dislodgement or Perforation. Infarct at Electrode Site. Electrolyte Imbalance. Inappropriate IPG Parameter Setting. Improper IPG/Lead Connection.	Sometimes Discernible
Loss of Capture	See "Pacing Thresholds" Above	See "Pacing Thresholds" Above	See "Pacing Thresholds" Above	

Mailer Kits Available for Returning Product

Medtronic urges all physicians to return explanted products and to notify Medtronic when a product is no longer in use, regardless of reason for explant or removal from use. The procedures for returning products vary by geographic location.

Mailer kits with prepaid US postage are available for use within the United States to send CRT, ICD, IPG, and leads to Medtronic's CRDM Returned Product Analysis Lab. These mailers are sized to accommodate the devices and leads from a single patient or clinical event and are designed to meet US postal regulations for mailing biohazard materials.

If the product being returned is located outside the United States, please contact your local Medtronic representative for instructions.

Medtronic also requests the return of devices from non-clinical sources, such as funeral homes, and will assume responsibility for storage and disposal of the product once received.

Mailer kits can be obtained by contacting the Returned Product Lab.

CRDM Returned Product Analysis Laboratory
Phone: 1 (800) 328-2518, ext. 44800
Email: crdm.returnedproduct@medtronic.com



www.medtronic.com

World Headquarters

Medtronic, Inc.
710 Medtronic Parkway
Minneapolis, MN 55432-5604
USA
Tel: (763) 514-4000
Fax: (763) 514-4879

Medtronic USA, Inc.
Toll-free: 1 (800) 328-2518
(24-hour technical support for
physicians and medical professionals)

Europe

Medtronic International Trading Sàrl
Route du Molliou 31
CH-1131 Tolochenaz
Switzerland
Tel: (41 21) 802 7000
Fax: (41 21) 802 7900

Canada

Medtronic of Canada Ltd.
6733 Kitimat Road
Mississauga, Ontario L5N 1W3
Canada
Tel: (905) 826-6020
Fax: (905) 826-6620
Toll-free: 1 (800) 268-5346

Asia Pacific

Medtronic International, Ltd.
16/F Manulife Plaza
The Lee Gardens, 33 Hysan Avenue
Causeway Bay
Hong Kong
Tel: (852) 2891 4456
Fax: (852) 2891 6830
enquiryap@medtronic.com

Latin America

Medtronic USA, Inc.
Doral Corporate Center II
3750 NW 87th Avenue Suite 700
Miami, FL 33178
USA
Tel: (305) 500-9328
Fax: (786) 709-4244



Medtronic