

CARDIAC RHYTHM & HEART FAILURE

Product Performance Report

Important Patient Management Information for Physicians

2020

2nd Edition – Issue 83

Medtronic

CRHF Product Performance Report

2020

2nd Edition

Issue 83

Introduction	3
Method for Estimating CRT, ICD, and IPG Device Performance	6
CRT-D	11
CRT-P	42
ICD	51
IPG	94
Method for Estimating Lead Performance	120
Pacing Leads	125
Defibrillation Leads	139
Left Heart Leads	152
Epi/Myocardial Leads	162
VDD Single Pass Lead	165
ICD and CRT-D Charge Time Performance	166
Customer Communications	173

Cutoff date for this edition is 31 Jul 2020 for Lead Study data and 09 October 2020 for all other data, unless otherwise stated.

Our Commitment to Quality

Medtronic was founded in 1949 and has grown to become a global leader in medical technology. Seeing what a difference medical technology could make in the lives of patients inspired our founder to develop the Medtronic Mission, which remains unchanged today.

The third tenet of the mission is all about quality:

"To strive without reserve for the greatest possible reliability and quality in our products, to be the unsurpassed standard of comparison, and to be recognized as a company of dedication, honesty, integrity, and service."

Regardless of function, all CRHF employees play a role in product quality. Whether designing new therapies, sourcing components, manufacturing products, hiring talented people, assigning financial resources to project teams, or serving in one of the hundreds of other roles, every employee has an influence on product quality.

Product performance information is received from many sources through various channels. Medtronic monitors information from many sources from Research and Development through Manufacturing and Field Performance Vigilance.

When a device is returned to Medtronic, laboratory technicians and engineers assess overall device function. Analysis of returned product is performed according to written procedures. These procedures determine the minimum analysis required. The analysis required varies depending on the type of device, age of the device, the associated information received with the device, actual experience with models of similar design, and other factors. Additional analysis is performed as necessary to investigate a performance concern from a customer, or to collect specific reliability data.

When a malfunction is identified, failure analysis is performed to provide the detailed information necessary to investigate possible causes and actions. Medtronic CRHF maintains in-house expertise and performs its failure analysis using facilities it owns and supports. This capability permits detailed failure analysis.

Analysis results are compared to original manufacturing records and design intent. Clinical observations are added to laboratory findings to help determine root cause. Each event is then compared to other events. If a pattern is detected, actions are taken to identify a common root cause, assess patient risk and an appropriate course of action.

Medtronic instituted the industry's first product performance reports in 1983 by publishing data on our chronic lead studies. Pacemakers and other devices followed as our performance reporting has constantly evolved based on customer needs and feedback. One thing has been a constant. It is our sincere commitment to communicate clearly, offering timely and appropriate product performance data and reliability information. This has always been and will continue to be our goal.

Contact Information

We invite our customers to use these telephone numbers to call with suggestions, inquiries, or specific problems related to our products.

US Technical Services Department

tshelp@medtronic.com

Phone:

1 (800) 723-4636 (Tachy)

1 (800) 505-4636 (Brady)

Fax:

1 (800) 824-2362

US Instrumental Technical Services

1 (800) 638-1991

Editorial Staff

IndependentPhysicianQualityPanel

David Cannom, MD, Los Angeles, CA
Steven Compton, MD, Anchorage, AK
James Daubert, MD, Durham, NC
N.A. Mark Estes, MD, Pittsburgh, PA
Kevin Hackett, MD, Columbus, OH
Andrew Krahn, MD, Vancouver, BC
Rachel Lampert, MD, New Haven, CT
R. Hardwin Mead, MD, Palo Alto, CA
Kevin Wheelan, MD, Dallas, TX

Editor

Kirk Hauge, Vice President, CRHF Quality and Regulatory

International Technical Centers

Europe (Heerlen NL) +31-45-566-8844

Japan (Tokyo) +81-3-6430-7026

rs.mst-techserviceseuropa@medtronic.com

Japan (Tokyo)

Please contact local Medtronic Representative.

Australia-New Zealand

au.crdmtechservices@medtronic.com

For questions related to returning explanted product or returning product that shows signs of malfunction, please contact:

Outside the United States:

Your Medtronic representative or international technical center at the number above.

Within the United States:

Your Medtronic representative or
CRHF Returned Product Analysis Laboratory
Phone: 1 (800) 328-2518, ext. 44800

Email:

crdm.returnedproduct@medtronic.com

Trademarks of Medtronic

Adapta®	Kappa®
Advisa®	Marquis®
Advisa DR MRI®	Maximo®
Amplia MRI™	Medtronic
Attain®	CareAlert®
Attain Ability®	Medtronic
Attain StarFix®	CareLink®
Attain Prevail®	MVP®
Brava™	Performa™
CapSure®	Protecta®
CapSure Sense®	Quick Look™
CapSureFix®	Relia™
CapSureFix Novus™	Reveal LINQ™
Capture	Revo MRI®
Management®	Secura®
Cardia™	SelectSecure®
CareLink®	Sensia®
Claria MRI™	Sensing Assurance
Compia MRI™	Sigma
Concerto®	Sprint®
Consulta®	Sprint Fidelis
Egida™	Sprint Quattro®
EnRhythm®	Sprint Quattro
EnRhythm MRI™	Secure®
Ensura MRI™	Surefix®
Evera™	Syncra®
InSync®	Transvene
	Versa®
	Virtuoso®
	Visia AF MRI™
	Viva™

Introduction

For 37 years, Medtronic has monitored performance via both returned product analysis and multicenter clinical studies.

This Product Performance Report (PPR) presents device survival estimates, advisory summaries, performance notes, and other information pertinent to assessing the performance of Medtronic implantable pulse generators (IPGs), implantable cardioverter defibrillators (ICDs), cardiac resynchronization therapy (CRT) devices, and implantable pacing and defibrillation leads.

This Product Performance Report has been prepared in accordance with International Standard ISO 5841- 2:2000(E).

The survival estimates provided in this report are considered to be representative of worldwide performance.

Survival Estimates

Medtronic, like other companies, monitors CRT, ICD, and IPG device performance using returned product analysis. We also monitor CRT, ICD, and IPG device performance using an active multicenter clinical study.

Returned product analysis is a passive approach to assessing product performance. This approach provides a suitable measure of product performance only when a significant number of explanted products are returned to the manufacturer. Returned product analysis provides a measure of hardware performance, but not necessarily the total clinical performance (e.g., the incidence of complications such as infection, erosion, muscle stimulation, etc. are not estimated).

The survival estimates provided in this report for CRT, ICD, and IPG devices are based on returned product analysis. This approach is suitable because a significant number of explanted generators are returned for analysis.

Lead performance is monitored differently. In contrast to CRT, ICD, and IPG devices, a very small percentage of leads are returned to the manufacturer due to the difficulty of explanting them. For leads, an active clinical study provides more accurate survival estimates compared to estimates based solely on returned product analysis.

Survival estimates for leads are based on clinical observations recorded via Medtronic's PAN Registry. This multicenter clinical study is designed to record clinical observations representative of the total clinical experience. Therefore, the lead survival estimates include both lead hardware failure and lead-related medical complications, and do not differentiate a lead hardware failure from other clinical events such as exit block, perforation, dislodgement, or concurrent pulse generator failure.

ICD Charge Times

Since May 2000, Medtronic has provided important information on charge time performance of ICDs. The information provided in this report shows how ICD charge time can vary during the time a device is implanted. The information is presented in graphical format showing charge time as a function of implant time. The data for charge times are collected from devices enrolled in the PAN registry.

Introduction continued

Customer Communications - Advisory Summaries

This Product Performance Report includes summaries of all advisories applicable to the performance of the products included in the report. An advisory is added to the report when any product affected by the advisory remains in service and at risk of experiencing the behavior described in the advisory. The advisory will remain in the report until Medtronic estimates no product affected by the advisory remains active, or the risk of experiencing the behavior described in the advisory has passed.

For most advisories, the products subject to the advisory retain essentially the same survival probability as the products of the same model(s) not affected by the advisory. For those advisories where the survival probabilities of the affected and non-affected populations do differ significantly, Medtronic will provide separate survival data for each population. The separate survival data will remain in the report until Medtronic estimates no affected product remains in active service.

Customer Communications- Performance Notes

This report concludes with a number of Performance Notes developed by Medtronic to provide additional product performance information relevant to follow-up practice and patient management.

How You Can Help

Medtronic urges all physicians to return explanted products and to notify Medtronic when a product is no longer in use, regardless of the reason for explant or removal from use. The procedures for returning products vary by geographic location.

Mailer kits with prepaid US postage are available for use within the United States to send CRTs, ICDs, IPGs, ICMs, and leads to Medtronic's Cardiac Rhythm and Heart Failure (CRHF) Returned Product Analysis Lab. These mailers are sized to accommodate the devices and leads from a single patient or clinical event and are designed to meet US postal regulations for mailing biohazard materials.

If the product being returned is located outside the United States, please contact your local Medtronic representative for instructions.

Medtronic also requests the return of explanted products from non-clinical sources, such as funeral homes, and will assume responsibility for storage and disposal of the product once received.

Mailer kits can be obtained by contacting the Returned Product Lab. For information on how to contact the Lab, refer to the Contact Information page of this report.

We continually strive to improve this CRHF Product Performance Report. In keeping with this philosophy, we ask for your suggestions on the content and format of this report, as well as any information you have regarding the performance of Medtronic products. For information on how to comment on this report, see the Contact Information page.

Introduction continued

Overview of Survival Analysis

Medtronic uses the Cutler-Ederer actuarial life table method for devices and Kaplan-Meier for leads to estimate the length of time over which they will perform within performance limits established by Medtronic. This probability to perform within performance limits over time is called the survival probability.

Devices and leads are followed until an event occurs where the device or lead ceases to operate within performance limits. The length of time from implant to the event is recorded for individual devices and leads in the population sample. The population sample for CRT, ICD, and IPG devices is made up of patients whose devices are registered as implanted in the United States. For leads, the population sample is the patients enrolled in our multicenter, international prospective Product Surveillance Registry.

For CRTs, IPGs and ICDs, the events can be normal battery depletion or a device malfunction. For leads, the events are complications as defined in the study protocol.

The actuarial life table method allows Medtronic to account for devices and leads removed from service for reasons unrelated to performance and for device and leads still in service. Devices and leads removed for reasons unrelated to performance or are still in service are said to be suspended. Examples of devices and leads removed from service for reasons unrelated to performance include:

- Removed to upgrade the device or lead
- No longer in service due to the death of the patient for reasons unrelated to the device or leads
- Implanted in patients who are lost to follow-up

For each suspension, the device or lead has performed within performance limits for a period of time, after which its performance is unknown.

Confidence Intervals

Since survival curves are based on a sample of the device and lead population, they are only estimates of survival. The larger the effective sample size, the more confident the estimate. A confidence interval can be calculated to assess the confidence in an estimate. In the Product Performance Report, Medtronic provides a 95% confidence interval. This can be interpreted as meaning that 95% of the time, the true survival of the device will fall somewhere in the interval.

Survival Curves in the Product Performance Report

Since the survival estimate can become very imprecise with small effective sample sizes, Medtronic truncates the survival curve when the effective sample size is less than 100 for CRTs, ICDs, and IPGs, and when the number entered is less than 50 for leads. The survival charts in the Product Performance Report show the effective sample size for each year interval where Medtronic has experience. When the effective sample size reaches 100 for CRTs, ICDs, and IPGs or when the number entered reaches 50 for leads, the next data point is added to the survival curve.

Although the report provides tabular data in one-year intervals, the device curves are actually computed and plotted using the Cutler-Ederer method and 1-month intervals (for CRT, ICD, and IPG devices) and leads curves are computed and plotted using Kaplan-Meier, which uses individual survival times.

A number of references are available for additional information on survival analysis using the Cutler-Ederer life table method¹ and for the Kaplan-Meier method.²

¹

Lee, Elisa T. (2003) *Statistical Methods for Survival Data Analysis – 3rd Edition* (Wiley Series in Probability and Statistics).

² Klein, John P., Moeschberger, Melvin L. *Survival Analysis Techniques for Censored and Truncated Data*, New York: Springer-Verlag New York, Inc., 1997.

Method for Estimating CRT, ICD, and IPG Device Performance

The performance of CRT, ICD, and IPG devices is expressed in terms of device survival estimates, where “survival” refers to the function of the device, not the survival of the patient. These survival estimates are intended to illustrate the probability that a device will survive for a given number of years without malfunction or battery depletion.

The survival estimates are determined from the analysis of Medtronic Cardiac Rhythm and Heart Failure (CRHF's) United States device registration data and US returned product analysis data. These data are presented graphically and numerically.

Because this analysis is based on returned product analysis, the performance data does not reflect any device-related medical complications such as erosion, infection, muscle stimulation, or muscle inhibition.

Categorization of Depleted and Malfunctioning Devices for Survival Analysis

For survival estimation, every device returned to Medtronic CRHF and analyzed in the CRHF Returned Product Analysis laboratory is assigned to one of three categories. The device 1) is functioning normally, 2) has reached normal battery depletion, or 3) has malfunctioned. This categorization is combined with data from our device registry for the total number of implants and the implant durations to create the survival curves presented on the following pages.

Definition of Malfunction

Medtronic CRHF considers a device as having malfunctioned whenever the analysis shows that any parameter was outside the performance limits established by Medtronic while implanted and in service. To be considered a malfunction or battery depletion, the device must have been returned to Medtronic and analyzed.

Devices damaged after explant, damaged due to failure to heed warnings or contraindications in the labeling, or damaged due to interaction with other implanted devices (including leads) are not considered device malfunctions.

A device subject to a safety advisory is not considered to have malfunctioned unless it has been returned to Medtronic CRHF and found, through analysis, to actually have performed outside the performance limits established by Medtronic.

Not all malfunctions expose the patient to a loss of therapy. Some malfunctions included in the following survival estimates may not have been detected at all by the physician or the patient. These malfunctions, however, are included in the survival estimates and provide important feedback to our product development organization.

To provide insight into the nature of malfunctions, each malfunction is categorized as Malfunction with Compromised Therapy Function or Malfunction without Compromised Therapy Function.

For this report, Normal Battery Depletion, Malfunction with Compromised Therapy Function, and Malfunction without Compromised Therapy Function are defined as follows:

Normal Battery Depletion – The condition when:

- (a) a device is returned with no associated complaint and the device has reached its elective replacement indicator(s) with implant time that meets or exceeds the nominal (50 percentile) predicted longevity at default (labeled) settings, or
- (b) a device is returned and the device has reached its elective replacement indicator(s) with implant time exceeding 80% of the expected longevity calculated using the available device setting information.
- (c) a device is taken out of service without an associated complaint and with evidence the battery reached its elective replacement indicator(s).

Medtronic CRHF establishes expected longevity by statistically characterizing the power consumed by the device and the power available from the device battery. This characterization is applied to a number of parameter configurations to derive a statistical mean longevity value and standard deviation for each parameter configuration. The statistical mean value minus three standard deviations is used as the expected longevity for determining if a battery depleted normally. The actual longevity achieved for any device while implanted will depend on the actual programmed parameters and patient factors, and may differ significantly from these estimates.

Malfunction with Compromised Therapy Function

The condition when a device is found to have malfunctioned in a manner that compromised pacing or defibrillation therapy (including complete loss or partial degradation), while implanted and in service, as confirmed by returned product analysis.

Examples: Sudden loss of battery voltage; accelerated current drain such that low battery was not detected before loss of therapy; sudden malfunction during defibrillation therapy resulting in aborted delivery of therapy, intermittent malfunction where therapy is compromised while in the malfunction state.

Malfunction without Compromised Therapy Function

The condition when a device is found to have malfunctioned in a manner that did not compromise pacing or defibrillation therapy, while implanted and in service, as confirmed by returned product analysis.

Examples: Error affecting diagnostic functions, telemetry function, data storage; malfunction of a component that causes battery to lose power quickly enough to cause premature battery depletion, but slowly enough that the condition is detected through normal follow-up before therapy is lost; mechanical problems with connector header that do not affect therapy.

Expanded Malfunction Detail

The malfunctions are further divided into categories that identify the subject area of the malfunction. The malfunctions are divided into the following subject areas:

Electrical Component – Findings linked to electrical components such as integrated circuits, resistors, capacitors, diodes, etc.

Electrical Interconnect – Findings linked to the connections between electrical components such as wires, solder joints, wire bonds, etc.

Battery – Findings linked to the battery and its components

Software/Firmware – Findings linked to software or firmware function

Possible Early Battery Depletion – Findings where the actual reported implant time is less than 80% of the expected longevity calculated using the available device setting information with no device malfunction observed. There may not be sufficient device setting information to determine conclusively if battery depletion was normal or premature in the absence of a specific root cause finding. However, returned devices meeting the above criteria are conservatively classified as Possible Early Battery Depletion malfunctions.

Other – Findings related to other components such as insulators, grommets, setscrews, and packaging, and findings where analysis is inconclusive.

Returned Product Analysis Process

Analysis of returned product is performed according to written procedures. These procedures determine the minimum analysis required. The analysis required varies depending on the type of device, age of the device, the associated information received with the device, actual experience with models of similar design, and other factors. Additional analysis is performed as necessary to investigate a performance concern from a customer, or to collect specific reliability data.

When a device is returned with a performance concern from a customer, the general analysis process includes a preliminary analysis of the device in its as-received condition, followed by an automated functional test using test equipment equivalent to the equipment used in manufacturing.

When a malfunction is identified, failure analysis is performed to provide the detailed information necessary to investigate possible causes and actions. Medtronic CRHF maintains in-house expertise and performs its failure analysis using facilities it owns and supports. This capability permits detailed failure analysis.

Statistical Methods for Survival Analysis

Of the several different statistical methods available for survival analysis, the Standard Actuarial Method, with suspensions assumed distributed evenly within the intervals (Cutler-Ederer Method), is used to determine survival estimates for CRT, IPG and ICD devices. Implant times are calculated from the implant date to the earlier of the explant date or the cutoff date of the report. From this data an estimate of the probability of device survival is calculated at each monthly interval.

On the following pages, each graph includes a survival curve where events include malfunctions and normal battery depletions. This survival curve is a good representation of the probability a device will survive a period of time without malfunction and without battery depletion. For example, if a device survival probability is 95% after 5 years of service, then the device has a 5% chance of being removed due to battery depletion or malfunction in the first 5 years following implant.

In addition, a second curve is included to show survival excluding normal battery depletion. This curve is a good representation of the probability for a device to survive without malfunction. This curve includes only malfunctions as events and excludes normal battery depletion.

Since the survival estimate can become very imprecise with small effective sample sizes, Medtronic truncates the survival curve when the effective sample size is less than 100 for CRT, ICD, and IPG devices. The survival charts in the Product Performance Report show the effective sample size for each year interval where we have experience. When the effective sample size reaches 100, the next data point is added to the survival curve.

Although the report provides tabular data in one-year intervals, the curves are actually computed and plotted using one-month intervals.

The data in the tables are rounded to the nearest tenth of one percent. Occasionally, a graph may show 100% survival, but have one or more malfunctions or battery depletions. This occurs because, even with the malfunctions or battery depletions, the data rounds to 100%.

Sample Size and How the Population and Population Samples Are Defined

The population sample from which the survival estimates are derived is comprised of the devices registered as implanted in the United States as of the report cutoff date. The number of registered implants, as well as an estimate of the number that remain in active service, is listed for each model. To be included in the population, the device must have been registered with Medtronic's registration system and implanted for at least one day.

This sample based on US implants is considered to be representative of the worldwide population, and therefore the survival estimates shown in this report should be representative of the performance worldwide of these models.

A CRT, ICD, or IPG model or model family will be included in this report when it has accumulated at least 10,000 implant months and will remain in the report as long as at least 500 devices remain active.

Methods Used to Adjust for Underreporting of Malfunction and Battery Depletion

The tables on the following pages show the actual number of malfunctions and battery depletions recorded by the analysis lab for US registered devices. Since not all devices are returned to Medtronic CRHF for analysis, these numbers underestimate the true number of malfunctions and battery depletions. To more accurately estimate the device survival probabilities, the number of malfunctions and battery depletions used to plot each interval of the "Including Normal Battery Depletion" survival curves is adjusted (multiplied) by a factor that is based on an estimate of the magnitude of underreporting. The magnitude of underreporting is estimated by comparing data in Medtronic's Device And Registrant Tracking (DART) system with data from Returned Product Analysis.

The DART system is an important element of Medtronic's Quality System. The DART system is designed to meet or exceed the US FDA's device tracking requirements set forth by the Safe Medical Devices Act. In the United States, over 98% of Medtronic's CRT, ICD, and IPG implants become registered in the DART system.

Because pacemakers do not cure the patient's underlying health problem, when a pacemaker stops functioning (due to either normal battery replacement or malfunction) it is replaced with a new pacemaker. Therefore, the replacement recorded in the DART system is a good indication that the previous pacemaker experienced either battery depletion or malfunction. The fraction of replaced devices that are subsequently returned can be used to estimate the correction factor for the under reporting of the combination of battery depletion and malfunction.

Note that devices of patients who have expired do not factor into the calculation of the correction. It is possible some proportion of these devices experienced battery depletion or malfunction. Since these are not counted into the correction factor based on the return rate of replaced devices, a correction factor based only on the return rate of replaced devices may still underestimate the true rate of battery depletion and malfunction. However, devices that are replaced because the patient is receiving a system upgrade or are removed because the patient no longer needs it (e.g., due to heart transplant) do contribute to the calculation of the correction factor and therefore impart an opposite bias.

Also note that this method of calculating the correction factor cannot distinguish between devices that are removed due to malfunction and those due to normal battery depletion. It might seem intuitive that devices that unexpectedly malfunction should be much more likely to be returned to the manufacturer than a device with ordinary normal battery depletion. But this has not been conclusively demonstrated. Therefore, this method only provides a correction factor reflecting the combination of battery depletion and malfunction.

No adjustment for underreporting is applied to the malfunction-free survival curve because a method for estimating malfunction-only underreporting has not been developed.


Adjustments to Registered Implants to Compensate for Unreported Devices Removed from Service

Devices are at times removed from service for reasons other than device malfunction or battery depletion. Examples are devices removed from service due to non-device related patient mortality and devices removed due to changes in the patient's medical condition. Because an accurate estimate of device survival depends on an accurate estimate of the number of devices in service, it is important not to overstate the number of devices in service.

Medtronic addresses this under reporting to ensure the number of devices in service is not overstated . Regular updates obtained from the Social Security Administration about deceased persons are used to update Medtronic's DART data about patients who have died but whose deaths had not been reported to Medtronic. In addition, the patient mortality rate derived from our DART system is monitored and compared to published mortality rates for comparable patient populations. If, during calculation of the survival curves, the patient mortality indicated by the data in DART is significantly different from published rates, an adjustment is applied to correct the difference. The correction factor for under reporting devices is also applied to account for devices that were removed and not reported or returned.

D204TRM Consulta CRT-D

US Market Release	Jan-12	Total Malfunctions	3
CE Approval Date		Therapy Function Not Compromised	3
Registered USA Implants	2,099	Battery Malfunction	1
Estimated Active USA Implants	552	Electrical Component	1
Normal Battery Depletions	713	Poss Early Battery Depltn	1
		Therapy Function Compromised	0




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 101 mo
Excluding NBD	100.0%	99.7%	99.3%	98.9%	98.8%	98.7%	98.7%	98.7%	98.7%
Including NBD	99.3%	97.9%	92.3%	79.4%	55.1%	23.4%	12.7%	10.0%	8.1%
Effective Sample Size	57376	52192	45293	34773	19534	6232	2403	816	147

D214TRM Consulta CRT-D

US Market Release		Total Malfunctions	
CE Approval Date	Jul-10	Therapy Function Not Compromised	
Registered USA Implants		Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 101 mo
Excluding NBD	100.0%	99.7%	99.3%	98.9%	98.8%	98.7%	98.7%	98.7%	98.7%
Including NBD	99.3%	97.9%	92.3%	79.4%	55.1%	23.4%	12.7%	10.0%	8.1%
Effective Sample Size	57376	52192	45293	34773	19534	6232	2403	816	147

D224TRK Consulta CRT-D

US Market Release	Sep-08	Total Malfunctions	602
CE Approval Date		Therapy Function Not Compromised	571
Registered USA Implants	65,980	Battery Malfunction	2
Estimated Active USA Implants	10,888	Electrical Component	65
Normal Battery Depletions	18,859	Electrical Interconnect	1
		Other Malfunction	1
		Poss Early Battery Depltn	496
		Software Malfunction	6
		Therapy Function Compromised	31
		Battery Malfunction	5
		Electrical Component	26




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 101 mo
Excluding NBD	100.0%	99.7%	99.3%	98.9%	98.8%	98.7%	98.7%	98.7%	98.7%
Including NBD	99.3%	97.9%	92.3%	79.4%	55.1%	23.4%	12.7%	10.0%	8.1%
Effective Sample Size	57376	52192	45293	34773	19534	6232	2403	816	147

D234TRK Consulta CRT-D

US Market Release		Total Malfunctions	
CE Approval Date	Mar-08	Therapy Function Not Compromised	
Registered USA Implants	3	Therapy Function Compromised	
Estimated Active USA Implants	1		
Normal Battery Depletions			




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 101 mo
Excluding NBD	100.0%	99.7%	99.3%	98.9%	98.8%	98.7%	98.7%	98.7%	98.7%
Including NBD	99.3%	97.9%	92.3%	79.4%	55.1%	23.4%	12.7%	10.0%	8.1%
Effective Sample Size	57376	52192	45293	34773	19534	6232	2403	816	147

D264TRM

Maximo II CRT-D

US Market Release	Jan-12	Total Malfunctions	1
CE Approval Date	Jul-10	Therapy Function Not Compromised	1
Registered USA Implants	15	Other Malfunction	1
Estimated Active USA Implants	4	Therapy Function Compromised	0
Normal Battery Depletions	5		




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 104 mo
Excluding NBD	100.0%	99.7%	99.3%	98.8%	98.7%	98.7%	98.7%	98.7%	98.7%
Including NBD	99.3%	97.7%	92.6%	79.7%	53.9%	26.1%	18.1%	16.2%	14.7%
Effective Sample Size	12810	11551	10047	7659	4051	1421	774	402	116

D274TRK

Concerto II CRT-D

US Market Release	Aug-09	Total Malfunctions	186
CE Approval Date		Therapy Function Not Compromised	175
Registered USA Implants	30,173	Battery Malfunction	1
Estimated Active USA Implants	5,709	Electrical Component	22
Normal Battery Depletions	7,973	Poss Early Battery Depltn	151
		Software Malfunction	1
		Therapy Function Compromised	11
		Battery Malfunction	1
		Electrical Component	10




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 111 mo
Excluding NBD	100.0%	99.8%	99.5%	99.2%	99.1%	99.1%	99.1%	99.1%	99.0%	99.0%
Including NBD	99.3%	98.0%	92.8%	80.1%	54.6%	23.3%	14.8%	13.1%	12.3%	11.9%
Effective Sample Size	25183	22995	20031	15275	8095	2546	1270	998	511	106

D284TRK

Maximo II CRT-D

US Market Release	Sep-08	Total Malfunctions	135
CE Approval Date	Mar-08	Therapy Function Not Compromised	130
Registered USA Implants	15,250	Electrical Component	6
Estimated Active USA Implants	2,775	Poss Early Battery Depltn	124
Normal Battery Depletions	4,068	Therapy Function Compromised	5
		Electrical Component	5




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 104 mo
Excluding NBD	100.0%	99.7%	99.3%	98.8%	98.7%	98.7%	98.7%	98.7%	98.7%
Including NBD	99.3%	97.7%	92.6%	79.7%	53.9%	26.1%	18.1%	16.2%	14.7%
Effective Sample Size	12810	11551	10047	7659	4051	1421	774	402	116

D294TRK

Concerto II CRT-D


US Market Release		Total Malfunctions	
CE Approval Date	Aug-08	Therapy Function Not Compromised	
Registered USA Implants		Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 111 mo
Excluding NBD	100.0%	99.8%	99.5%	99.2%	99.1%	99.1%	99.1%	99.1%	99.0%	99.0%
Including NBD	99.3%	98.0%	92.8%	80.1%	54.6%	23.3%	14.8%	13.1%	12.3%	11.9%
Effective Sample Size	25183	22995	20031	15275	8095	2546	1270	998	511	106

US Market Release	Mar-11	Total Malfunctions	92
CE Approval Date		Therapy Function Not Compromised	73
Registered USA Implants	42,520	Battery Malfunction	7
Estimated Active USA Implants	11,179	Electrical Component	39
Normal Battery Depletions	10,299	Other Malfunction	2
		Poss Early Battery Depltn	25
		Therapy Function Compromised	19
		Battery Malfunction	11
		Electrical Component	8




- Excluding Normal Battery Depletion
- Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 90 mo
Excluding NBD	100.0%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.6%	98.7%	94.7%	85.5%	64.5%	31.0%	18.0%	14.8%
Effective Sample Size	55132	50767	44650	35273	21266	7825	2903	257

D314TRM Protecta XT CRT-D

US Market Release	Nov-11	Total Malfunctions	20
CE Approval Date		Therapy Function Not Compromised	17
Registered USA Implants	12,259	Battery Malfunction	4
Estimated Active USA Implants	3,307	Electrical Component	8
Normal Battery Depletions	3,449	Poss Early Battery Depltn	5
		Therapy Function Compromised	3
		Battery Malfunction	1
		Electrical Component	2




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 90 mo
Excluding NBD	100.0%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.6%	98.7%	94.7%	85.5%	64.5%	31.0%	18.0%	14.8%
Effective Sample Size	55132	50767	44650	35273	21266	7825	2903	257

D334TRG Protecta CRT-D

US Market Release	Mar-11	Total Malfunctions	13
CE Approval Date		Therapy Function Not Compromised	11
Registered USA Implants	8,100	Electrical Component	8
Estimated Active USA Implants	2,350	Poss Early Battery Depltn	3
Normal Battery Depletions	2,074	Therapy Function Compromised	2
		Electrical Component	1
		Electrical Interconnect	1




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 90 mo
Excluding NBD	100.0%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.6%	98.7%	94.7%	85.5%	64.5%	31.0%	18.0%	14.8%
Effective Sample Size	55132	50767	44650	35273	21266	7825	2903	257

D334TRM Protecta CRT-D

US Market Release	Nov-11	Total Malfunctions	8
CE Approval Date		Therapy Function Not Compromised	6
Registered USA Implants	1,786	Battery Malfunction	3
Estimated Active USA Implants	516	Electrical Component	1
Normal Battery Depletions	550	Poss Early Battery Depltn	2
		Therapy Function Compromised	2
		Battery Malfunction	2




• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 90 mo
Excluding NBD	100.0%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.6%	98.7%	94.7%	85.5%	64.5%	31.0%	18.0%	14.8%
Effective Sample Size	55132	50767	44650	35273	21266	7825	2903	257

D354TRG Protecta XT CRT-D

US Market Release		Total Malfunctions	
CE Approval Date	Mar-10	Therapy Function Not Compromised	
Registered USA Implants	5	Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions	1		



• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 90 mo
Excluding NBD	100.0%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.6%	98.7%	94.7%	85.5%	64.5%	31.0%	18.0%	14.8%
Effective Sample Size	55132	50767	44650	35273	21266	7825	2903	257

D354TRM

Protecta XT CRT-D

US Market Release

Total Malfunctions

CE Approval Date

Jul-10

Therapy Function Not Compromised


Registered USA Implants

2

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 90 mo
Excluding NBD	100.0%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.6%	98.7%	94.7%	85.5%	64.5%	31.0%	18.0%	14.8%
Effective Sample Size	55132	50767	44650	35273	21266	7825	2903	257

D364TRG

Protecta CRT-D

US Market Release

Total Malfunctions

CE Approval Date

Mar-10

Therapy Function Not Compromised


Registered USA Implants

1

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 90 mo
Excluding NBD	100.0%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.6%	98.7%	94.7%	85.5%	64.5%	31.0%	18.0%	14.8%
Effective Sample Size	55132	50767	44650	35273	21266	7825	2903	257

D364TRM

Protecta CRT-D

US Market Release

Total Malfunctions

CE Approval Date

Jul-10

Therapy Function Not Compromised


Registered USA Implants

1

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 90 mo
Excluding NBD	100.0%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.6%	98.7%	94.7%	85.5%	64.5%	31.0%	18.0%	14.8%
Effective Sample Size	55132	50767	44650	35273	21266	7825	2903	257

D384TRG

Cardia CRT-D

US Market Release

Total Malfunctions

CE Approval Date

Jan-11


Therapy Function Not Compromised

Registered USA Implants

Estimated Active USA Implants

Therapy Function Compromised

Normal Battery Depletions




• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 90 mo
Excluding NBD	100.0%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.6%	98.7%	94.7%	85.5%	64.5%	31.0%	18.0%	14.8%
Effective Sample Size	55132	50767	44650	35273	21266	7825	2903	257

D394TRG Egida CRT-D

US Market Release
CE Approval Date Jan-11
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 90 mo
Excluding NBD	100.0%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.6%	98.7%	94.7%	85.5%	64.5%	31.0%	18.0%	14.8%
Effective Sample Size	55132	50767	44650	35273	21266	7825	2903	257

DTBA1D1 Viva XT

US Market Release Jan-13
CE Approval Date
Registered USA Implants 57,340
Estimated Active USA Implants 40,182
Normal Battery Depletions 2,461

Total Malfunctions 55
Therapy Function Not Compromised 41
 Battery Malfunction 7
 Electrical Component 30
 Other Malfunction 3
 Poss Early Battery Depltn 1
Therapy Function Compromised 14
 Battery Malfunction 11
 Electrical Component 3




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 82 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.7%	98.9%	96.8%	92.4%	84.3%	65.0%
Effective Sample Size	89737	82978	74389	60065	39113	18300	658

DTBA1D4 Viva XT

US Market Release	Jan-13	Total Malfunctions	29
CE Approval Date		Therapy Function Not Compromised	23
Registered USA Implants	20,304	Battery Malfunction	3
Estimated Active USA Implants	14,757	Electrical Component	17
Normal Battery Depletions	1,052	Poss Early Battery Depltn	3
		Therapy Function Compromised	6
		Battery Malfunction	3
		Electrical Component	3




• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 82 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.7%	98.9%	96.8%	92.4%	84.3%	65.0%
Effective Sample Size	89737	82978	74389	60065	39113	18300	658

DTBA1Q1 Viva Quad XT

US Market Release	Jul-14	Total Malfunctions	7
CE Approval Date		Therapy Function Not Compromised	6
Registered USA Implants	11,081	Battery Malfunction	1
Estimated Active USA Implants	8,595	Electrical Component	3
Normal Battery Depletions	257	Other Malfunction	1
		Poss Early Battery Depltn	1
		Therapy Function Compromised	1
		Electrical Component	1




• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	at 71 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.8%	99.4%	98.3%	95.3%	83.0%
Effective Sample Size	34822	32321	29671	24946	13508	407

DTBA1QQ Viva Quad XT

US Market Release	Jul-14	Total Malfunctions	34
CE Approval Date		Therapy Function Not Compromised	27
Registered USA Implants	27,720	Battery Malfunction	6
Estimated Active USA Implants	23,490	Electrical Component	16
Normal Battery Depletions	603	Electrical Interconnect	1
		Other Malfunction	2
		Poss Early Battery Depltn	2
		Therapy Function Compromised	7
		Battery Malfunction	5
		Electrical Component	2




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	at 71 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.8%	99.4%	98.3%	95.3%	83.0%
Effective Sample Size	34822	32321	29671	24946	13508	407

DTBA2D1 Viva XT

US Market Release		Total Malfunctions	
CE Approval Date	Aug-16	Therapy Function Not Compromised	
Registered USA Implants	1	Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions	1		



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 82 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.7%	98.9%	96.8%	92.4%	84.3%	65.0%
Effective Sample Size	89737	82978	74389	60065	39113	18300	658

DTBA2D4

Viva XT

US Market Release

Total Malfunctions

CE Approval Date

Aug-12


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 82 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.7%	98.9%	96.8%	92.4%	84.3%	65.0%
Effective Sample Size	89737	82978	74389	60065	39113	18300	658

DTBA2Q1

Viva Quad XT

US Market Release

Total Malfunctions

CE Approval Date

Sep-13


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	at 71 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.8%	99.4%	98.3%	95.3%	83.0%
Effective Sample Size	34822	32321	29671	24946	13508	407

US Market Release
 CE Approval Date Aug-12
 Registered USA Implants
 Estimated Active USA Implants
 Normal Battery Depletions

Total Malfunctions
 Therapy Function Not Compromised
 Therapy Function Compromised




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	at 71 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.8%	99.4%	98.3%	95.3%	83.0%
Effective Sample Size	34822	32321	29671	24946	13508	407

US Market Release Jan-13
 CE Approval Date
 Registered USA Implants 14,126
 Estimated Active USA Implants 9,215
 Normal Battery Depletions 903

Total Malfunctions 16
 Therapy Function Not Compromised 12
 Battery Malfunction 6
 Electrical Component 3
 Other Malfunction 1
 Poss Early Battery Depltn 2
 Therapy Function Compromised 4
 Battery Malfunction 3
 Electrical Component 1




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 82 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.7%	98.9%	96.8%	92.4%	84.3%	65.0%
Effective Sample Size	89737	82978	74389	60065	39113	18300	658

DTBB1D4 Viva S

US Market Release	Jan-13	Total Malfunctions	6
CE Approval Date		Therapy Function Not Compromised	4
Registered USA Implants	4,654	Battery Malfunction	1
Estimated Active USA Implants	3,238	Electrical Component	2
Normal Battery Depletions	322	Other Malfunction	1
		Therapy Function Compromised	2
		Battery Malfunction	2




• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 82 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.7%	98.9%	96.8%	92.4%	84.3%	65.0%
Effective Sample Size	89737	82978	74389	60065	39113	18300	658

DTBB1Q1 Viva Quad S

US Market Release	Jul-14	Total Malfunctions	1
CE Approval Date		Therapy Function Not Compromised	1
Registered USA Implants	2,361	Electrical Component	1
Estimated Active USA Implants	1,833	Therapy Function Compromised	0
Normal Battery Depletions	53		




• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 82 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.7%	98.9%	96.8%	92.4%	84.3%	65.0%
Effective Sample Size	89737	82978	74389	60065	39113	18300	658

DTBB1QQ Viva Quad S

US Market Release	Jul-14	Total Malfunctions	9
CE Approval Date		Therapy Function Not Compromised	7
Registered USA Implants	5,118	Battery Malfunction	1
Estimated Active USA Implants	4,293	Electrical Component	2
Normal Battery Depletions	156	Other Malfunction	2
		Poss Early Battery Depltn	2
		Therapy Function Compromised	2
		Electrical Component	2




• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 82 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.7%	98.9%	96.8%	92.4%	84.3%	65.0%
Effective Sample Size	89737	82978	74389	60065	39113	18300	658

DTBB2D1 Viva S

US Market Release		Total Malfunctions	
CE Approval Date	Aug-12	Therapy Function Not Compromised	
Registered USA Implants		Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			



• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 82 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.7%	98.9%	96.8%	92.4%	84.3%	65.0%
Effective Sample Size	89737	82978	74389	60065	39113	18300	658

DTBB2D4

Viva S

US Market Release

Total Malfunctions

CE Approval Date

Aug-12


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 82 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.7%	98.9%	96.8%	92.4%	84.3%	65.0%
Effective Sample Size	89737	82978	74389	60065	39113	18300	658

DTBB2QQ

Viva Quad S

US Market Release

Total Malfunctions

CE Approval Date

Aug-12


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 82 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.7%	98.9%	96.8%	92.4%	84.3%	65.0%
Effective Sample Size	89737	82978	74389	60065	39113	18300	658

DTBC2D1 Brava

US Market Release
CE Approval Date Aug-12
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 82 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.7%	98.9%	96.8%	92.4%	84.3%	65.0%
Effective Sample Size	89737	82978	74389	60065	39113	18300	658

DTBC2D4 Brava

US Market Release
CE Approval Date Aug-12
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 82 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.7%	98.9%	96.8%	92.4%	84.3%	65.0%
Effective Sample Size	89737	82978	74389	60065	39113	18300	658

DTBC2Q1 Brava Quad

US Market Release
CE Approval Date Sep-13
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 82 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.7%	98.9%	96.8%	92.4%	84.3%	65.0%
Effective Sample Size	89737	82978	74389	60065	39113	18300	658

DTBC2QQ Brava Quad

US Market Release
CE Approval Date Aug-12
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 82 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.7%	98.9%	96.8%	92.4%	84.3%	65.0%
Effective Sample Size	89737	82978	74389	60065	39113	18300	658

DTBX1QQ Viva Quad C

US Market Release	Jul-14	Total Malfunctions	1
CE Approval Date		Therapy Function Not Compromised	1
Registered USA Implants	637	Electrical Component	1
Estimated Active USA Implants	315	Therapy Function Compromised	0
Normal Battery Depletions	103		




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	at 71 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.8%	99.4%	98.3%	95.3%	83.0%
Effective Sample Size	34822	32321	29671	24946	13508	407

DTBX2QQ Viva Quad C

US Market Release	Jul-14	Total Malfunctions	
CE Approval Date		Therapy Function Not Compromised	
Registered USA Implants		Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	at 71 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.8%	99.4%	98.3%	95.3%	83.0%
Effective Sample Size	34822	32321	29671	24946	13508	407

DTMA1D1 Claria MRI

US Market Release	Dec-16	Total Malfunctions	
CE Approval Date		Therapy Function Not Compromised	
Registered USA Implants	10,276	Therapy Function Compromised	
Estimated Active USA Implants	9,492		
Normal Battery Depletions	18		




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	at 52 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.1%	96.9%	96.9%
Effective Sample Size	22390	14415	5891	530	154

DTMA1D4 Claria MRI

US Market Release	Dec-16	Total Malfunctions	2
CE Approval Date		Therapy Function Not Compromised	1
Registered USA Implants	9,641	Electrical Component	1
Estimated Active USA Implants	9,018	Therapy Function Compromised	1
Normal Battery Depletions	15	Electrical Interconnect	1




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	at 52 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.1%	96.9%	96.9%
Effective Sample Size	22390	14415	5891	530	154

DTMA1Q1 Claria MRI

US Market Release	Dec-16	Total Malfunctions	2
CE Approval Date		Therapy Function Not Compromised	2
Registered USA Implants	7,405	Electrical Interconnect	1
Estimated Active USA Implants	6,964	Other Malfunction	1
Normal Battery Depletions	1	Therapy Function Compromised	0




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	at 55 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%
Including NBD	99.9%	99.9%	99.7%	99.0%	98.3%
Effective Sample Size	65154	42381	22802	6320	321

DTMA1QQ Claria MRI

US Market Release	Dec-16	Total Malfunctions	7
CE Approval Date		Therapy Function Not Compromised	5
Registered USA Implants	49,509	Electrical Component	5
Estimated Active USA Implants	47,443	Therapy Function Compromised	2
Normal Battery Depletions	32	Electrical Component	2



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion


Years	1	2	3	4	at 55 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%
Including NBD	99.9%	99.9%	99.7%	99.0%	98.3%
Effective Sample Size	65154	42381	22802	6320	321

DTMA2D1

Claria MRI

US Market Release
CE Approval Date Aug-16
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion


Years	1	2	3	4	at 52 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.1%	96.9%	96.9%
Effective Sample Size	22390	14415	5891	530	154

DTMA2D4

Claria MRI

US Market Release
CE Approval Date Feb-16
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	at 52 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.1%	96.9%	96.9%
Effective Sample Size	22390	14415	5891	530	154

DTMA2Q1 Claria MRI

US Market Release
CE Approval Date Aug-16
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	at 55 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%
Including NBD	99.9%	99.9%	99.7%	99.0%	98.3%
Effective Sample Size	65154	42381	22802	6320	321

DTMA2QQ Claria MRI

US Market Release
CE Approval Date Feb-16
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	at 55 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%
Including NBD	99.9%	99.9%	99.7%	99.0%	98.3%
Effective Sample Size	65154	42381	22802	6320	321

DTMB1D1 Amplia MRI

US Market Release	Dec-16	Total Malfunctions	2
CE Approval Date		Therapy Function Not Compromised	1
Registered USA Implants	5,212	Other Malfunction	1
Estimated Active USA Implants	4,761	Therapy Function Compromised	1
Normal Battery Depletions	10	Battery Malfunction	1




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	at 52 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.1%	96.9%	96.9%
Effective Sample Size	22390	14415	5891	530	154

DTMB1D4 Amplia MRI

US Market Release	Feb-16	Total Malfunctions	2
CE Approval Date		Therapy Function Not Compromised	2
Registered USA Implants	5,560	Electrical Component	2
Estimated Active USA Implants	4,975	Therapy Function Compromised	0
Normal Battery Depletions	26		




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	at 52 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.1%	96.9%	96.9%
Effective Sample Size	22390	14415	5891	530	154

DTMB1Q1 Amplia MRI

US Market Release	Dec-16	Total Malfunctions	
CE Approval Date		Therapy Function Not Compromised	
Registered USA Implants	3,212	Therapy Function Compromised	
Estimated Active USA Implants	2,943		
Normal Battery Depletions	3		




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	at 55 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%
Including NBD	99.9%	99.9%	99.7%	99.0%	98.3%
Effective Sample Size	65154	42381	22802	6320	321

DTMB1QQ Amplia MRI

US Market Release	Feb-16	Total Malfunctions	19
CE Approval Date		Therapy Function Not Compromised	17
Registered USA Implants	27,662	Battery Malfunction	3
Estimated Active USA Implants	25,835	Electrical Component	8
Normal Battery Depletions	73	Other Malfunction	5
		Poss Early Battery Depltn	1
		Therapy Function Compromised	2
		Battery Malfunction	2




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	at 55 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%
Including NBD	99.9%	99.9%	99.7%	99.0%	98.3%
Effective Sample Size	65154	42381	22802	6320	321

DTMB2D1 Amplia MRI

US Market Release
CE Approval Date Aug-16
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	at 52 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.1%	96.9%	96.9%
Effective Sample Size	22390	14415	5891	530	154

DTMB2D4 Amplia MRI

US Market Release
CE Approval Date Feb-16
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	at 52 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.1%	96.9%	96.9%
Effective Sample Size	22390	14415	5891	530	154

DTMB2Q1 Amplia MRI

US Market Release
CE Approval Date Aug-16
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	at 55 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%
Including NBD	99.9%	99.9%	99.7%	99.0%	98.3%
Effective Sample Size	65154	42381	22802	6320	321

DTMB2QQ Amplia MRI

US Market Release
CE Approval Date Feb-16
Registered USA Implants 1
Estimated Active USA Implants 1
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	at 55 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%
Including NBD	99.9%	99.9%	99.7%	99.0%	98.3%
Effective Sample Size	65154	42381	22802	6320	321

DTMC1D1 Compia MRI

US Market Release	Dec-16	Total Malfunctions	
CE Approval Date		Therapy Function Not Compromised	
Registered USA Implants	701	Therapy Function Compromised	
Estimated Active USA Implants	639		
Normal Battery Depletions	3		




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	at 52 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.1%	96.9%	96.9%
Effective Sample Size	22390	14415	5891	530	154

DTMC1QQ Compia MRI

US Market Release	Feb-16	Total Malfunctions	2
CE Approval Date		Therapy Function Not Compromised	2
Registered USA Implants	3,524	Electrical Component	2
Estimated Active USA Implants	3,287	Therapy Function Compromised	0
Normal Battery Depletions	9		



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion


Years	1	2	3	4	at 55 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%
Including NBD	99.9%	99.9%	99.7%	99.0%	98.3%
Effective Sample Size	65154	42381	22802	6320	321

DTMC2D1

Compia MRI

US Market Release
CE Approval Date Aug-16
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised



● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion


Years	1	2	3	4	at 52 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.1%	96.9%	96.9%
Effective Sample Size	22390	14415	5891	530	154

DTMC2D4

Compia MRI

US Market Release
CE Approval Date Feb-16
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	at 52 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.1%	96.9%	96.9%
Effective Sample Size	22390	14415	5891	530	154

US Market Release
CE Approval Date Feb-16
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions


Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	at 55 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%
Including NBD	99.9%	99.9%	99.7%	99.0%	98.3%
Effective Sample Size	65154	42381	22802	6320	321


US Market Release	Feb-03	Total Malfunctions	114
CE Approval Date	Feb-01	Therapy Function Not Compromised	65
Registered USA Implants	39,440	Battery Malfunction	53
Estimated Active USA Implants	4,354	Electrical Component	2
Normal Battery Depletions	5,183	Electrical Interconnect	3
		Other Malfunction	5
		Poss Early Battery Depltn	2
		Therapy Function Compromised	49
		Battery Malfunction	37
		Electrical Interconnect	12



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 134 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.7%	99.4%	99.0%	98.7%	98.6%	98.6%
Including NBD	99.2%	98.9%	98.0%	95.8%	92.0%	85.2%	73.2%	55.2%	37.1%	26.7%	21.1%	18.1%
Effective Sample Size	30288	25920	22239	18994	15837	12106	8593	5517	3109	1799	326	136

US Market Release	Mar-11	Total Malfunctions	6
CE Approval Date	May-10	Therapy Function Not Compromised	6
Registered USA Implants	10,230	Other Malfunction	1
Estimated Active USA Implants	5,608	Poss Early Battery Depltn	5
Normal Battery Depletions	508	Therapy Function Compromised	0




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 93 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.8%	99.5%	98.1%	95.8%	91.8%	84.2%	58.7%
Effective Sample Size	26671	23951	21390	18233	14209	9522	3922	374

C3TR01

Consulta CRT-P

US Market Release		Total Malfunctions	
CE Approval Date	May-10	Therapy Function Not Compromised	
Registered USA Implants	2	Therapy Function Compromised	
Estimated Active USA Implants	2		
Normal Battery Depletions			




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 93 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.8%	99.5%	98.1%	95.8%	91.8%	84.2%	58.7%
Effective Sample Size	26671	23951	21390	18233	14209	9522	3922	374

C4TR01

Consulta CRT-P

US Market Release	Mar-11	Total Malfunctions	7
CE Approval Date		Therapy Function Not Compromised	5
Registered USA Implants	23,541	Poss Early Battery Depltn	5
Estimated Active USA Implants	14,824	Therapy Function Compromised	2
Normal Battery Depletions	966	Electrical Component	1
		Poss Early Battery Depltn	1



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion


Years	1	2	3	4	5	6	7	at 93 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.8%	99.5%	98.1%	95.8%	91.8%	84.2%	58.7%
Effective Sample Size	26671	23951	21390	18233	14209	9522	3922	374

C5TR01

Viva CRT-P

US Market Release
CE Approval Date Apr-14
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised



● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion


Years	1	2	3	4	at 60 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%
Including NBD	99.9%	99.7%	99.5%	98.1%	94.3%
Effective Sample Size	7633	6853	6103	3325	183

C6TR01

Viva CRT-P

US Market Release Jul-14
CE Approval Date
Registered USA Implants 9,298
Estimated Active USA Implants 7,689
Normal Battery Depletions 86

Total Malfunctions 4
Therapy Function Not Compromised 4
 Poss Early Battery Depltn 4
Therapy Function Compromised 0




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	at 60 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%
Including NBD	99.9%	99.7%	99.5%	98.1%	94.3%
Effective Sample Size	7633	6853	6103	3325	183

W1TR01 Percepta CRTP MRI

US Market Release	May-17	Total Malfunctions	2
CE Approval Date		Therapy Function Not Compromised	1
Registered USA Implants	6,100	Other Malfunction	1
Estimated Active USA Implants	5,754	Therapy Function Compromised	1
Normal Battery Depletions	2	Electrical Component	1




- Excluding Normal Battery Depletion
- Including Normal Battery Depletion

Years	1	2	3	at 38 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.6%	99.6%
Effective Sample Size	5760	2514	304	124

W1TR02 Serena CRTP MRI

US Market Release	May-17	Total Malfunctions	1
CE Approval Date		Therapy Function Not Compromised	1
Registered USA Implants	1,357	Other Malfunction	1
Estimated Active USA Implants	1,263	Therapy Function Compromised	0
Normal Battery Depletions	1		




- Excluding Normal Battery Depletion
- Including Normal Battery Depletion

Years	1	2	3	at 38 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.6%	99.6%
Effective Sample Size	5760	2514	304	124

W1TR03

Solara CRTP MRI

US Market Release	May-17	Total Malfunctions	1
CE Approval Date		Therapy Function Not Compromised	1
Registered USA Implants	2,254	Electrical Component	1
Estimated Active USA Implants	2,088	Therapy Function Compromised	0
Normal Battery Depletions			




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	at 38 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.6%	99.6%
Effective Sample Size	5760	2514	304	124

W1TR04

Percepta CRTP MRI

US Market Release		Total Malfunctions	
CE Approval Date	Feb-17	Therapy Function Not Compromised	
Registered USA Implants		Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion


Years	1	2	3	at 38 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.6%	99.6%
Effective Sample Size	5760	2514	304	124

W1TR05

Serena CRTP MRI

US Market Release
 CE Approval Date Feb-17
 Registered USA Implants
 Estimated Active USA Implants
 Normal Battery Depletions

Total Malfunctions
 Therapy Function Not Compromised
 Therapy Function Compromised



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion


Years	1	2	3	at 38 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.6%	99.6%
Effective Sample Size	5760	2514	304	124

W1TR06

Solara CRTP MRI

US Market Release
 CE Approval Date Feb-17
 Registered USA Implants
 Estimated Active USA Implants
 Normal Battery Depletions

Total Malfunctions
 Therapy Function Not Compromised
 Therapy Function Compromised




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	at 38 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.6%	99.6%
Effective Sample Size	5760	2514	304	124

W4TR01

Percepta Quad CRTP MRI SureScan

US Market Release	May-17	Total Malfunctions	3
CE Approval Date		Therapy Function Not Compromised	3
Registered USA Implants	23,245	Electrical Component	2
Estimated Active USA Implants	21,982	Other Malfunction	1
Normal Battery Depletions	4	Therapy Function Compromised	0




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	at 40 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.7%	99.7%
Effective Sample Size	20580	10097	1975	114

W4TR02

Serena Quad CRTP MRI SureScan

US Market Release	May-17	Total Malfunctions	
CE Approval Date		Therapy Function Not Compromised	
Registered USA Implants	4,212	Therapy Function Compromised	
Estimated Active USA Implants	3,960		
Normal Battery Depletions	1		




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	at 40 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.7%	99.7%
Effective Sample Size	20580	10097	1975	114

W4TR03

Solara Quad CRTP MRI SureScan

US Market Release	May-17	Total Malfunctions	2
CE Approval Date		Therapy Function Not Compromised	0
Registered USA Implants	6,184	Therapy Function Compromised	2
Estimated Active USA Implants	5,781	Electrical Component	1
Normal Battery Depletions	1	Poss Early Battery Depltn	1




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	at 40 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.7%	99.7%
Effective Sample Size	20580	10097	1975	114

W4TR04

Percepta Quad CRT-P MRI SureScan

US Market Release		Total Malfunctions	
CE Approval Date	Feb-17	Therapy Function Not Compromised	
Registered USA Implants		Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion


Years	1	2	3	at 40 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.7%	99.7%
Effective Sample Size	20580	10097	1975	114

W4TR05

Serena Quad CRTP MRI SureScan

US Market Release
CE Approval Date Feb-17
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised



● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion


Years	1	2	3	at 40 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.7%	99.7%
Effective Sample Size	20580	10097	1975	114

W4TR06

Solara Quad CRTP MRI SureScan

US Market Release
CE Approval Date Feb-17
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	at 40 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.7%	99.7%
Effective Sample Size	20580	10097	1975	114

7232Cx

Maximo VR

US Market Release	Oct-03	Total Malfunctions	72
CE Approval Date	Oct-03	Therapy Function Not Compromised	57
Registered USA Implants	43,490	Electrical Component	28
Estimated Active USA Implants	4,549	Other Malfunction	2
Normal Battery Depletions	10,284	Poss Early Battery Depltn	25
		Software Malfunction	2
		Therapy Function Compromised	15
		Electrical Component	12
		Electrical Interconnect	1
		Other Malfunction	1
		Poss Early Battery Depltn	1




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 170 mo
Excluding NBD	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%	99.8%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.3%	99.0%	98.6%	98.2%	96.6%	90.6%	82.1%	69.9%	44.3%	18.0%	13.3%	12.4%	11.7%	10.9%	10.5%
Effective Sample Size	37910	33905	30200	26602	23407	20325	17148	13677	8062	2735	1656	1295	919	304	127

D164AWG

Virtuoso DR

US Market Release		Total Malfunctions	
CE Approval Date	Mar-06	Therapy Function Not Compromised	
Registered USA Implants	10	Therapy Function Compromised	
Estimated Active USA Implants	3		
Normal Battery Depletions	3		




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 135 mo
Excluding NBD	100.0%	99.9%	99.9%	99.4%	97.1%	96.9%	96.9%	96.8%	96.8%	96.7%	96.7%	96.7%
Including NBD	99.4%	99.1%	98.7%	97.1%	88.8%	69.5%	42.7%	22.7%	10.7%	6.2%	5.0%	4.1%
Effective Sample Size	62983	57728	52553	47696	40476	29339	16198	7380	2892	1318	639	134

D164VWC

Virtuoso VR

US Market Release		Total Malfunctions	1
CE Approval Date	Mar-06	Therapy Function Not Compromised	1
Registered USA Implants	6	Electrical Component	1
Estimated Active USA Implants	2	Therapy Function Compromised	0
Normal Battery Depletions			




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 140 mo
Excluding NBD	100.0%	99.9%	99.9%	99.7%	98.1%	96.8%	96.8%	96.6%	96.6%	96.6%	96.6%	96.6%
Including NBD	99.5%	99.2%	99.0%	98.5%	95.6%	86.4%	74.8%	55.8%	38.8%	25.7%	16.1%	10.6%
Effective Sample Size	28341	25818	23500	21476	19078	15942	12853	8833	5507	3235	1634	250

D204DRM

Secura DR

US Market Release	Jan-12	Total Malfunctions	4
CE Approval Date		Therapy Function Not Compromised	1
Registered USA Implants	1,879	Other Malfunction	1
Estimated Active USA Implants	786	Therapy Function Compromised	3
Normal Battery Depletions	233	Battery Malfunction	1
		Electrical Component	2




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 115 mo
Excluding NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.7%	99.6%	99.5%	99.5%	99.5%
Including NBD	99.5%	99.3%	98.9%	98.0%	95.2%	87.7%	69.6%	37.8%	20.7%	17.4%
Effective Sample Size	44696	41812	39249	36416	32590	26312	16979	6468	2058	404

D204VRM

Secura VR

US Market Release	May-12	Total Malfunctions	2
CE Approval Date		Therapy Function Not Compromised	1
Registered USA Implants	1,184	Electrical Component	1
Estimated Active USA Implants	809	Therapy Function Compromised	1
Normal Battery Depletions	6	Battery Malfunction	1




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 122 mo
Excluding NBD	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.6%	99.6%	99.6%
Including NBD	99.7%	99.5%	99.3%	99.0%	98.6%	98.1%	96.2%	91.2%	76.4%	43.8%	36.5%
Effective Sample Size	17912	16706	15718	14644	13440	12196	10777	7948	4358	651	244

D214DRM

Secura DR

US Market Release		Total Malfunctions	
CE Approval Date	Jul-10	Therapy Function Not Compromised	
Registered USA Implants	1	Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 115 mo
Excluding NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.7%	99.6%	99.5%	99.5%	99.5%
Including NBD	99.5%	99.3%	98.9%	98.0%	95.2%	87.7%	69.6%	37.8%	20.7%	17.4%
Effective Sample Size	44696	41812	39249	36416	32590	26312	16979	6468	2058	404

US Market Release

Total Malfunctions

CE Approval Date

Dec-10


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants


Normal Battery Depletions



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 122 mo
Excluding NBD	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.6%	99.6%	99.6%
Including NBD	99.7%	99.5%	99.3%	99.0%	98.6%	98.1%	96.2%	91.2%	76.4%	43.8%	36.5%
Effective Sample Size	17912	16706	15718	14644	13440	12196	10777	7948	4358	651	244


US Market Release	Sep-08	Total Malfunctions	151
CE Approval Date		Therapy Function Not Compromised	115
Registered USA Implants	49,915	Battery Malfunction	14
Estimated Active USA Implants	10,893	Electrical Component	38
Normal Battery Depletions	9,979	Other Malfunction	4
		Poss Early Battery Depltn	50
		Software Malfunction	9
		Therapy Function Compromised	36
		Battery Malfunction	20
		Electrical Component	13
		Other Malfunction	1
		Poss Early Battery Depltn	1
		Software Malfunction	1



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 115 mo
Excluding NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.7%	99.6%	99.5%	99.5%	99.5%
Including NBD	99.5%	99.3%	98.9%	98.0%	95.2%	87.7%	69.6%	37.8%	20.7%	17.4%
Effective Sample Size	44696	41812	39249	36416	32590	26312	16979	6468	2058	404

US Market Release	Sep-08	Total Malfunctions	52
CE Approval Date		Therapy Function Not Compromised	35
Registered USA Implants	20,048	Battery Malfunction	14
Estimated Active USA Implants	6,353	Electrical Component	10
Normal Battery Depletions	1,766	Other Malfunction	1
		Poss Early Battery Depltn	8
		Software Malfunction	2
		Therapy Function Compromised	17
		Battery Malfunction	9
		Electrical Component	6
		Poss Early Battery Depltn	1
		Software Malfunction	1



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 122 mo
Excluding NBD	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.6%	99.6%	99.6%
Including NBD	99.7%	99.5%	99.3%	99.0%	98.6%	98.1%	96.2%	91.2%	76.4%	43.8%	36.5%
Effective Sample Size	17912	16706	15718	14644	13440	12196	10777	7948	4358	651	244

D234DRG

Secura DR

US Market Release

Total Malfunctions

CE Approval Date

Mar-08

Therapy Function Not Compromised

Registered USA Implants

4


Therapy Function Compromised

Estimated Active USA Implants

2

Normal Battery Depletions

1



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 115 mo
Excluding NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.7%	99.6%	99.5%	99.5%	99.5%
Including NBD	99.5%	99.3%	98.9%	98.0%	95.2%	87.7%	69.6%	37.8%	20.7%	17.4%
Effective Sample Size	44696	41812	39249	36416	32590	26312	16979	6468	2058	404

D234VRC

Secura VR

US Market Release

Total Malfunctions

CE Approval Date

Mar-08

Therapy Function Not Compromised


Registered USA Implants

3

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 122 mo
Excluding NBD	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.6%	99.6%	99.6%
Including NBD	99.7%	99.5%	99.3%	99.0%	98.6%	98.1%	96.2%	91.2%	76.4%	43.8%	36.5%
Effective Sample Size	17912	16706	15718	14644	13440	12196	10777	7948	4358	651	244

D264DRM

Maximo II DR

US Market Release	Jan-12	Total Malfunctions
CE Approval Date	Jul-10	Therapy Function Not Compromised
Registered USA Implants	7	
Estimated Active USA Implants		Therapy Function Compromised
Normal Battery Depletions	2	




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 120 mo
Excluding NBD	100.0%	100.0%	99.9%	99.8%	99.7%	99.7%	99.6%	99.4%	99.4%	99.4%
Including NBD	99.6%	99.4%	99.0%	98.2%	95.4%	87.2%	71.4%	48.2%	33.4%	27.7%
Effective Sample Size	17322	16154	15177	14093	12604	9997	6237	2731	1171	102

D264VRM

Maximo II VR


US Market Release	May-12	Total Malfunctions
CE Approval Date	Dec-10	Therapy Function Not Compromised
Registered USA Implants	1	
Estimated Active USA Implants	1	Therapy Function Compromised
Normal Battery Depletions		



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 123 mo
Excluding NBD	100.0%	99.9%	99.9%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%	99.6%	99.6%
Including NBD	99.8%	99.6%	99.4%	99.2%	98.8%	98.1%	95.3%	89.1%	74.6%	46.7%	37.6%
Effective Sample Size	11020	10322	9706	9023	8291	7509	6546	4887	2615	494	167


US Market Release	Aug-09	Total Malfunctions	46
CE Approval Date		Therapy Function Not Compromised	29
Registered USA Implants	22,234	Battery Malfunction	10
Estimated Active USA Implants	5,147	Electrical Component	11
Normal Battery Depletions	4,249	Poss Early Battery Depltn	7
		Software Malfunction	1
		Therapy Function Compromised	17
		Battery Malfunction	14
		Electrical Component	2
		Other Malfunction	1



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 116 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.7%	99.6%	99.6%	99.6%
Including NBD	99.7%	99.5%	99.1%	98.5%	95.9%	89.0%	70.2%	37.8%	24.7%	22.1%
Effective Sample Size	19039	17868	16815	15621	13916	11174	7165	3097	1423	115


US Market Release	Aug-09	Total Malfunctions	20
CE Approval Date		Therapy Function Not Compromised	13
Registered USA Implants	9,127	Battery Malfunction	6
Estimated Active USA Implants	3,046	Electrical Component	4
Normal Battery Depletions	674	Poss Early Battery Depltn	2
		Software Malfunction	1
		Therapy Function Compromised	7
		Battery Malfunction	6
		Electrical Component	1



- Excluding Normal Battery Depletion
- Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 120 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%	99.6%
Including NBD	99.6%	99.6%	99.4%	99.3%	98.9%	98.5%	96.9%	92.2%	80.9%	37.2%
Effective Sample Size	7630	7157	6749	6283	5797	5269	4677	3824	2360	124

US Market Release	Sep-08	Total Malfunctions	71
CE Approval Date	Mar-08	Therapy Function Not Compromised	54
Registered USA Implants	20,097	Battery Malfunction	7
Estimated Active USA Implants	4,704	Electrical Component	15
Normal Battery Depletions	3,506	Other Malfunction	2
		Poss Early Battery Depltn	30
		Therapy Function Compromised	17
		Battery Malfunction	11
		Electrical Component	5
		Poss Early Battery Depltn	1




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 120 mo
Excluding NBD	100.0%	100.0%	99.9%	99.8%	99.7%	99.7%	99.6%	99.4%	99.4%	99.4%
Including NBD	99.6%	99.4%	99.0%	98.2%	95.4%	87.2%	71.4%	48.2%	33.4%	27.7%
Effective Sample Size	17322	16154	15177	14093	12604	9997	6237	2731	1171	102

D284VRC

Maximo II VR

US Market Release	Sep-08	Total Malfunctions	31
CE Approval Date	Mar-08	Therapy Function Not Compromised	22
Registered USA Implants	13,033	Battery Malfunction	9
Estimated Active USA Implants	4,453	Electrical Component	6
Normal Battery Depletions	1,283	Poss Early Battery Depltn	4
		Software Malfunction	3
		Therapy Function Compromised	9
		Battery Malfunction	6
		Electrical Component	2
		Software Malfunction	1




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 123 mo
Excluding NBD	100.0%	99.9%	99.9%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%	99.6%	99.6%
Including NBD	99.8%	99.6%	99.4%	99.2%	98.8%	98.1%	95.3%	89.1%	74.6%	46.7%	37.6%
Effective Sample Size	11020	10322	9706	9023	8291	7509	6546	4887	2615	494	167

D294DRG

Virtuoso II DR

US Market Release		Total Malfunctions	
CE Approval Date	Aug-08	Therapy Function Not Compromised	
Registered USA Implants	3	Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions	1		



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 116 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.7%	99.6%	99.6%	99.6%
Including NBD	99.7%	99.5%	99.1%	98.5%	95.9%	89.0%	70.2%	37.8%	24.7%	22.1%
Effective Sample Size	19039	17868	16815	15621	13916	11174	7165	3097	1423	115

D294VRC

Virtuoso II VR

US Market Release

Aug-08

Total Malfunctions


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 120 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%	99.6%
Including NBD	99.6%	99.6%	99.4%	99.3%	98.9%	98.5%	96.9%	92.2%	80.9%	37.2%
Effective Sample Size	7630	7157	6749	6283	5797	5269	4677	3824	2360	124

D314DRG

Protecta XT DR

US Market Release

Mar-11

Total Malfunctions

71

CE Approval Date

Therapy Function Not Compromised

39

Registered USA Implants

34,845

Battery Malfunction

7

Estimated Active USA Implants

12,032

Electrical Component

26

Normal Battery Depletions

3,673

Electrical Interconnect

1

Other Malfunction

1

Poss Early Battery Depltn

4

Therapy Function Compromised


32

Battery Malfunction

25

Electrical Component

7




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 99 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%
Including NBD	99.8%	99.7%	99.5%	99.0%	97.3%	91.5%	78.9%	52.1%	21.7%
Effective Sample Size	54479	51185	48029	44556	40080	32665	20939	3334	271

D314DRM Protecta XT DR

US Market Release	Nov-11	Total Malfunctions	22
CE Approval Date		Therapy Function Not Compromised	14
Registered USA Implants	13,925	Battery Malfunction	1
Estimated Active USA Implants	5,906	Electrical Component	12
Normal Battery Depletions	1,259	Other Malfunction	1
		Therapy Function Compromised	8
		Battery Malfunction	7
		Electrical Component	1




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 99 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%
Including NBD	99.8%	99.7%	99.5%	99.0%	97.3%	91.5%	78.9%	52.1%	21.7%
Effective Sample Size	54479	51185	48029	44556	40080	32665	20939	3334	271

D314VRG Protecta XT VR

US Market Release	Mar-11	Total Malfunctions	25
CE Approval Date		Therapy Function Not Compromised	17
Registered USA Implants	14,226	Battery Malfunction	7
Estimated Active USA Implants	8,108	Electrical Component	9
Normal Battery Depletions	366	Other Malfunction	1
		Therapy Function Compromised	8
		Battery Malfunction	7
		Electrical Component	1




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 103 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.6%	99.3%	98.9%	97.7%	93.2%	80.5%
Effective Sample Size	25992	24355	22948	21358	19640	17843	14817	5518	509

D314VRM Protecta XT VR

US Market Release	May-12	Total Malfunctions	8
CE Approval Date		Therapy Function Not Compromised	4
Registered USA Implants	7,376	Battery Malfunction	1
Estimated Active USA Implants	5,077	Electrical Component	2
Normal Battery Depletions	76	Poss Early Battery Depltn	1
		Therapy Function Compromised	4
		Battery Malfunction	2
		Electrical Component	2




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 103 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.6%	99.3%	98.9%	97.7%	93.2%	80.5%
Effective Sample Size	25992	24355	22948	21358	19640	17843	14817	5518	509

D334DRG Protecta DR

US Market Release	Mar-11	Total Malfunctions	19
CE Approval Date		Therapy Function Not Compromised	9
Registered USA Implants	10,689	Battery Malfunction	2
Estimated Active USA Implants	3,712	Electrical Component	6
Normal Battery Depletions	1,493	Poss Early Battery Depltn	1
		Therapy Function Compromised	10
		Battery Malfunction	7
		Electrical Component	3




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 99 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%
Including NBD	99.8%	99.7%	99.5%	99.0%	97.3%	91.5%	78.9%	52.1%	21.7%
Effective Sample Size	54479	51185	48029	44556	40080	32665	20939	3334	271

D334DRM Protecta DR

US Market Release	Nov-11	Total Malfunctions	
CE Approval Date		Therapy Function Not Compromised	
Registered USA Implants	2,991	Therapy Function Compromised	
Estimated Active USA Implants	1,222		
Normal Battery Depletions	428		




• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 99 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%
Including NBD	99.8%	99.7%	99.5%	99.0%	97.3%	91.5%	78.9%	52.1%	21.7%
Effective Sample Size	54479	51185	48029	44556	40080	32665	20939	3334	271

D334VRG Protecta VR

US Market Release	Mar-11	Total Malfunctions	11
CE Approval Date		Therapy Function Not Compromised	6
Registered USA Implants	6,481	Battery Malfunction	2
Estimated Active USA Implants	3,832	Electrical Component	4
Normal Battery Depletions	174	Therapy Function Compromised	5
		Battery Malfunction	3
		Electrical Component	2




• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 103 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.6%	99.3%	98.9%	97.7%	93.2%	80.5%
Effective Sample Size	25992	24355	22948	21358	19640	17843	14817	5518	509

D334VRM

Protecta VR

US Market Release	May-12	Total Malfunctions	2
CE Approval Date		Therapy Function Not Compromised	1
Registered USA Implants	2,162	Other Malfunction	1
Estimated Active USA Implants	1,497	Therapy Function Compromised	1
Normal Battery Depletions	36	Battery Malfunction	1




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 103 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.6%	99.3%	98.9%	97.7%	93.2%	80.5%
Effective Sample Size	25992	24355	22948	21358	19640	17843	14817	5518	509

D354DRG

Protecta XT DR

US Market Release		Total Malfunctions	
CE Approval Date	Mar-10	Therapy Function Not Compromised	
Registered USA Implants	4	Therapy Function Compromised	
Estimated Active USA Implants	2		
Normal Battery Depletions			



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion


Years	1	2	3	4	5	6	7	8	at 99 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%
Including NBD	99.8%	99.7%	99.5%	99.0%	97.3%	91.5%	78.9%	52.1%	21.7%
Effective Sample Size	54479	51185	48029	44556	40080	32665	20939	3334	271

D354DRM

Protecta XT DR

US Market Release
CE Approval Date Jul-10
Registered USA Implants 1
Estimated Active USA Implants 1
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised



● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion


Years	1	2	3	4	5	6	7	8	at 99 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%
Including NBD	99.8%	99.7%	99.5%	99.0%	97.3%	91.5%	78.9%	52.1%	21.7%
Effective Sample Size	54479	51185	48029	44556	40080	32665	20939	3334	271

D354VRG

Protecta XT VR

US Market Release
CE Approval Date Mar-10
Registered USA Implants 1
Estimated Active USA Implants 1
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised



● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 103 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.6%	99.3%	98.9%	97.7%	93.2%	80.5%
Effective Sample Size	25992	24355	22948	21358	19640	17843	14817	5518	509

D354VRM

Protecta XT VR

US Market Release

Total Malfunctions

CE Approval Date

Dec-10

Therapy Function Not Compromised

Registered USA Implants


1

Therapy Function Compromised

Estimated Active USA Implants

0

Normal Battery Depletions



• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 103 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.6%	99.3%	98.9%	97.7%	93.2%	80.5%
Effective Sample Size	25992	24355	22948	21358	19640	17843	14817	5518	509

D364DRG

Protecta DR

US Market Release

Total Malfunctions

CE Approval Date

Mar-10

Therapy Function Not Compromised

Registered USA Implants


3

Therapy Function Compromised

Estimated Active USA Implants

2

Normal Battery Depletions



• Excluding Normal Battery Depletion • Including Normal Battery Depletion


Years	1	2	3	4	5	6	7	8	at 99 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%
Including NBD	99.8%	99.7%	99.5%	99.0%	97.3%	91.5%	78.9%	52.1%	21.7%
Effective Sample Size	54479	51185	48029	44556	40080	32665	20939	3334	271

D364DRM

Protecta DR

US Market Release
CE Approval Date Jul-10
Registered USA Implants 1
Estimated Active USA Implants
Normal Battery Depletions 1

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised



● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion


Years	1	2	3	4	5	6	7	8	at 99 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%
Including NBD	99.8%	99.7%	99.5%	99.0%	97.3%	91.5%	78.9%	52.1%	21.7%
Effective Sample Size	54479	51185	48029	44556	40080	32665	20939	3334	271

D364VRG

Protecta VR

US Market Release
CE Approval Date Mar-10
Registered USA Implants 1
Estimated Active USA Implants 1
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised



● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion


Years	1	2	3	4	5	6	7	8	at 103 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.6%	99.3%	98.9%	97.7%	93.2%	80.5%
Effective Sample Size	25992	24355	22948	21358	19640	17843	14817	5518	509

D364VRM

Protecta VR

US Market Release
CE Approval Date Dec-10
Registered USA Implants 4
Estimated Active USA Implants 2
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised



● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion


Years	1	2	3	4	5	6	7	8	at 103 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.6%	99.3%	98.9%	97.7%	93.2%	80.5%
Effective Sample Size	25992	24355	22948	21358	19640	17843	14817	5518	509

D384DRG

Cardia DR

US Market Release
CE Approval Date Jan-11
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised



● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 99 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%
Including NBD	99.8%	99.7%	99.5%	99.0%	97.3%	91.5%	78.9%	52.1%	21.7%
Effective Sample Size	54479	51185	48029	44556	40080	32665	20939	3334	271

D384VRG

Cardia VR

US Market Release

Total Malfunctions

CE Approval Date

Jan-11


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 103 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.6%	99.3%	98.9%	97.7%	93.2%	80.5%
Effective Sample Size	25992	24355	22948	21358	19640	17843	14817	5518	509

D394DRG

Egida DR

US Market Release

Total Malfunctions

CE Approval Date

Jan-11


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



• Excluding Normal Battery Depletion • Including Normal Battery Depletion


Years	1	2	3	4	5	6	7	8	at 99 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%
Including NBD	99.8%	99.7%	99.5%	99.0%	97.3%	91.5%	78.9%	52.1%	21.7%
Effective Sample Size	54479	51185	48029	44556	40080	32665	20939	3334	271

D394VRG

Egida VR

US Market Release
CE Approval Date Jan-11
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised



● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion


Years	1	2	3	4	5	6	7	8	at 103 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.6%	99.3%	98.9%	97.7%	93.2%	80.5%
Effective Sample Size	25992	24355	22948	21358	19640	17843	14817	5518	509

DDBB1D1

Evera XT

US Market Release Apr-13
CE Approval Date
Registered USA Implants 43,268
Estimated Active USA Implants 34,702
Normal Battery Depletions 150

Total Malfunctions 49
Therapy Function Not Compromised 26
 Battery Malfunction 15
 Electrical Component 10
 Other Malfunction 1
Therapy Function Compromised 23
 Battery Malfunction 18
 Electrical Component 2
 Electrical Interconnect 1
 Other Malfunction 2




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 88 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.4%	99.0%	98.1%	97.9%
Effective Sample Size	174130	140248	106873	74390	46724	23835	4375	114

DDBB1D4 Evera XT

US Market Release	Apr-13	Total Malfunctions	42
CE Approval Date		Therapy Function Not Compromised	26
Registered USA Implants	30,456	Battery Malfunction	17
Estimated Active USA Implants	25,635	Electrical Component	6
Normal Battery Depletions	85	Electrical Interconnect	1
		Other Malfunction	1
		Poss Early Battery Depltn	1
		Therapy Function Compromised	16
		Battery Malfunction	13
		Electrical Component	3




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 88 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.4%	99.0%	98.1%	97.9%
Effective Sample Size	174130	140248	106873	74390	46724	23835	4375	114

DDBB2D1 Evera XT

US Market Release		Total Malfunctions	
CE Approval Date	Dec-12	Therapy Function Not Compromised	
Registered USA Implants	2	Therapy Function Compromised	
Estimated Active USA Implants	1		
Normal Battery Depletions			




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 88 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.4%	99.0%	98.1%	97.9%
Effective Sample Size	174130	140248	106873	74390	46724	23835	4375	114

DDBB2D4 Evera XT

US Market Release
CE Approval Date Dec-12
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 88 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.4%	99.0%	98.1%	97.9%
Effective Sample Size	174130	140248	106873	74390	46724	23835	4375	114

DDBC3D1 Evera S

US Market Release Apr-13 **Total Malfunctions** 12
CE Approval Date Dec-12 **Therapy Function Not Compromised** 6
Registered USA Implants 8,450 Battery Malfunction 4
Estimated Active USA Implants 6,795 Electrical Component 2
Normal Battery Depletions 31 **Therapy Function Compromised** 6
 Battery Malfunction 5
 Electrical Component 1




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 88 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.4%	99.0%	98.1%	97.9%
Effective Sample Size	174130	140248	106873	74390	46724	23835	4375	114

DDBC3D4 Evera S

US Market Release	Apr-13	Total Malfunctions	8
CE Approval Date	Dec-13	Therapy Function Not Compromised	4
Registered USA Implants	6,082	Battery Malfunction	2
Estimated Active USA Implants	5,081	Electrical Component	2
Normal Battery Depletions	17	Therapy Function Compromised	4
		Battery Malfunction	3
		Poss Early Battery Depltn	1




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 88 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.4%	99.0%	98.1%	97.9%
Effective Sample Size	174130	140248	106873	74390	46724	23835	4375	114

DDMB1D1 Evera MRI XT

US Market Release	Oct-16	Total Malfunctions	11
CE Approval Date		Therapy Function Not Compromised	8
Registered USA Implants	32,110	Battery Malfunction	4
Estimated Active USA Implants	29,926	Electrical Component	2
Normal Battery Depletions	9	Electrical Interconnect	1
		Other Malfunction	1
		Therapy Function Compromised	3
		Electrical Component	3




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 88 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.4%	99.0%	98.1%	97.9%
Effective Sample Size	174130	140248	106873	74390	46724	23835	4375	114

DDMB1D4 Evera MRI XT

US Market Release	Sep-15	Total Malfunctions	32
CE Approval Date		Therapy Function Not Compromised	25
Registered USA Implants	81,539	Battery Malfunction	8
Estimated Active USA Implants	76,508	Electrical Component	14
Normal Battery Depletions	44	Electrical Interconnect	2
		Other Malfunction	1
		Therapy Function Compromised	7
		Battery Malfunction	6
		Electrical Component	1




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 88 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.4%	99.0%	98.1%	97.9%
Effective Sample Size	174130	140248	106873	74390	46724	23835	4375	114

DDMB2D1 Evera MRI XT

US Market Release		Total Malfunctions	
CE Approval Date	Sep-16	Therapy Function Not Compromised	
Registered USA Implants	1	Therapy Function Compromised	
Estimated Active USA Implants	1		
Normal Battery Depletions			




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 88 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.4%	99.0%	98.1%	97.9%
Effective Sample Size	174130	140248	106873	74390	46724	23835	4375	114

DDMB2D4 Evera MRI XT

US Market Release
CE Approval Date Mar-14
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 88 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.4%	99.0%	98.1%	97.9%
Effective Sample Size	174130	140248	106873	74390	46724	23835	4375	114

DDMC3D1 Evera MRI S

US Market Release Oct-16
CE Approval Date Sep-16
Registered USA Implants 3,058
Estimated Active USA Implants 2,862
Normal Battery Depletions 1

Total Malfunctions 1
Therapy Function Not Compromised 1
 Electrical Component 1
Therapy Function Compromised 0




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 88 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.4%	99.0%	98.1%	97.9%
Effective Sample Size	174130	140248	106873	74390	46724	23835	4375	114

DDMC3D4 Evera MRI

US Market Release	Sep-15	Total Malfunctions	3
CE Approval Date	Mar-14	Therapy Function Not Compromised	1
Registered USA Implants	5,735	Electrical Component	1
Estimated Active USA Implants	5,360	Therapy Function Compromised	2
Normal Battery Depletions	1	Battery Malfunction	1
		Electrical Component	1




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 88 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.4%	99.0%	98.1%	97.9%
Effective Sample Size	174130	140248	106873	74390	46724	23835	4375	114

DDMD3D1 Primo

US Market Release	Mar-18	Total Malfunctions	
CE Approval Date	Nov-17	Therapy Function Not Compromised	
Registered USA Implants	189	Therapy Function Compromised	
Estimated Active USA Implants	177		
Normal Battery Depletions			




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 88 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.4%	99.0%	98.1%	97.9%
Effective Sample Size	174130	140248	106873	74390	46724	23835	4375	114

DDMD3D4 Primo

US Market Release Mar-18 **Total Malfunctions**
CE Approval Date Nov-17 **Therapy Function Not Compromised**
Registered USA Implants 374
Estimated Active USA Implants 364 **Therapy Function Compromised**
Normal Battery Depletions




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 88 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.4%	99.0%	98.1%	97.9%
Effective Sample Size	174130	140248	106873	74390	46724	23835	4375	114

DDME3D1 Mirro

US Market Release Mar-18 **Total Malfunctions**
CE Approval Date Nov-17 **Therapy Function Not Compromised**
Registered USA Implants
Estimated Active USA Implants **Therapy Function Compromised**
Normal Battery Depletions




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 88 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.4%	99.0%	98.1%	97.9%
Effective Sample Size	174130	140248	106873	74390	46724	23835	4375	114

DDME3D4 Mirro

US Market Release Mar-18 **Total Malfunctions**
CE Approval Date Nov-17 **Therapy Function Not Compromised**
Registered USA Implants
Estimated Active USA Implants **Therapy Function Compromised**
Normal Battery Depletions




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 88 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.4%	99.0%	98.1%	97.9%
Effective Sample Size	174130	140248	106873	74390	46724	23835	4375	114

DVAB1D1 Visia AF

US Market Release Jan-16 **Total Malfunctions** 1
CE Approval Date **Therapy Function Not Compromised** 1
Registered USA Implants 3,057 **Battery Malfunction** 1
Estimated Active USA Implants 2,718 **Therapy Function Compromised** 0
Normal Battery Depletions 5




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	at 52 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.8%	99.7%	99.7%
Effective Sample Size	49419	33285	17502	2982	179

DVAB1D4 Visia AF

US Market Release Jan-16 **Total Malfunctions**
CE Approval Date **Therapy Function Not Compromised**
Registered USA Implants 2,064 **Therapy Function Compromised**
Estimated Active USA Implants 1,900
Normal Battery Depletions




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	at 52 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.8%	99.7%	99.7%
Effective Sample Size	49419	33285	17502	2982	179

DVAB2D1 Visia AF XT

US Market Release **Total Malfunctions**
CE Approval Date Oct-15 **Therapy Function Not Compromised**
Registered USA Implants **Therapy Function Compromised**
Estimated Active USA Implants
Normal Battery Depletions




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	at 52 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.8%	99.7%	99.7%
Effective Sample Size	49419	33285	17502	2982	179

DVAC3D1 Visia AF S

US Market Release Jan-16 **Total Malfunctions**
CE Approval Date Oct-15 **Therapy Function Not Compromised**
Registered USA Implants
Estimated Active USA Implants **Therapy Function Compromised**
Normal Battery Depletions




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	at 52 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.8%	99.7%	99.7%
Effective Sample Size	49419	33285	17502	2982	179

DVBB1D1 Evera XT

US Market Release Apr-13 **Total Malfunctions** 29
CE Approval Date **Therapy Function Not Compromised** 22
Registered USA Implants 16,122 **Battery Malfunction** 16
Estimated Active USA Implants 12,721 **Electrical Component** 6
Normal Battery Depletions 18 **Therapy Function Compromised** 7
Battery Malfunction 5
Electrical Component 2




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 81 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.7%
Including NBD	100.0%	99.9%	99.8%	99.6%	99.5%	99.2%	98.8%
Effective Sample Size	53488	50193	47288	42727	27410	10891	906

DVBB1D4 Evera XT

US Market Release	Apr-13	Total Malfunctions	53
CE Approval Date		Therapy Function Not Compromised	37
Registered USA Implants	22,381	Battery Malfunction	24
Estimated Active USA Implants	18,893	Electrical Component	7
Normal Battery Depletions	28	Other Malfunction	5
		Poss Early Battery Depltn	1
		Therapy Function Compromised	16
		Battery Malfunction	15
		Electrical Component	1




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 81 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.7%
Including NBD	100.0%	99.9%	99.8%	99.6%	99.5%	99.2%	98.8%
Effective Sample Size	53488	50193	47288	42727	27410	10891	906

DVBB2D1 Evera XT

US Market Release		Total Malfunctions	
CE Approval Date	Dec-12	Therapy Function Not Compromised	
Registered USA Implants		Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 81 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.7%
Including NBD	100.0%	99.9%	99.8%	99.6%	99.5%	99.2%	98.8%
Effective Sample Size	53488	50193	47288	42727	27410	10891	906

DVBB2D4 Evera XT

US Market Release		Total Malfunctions	
CE Approval Date	Dec-12	Therapy Function Not Compromised	
Registered USA Implants	1	Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			




• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 81 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.7%
Including NBD	100.0%	99.9%	99.8%	99.6%	99.5%	99.2%	98.8%
Effective Sample Size	53488	50193	47288	42727	27410	10891	906

DVBC3D1 Evera S

US Market Release	Apr-13	Total Malfunctions	18
CE Approval Date	Dec-12	Therapy Function Not Compromised	13
Registered USA Implants	4,645	Battery Malfunction	12
Estimated Active USA Implants	3,693	Electrical Component	1
Normal Battery Depletions	5	Therapy Function Compromised	5
		Battery Malfunction	4
		Electrical Component	1




• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 81 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.7%
Including NBD	100.0%	99.9%	99.8%	99.6%	99.5%	99.2%	98.8%
Effective Sample Size	53488	50193	47288	42727	27410	10891	906

DVBC3D4 Evera S

US Market Release	Apr-13	Total Malfunctions	10
CE Approval Date	Dec-12	Therapy Function Not Compromised	7
Registered USA Implants	5,646	Battery Malfunction	5
Estimated Active USA Implants	4,763	Electrical Component	2
Normal Battery Depletions	5	Therapy Function Compromised	3
		Battery Malfunction	3




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 81 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.7%
Including NBD	100.0%	99.9%	99.8%	99.6%	99.5%	99.2%	98.8%
Effective Sample Size	53488	50193	47288	42727	27410	10891	906

DVFB1D1 Visia MRI AF

US Market Release	Oct-16	Total Malfunctions	4
CE Approval Date		Therapy Function Not Compromised	4
Registered USA Implants	14,014	Battery Malfunction	1
Estimated Active USA Implants	13,188	Electrical Component	1
Normal Battery Depletions	6	Other Malfunction	2
		Therapy Function Compromised	0




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	at 52 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.8%	99.7%	99.7%
Effective Sample Size	49419	33285	17502	2982	179

DVFB1D4

Visia MRI AF

US Market Release	Jan-16	Total Malfunctions	20
CE Approval Date		Therapy Function Not Compromised	17
Registered USA Implants	43,005	Battery Malfunction	8
Estimated Active USA Implants	40,392	Electrical Component	7
Normal Battery Depletions	5	Other Malfunction	2
		Therapy Function Compromised	3
		Battery Malfunction	2
		Electrical Component	1




- Excluding Normal Battery Depletion
- Including Normal Battery Depletion

Years	1	2	3	4	at 52 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.8%	99.7%	99.7%
Effective Sample Size	49419	33285	17502	2982	179

DVFB2D1

Visia MRI AF XT

US Market Release		Total Malfunctions	
CE Approval Date	Sep-16	Therapy Function Not Compromised	
Registered USA Implants		Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			



- Excluding Normal Battery Depletion
- Including Normal Battery Depletion


Years	1	2	3	4	at 52 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.8%	99.7%	99.7%
Effective Sample Size	49419	33285	17502	2982	179

DVFB2D4

Visia MRI AF XT

US Market Release
CE Approval Date Oct-15
Registered USA Implants 1
Estimated Active USA Implants 1
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised



● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion


Years	1	2	3	4	at 52 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.8%	99.7%	99.7%
Effective Sample Size	49419	33285	17502	2982	179

DVFC3D1

Visia MRI AF S

US Market Release Oct-16
CE Approval Date Sep-16
Registered USA Implants 1,168
Estimated Active USA Implants 1,121
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	at 52 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.8%	99.7%	99.7%
Effective Sample Size	49419	33285	17502	2982	179

DVFC3D4

Visia MRI AF S

US Market Release	Jan-16	Total Malfunctions	1
CE Approval Date	Oct-15	Therapy Function Not Compromised	1
Registered USA Implants	2,534	Battery Malfunction	1
Estimated Active USA Implants	2,414	Therapy Function Compromised	0
Normal Battery Depletions	2		




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	at 52 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.8%	99.7%	99.7%
Effective Sample Size	49419	33285	17502	2982	179

DVMB1D4

Evera MRI XT

US Market Release	Sep-15	Total Malfunctions	10
CE Approval Date		Therapy Function Not Compromised	7
Registered USA Implants	10,625	Battery Malfunction	2
Estimated Active USA Implants	9,501	Electrical Component	3
Normal Battery Depletions	7	Other Malfunction	2
		Therapy Function Compromised	3
		Battery Malfunction	3



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion


Years	1	2	3	4	5	6	at 81 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.7%
Including NBD	100.0%	99.9%	99.8%	99.6%	99.5%	99.2%	98.8%
Effective Sample Size	53488	50193	47288	42727	27410	10891	906

DVMB2D1

Evera MRI XT

US Market Release
 CE Approval Date Sep-16
 Registered USA Implants
 Estimated Active USA Implants
 Normal Battery Depletions

Total Malfunctions
 Therapy Function Not Compromised
 Therapy Function Compromised



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion


Years	1	2	3	4	5	6	at 81 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.7%
Including NBD	100.0%	99.9%	99.8%	99.6%	99.5%	99.2%	98.8%
Effective Sample Size	53488	50193	47288	42727	27410	10891	906

DVMB2D4

Evera MRI XT

US Market Release
 CE Approval Date Mar-14
 Registered USA Implants 1
 Estimated Active USA Implants
 Normal Battery Depletions

Total Malfunctions
 Therapy Function Not Compromised
 Therapy Function Compromised




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 81 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.7%
Including NBD	100.0%	99.9%	99.8%	99.6%	99.5%	99.2%	98.8%
Effective Sample Size	53488	50193	47288	42727	27410	10891	906

DVMC3D1 Evera MRI S

US Market Release Oct-16 **Total Malfunctions**
CE Approval Date Sep-16 **Therapy Function Not Compromised**
Registered USA Implants
Estimated Active USA Implants **Therapy Function Compromised**
Normal Battery Depletions




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 81 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.7%
Including NBD	100.0%	99.9%	99.8%	99.6%	99.5%	99.2%	98.8%
Effective Sample Size	53488	50193	47288	42727	27410	10891	906

DVMC3D4 Evera MRI S

US Market Release Sep-15 **Total Malfunctions**
CE Approval Date Mar-14 **Therapy Function Not Compromised**
Registered USA Implants 3
Estimated Active USA Implants 2 **Therapy Function Compromised**
Normal Battery Depletions




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 81 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.7%
Including NBD	100.0%	99.9%	99.8%	99.6%	99.5%	99.2%	98.8%
Effective Sample Size	53488	50193	47288	42727	27410	10891	906

DVMD3D1 Primo

US Market Release Mar-18 **Total Malfunctions**
CE Approval Date Nov-17 **Therapy Function Not Compromised**
Registered USA Implants 99
Estimated Active USA Implants 96 **Therapy Function Compromised**
Normal Battery Depletions




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 81 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.7%
Including NBD	100.0%	99.9%	99.8%	99.6%	99.5%	99.2%	98.8%
Effective Sample Size	53488	50193	47288	42727	27410	10891	906

DVMD3D4 Primo

US Market Release Mar-18 **Total Malfunctions**
CE Approval Date Nov-17 **Therapy Function Not Compromised**
Registered USA Implants 160
Estimated Active USA Implants 158 **Therapy Function Compromised**
Normal Battery Depletions




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 81 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.7%
Including NBD	100.0%	99.9%	99.8%	99.6%	99.5%	99.2%	98.8%
Effective Sample Size	53488	50193	47288	42727	27410	10891	906

DVME3D1 Mirro

US Market Release Mar-18 **Total Malfunctions**
CE Approval Date Nov-17 **Therapy Function Not Compromised**
Registered USA Implants
Estimated Active USA Implants **Therapy Function Compromised**
Normal Battery Depletions




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 81 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.7%
Including NBD	100.0%	99.9%	99.8%	99.6%	99.5%	99.2%	98.8%
Effective Sample Size	53488	50193	47288	42727	27410	10891	906

DVME3D4 Mirro

US Market Release Mar-18 **Total Malfunctions**
CE Approval Date Nov-17 **Therapy Function Not Compromised**
Registered USA Implants
Estimated Active USA Implants **Therapy Function Compromised**
Normal Battery Depletions




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 81 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.7%
Including NBD	100.0%	99.9%	99.8%	99.6%	99.5%	99.2%	98.8%
Effective Sample Size	53488	50193	47288	42727	27410	10891	906

A2DR01

Advisa DR MRI

US Market Release	Jan-13	Total Malfunctions	58
CE Approval Date		Therapy Function Not Compromised	54
Registered USA Implants	347,156	Battery Malfunction	1
Estimated Active USA Implants	313,709	Electrical Component	32
Normal Battery Depletions	373	Electrical Interconnect	2
		Other Malfunction	1
		Poss Early Battery Depltn	16
		Software Malfunction	2
		Therapy Function Compromised	4
		Electrical Component	4




- Excluding Normal Battery Depletion
- Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 79 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	99.9%	99.9%	99.7%	99.5%	98.2%
Effective Sample Size	315973	297844	254069	165781	87644	23651	783

A3DR01

Advisa DR MRI

US Market Release		Total Malfunctions	
CE Approval Date	Jun-09	Therapy Function Not Compromised	
Registered USA Implants	17	Therapy Function Compromised	
Estimated Active USA Implants	9		
Normal Battery Depletions	1		




- Excluding Normal Battery Depletion
- Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 79 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	99.9%	99.9%	99.7%	99.5%	98.2%
Effective Sample Size	315973	297844	254069	165781	87644	23651	783

A3SR01

Advisa SR MRI

US Market Release	Mar-15	Total Malfunctions	9
CE Approval Date	Apr-14	Therapy Function Not Compromised	8
Registered USA Implants	28,591	Electrical Component	3
Estimated Active USA Implants	25,284	Electrical Interconnect	1
Normal Battery Depletions	17	Other Malfunction	2
		Poss Early Battery Depltn	2
		Therapy Function Compromised	1
		Electrical Component	1




• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	at 54 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.8%	99.8%	99.7%
Effective Sample Size	23034	20155	14857	4893	313

A5DR01

Advisa DR

US Market Release		Total Malfunctions	
CE Approval Date	Jun-09	Therapy Function Not Compromised	
Registered USA Implants	1	Therapy Function Compromised	
Estimated Active USA Implants	1		
Normal Battery Depletions			




• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 79 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	99.9%	99.9%	99.7%	99.5%	98.2%
Effective Sample Size	315973	297844	254069	165781	87644	23651	783

ADD01 Adapta D

US Market Release	Jul-06	Total Malfunctions	
CE Approval Date	Sep-05	Therapy Function Not Compromised	
Registered USA Implants	1	Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 140 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.8%	99.8%	99.7%	99.6%	99.4%	98.9%	98.0%	94.3%	84.6%	67.7%	53.6%	35.6%
Effective Sample Size	400446	377853	355024	327962	298069	266684	225903	178531	120995	62839	19932	370

ADDR01 Adapta DR

US Market Release	Jul-06	Total Malfunctions	93
CE Approval Date	Sep-05	Therapy Function Not Compromised	65
Registered USA Implants	460,243	Electrical Component	56
Estimated Active USA Implants	255,448	Electrical Interconnect	1
Normal Battery Depletions	29,820	Other Malfunction	1
		Poss Early Battery Depltn	7
		Therapy Function Compromised	28
		Electrical Component	23
		Electrical Interconnect	3
		Other Malfunction	2




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 140 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.8%	99.8%	99.7%	99.6%	99.4%	98.9%	98.0%	94.3%	84.6%	67.7%	53.6%	35.6%
Effective Sample Size	400446	377853	355024	327962	298069	266684	225903	178531	120995	62839	19932	370

ADDR03 Adapta DR

US Market Release	Jul-06	Total Malfunctions	2
CE Approval Date	Sep-05	Therapy Function Not Compromised	1
Registered USA Implants	4,511	Electrical Component	1
Estimated Active USA Implants	2,291	Therapy Function Compromised	1
Normal Battery Depletions	410	Electrical Component	1




• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 140 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.8%	99.8%	99.7%	99.6%	99.4%	98.9%	98.0%	94.3%	84.6%	67.7%	53.6%	35.6%
Effective Sample Size	400446	377853	355024	327962	298069	266684	225903	178531	120995	62839	19932	370

ADDR06 Adapta DR

US Market Release	Jul-06	Total Malfunctions	1
CE Approval Date	Sep-05	Therapy Function Not Compromised	1
Registered USA Implants	3,487	Electrical Component	1
Estimated Active USA Implants	1,417	Therapy Function Compromised	0
Normal Battery Depletions	342		




• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 140 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.8%	99.8%	99.7%	99.6%	99.4%	98.9%	98.0%	94.3%	84.6%	67.7%	53.6%	35.6%
Effective Sample Size	400446	377853	355024	327962	298069	266684	225903	178531	120995	62839	19932	370

ADDRL1 Adapta L DR

US Market Release	Jul-06	Total Malfunctions	23
CE Approval Date	Sep-05	Therapy Function Not Compromised	16
Registered USA Implants	138,434	Electrical Component	13
Estimated Active USA Implants	100,084	Electrical Interconnect	1
Normal Battery Depletions	2,554	Poss Early Battery Depltn	2
		Therapy Function Compromised	7
		Electrical Component	4
		Electrical Interconnect	1
		Other Malfunction	2




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 134 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.5%	99.2%	98.9%	98.0%	94.4%	74.1%	58.4%
Effective Sample Size	120669	113970	107136	96912	84168	70172	53843	38344	24353	12565	2245	750

ADDRS1 Adapta S DR

US Market Release	Jul-06	Total Malfunctions	14
CE Approval Date	Sep-05	Therapy Function Not Compromised	8
Registered USA Implants	49,193	Electrical Component	5
Estimated Active USA Implants	23,076	Poss Early Battery Depltn	3
Normal Battery Depletions	5,047	Therapy Function Compromised	6
		Electrical Component	4
		Other Malfunction	2




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 132 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.5%	99.4%	99.2%	98.6%	97.2%	92.0%	79.3%	63.5%	50.7%	42.7%	38.7%
Effective Sample Size	41251	37994	34799	31500	27849	22858	15712	9377	4884	2070	206

ADSR01 Adapta SR

US Market Release	Jul-06	Total Malfunctions	17
CE Approval Date	Sep-05	Therapy Function Not Compromised	11
Registered USA Implants	93,242	Electrical Component	6
Estimated Active USA Implants	45,665	Electrical Interconnect	1
Normal Battery Depletions	4,241	Poss Early Battery Depltn	4
		Therapy Function Compromised	6
		Electrical Component	5
		Electrical Interconnect	1




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 114 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.8%	99.7%	99.5%	99.1%	98.5%	97.4%	94.5%	85.5%	56.6%	20.8%
Effective Sample Size	73578	64702	57261	49709	42373	33271	23784	14074	4293	415

ADSR03 Adapta SR

US Market Release	Jul-06	Total Malfunctions	
CE Approval Date	Sep-05	Therapy Function Not Compromised	
Registered USA Implants	2,098	Therapy Function Compromised	
Estimated Active USA Implants	881		
Normal Battery Depletions	139		




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 114 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.8%	99.7%	99.5%	99.1%	98.5%	97.4%	94.5%	85.5%	56.6%	20.8%
Effective Sample Size	73578	64702	57261	49709	42373	33271	23784	14074	4293	415

ADSR06 Adapta SR

US Market Release	Jul-06	Total Malfunctions	2
CE Approval Date	Sep-05	Therapy Function Not Compromised	2
Registered USA Implants	2,854	Electrical Component	2
Estimated Active USA Implants	1,109	Therapy Function Compromised	0
Normal Battery Depletions	216		




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 114 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.8%	99.7%	99.5%	99.1%	98.5%	97.4%	94.5%	85.5%	56.6%	20.8%
Effective Sample Size	73578	64702	57261	49709	42373	33271	23784	14074	4293	415

ADVDD01 Adapta VDD

US Market Release	Jul-06	Total Malfunctions	
CE Approval Date	Sep-05	Therapy Function Not Compromised	
Registered USA Implants	1,421	Therapy Function Compromised	
Estimated Active USA Implants	648		
Normal Battery Depletions	94		




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 109 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	100.0%	100.0%	100.0%	97.3%	87.0%	64.7%	50.7%	50.7%
Effective Sample Size	1223	1144	1017	921	826	703	513	248	111	100

ATDR01

Attesta DR MRI

US Market Release	Aug-17	Total Malfunctions
CE Approval Date	Jun-17	Therapy Function Not Compromised
Registered USA Implants	11	Therapy Function Compromised
Estimated Active USA Implants	10	
Normal Battery Depletions		




- Years
- Excluding NBD
- Including NBD
- Effective
- Sample Size

ATDRL1

Attesta L DR MRI

US Market Release	Aug-17	Total Malfunctions
CE Approval Date	Jun-17	Therapy Function Not Compromised
Registered USA Implants	1	Therapy Function Compromised
Estimated Active USA Implants		
Normal Battery Depletions		




- Years
- Excluding NBD
- Including NBD
- Effective
- Sample Size

ATDRS1

Attestata S DR MRI

US Market Release	Aug-17	Total Malfunctions
CE Approval Date	Jun-17	Therapy Function Not Compromised
Registered USA Implants		
Estimated Active USA Implants		Therapy Function Compromised
Normal Battery Depletions		




- Years
- Excluding NBD
- Including NBD
- Effective
- Sample Size

ATSR01

Attestata SR MRI

US Market Release	Aug-17	Total Malfunctions
CE Approval Date	Jun-17	Therapy Function Not Compromised
Registered USA Implants	2	
Estimated Active USA Implants	2	Therapy Function Compromised
Normal Battery Depletions		




- Years
- Excluding NBD
- Including NBD
- Effective
- Sample Size

EN1DR01 Ensura MRI

US Market Release
CE Approval Date Jun-10
Registered USA Implants 18
Estimated Active USA Implants 14
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	at 79 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	99.9%	99.9%	99.7%	99.5%	98.2%
Effective Sample Size	315973	297844	254069	165781	87644	23651	783

EN1SR01 Ensura SR MRI

US Market Release
CE Approval Date Apr-14
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions


Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised



● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	at 54 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.8%	99.8%	99.7%
Effective Sample Size	23034	20155	14857	4893	313

US Market Release	May-05	Total Malfunctions	15,074
CE Approval Date	Aug-04	Therapy Function Not Compromised	15,019
Registered USA Implants	109,829	Battery Malfunction	14,888
Estimated Active USA Implants	17,101	Electrical Component	59
Normal Battery Depletions	17,172	Electrical Interconnect	2
		Other Malfunction	1
		Poss Early Battery Depltn	69
		Therapy Function Compromised	55
		Battery Malfunction	6
		Electrical Component	38
		Electrical Interconnect	4
		Other Malfunction	5
		Poss Early Battery Depltn	2




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 129 mo
Excluding NBD	99.9%	99.9%	99.7%	98.0%	93.6%	86.9%	81.5%	78.0%	76.0%	75.6%	75.6%
Including NBD	99.6%	99.5%	99.0%	96.5%	90.1%	77.9%	62.4%	45.3%	29.6%	15.6%	2.7%
Effective Sample Size	94470	88156	82115	75107	65154	51147	36780	23442	13337	4914	272

RED01 Relia D

US Market Release
CE Approval Date May-08
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 140 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.8%	99.8%	99.7%	99.6%	99.4%	98.9%	98.0%	94.3%	84.6%	67.7%	53.6%	35.6%
Effective Sample Size	400446	377853	355024	327962	298069	266684	225903	178531	120995	62839	19932	370

REDR01 Relia DR

US Market Release
CE Approval Date May-08
Registered USA Implants 6
Estimated Active USA Implants 4
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 140 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.8%	99.8%	99.7%	99.6%	99.4%	98.9%	98.0%	94.3%	84.6%	67.7%	53.6%	35.6%
Effective Sample Size	400446	377853	355024	327962	298069	266684	225903	178531	120995	62839	19932	370

RES01 Relia S

US Market Release
CE Approval Date May-08
Registered USA Implants 3
Estimated Active USA Implants 2
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 114 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.8%	99.7%	99.5%	99.1%	98.5%	97.4%	94.5%	85.5%	56.6%	20.8%
Effective Sample Size	73578	64702	57261	49709	42373	33271	23784	14074	4293	415

RESR01 Relia SR

US Market Release
CE Approval Date May-08
Registered USA Implants 4
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 114 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.8%	99.7%	99.5%	99.1%	98.5%	97.4%	94.5%	85.5%	56.6%	20.8%
Effective Sample Size	73578	64702	57261	49709	42373	33271	23784	14074	4293	415

REVDD01 Relia VDD

US Market Release
CE Approval Date May-08
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




• Excluding Normal Battery Depletion
 • Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 109 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	100.0%	100.0%	100.0%	97.3%	87.0%	64.7%	50.7%	50.7%
Effective Sample Size	1223	1144	1017	921	826	703	513	248	111	100

RVDR01 Revo MRI SureScan

US Market Release Feb-11 **Total Malfunctions** 107
CE Approval Date **Therapy Function Not Compromised** 104
Registered USA Implants 69,172 Battery Malfunction 1
Estimated Active USA Implants 47,066 Electrical Component 39
Normal Battery Depletions 3,119 Electrical Interconnect 1
 Other Malfunction 1
 Poss Early Battery Depltn 59
 Software Malfunction 3
Therapy Function Compromised 3
 Electrical Component 3




• Excluding Normal Battery Depletion
 • Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 107 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.5%	98.9%	96.1%	89.3%	62.9%
Effective Sample Size	59623	56286	53460	50185	46291	42108	36673	20362	1059

SD303 Sigma 300 D

US Market Release	Aug-99	Total Malfunctions	2
CE Approval Date	Dec-98	Therapy Function Not Compromised	0
Registered USA Implants	123		
Estimated Active USA Implants	19	Therapy Function Compromised	2
Normal Battery Depletions	8	Electrical Interconnect	2




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 177 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.6%	99.5%	99.4%	99.3%	99.3%	99.2%	99.2%	99.2%
Including NBD	99.6%	99.5%	99.3%	99.0%	98.5%	97.8%	96.6%	93.9%	89.3%	81.8%	70.3%	53.5%	36.2%	21.6%	10.2%
Effective Sample Size	86857	76874	67893	59639	52212	45631	39470	34029	29165	24192	18498	11250	5524	2154	241

SDR303 Sigma 300 DR

US Market Release	Aug-99	Total Malfunctions	288
CE Approval Date	Dec-98	Therapy Function Not Compromised	62
Registered USA Implants	104,543	Electrical Component	9
Estimated Active USA Implants	10,253	Electrical Interconnect	51
Normal Battery Depletions	11,000	Other Malfunction	1
		Poss Early Battery Depltn	1
		Therapy Function Compromised	226
		Electrical Component	7
		Electrical Interconnect	218
		Other Malfunction	1




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 177 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.6%	99.5%	99.4%	99.3%	99.3%	99.2%	99.2%	99.2%
Including NBD	99.6%	99.5%	99.3%	99.0%	98.5%	97.8%	96.6%	93.9%	89.3%	81.8%	70.3%	53.5%	36.2%	21.6%	10.2%
Effective Sample Size	86857	76874	67893	59639	52212	45631	39470	34029	29165	24192	18498	11250	5524	2154	241

SED01 Sensia D

US Market Release	Jul-06	Total Malfunctions	
CE Approval Date	Sep-05	Therapy Function Not Compromised	
Registered USA Implants	9		
Estimated Active USA Implants	5	Therapy Function Compromised	
Normal Battery Depletions	1		




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 153 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.8%	99.8%	99.7%	99.5%	99.3%	98.7%	97.5%	93.0%	82.9%	70.6%	62.1%	58.8%	55.7%
Effective Sample Size	123535	114163	106118	99236	90534	80030	67770	53678	37236	22207	11116	3536	248

SEDR01 Sensia DR

US Market Release	Jul-06	Total Malfunctions	32
CE Approval Date	Sep-05	Therapy Function Not Compromised	17
Registered USA Implants	149,298	Electrical Component	15
Estimated Active USA Implants	67,591	Electrical Interconnect	1
Normal Battery Depletions	10,248	Other Malfunction	1
		Therapy Function Compromised	15
		Electrical Component	6
		Electrical Interconnect	3
		Other Malfunction	5
		Poss Early Battery Depltn	1




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 153 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.8%	99.8%	99.7%	99.5%	99.3%	98.7%	97.5%	93.0%	82.9%	70.6%	62.1%	58.8%	55.7%
Effective Sample Size	123535	114163	106118	99236	90534	80030	67770	53678	37236	22207	11116	3536	248

SEDRL1

Sensia L DR

US Market Release Jul-06 **Total Malfunctions**
CE Approval Date Sep-05 **Therapy Function Not Compromised**
Registered USA Implants 3
Estimated Active USA Implants 2 **Therapy Function Compromised**
Normal Battery Depletions




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 134 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.5%	99.2%	98.9%	98.0%	94.4%	74.1%	58.4%
Effective Sample Size	120669	113970	107136	96912	84168	70172	53843	38344	24353	12565	2245	750

SES01

Sensia S

US Market Release Jul-06 **Total Malfunctions**
CE Approval Date Sep-05 **Therapy Function Not Compromised**
Registered USA Implants 10
Estimated Active USA Implants 3 **Therapy Function Compromised**
Normal Battery Depletions




● Excluding Normal Battery Depletion
 ● Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 122 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.8%	99.7%	99.6%	99.2%	98.7%	97.9%	95.9%	89.8%	72.8%	44.8%	34.2%
Effective Sample Size	86467	76337	67811	59102	50628	41086	31239	20994	10218	1732	593

SESR01 Sensia SR

US Market Release	Jul-06	Total Malfunctions	16
CE Approval Date	Sep-05	Therapy Function Not Compromised	12
Registered USA Implants	117,244	Electrical Component	7
Estimated Active USA Implants	52,622	Other Malfunction	1
Normal Battery Depletions	5,626	Poss Early Battery Depltn	4
		Therapy Function Compromised	4
		Electrical Component	3
		Electrical Interconnect	1




• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 122 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.8%	99.7%	99.6%	99.2%	98.7%	97.9%	95.9%	89.8%	72.8%	44.8%	34.2%
Effective Sample Size	86467	76337	67811	59102	50628	41086	31239	20994	10218	1732	593

SPDR01 Sphera DR MRI

US Market Release	Aug-17	Total Malfunctions	
CE Approval Date	Jun-17	Therapy Function Not Compromised	
Registered USA Implants	1	Therapy Function Compromised	
Estimated Active USA Implants	1		
Normal Battery Depletions			




Years	_____
Excluding NBD	_____
Including NBD	_____
Effective Sample Size	_____

SPDRL1

Sphera L DR MRI

US Market Release	Aug-17	Total Malfunctions
CE Approval Date	Jun-17	Therapy Function Not Compromised
Registered USA Implants		
Estimated Active USA Implants		Therapy Function Compromised
Normal Battery Depletions		




Years _____
Excluding NBD _____
Including NBD _____
Effective _____
Sample Size _____

SPSR01

Sphera SR MRI

US Market Release	Aug-17	Total Malfunctions
CE Approval Date	Jun-17	Therapy Function Not Compromised
Registered USA Implants		
Estimated Active USA Implants		Therapy Function Compromised
Normal Battery Depletions		




Years _____
Excluding NBD _____
Including NBD _____
Effective _____
Sample Size _____

SS303

Sigma 300 S

US Market Release	Sep-99	Total Malfunctions	
CE Approval Date	Dec-98	Therapy Function Not Compromised	
Registered USA Implants	248		
Estimated Active USA Implants	46	Therapy Function Compromised	
Normal Battery Depletions			




• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 171 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	99.9%	99.8%	99.7%	99.7%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%
Including NBD	99.7%	99.5%	99.2%	98.8%	98.3%	97.4%	95.8%	93.2%	88.2%	79.7%	65.2%	47.6%	34.2%	20.5%	15.7%
Effective Sample Size	40516	33425	27628	22928	19050	15806	13106	10867	8818	6765	4593	2576	1355	360	121

SSR303

Sigma 300 SR

US Market Release	Aug-99	Total Malfunctions	58
CE Approval Date	Dec-98	Therapy Function Not Compromised	11
Registered USA Implants	51,247	Electrical Interconnect	10
Estimated Active USA Implants	3,980	Other Malfunction	1
Normal Battery Depletions	3,031	Therapy Function Compromised	47
		Electrical Component	3
		Electrical Interconnect	44




• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 171 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	99.9%	99.8%	99.7%	99.7%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%
Including NBD	99.7%	99.5%	99.2%	98.8%	98.3%	97.4%	95.8%	93.2%	88.2%	79.7%	65.2%	47.6%	34.2%	20.5%	15.7%
Effective Sample Size	40516	33425	27628	22928	19050	15806	13106	10867	8818	6765	4593	2576	1355	360	121

VEDR01 Versa DR

US Market Release	Jul-06	Total Malfunctions	24
CE Approval Date	Sep-05	Therapy Function Not Compromised	11
Registered USA Implants	118,769	Electrical Component	7
Estimated Active USA Implants	55,438	Electrical Interconnect	2
Normal Battery Depletions	9,803	Poss Early Battery Depltn	2
		Therapy Function Compromised	13
		Electrical Component	9
		Other Malfunction	4




• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 150 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.8%	99.7%	99.6%	99.5%	99.2%	98.5%	97.1%	91.7%	78.8%	63.9%	54.3%	50.7%	49.3%
Effective Sample Size	100113	93306	86234	78259	70885	63810	55186	44153	29621	16648	7679	1891	170

W1DR01 Azure XT DR

US Market Release	Aug-17	Total Malfunctions	42
CE Approval Date	Mar-17	Therapy Function Not Compromised	37
Registered USA Implants	258,514	Battery Malfunction	1
Estimated Active USA Implants	248,456	Electrical Component	18
Normal Battery Depletions	3	Other Malfunction	18
		Therapy Function Compromised	5
		Electrical Component	5




• Excluding Normal Battery Depletion • Including Normal Battery Depletion

Years	1	2	at 34 mo
Excluding NBD	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	100.0%
Effective Sample Size	158062	56044	549

W1SR01

Azure XT SR

US Market Release	Aug-17	Total Malfunctions	7
CE Approval Date	Mar-17	Therapy Function Not Compromised	7
Registered USA Implants	22,582	Battery Malfunction	1
Estimated Active USA Implants	21,030	Electrical Component	2
Normal Battery Depletions	2	Other Malfunction	4
		Therapy Function Compromised	0




Years	1	2	at 33 mo
Excluding NBD	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.9%
Effective Sample Size	14965	5088	203

W2DR01

Azure XT DR

US Market Release		Total Malfunctions	
CE Approval Date	Mar-17	Therapy Function Not Compromised	
Registered USA Implants	2	Therapy Function Compromised	
Estimated Active USA Implants	2		
Normal Battery Depletions			




Years	1	2	at 34 mo
Excluding NBD	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	100.0%
Effective Sample Size	158062	56044	549

W2SR01

Azure XT SR

US Market Release
 CE Approval Date Mar-17
 Registered USA Implants
 Estimated Active USA Implants
 Normal Battery Depletions

Total Malfunctions
 Therapy Function Not Compromised
 Therapy Function Compromised



● Excluding Normal Battery Depletion ● Including Normal Battery Depletion


Years	1	2	at 33 mo
Excluding NBD	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.9%
Effective Sample Size	14965	5088	203

W3DR01

Azure S DR

US Market Release Aug-17
 CE Approval Date Mar-17
 Registered USA Implants 30,376
 Estimated Active USA Implants 29,160
 Normal Battery Depletions 1

Total Malfunctions 3
 Therapy Function Not Compromised 2
 Electrical Component 2
 Therapy Function Compromised 1
 Electrical Component 1




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	at 34 mo
Excluding NBD	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	100.0%
Effective Sample Size	158062	56044	549

W3SR01

Azure S SR

US Market Release	Aug-17	Total Malfunctions	1
CE Approval Date	Mar-17	Therapy Function Not Compromised	1
Registered USA Implants	5,398	Electrical Component	1
Estimated Active USA Implants	5,034	Therapy Function Compromised	0
Normal Battery Depletions			




● Excluding Normal Battery Depletion ● Including Normal Battery Depletion

Years	1	2	at 33 mo
Excluding NBD	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.9%
Effective Sample Size	14965	5088	203

X2DR01

Astra XT DR MRI SureScan

US Market Release		Total Malfunctions	
CE Approval Date	Mar-17	Therapy Function Not Compromised	
Registered USA Implants	2	Therapy Function Compromised	
Estimated Active USA Implants	1		
Normal Battery Depletions			




Years	1	2	at 33 mo
Excluding NBD			
Including NBD			
Effective Sample Size			

X2SR01

Astra XT SR MRI SureScan

US Market Release
CE Approval Date Mar-17
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised




Years
Excluding NBD
Including NBD
Effective
Sample Size

X3DR01

Astra S DR

US Market Release
CE Approval Date Mar-17
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions
Therapy Function Not Compromised
Therapy Function Compromised



Years
Excluding NBD
Including NBD
Effective
Sample Size

US Market Release

Total Malfunctions

CE Approval Date

Mar-17


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



Method for Estimating Lead Performance

Medtronic Cardiac Rhythm and Heart Failure (CRHF) has tracked lead survival for over 36 years with its multicenter, global chronic lead studies.

Leads Performance Analysis

Implanted leads operate in the challenging biochemical environment of the human body and the body's response to foreign objects. Implanted leads are also subject to mechanical stresses associated with heart motion, body motion, and patient anatomy.

In this environment, pacemaker and defibrillation leads cannot be expected to last forever. While IPGs and ICDs have a battery that will deplete after a predictable length of time, a lead's longevity cannot be predicted easily based on mechanical measurements, nor are there simple indicators that a lead is approaching the end of its service life. Therefore, regular monitoring while implanted, and evaluation of lead integrity upon IPG or ICD replacement, is necessary to determine if a lead may be approaching the end of its service life.

Shortfalls Of Using Returned Product And Complaints To Estimate Lead Performance

Leads and lead segments returned to Medtronic are analyzed to determine whether or not they meet performance limits established by Medtronic. Although returned product analyses are valuable for gaining insight into lead failure mechanisms, this data cannot be used by itself for determining the survival probability of leads because only a small fraction of leads are explanted and returned for analysis. Some leads are modified due to adverse device effect, however may not be explanted. Additionally, those leads that are returned cannot be assumed to be statistically representative of the performance of the total population for a given lead model. Partial or total lead extraction can result in significant damage to a lead, making a definitive analysis of a suspected failure, and its cause, impossible.

To account for the under reporting inherent with lead survival analysis based solely on returned product, some manufacturers add reported complaints where adverse product performance is evident but the product itself has not been returned. The improvement to the accuracy of survival estimates depends on the degree to which all complaints are actually communicated to the manufacturer. Since not all complaints are communicated to the manufacturer, adding complaints to the survival analysis does not completely solve the under reporting problem.

Lead survival probabilities are more appropriately determined through a prospective clinical surveillance study that includes active follow up with the patients. Although Medtronic monitors returned product analysis and complaints, these are not used to determine lead survival estimates.

Medtronic consolidated all cardiac rhythm surveillance registries into the PAN Registry. The PAN Registry is a patient centric surveillance platform which follows patients implanted with Medtronic cardiac rhythm product(s). The Product Performance Report (PPR) tracks PAN Registry enrolled patients to monitor lead performance status in vivo. The PAN Registry is designed to record clinical observations representative of the total clinical experience. Lead survival estimates include both lead hardware failure and lead-related clinical events that are classified as product performance events, and do not differentiate a lead hardware failure from other clinical events such as Failure to capture, perforation, dislodgement, or concurrent pulse generator failure.

PAN Registry

Medtronic has been monitoring the performance of its cardiac therapy products with a multicenter study since 1983 and has evaluated the performance of more than 131,000 leads, with data reported from countries around the world. Throughout this time period, Medtronic has continually worked to adapt systems and processes to more effectively monitor product performance following market release. The following summarizes current registry requirements.

Method for Estimating Lead Performance continued

Medtronic's product surveillance registry is a world-wide study that has a prospective, non-randomized, observational design. A key purpose of the registry is to provide continuing evaluation and periodic reporting of the long-term reliability and performance of Medtronic market-released cardiac rhythm therapy products. Product-related adverse events, indicating the status of the product, are collected to measure product survival probabilities. The data gathered may also be used to support the design and development of new cardiac therapy products. The registry is designed to continue indefinitely, encompassing new products as they become commercially available.

To ensure a sufficiently large and representative source of data, participating clinical sites must meet pre-specified selection criteria. Patients are enrolled upon implantation of a Medtronic Cardiac rhythm product. Every effort is made to ensure participants are representative of the range of clinical environments in which Medtronic cardiac rhythm products are used. Eligible products for enrollment include Medtronic market-released cardiac rhythm therapy products for which additional information to further characterize product performance following market release is desired. Number of enrollments is reviewed regularly to ensure adequate sample size is obtained for each individual product. Enrollment may be capped and follow-up discontinued when sufficient duration and precision is achieved to effectively characterize product survivability.

Enrolled patients are followed in accordance with the standard care practices of their care provider from their implant date until they can no longer be followed (e.g., death, lost to follow-up, etc.). However, to ensure regular patient status assessments are completed, follow-up windows consistent with typical care practices have been established with a minimum annual follow-up requirement. Product-related adverse events, system modifications and changes in patient status (e.g. death and withdrawal from the study) are required to be reported upon occurrence. This active surveillance model ensures a robust dataset for effectively monitoring product performance.

Patients are eligible for enrollment if:

- Patient is intended to be implanted or is within 30 days post-implant of a Medtronic market-released cardiac lead connected to a market-released CRT, ICD, or IPG device, and the lead is used for a pacing, sensing, or defibrillation application, or
- Patient participated in a qualifying investigational study of a Medtronic cardiac rhythm product that is now market-released; complete implant and follow-up data are available; and the data can be appropriately and legally released

Each site is required to inform Medtronic whenever a lead event has occurred, a lead is modified, or when a patient is no longer participating. Timely, accurate, and complete reporting and analysis of safety information for surveillance is crucial for the protection of patients, clinicians, and the sponsor Medtronic continually evaluates the quality and integrity of the data through a combination of on-site and centralized monitoring activities.

Lead Complications

Chronic lead performance is characterized by estimating lead related complication free survival probabilities. For analysis purposes, the complication criteria, which align with the AdvaMed 'Industry Guidance for Uniform Reporting of Clinical Performance of Cardiac Rhythm Management Pulse Generators and Leads', are defined below. These criteria do not, however, enable a lead integrity or "hardware" failure to be conclusively differentiated from other clinical events such as an undetected lead dislodgement, perforation, or concurrent pulse generator failure manifested as a sensing or capture problem.

Method for Estimating Lead Performance continued

All reported lead-related adverse events are classified by the reporting investigator and are adjudicated by an independent event adjudication committee¹. A lead-related event with at least one of the following classifications that is adjudicated by the committee as a complication and occurs more than 30 days after implant is considered a product performance event and will contribute to the survival analysis endpoint. Events with an onset date of 30 days or less after the implant are considered procedure related and therefore are not included as product performance events. Product performance events include, but are not limited to:

- Failure to capture
- Failure to sense/undersensing
- Oversensing
- Elevated pacing thresholds
- Abnormal pacing impedance (based on lead model, but normal range is typically 200 - 2,000 ohms)
- Abnormal defibrillation impedance (based on lead model, but normal range is typically 20 - 200 ohms)
- Lead Insulation breach
- Lead Conductor fracture, confirmed electrically, visually or radiographically
- Extracardiac stimulation
- Cardiac perforation
- Lead dislodgement
- Structural Lead Failure

Data Analysis Methods

The performance of leads is expressed in terms of lead survival estimates, where "survival" refers to the function of the lead, not the survival of the patient.. These survival estimates are intended to illustrate the probability that a lead will survive for a given number of years without a chronic lead-related complication.

Active surveillance normally begins at the time of implant and continues until a product performance or censoring event occurs. In some cases in the PAN Registry, active surveillance of a device starts after the device was implanted. The survival probability of such device is conditional on survival to the time when the device enters the Registry. This phenomenon is called Left-truncation². PPR lead survival analysis is estimated using the Kaplan-Meier method, a statistical method to incorporate data from these retrospectively enrolled devices, left-truncated data, was applied. The statistical technique uses data from existing devices while appropriately adjusting the device survival curves for the time the device was not actively followed in the registry. Thus, in some cases sample sizes may fluctuate from one time interval to the next interval.

On the following pages, each graph includes a survival curve for each lead model. The survival estimates is the probability that a lead is free of a product performance event at a given time point. For example, if a survival probability is 95% after 5 years of service, then the lead has a 5% chance of experiencing a lead-related complication in the first 5 years following implant.

The data in the tables is rounded to the nearest tenth of one percent. Occasionally, a graph may show 100% survival, but have one or more complications. This occurs because even with the complications, the data rounds to 100%.

The survival curves are statistical estimates. As sample size increases and performance experience accumulates, the estimation improves. Confidence intervals are provided as a way to indicate the degree of certainty of the estimates. Greenwood's formula is used to calculate the standard errors, and the log-log method is used to produce the 2-sided 95% confidence bounds.

Since the survival estimate can become very imprecise with small effective sample sizes, Medtronic truncates the survival curve when the number of leads entering an interval is less than 50 leads. When the number of leads entering an interval reaches 50, the next data point is added to the survival

Method for Estimating Lead Performance continued

curve. For those lead models that do not have sufficient sample size, a survival curve will not be presented.

Definition of Analysis Dataset

The survival estimates are derived from all device components successfully enrolled as of the data received cut-off date (e.g. date of data entry at a study site). The number of enrollments is listed for each lead model.

This sample is considered to be representative of the worldwide population, and therefore the survival estimates shown should be representative of the performance worldwide of these models.

Criteria for Model Inclusion

Performance information for a model or model family will be published when more than 100 leads have been enrolled and no fewer than 50 leads followed for at least 6 months. Medtronic, at its discretion, may stop providing updated performance information on lead models that received original US market-release approval 20 or more years ago.

Returned Product Analysis Results

Although the returned product analysis data is not used to generate the survival estimates, the data provides valuable insight into the causes of lead malfunction.

For reporting returned product analysis results, Medtronic CRHF considers a lead as having malfunctioned whenever the analysis shows that any parameter was outside the performance limits established by Medtronic while implanted and in service. To be considered a malfunction for returned product analysis reporting, the lead must have been returned to Medtronic and analyzed.

The results of the analysis is presented in four categories. The lead reporting categories are:

Conductor Fracture: Conductor malfunction with complete or intermittent loss of continuity that could interrupt current flow (e.g., fractured conductors), including those associated with clavicle flex fatigue or crush damage.

Insulation Breach: A malfunction of the insulation allowing inappropriate entry of body fluids or inappropriate current flow between the conductors, or between the conductor and the body. Examples include cuts, tears, depressions, abrasions, and material degradation.

Crimps/Welds/Bonds: Any malfunction in a conductor or lead body associated with a point of connection.

Other: Malfunctions of specific lead mechanical attributes, such as sensors, connectors, seal rings, or malfunction modes not included in the three categories above.

A lead subject to a safety advisory is not considered to have malfunctioned unless it has been returned to Medtronic CRHF and found, through analysis, to actually have performed outside the performance limits established by Medtronic.

For leads designed for either ventricular or atrial use, the numbers listed in the Returned Product Analysis tables include both.

The numbers of malfunctions listed in the Returned Product Analysis tables are the actual numbers confirmed in the returned product analysis. The numbers of complications listed in the complications tables are the actual numbers observed in the PSR centers around the world.

Method for Estimating Lead Performance continued

US Reports of Acute Lead Observations (Occurring within First Month of Service)

In the first weeks following lead implantation, physiologic responses and lead performance can vary until long-term lead stability is attained. Acute (defined as the first month after implant) lead performance may be subject to a number of factors, including patient-specific anatomy, clinical conditions and/or varying implant conditions/techniques. After a period of time, the implant and the lead performance stabilizes. It is for this reason that the Product Surveillance Registry results, which are intended to measure long-term performance, do not include complications that occur within the first 30 days after implant.

Information about the clinical experience in the first month of service is included in our reporting. The source for this information is Medtronic's complaint handling system database. The information is summarized in tables titled "US Reports of Acute Lead Observations."

Each Event Report received by Medtronic's complaint handling system is assigned one or more Reason for Report codes based on the information received. The Reason for Report codes have been grouped into Acute Lead Observation categories. The categories used for this product performance reporting are drawn from the "FDA Guidance for Submission of Research and Marketing Applications for Permanent Pacemaker Leads and for Pacemaker Lead Adapter 510(k) Submissions." The categories are:

1. Cardiac Perforation
2. Conductor Fracture
3. Lead Dislodgement
4. Failure to Capture
5. Oversensing
6. Failure to Sense
7. Insulation Breach
8. Impedance Abnormal
9. Extracardiac Stimulation
10. Unspecified

Although multiple observations are possible for any given lead, only one observation is reported per lead. The observation reported is the observation highest on the list. For example, if an Event Report includes observations for both Lead Dislodgement and Failure to Sense, Lead Dislodgement is reported.

The lead event reported to Medtronic may or may not have involved clinical action or product returned to Medtronic. The lead may have remained implanted and in service.

Estimated Number of Implanted and Active Leads in the United States

In addition to providing the number of leads enrolled in the PSR, we also provide the number of leads registered as implanted and the number remaining active in the United States based on the status recorded in the Medtronic Device and Registrant Tracking system.

Footnotes:

1: During the evolution of SLS, event adjudication was transitioned from a Medtronic technical review committee to an independent event adjudication committee in 2011. Data analyses include adjudication using both methods.

2: Klein, John P., Moeschberger, Melvin L. Survival Analysis Techniques for Censored and Truncated Data, New York: Springer-Verlag New York, Inc., 1997.

US Market Release	03Aug2005
CE Approval	31Jan2003
Registered USA Implants	73,220
Estimated Active USA Implants	59,861
Fixation Type	Fixed Screw
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	27
Crimp Weld Bond	
Insulation Breach	45
Other	8

US Acute Lead Observations

Cardiac Perforation	26
Conductor Fracture	2
Extracardiac Stimulation	6
Failure To Capture	247
Failure To Sense	20
Impedance Abnormal	2
Insulation Breach	1
Lead Dislodgement	288
Oversensing	56
Unspecified	2

Atrial Placement

Product Surveillance Registry Results


Number of Leads Enrolled in Study	1,266
Cumulative Months of Followup	63,202
Number of Leads Active in Study	551

Qualifying Complications

Cardiac Perforation	1
Conductor Fracture	3
Extracardiac Stimulation	1
Failure To Capture	4
Failure To Sense	3

19

Impedance Abnormal	2
Insulation Breach	1
Lead Dislodgement	4



Years	1	2	3	4	5	6	7	8	9	10	at 126 mo
%	99.4%	99.2%	99.2%	99.0%	98.8%	98.6%	98.0%	97.4%	97.0%	95.2%	95.2%
#	961	781	673	540	458	386	324	279	178	74	59

His Bundle Placement

Product Surveillance Registry Results


Number of Leads Enrolled in Study	841
Cumulative Months of Followup	12,014
Number of Leads Active in Study	695

Qualifying Complications

Failure To Capture	14
Failure To Sense	1

18

Lead Dislodgement	2
Oversensing	1



Years	1	2	3	at 48 mo
%	98.2%	96.6%	96.6%	94.4%
#	366	169	98	61


Ventricular Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,118
Cumulative Months of Followup	41,812
Number of Leads Active in Study	607

Qualifying Complications

Failure To Capture	7	Impedance Abnormal	1
		Lead Dislodgement	5
		Other Complication	1




Years	1	2	3	4	5	6	7	8	9	at 114 mo
%	99.2%	98.7%	98.5%	98.3%	98.3%	97.5%	97.5%	96.9%	96.9%	96.9%
#	684	548	461	346	288	231	183	151	98	69

4073 CapSure Sense

US Market Release	23Jun2002
CE Approval	01Feb2002
Registered USA Implants	771
Estimated Active USA Implants	235
Fixation Type	Tines
Pace Sense Polarity	Unipolar
Steroid Indicator	Yes

US Returned Product Analysis

US Acute Lead Observations



Years	at mo
%	
#	

US Market Release	23Jun2002
CE Approval	01Feb2002
Registered USA Implants	139,439
Estimated Active USA Implants	83,168
Fixation Type	Tines
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	12
Crimp Weld Bond	
Insulation Breach	46
Other	

US Acute Lead Observations

Cardiac Perforation	27
Conductor Fracture	2
Extracardiac Stimulation	3
Failure To Capture	126
Failure To Sense	8
Impedance Abnormal	4
Insulation Breach	
Lead Dislodgement	164
Oversensing	7
Unspecified	


Atrial Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	227
Cumulative Months of Followup	26,032
Number of Leads Active in Study	94

Qualifying Complications

Failure To Sense	1	Lead Dislodgement	1
------------------	---	-------------------	---



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 174 mo
%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%
#	214	205	198	183	167	158	148	136	127	117	106	103	86	59	51


Ventricular Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,177
Cumulative Months of Followup	70,949
Number of Leads Active in Study	288

Qualifying Complications

Conductor Fracture	1	Impedance Abnormal	2
Failure To Capture	3	Insulation Breach	2
		Lead Dislodgement	2
		Other Complication	1



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 174 mo
%	99.4%	99.3%	99.2%	99.2%	99.2%	99.2%	99.2%	99.2%	99.2%	98.7%	98.7%	97.9%	97.9%	96.8%	96.8%
#	1,010	838	697	592	461	371	296	243	198	162	150	138	102	64	52

US Market Release	25Feb2004
CE Approval	14Jun2004
Registered USA Implants	678,898
Estimated Active USA Implants	464,630
Fixation Type	Active Screw In
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	106
Crimp Weld Bond	1
Insulation Breach	162
Other	20

US Acute Lead Observations

Cardiac Perforation	164
Conductor Fracture	10
Extracardiac Stimulation	23
Failure To Capture	220
Failure To Sense	91
Impedance Abnormal	37
Insulation Breach	1
Lead Dislodgement	583
Oversensing	76
Unspecified	10

Atrial Placement


Product Surveillance Registry Results

Number of Leads Enrolled in Study	3,958
Cumulative Months of Followup	215,523
Number of Leads Active in Study	1,538

Qualifying Complications

29

Cardiac Perforation	1	Insulation Breach	2
Conductor Fracture	2	Lead Dislodgement	8
Failure To Capture	10	Oversensing	1
Failure To Sense	3	Other Complication	2



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	at 162 mo
%	99.8%	99.7%	99.6%	99.5%	99.3%	98.9%	98.7%	98.6%	98.6%	98.6%	98.3%	98.3%	98.3%	98.3%
#	2,934	2,531	2,243	1,967	1,667	1,399	1,082	841	661	467	283	167	92	66

Ventricular Placement


Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,684
Cumulative Months of Followup	101,093
Number of Leads Active in Study	444

Qualifying Complications

11

Conductor Fracture	1	Impedance Abnormal	2
Extracardiac Stimulation	1	Lead Dislodgement	1
Failure To Capture	5	Other Complication	1



Years	1	2	3	4	5	6	7	8	9	10	11	12	at 156 mo
%	99.7%	99.7%	99.7%	99.6%	99.4%	99.4%	99.0%	99.0%	98.8%	98.8%	98.8%	98.8%	97.6%
#	1,359	1,174	1,042	888	723	623	484	406	340	254	169	116	66

US Market Release	17Sep1998
CE Approval	15Apr1998
Registered USA Implants	185,538
Estimated Active USA Implants	60,015
Fixation Type	Tines
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	19
Crimp Weld Bond	
Insulation Breach	91
Other	

US Acute Lead Observations


Cardiac Perforation	4
Conductor Fracture	4
Extracardiac Stimulation	1
Failure To Capture	35
Failure To Sense	1
Impedance Abnormal	2
Insulation Breach	1
Lead Dislodgement	35
Oversensing	1
Unspecified	1

Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,200
Cumulative Months of Followup	69,410
Number of Leads Active in Study	31

Qualifying Complications

Conductor Fracture	3	Impedance Abnormal	1
Extracardiac Stimulation	1	Lead Dislodgement	4
Failure To Capture	12		



Years	1	2	3	4	5	6	7	8	9	10	11	at 144 mo
%	98.8%	98.7%	98.5%	98.1%	97.8%	97.4%	97.4%	97.4%	97.4%	97.4%	97.4%	97.4%
#	921	822	734	629	515	402	331	277	237	153	89	52

US Market Release	23Jun2002
CE Approval	01Feb2002
Registered USA Implants	98,359
Estimated Active USA Implants	63,677
Fixation Type	J-shape, tines
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	11
Crimp Weld Bond	
Insulation Breach	17
Other	

US Acute Lead Observations


Cardiac Perforation	1
Conductor Fracture	1
Extracardiac Stimulation	1
Failure To Capture	84
Failure To Sense	37
Impedance Abnormal	4
Insulation Breach	
Lead Dislodgement	189
Oversensing	6
Unspecified	4

Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,276
Cumulative Months of Followup	53,137
Number of Leads Active in Study	612

Qualifying Complications

Conductor Fracture	2	Lead Dislodgement	7
Failure To Capture	2		



Years	1	2	3	4	5	6	7	8	9	at 114 mo
%	99.2%	99.2%	99.2%	99.0%	98.7%	98.7%	98.7%	98.7%	98.7%	98.7%
#	962	748	604	467	345	251	188	116	64	53

US Market Release	05Oct1998
CE Approval	15Apr1998
Registered USA Implants	88,230
Estimated Active USA Implants	30,299
Fixation Type	J-shape, tines
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	10
Crimp Weld Bond	
Insulation Breach	30
Other	

US Acute Lead Observations


Cardiac Perforation	
Conductor Fracture	
Extracardiac Stimulation	
Failure To Capture	10
Failure To Sense	2
Impedance Abnormal	
Insulation Breach	1
Lead Dislodgement	37
Oversensing	2
Unspecified	2

Product Surveillance Registry Results

Number of Leads Enrolled in Study	358
Cumulative Months of Followup	20,603
Number of Leads Active in Study	37

Qualifying Complications

Failure To Capture	5	Lead Dislodgement	2
Failure To Sense	1	Other Complication	1



Years	1	2	3	4	5	6	7	8	9	10	11	12	at 150 mo
%	97.7%	97.7%	97.7%	97.7%	97.0%	97.0%	97.0%	96.0%	96.0%	96.0%	96.0%	94.5%	94.5%
#	201	178	162	152	129	121	104	100	88	75	70	65	54

US Market Release	03Jun1998
CE Approval	05Jun1997
Registered USA Implants	98,908
Estimated Active USA Implants	30,334
Fixation Type	Tines
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	15
Crimp Weld Bond	1
Insulation Breach	43
Other	

US Acute Lead Observations

Cardiac Perforation	2
Conductor Fracture	2
Extracardiac Stimulation	
Failure To Capture	23
Failure To Sense	
Impedance Abnormal	4
Insulation Breach	1
Lead Dislodgement	30
Oversensing	
Unspecified	9


Atrial Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	426
Cumulative Months of Followup	40,703
Number of Leads Active in Study	44

Qualifying Complications

Failure To Capture	1	Lead Dislodgement	1
--------------------	---	-------------------	---



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 180 mo
%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%
#	411	391	358	322	289	252	219	186	153	129	108	93	75	64	54


Ventricular Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	988
Cumulative Months of Followup	34,468
Number of Leads Active in Study	26

Qualifying Complications

Failure To Capture	7	Impedance Abnormal	1
Failure To Sense	2	Lead Dislodgement	1



Years	1	2	3	4	5	6	7	8	9	10	11	at 144 mo
%	99.3%	99.1%	98.8%	98.4%	98.4%	97.0%	97.0%	96.3%	96.3%	96.3%	96.3%	96.3%
#	473	390	302	262	228	190	166	140	108	89	69	51

US Market Release	31Aug2000
CE Approval	12Aug1999
Registered USA Implants	2,777,413
Estimated Active USA Implants	1,841,973
Fixation Type	Active Screw In
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	1,146
Crimp Weld Bond	
Insulation Breach	1,223
Other	185

US Acute Lead Observations

Cardiac Perforation	1,212
Conductor Fracture	25
Extracardiac Stimulation	92
Failure To Capture	1,513
Failure To Sense	632
Impedance Abnormal	192
Insulation Breach	11
Lead Dislodgement	3,752
Oversensing	473
Unspecified	26

Atrial Placement


Product Surveillance Registry Results

Number of Leads Enrolled in Study	9,874
Cumulative Months of Followup	446,886
Number of Leads Active in Study	4,321

Qualifying Complications

86

Cardiac Perforation	2	Impedance Abnormal	7
Conductor Fracture	11	Insulation Breach	3
Extracardiac Stimulation	3	Lead Dislodgement	31
Failure To Capture	14	Oversensing	4
Failure To Sense	6	Other Complication	5



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	at 204 mo
%	99.6%	99.5%	99.4%	99.1%	98.8%	98.6%	98.4%	98.2%	98.2%	98.1%	98.0%	97.7%	97.6%	97.3%	97.3%	97.3%	97.3%
#	6,521	5,407	4,548	3,605	2,869	2,366	1,872	1,489	1,239	999	805	644	489	330	206	132	69

Ventricular Placement


Product Surveillance Registry Results

Number of Leads Enrolled in Study	3,108
Cumulative Months of Followup	129,553
Number of Leads Active in Study	934

Qualifying Complications

31

Cardiac Perforation	1	Impedance Abnormal	4
Conductor Fracture	6	Lead Dislodgement	5
Failure To Capture	12	Oversensing	1
Failure To Sense	1	Other Complication	1



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	at 192 mo
%	99.5%	99.3%	99.2%	99.0%	98.8%	98.4%	98.3%	97.8%	97.3%	97.3%	97.0%	97.0%	96.4%	95.6%	95.6%	95.6%
#	1,929	1,490	1,205	976	836	703	526	421	362	305	251	201	150	115	79	51

5086MRI CapsureFix Novus MRI

US Market Release	08Feb2011
CE Approval	21Jan2009
Registered USA Implants	208,634
Estimated Active USA Implants	172,659
Fixation Type	Active Screw In
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	88
Crimp Weld Bond	
Insulation Breach	151
Other	11

US Acute Lead Observations

Cardiac Perforation	213
Conductor Fracture	2
Extracardiac Stimulation	18
Failure To Capture	142
Failure To Sense	28
Impedance Abnormal	9
Insulation Breach	1
Lead Dislodgement	309
Oversensing	31
Unspecified	


Atrial Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	3,117
Cumulative Months of Followup	132,521
Number of Leads Active in Study	1,450

Qualifying Complications

Conductor Fracture	3	Lead Dislodgement	11
Failure To Capture	3	Oversensing	2
		Other Complication	1



Years	1	2	3	4	5	6	7	at 96 mo
%	99.8%	99.6%	99.6%	99.4%	99.3%	98.9%	98.3%	98.3%
#	2,535	2,208	1,883	1,465	758	418	268	124


Ventricular Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	3,057
Cumulative Months of Followup	130,875
Number of Leads Active in Study	1,432

Qualifying Complications

Conductor Fracture	2	Impedance Abnormal	2
Failure To Capture	9	Lead Dislodgement	3
Failure To Sense	1	Oversensing	2
		Other Complication	1



Years	1	2	3	4	5	6	7	at 96 mo
%	99.7%	99.7%	99.6%	99.6%	99.6%	99.1%	97.9%	97.3%
#	2,532	2,188	1,856	1,434	733	394	263	116

US Market Release	03Jun1998
CE Approval	25Sep1997
Registered USA Implants	140,174
Estimated Active USA Implants	47,796
Fixation Type	Tines
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	24
Crimp Weld Bond	
Insulation Breach	65
Other	1

US Acute Lead Observations


Cardiac Perforation	7
Conductor Fracture	2
Extracardiac Stimulation	3
Failure To Capture	49
Failure To Sense	7
Impedance Abnormal	1
Insulation Breach	3
Lead Dislodgement	72
Oversensing	1
Unspecified	8

Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,213
Cumulative Months of Followup	53,615
Number of Leads Active in Study	30

Qualifying Complications

Extracardiac Stimulation	1	Impedance Abnormal	1
Failure To Capture	3	Lead Dislodgement	5



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	at 162 mo
%	99.5%	99.3%	99.2%	98.9%	98.9%	98.6%	98.6%	98.6%	98.6%	98.6%	97.8%	97.8%	97.8%	97.8%
#	814	652	517	420	333	262	213	170	145	129	106	80	55	51

US Market Release	03Jun1998
CE Approval	05Jun1997
Registered USA Implants	64,448
Estimated Active USA Implants	21,895
Fixation Type	Tines
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	21
Crimp Weld Bond	
Insulation Breach	39
Other	

US Acute Lead Observations


Cardiac Perforation	
Conductor Fracture	1
Extracardiac Stimulation	
Failure To Capture	31
Failure To Sense	2
Impedance Abnormal	1
Insulation Breach	
Lead Dislodgement	38
Oversensing	
Unspecified	3

Product Surveillance Registry Results

Number of Leads Enrolled in Study	364
Cumulative Months of Followup	9,143
Number of Leads Active in Study	9

Qualifying Complications

Failure To Capture	2	5
Impedance Abnormal		1
Lead Dislodgement		1
Oversensing		1



Years	1	2	3	4	5	6	at 84 mo
%	100.0%	98.2%	97.2%	96.0%	96.0%	96.0%	96.0%
#	141	107	84	77	63	55	55

US Market Release	03Jun1998
CE Approval	25Sep1997
Registered USA Implants	36,938
Estimated Active USA Implants	15,431
Fixation Type	Tines
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	6
Crimp Weld Bond	
Insulation Breach	7
Other	

US Acute Lead Observations


Cardiac Perforation	1
Conductor Fracture	
Extracardiac Stimulation	
Failure To Capture	4
Failure To Sense	3
Impedance Abnormal	
Insulation Breach	
Lead Dislodgement	43
Oversensing	1
Unspecified	1

Product Surveillance Registry Results

Number of Leads Enrolled in Study	718
Cumulative Months of Followup	37,927
Number of Leads Active in Study	40

Qualifying Complications

Failure To Capture	3	Lead Dislodgement	2
--------------------	---	-------------------	---



Years	1	2	3	4	5	6	7	8	9	10	11	12	at 150 mo
%	99.6%	99.3%	99.3%	98.9%	98.9%	98.9%	98.9%	98.9%	98.9%	98.9%	98.9%	98.9%	98.9%
#	523	432	351	298	248	194	157	141	118	106	94	66	56

US Market Release	25Jun2001
CE Approval	23Mar2001
Registered USA Implants	17,590
Estimated Active USA Implants	8,607
Fixation Type	Tines
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	15
Crimp Weld Bond	
Insulation Breach	17
Other	

US Acute Lead Observations


Cardiac Perforation	
Conductor Fracture	
Extracardiac Stimulation	
Failure To Capture	4
Failure To Sense	
Impedance Abnormal	
Insulation Breach	
Lead Dislodgement	14
Oversensing	
Unspecified	2

Product Surveillance Registry Results

Number of Leads Enrolled in Study	39
Cumulative Months of Followup	3,628
Number of Leads Active in Study	12

Qualifying Complications

3	
Conductor Fracture	1
Insulation Breach	1
Oversensing	1



Years	at 0 mo
%	100.0%
#	0

6721 Epicardial Patch

US Market Release	31Mar1994
CE Approval	01Jan1993
Registered USA Implants	3,260
Estimated Active USA Implants	1,107
Fixation Type	Suture
Pace Sense Polarity	n/a
Steroid Indicator	None

US Returned Product Analysis

Conductor Fracture	14
Crimp Weld Bond	
Insulation Breach	1
Other	

US Acute Lead Observations


Cardiac Perforation	1
Conductor Fracture	2
Extracardiac Stimulation	
Failure To Capture	3
Failure To Sense	1
Impedance Abnormal	17
Insulation Breach	
Lead Dislodgement	
Oversensing	1
Unspecified	

Product Surveillance Registry Results

Number of Leads Enrolled in Study	417
Cumulative Months of Followup	23,972
Number of Leads Active in Study	7

Qualifying Complications

Conductor Fracture	21	Impedance Abnormal	4
Failure To Capture	8	Insulation Breach	2
		Oversensing	12



Years	1	2	3	4	5	6	7	8	at 108 mo
%	96.6%	94.6%	92.1%	91.0%	89.2%	84.6%	83.2%	83.2%	83.2%
#	348	319	272	219	186	133	99	64	56

US Market Release	02Sep2004
CE Approval	
Registered USA Implants	350
Estimated Active USA Implants	107
Fixation Type	Tines
Pace Sense Polarity	True Bipolar/One Coil
Steroid Indicator	Yes

US Returned Product Analysis


Conductor Fracture	5
Crimp Weld Bond	
Insulation Breach	
Other	

US Acute Lead Observations

Cardiac Perforation	
Conductor Fracture	
Extracardiac Stimulation	
Failure To Capture	
Failure To Sense	
Impedance Abnormal	
Insulation Breach	
Lead Dislodgement	
Oversensing	
Unspecified	1

Product Surveillance Registry Results

Number of Leads Enrolled in Study	4
Cumulative Months of Followup	293
Number of Leads Active in Study	1



Years	at 0 mo
%	100.0%
#	0

US Market Release	02Sep2004
CE Approval	
Registered USA Implants	8,060
Estimated Active USA Implants	1,876
Fixation Type	Active Screw In
Pace Sense Polarity	True Bipolar/One Coil
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	652
Crimp Weld Bond	
Insulation Breach	1
Other	5

US Acute Lead Observations


Cardiac Perforation	1
Conductor Fracture	2
Extracardiac Stimulation	
Failure To Capture	1
Failure To Sense	1
Impedance Abnormal	
Insulation Breach	
Lead Dislodgement	1
Oversensing	3
Unspecified	1

Product Surveillance Registry Results

Number of Leads Enrolled in Study	311
Cumulative Months of Followup	17,815
Number of Leads Active in Study	14

Qualifying Complications

Conductor Fracture	35	Impedance Abnormal	10
Failure To Capture	3	Lead Dislodgement	2
Failure To Sense	1	Oversensing	7



Years	1	2	3	4	5	6	7	at 90 mo
%	98.2%	96.2%	93.1%	88.3%	82.2%	74.3%	72.3%	72.3%
#	261	232	204	167	138	104	70	56

US Market Release	01Nov2008
CE Approval	31Mar2008
Registered USA Implants	62,149
Estimated Active USA Implants	46,413
Fixation Type	Active Screw In
Pace Sense Polarity	True Bipolar/One Coil
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	371
Crimp Weld Bond	
Insulation Breach	12
Other	41

US Acute Lead Observations


Cardiac Perforation	26
Conductor Fracture	4
Extracardiac Stimulation	1
Failure To Capture	27
Failure To Sense	12
Impedance Abnormal	26
Insulation Breach	1
Lead Dislodgement	60
Oversensing	60
Unspecified	5

Product Surveillance Registry Results

Number of Leads Enrolled in Study	2,760
Cumulative Months of Followup	135,868
Number of Leads Active in Study	864

Qualifying Complications

Cardiac Perforation	1	Impedance Abnormal	7
Conductor Fracture	20	Lead Dislodgement	7
Extracardiac Stimulation	1	Oversensing	7
Failure To Capture	6	Unspecified	1
Failure To Sense	1	Other Complication	2



Years	1	2	3	4	5	6	7	8	9	at 120 mo
%	99.5%	99.2%	98.9%	98.6%	98.4%	98.0%	97.3%	96.6%	95.0%	95.0%
#	2,274	1,865	1,516	1,219	1,029	874	652	383	220	78

US Market Release	02Aug2012
CE Approval	12Jul2012
Registered USA Implants	254,918
Estimated Active USA Implants	235,866
Fixation Type	Active Screw In
Pace Sense Polarity	True Bipolar/One Coil
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	391
Crimp Weld Bond	
Insulation Breach	17
Other	64

US Acute Lead Observations


Cardiac Perforation	123
Conductor Fracture	11
Extracardiac Stimulation	22
Failure To Capture	235
Failure To Sense	72
Impedance Abnormal	71
Insulation Breach	2
Lead Dislodgement	419
Oversensing	203
Unspecified	

Product Surveillance Registry Results

Number of Leads Enrolled in Study	6,768
Cumulative Months of Followup	227,023
Number of Leads Active in Study	4,044

Qualifying Complications

Cardiac Perforation	1	Impedance Abnormal	5
Conductor Fracture	19	Insulation Breach	2
Extracardiac Stimulation	1	Lead Dislodgement	15
Failure To Capture	13	Oversensing	2
Failure To Sense	1	Unspecified	1
		Other Complication	2



Years	1	2	3	4	5	6	at 84 mo
%	99.6%	99.5%	99.2%	98.9%	98.4%	98.0%	98.0%
#	5,238	4,125	3,152	1,952	1,007	403	52

6937A Transvene SVC-CS

US Market Release	06Apr2001
CE Approval	
Registered USA Implants	2,688
Estimated Active USA Implants	1,577
Fixation Type	Passive
Pace Sense Polarity	One Coil
Steroid Indicator	None

US Returned Product Analysis

Conductor Fracture	5
Crimp Weld Bond	
Insulation Breach	
Other	

US Acute Lead Observations


Cardiac Perforation	
Conductor Fracture	3
Extracardiac Stimulation	
Failure To Capture	
Failure To Sense	
Impedance Abnormal	
Insulation Breach	
Lead Dislodgement	
Oversensing	
Unspecified	2

Product Surveillance Registry Results

Number of Leads Enrolled in Study	122
Cumulative Months of Followup	13,760
Number of Leads Active in Study	9

Qualifying Complications

Conductor Fracture	5	Impedance Abnormal	1
		Insulation Breach	2
		Lead Dislodgement	1
		Unspecified	4
		Other Complication	1



Years	1	2	3	4	5	6	7	8	9	at 114 mo
%	99.1%	99.1%	99.1%	98.2%	95.4%	94.3%	93.2%	91.9%	89.1%	89.1%
#	115	113	109	103	91	82	75	69	56	51

US Market Release	13Dec2000
CE Approval	05Nov1999
Registered USA Implants	44,787
Estimated Active USA Implants	17,965
Fixation Type	Tines
Pace Sense Polarity	True Bipolar/Two Coils
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	209
Crimp Weld Bond	1
Insulation Breach	5
Other	4

US Acute Lead Observations


Cardiac Perforation	
Conductor Fracture	2
Extracardiac Stimulation	
Failure To Capture	17
Failure To Sense	3
Impedance Abnormal	10
Insulation Breach	
Lead Dislodgement	24
Oversensing	18
Unspecified	6

Product Surveillance Registry Results

Number of Leads Enrolled in Study	621
Cumulative Months of Followup	34,575
Number of Leads Active in Study	113

Qualifying Complications

Conductor Fracture	16	Impedance Abnormal	4
Failure To Capture	4	Oversensing	3
Failure To Sense	1	Unspecified	1




Years	1	2	3	4	5	6	7	8	9	10	at 132 mo
%	100.0%	99.8%	99.2%	97.3%	95.1%	91.9%	91.3%	90.7%	90.7%	90.7%	89.2%
#	500	415	349	288	226	184	150	128	94	74	52

US Market Release	05Jan2016
CE Approval	12Sep2013
Registered USA Implants	2,468
Estimated Active USA Implants	2,360
Fixation Type	Tines
Pace Sense Polarity	True Bipolar/Two Coils
Steroid Indicator	Yes

US Returned Product Analysis

US Acute Lead Observations

Cardiac Perforation	
Conductor Fracture	
Extracardiac Stimulation	
Failure To Capture	2
Failure To Sense	
Impedance Abnormal	
Insulation Breach	
Lead Dislodgement	5
Oversensing	5
Unspecified	



Years	at mo
%	
#	

US Market Release	12Nov2001
CE Approval	04Oct2001
Registered USA Implants	375,413
Estimated Active USA Implants	185,409
Fixation Type	Active Screw In
Pace Sense Polarity	True Bipolar/Two Coils
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	1,238
Crimp Weld Bond	4
Insulation Breach	98
Other	190

US Acute Lead Observations

Cardiac Perforation	29
Conductor Fracture	26
Extracardiac Stimulation	2
Failure To Capture	82
Failure To Sense	34
Impedance Abnormal	59
Insulation Breach	4
Lead Dislodgement	123
Oversensing	140
Unspecified	20

Product Surveillance Registry Results


Number of Leads Enrolled in Study	4,458
Cumulative Months of Followup	263,968
Number of Leads Active in Study	965

Qualifying Complications

Conductor Fracture	33
Failure To Capture	7
Failure To Sense	2

91

Impedance Abnormal	13
Insulation Breach	5
Lead Dislodgement	5
Oversensing	19
Unspecified	3
Other Complication	4



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	at 198 mo
%	99.5%	99.3%	99.0%	98.7%	98.2%	97.9%	97.5%	97.0%	96.6%	96.0%	95.6%	95.0%	95.0%	94.4%	93.7%	90.5%	90.5%
#	3,273	2,875	2,519	2,223	1,976	1,728	1,456	1,240	976	699	436	256	197	166	132	84	64

US Market Release	13Feb2012
CE Approval	12Mar2010
Registered USA Implants	121,904
Estimated Active USA Implants	106,264
Fixation Type	Active Screw In
Pace Sense Polarity	True Bipolar/Two Coils
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	173
Crimp Weld Bond	
Insulation Breach	12
Other	30

US Acute Lead Observations

Cardiac Perforation	33
Conductor Fracture	9
Extracardiac Stimulation	11
Failure To Capture	99
Failure To Sense	37
Impedance Abnormal	28
Insulation Breach	
Lead Dislodgement	208
Oversensing	74
Unspecified	

Product Surveillance Registry Results


Number of Leads Enrolled in Study	2,166
Cumulative Months of Followup	102,428
Number of Leads Active in Study	851

Qualifying Complications

Conductor Fracture	9
Failure To Capture	4
Failure To Sense	2

17

Lead Dislodgement	1
Other Complication	1



Years	1	2	3	4	5	6	7	at 90 mo
%	99.7%	99.5%	99.3%	99.3%	99.0%	99.0%	98.5%	98.5%
#	1,730	1,459	1,288	1,074	886	706	382	167

US Market Release	02Sep2004
CE Approval	
Registered USA Implants	10,343
Estimated Active USA Implants	2,699
Fixation Type	Tines
Pace Sense Polarity	True Bipolar/Two Coils
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	210
Crimp Weld Bond	
Insulation Breach	3
Other	4

US Acute Lead Observations


Cardiac Perforation	
Conductor Fracture	2
Extracardiac Stimulation	
Failure To Capture	7
Failure To Sense	
Impedance Abnormal	
Insulation Breach	
Lead Dislodgement	7
Oversensing	1
Unspecified	3

Product Surveillance Registry Results

Number of Leads Enrolled in Study	39
Cumulative Months of Followup	2,298
Number of Leads Active in Study	4

Qualifying Complications

Conductor Fracture	3	Impedance Abnormal	1
--------------------	---	--------------------	---



Years	at 0 mo
%	100.0%
#	0

US Market Release	02Sep2004
CE Approval	
Registered USA Implants	186,160
Estimated Active USA Implants	40,652
Fixation Type	Active Screw In
Pace Sense Polarity	True Bipolar/Two Coils
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	7,983
Crimp Weld Bond	3
Insulation Breach	37
Other	104

US Acute Lead Observations


Cardiac Perforation	10
Conductor Fracture	48
Extracardiac Stimulation	
Failure To Capture	32
Failure To Sense	19
Impedance Abnormal	19
Insulation Breach	5
Lead Dislodgement	22
Oversensing	35
Unspecified	25

Product Surveillance Registry Results

Number of Leads Enrolled in Study	980
Cumulative Months of Followup	55,832
Number of Leads Active in Study	73

Qualifying Complications

Conductor Fracture	73	Impedance Abnormal	19
Failure To Capture	5	Insulation Breach	2
Failure To Sense	6	Lead Dislodgement	1
		Oversensing	21
		Other Complication	2



Years	1	2	3	4	5	6	7	8	9	10	11	12	at 156 mo
%	98.5%	96.5%	93.4%	91.0%	88.2%	84.4%	81.5%	79.3%	78.6%	77.2%	71.5%	69.0%	66.7%
#	717	624	530	456	390	342	280	236	188	153	126	94	65

US Market Release	11Jun2001
CE Approval	19Dec1997
Registered USA Implants	5,275
Estimated Active USA Implants	2,841
Fixation Type	Suture on Anchor Sleeve
Pace Sense Polarity	One Coil
Steroid Indicator	None

US Returned Product Analysis

Conductor Fracture	33
Crimp Weld Bond	
Insulation Breach	
Other	

US Acute Lead Observations


Cardiac Perforation	1
Conductor Fracture	
Extracardiac Stimulation	
Failure To Capture	1
Failure To Sense	
Impedance Abnormal	13
Insulation Breach	1
Lead Dislodgement	2
Oversensing	1
Unspecified	

Product Surveillance Registry Results

Number of Leads Enrolled in Study	53
Cumulative Months of Followup	2,338
Number of Leads Active in Study	6

Qualifying Complications

Conductor Fracture	1	Impedance Abnormal	2
--------------------	---	--------------------	---



Years	at 0 mo
%	100.0%
#	0

US Market Release	28Aug2001
CE Approval	
Registered USA Implants	11,925
Estimated Active USA Implants	1,617
Fixation Type	Distal Continuous Curve
Pace Sense Polarity	Unipolar
Steroid Indicator	None

US Returned Product Analysis

Conductor Fracture	1
Crimp Weld Bond	
Insulation Breach	3
Other	2

US Acute Lead Observations


Cardiac Perforation	
Conductor Fracture	
Extracardiac Stimulation	1
Failure To Capture	3
Failure To Sense	1
Impedance Abnormal	
Insulation Breach	
Lead Dislodgement	9
Oversensing	
Unspecified	

Product Surveillance Registry Results

Number of Leads Enrolled in Study	140
Cumulative Months of Followup	6,989
Number of Leads Active in Study	6

Qualifying Complications

Failure To Capture	3
--------------------	---



Years	1	2	3	at 48 mo
%	99.1%	98.0%	98.0%	98.0%
#	101	85	65	52

US Market Release	03May2002
CE Approval	22Dec2000
Registered USA Implants	100,523
Estimated Active USA Implants	21,119
Fixation Type	Double Curve
Pace Sense Polarity	Unipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	88
Crimp Weld Bond	
Insulation Breach	31
Other	12

US Acute Lead Observations


Cardiac Perforation	
Conductor Fracture	
Extracardiac Stimulation	18
Failure To Capture	11
Failure To Sense	
Impedance Abnormal	
Insulation Breach	
Lead Dislodgement	45
Oversensing	1
Unspecified	2

Product Surveillance Registry Results

Number of Leads Enrolled in Study	803
Cumulative Months of Followup	40,834
Number of Leads Active in Study	49

Qualifying Complications

Conductor Fracture	1	Impedance Abnormal	2
Extracardiac Stimulation	10	Lead Dislodgement	14
Failure To Capture	19	Unspecified	3



Years	1	2	3	4	5	6	7	8	9	10	11	12	at 150 mo
%	96.0%	95.1%	94.4%	94.1%	93.4%	92.6%	92.2%	91.7%	91.1%	90.4%	90.4%	88.1%	88.1%
#	567	442	374	301	249	225	191	169	137	114	93	66	57

US Market Release	24Aug2004
CE Approval	14Jul2003
Registered USA Implants	114,938
Estimated Active USA Implants	47,538
Fixation Type	Double Curve
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	44
Crimp Weld Bond	
Insulation Breach	151
Other	2

US Acute Lead Observations


Cardiac Perforation	2
Conductor Fracture	2
Extracardiac Stimulation	49
Failure To Capture	42
Failure To Sense	
Impedance Abnormal	9
Insulation Breach	
Lead Dislodgement	152
Oversensing	2
Unspecified	4

Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,645
Cumulative Months of Followup	90,692
Number of Leads Active in Study	288

Qualifying Complications

Conductor Fracture	2
Extracardiac Stimulation	11
Failure To Capture	19
Insulation Breach	2
Lead Dislodgement	30
Insulation Breach Esc	1



Years	1	2	3	4	5	6	7	8	9	10	11	12	at 156 mo
%	98.6%	97.4%	96.7%	96.1%	95.6%	94.4%	94.2%	93.5%	93.5%	93.2%	92.7%	92.7%	92.7%
#	1,237	1,045	896	767	693	605	467	366	283	202	143	81	54

US Market Release	15Aug2008
CE Approval	13May2005
Registered USA Implants	17,420
Estimated Active USA Implants	10,015
Fixation Type	Deployable Lobe Fixation
Pace Sense Polarity	Unipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	10
Crimp Weld Bond	
Insulation Breach	3
Other	2

US Acute Lead Observations


Cardiac Perforation	
Conductor Fracture	
Extracardiac Stimulation	29
Failure To Capture	21
Failure To Sense	
Impedance Abnormal	4
Insulation Breach	
Lead Dislodgement	29
Oversensing	
Unspecified	1

Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,486
Cumulative Months of Followup	78,948
Number of Leads Active in Study	301

Qualifying Complications

Conductor Fracture	4	Impedance Abnormal	2
Extracardiac Stimulation	14	Insulation Breach	5
Failure To Capture	8	Lead Dislodgement	5
		Other Complication	1



Years	1	2	3	4	5	6	7	8	9	at 120 mo
%	99.2%	98.6%	98.2%	97.7%	97.3%	96.9%	96.4%	95.5%	95.1%	95.1%
#	1,244	1,073	924	740	608	483	347	233	138	58

US Market Release	15May2009
CE Approval	24Jul2007
Registered USA Implants	69,379
Estimated Active USA Implants	43,846
Fixation Type	Double Curve
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	24
Crimp Weld Bond	
Insulation Breach	2
Other	9

US Acute Lead Observations


Cardiac Perforation	3
Conductor Fracture	2
Extracardiac Stimulation	95
Failure To Capture	63
Failure To Sense	1
Impedance Abnormal	10
Insulation Breach	1
Lead Dislodgement	218
Oversensing	1
Unspecified	2

Product Surveillance Registry Results

Number of Leads Enrolled in Study	2,299
Cumulative Months of Followup	105,677
Number of Leads Active in Study	397

Qualifying Complications

Conductor Fracture	3	Impedance Abnormal	2
Extracardiac Stimulation	14	Insulation Breach	1
Failure To Capture	39	Lead Dislodgement	23
		Other Complication	4



Years	1	2	3	4	5	6	7	8	9	at 114 mo
%	98.0%	97.3%	96.6%	95.9%	95.5%	94.8%	94.4%	94.1%	93.3%	92.7%
#	1,868	1,478	1,167	927	751	580	428	310	181	107

US Market Release	01Apr2011
CE Approval	18Dec2009
Registered USA Implants	34,783
Estimated Active USA Implants	26,074
Fixation Type	Double Curve
Pace Sense Polarity	Dual Electrodes
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	3
Crimp Weld Bond	2
Insulation Breach	
Other	4

US Acute Lead Observations

Cardiac Perforation	2
Conductor Fracture	1
Extracardiac Stimulation	61
Failure To Capture	30
Failure To Sense	
Impedance Abnormal	11
Insulation Breach	4
Lead Dislodgement	115
Oversensing	
Unspecified	


Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,459
Cumulative Months of Followup	62,973
Number of Leads Active in Study	415

Qualifying Complications

Extracardiac Stimulation	12	Lead Dislodgement	13
Failure To Capture	9	Other Complication	1

35



Years	1	2	3	4	5	6	7	at 90 mo
%	98.7%	97.9%	97.6%	97.2%	96.7%	96.5%	96.5%	96.5%
#	1,150	924	751	630	519	377	184	102

US Market Release	01Aug2014
CE Approval	01Jan2013
Registered USA Implants	86,244
Estimated Active USA Implants	78,936
Fixation Type	Double Curve
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	5
Crimp Weld Bond	
Insulation Breach	
Other	18

US Acute Lead Observations


Cardiac Perforation	6
Conductor Fracture	1
Extracardiac Stimulation	188
Failure To Capture	103
Failure To Sense	1
Impedance Abnormal	28
Insulation Breach	
Lead Dislodgement	181
Oversensing	
Unspecified	

Product Surveillance Registry Results

Number of Leads Enrolled in Study	2,110
Cumulative Months of Followup	61,103
Number of Leads Active in Study	1,336

Qualifying Complications

Extracardiac Stimulation	4	Lead Dislodgement	12
		Other Complication	1



Years	1	2	3	4	at 60 mo
%	99.3%	99.0%	98.9%	98.9%	98.9%
#	1,603	1,105	775	460	159

4396 Attain Ability Straight

US Market Release	31Mar2011
CE Approval	18Dec2009
Registered USA Implants	8,157
Estimated Active USA Implants	6,111
Fixation Type	Tines
Pace Sense Polarity	Dual Electrodes
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	5
Crimp Weld Bond	
Insulation Breach	
Other	

US Acute Lead Observations


Cardiac Perforation	1
Conductor Fracture	2
Extracardiac Stimulation	20
Failure To Capture	10
Failure To Sense	
Impedance Abnormal	
Insulation Breach	
Lead Dislodgement	34
Oversensing	
Unspecified	

Product Surveillance Registry Results

Number of Leads Enrolled in Study	467
Cumulative Months of Followup	20,453
Number of Leads Active in Study	160

Qualifying Complications

Extracardiac Stimulation	1	Insulation Breach	1
Failure To Capture	3	Lead Dislodgement	2



Years	1	2	3	4	5	6	at 84 mo
%	99.8%	99.5%	98.4%	98.4%	98.4%	98.4%	98.4%
#	370	293	253	215	171	110	50

4398 Attain Performa Straight

US Market Release	10Dec2014
CE Approval	01Jan2013
Registered USA Implants	27,306
Estimated Active USA Implants	25,351
Fixation Type	Tines
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	2
Crimp Weld Bond	
Insulation Breach	
Other	5

US Acute Lead Observations


Cardiac Perforation	6
Conductor Fracture	
Extracardiac Stimulation	80
Failure To Capture	43
Failure To Sense	
Impedance Abnormal	6
Insulation Breach	
Lead Dislodgement	34
Oversensing	
Unspecified	

Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,377
Cumulative Months of Followup	25,935
Number of Leads Active in Study	1,040

Qualifying Complications

Failure To Capture	3	8
Impedance Abnormal		1
Lead Dislodgement		4



Years	1	2	3	4	at 54 mo
%	99.7%	99.7%	99.0%	99.0%	97.7%
#	802	447	210	102	58

US Market Release	10Dec2014
CE Approval	01Jan2013
Registered USA Implants	49,758
Estimated Active USA Implants	46,375
Fixation Type	S-shape
Pace Sense Polarity	Quad Pole
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	4
Crimp Weld Bond	
Insulation Breach	
Other	7

US Acute Lead Observations


Cardiac Perforation	9
Conductor Fracture	1
Extracardiac Stimulation	83
Failure To Capture	46
Failure To Sense	
Impedance Abnormal	16
Insulation Breach	
Lead Dislodgement	51
Oversensing	1
Unspecified	

Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,247
Cumulative Months of Followup	31,812
Number of Leads Active in Study	825

Qualifying Complications

Extracardiac Stimulation	2	Lead Dislodgement	9
Failure To Sense	1		



Years	1	2	3	4	at 54 mo
%	99.3%	98.9%	98.9%	98.5%	98.5%
#	936	586	353	168	91

US Market Release	06Sep1996
CE Approval	01Jan1993
Registered USA Implants	23,466
Estimated Active USA Implants	8,162
Fixation Type	Suture
Pace Sense Polarity	Unipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	286
Crimp Weld Bond	1
Insulation Breach	63
Other	

US Acute Lead Observations


Cardiac Perforation	1
Conductor Fracture	1
Extracardiac Stimulation	
Failure To Capture	10
Failure To Sense	5
Impedance Abnormal	16
Insulation Breach	
Lead Dislodgement	
Oversensing	1
Unspecified	3

Product Surveillance Registry Results

Number of Leads Enrolled in Study	234
Cumulative Months of Followup	7,389
Number of Leads Active in Study	6

Qualifying Complications

Conductor Fracture	10
Failure To Capture	3
Failure To Sense	1
Insulation Breach	1
Oversensing	2



Years	1	2	3	at 48 mo
%	98.6%	95.8%	94.8%	86.4%
#	119	101	83	61

US Market Release	16Sep1999
CE Approval	21Apr1998
Registered USA Implants	51,723
Estimated Active USA Implants	31,246
Fixation Type	Suture
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	109
Crimp Weld Bond	
Insulation Breach	56
Other	1

US Acute Lead Observations


Cardiac Perforation	1
Conductor Fracture	3
Extracardiac Stimulation	4
Failure To Capture	62
Failure To Sense	7
Impedance Abnormal	10
Insulation Breach	1
Lead Dislodgement	7
Oversensing	24
Unspecified	

Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,027
Cumulative Months of Followup	61,494
Number of Leads Active in Study	221

Qualifying Complications

Conductor Fracture	24	Impedance Abnormal	5
Extracardiac Stimulation	2	Insulation Breach	4
Failure To Capture	29	Lead Dislodgement	1
Failure To Sense	3	Oversensing	25
		Other Complication	2



Years	1	2	3	4	5	6	7	8	9	10	11	12	at 150 mo
%	99.5%	97.6%	96.1%	94.3%	93.2%	91.0%	89.1%	89.1%	84.5%	83.5%	79.9%	75.4%	72.0%
#	799	716	619	529	452	367	316	266	186	138	100	74	61

US Market Release	03Dec1992
CE Approval	01Jan1993
Registered USA Implants	54,914
Estimated Active USA Implants	16,433
Fixation Type	Fixed Screw
Pace Sense Polarity	Unipolar
Steroid Indicator	None

US Returned Product Analysis

Conductor Fracture	27
Crimp Weld Bond	
Insulation Breach	2
Other	1

US Acute Lead Observations


Cardiac Perforation	1
Conductor Fracture	
Extracardiac Stimulation	6
Failure To Capture	85
Failure To Sense	3
Impedance Abnormal	8
Insulation Breach	
Lead Dislodgement	2
Oversensing	1
Unspecified	1

Product Surveillance Registry Results

Number of Leads Enrolled in Study	453
Cumulative Months of Followup	14,668
Number of Leads Active in Study	92

Qualifying Complications

Conductor Fracture	3	Impedance Abnormal	1
Extracardiac Stimulation	1	Lead Dislodgement	2
Failure To Capture	20	Oversensing	2
Failure To Sense	2	Other Complication	1



Years	1	2	3	4	5	6	at 84 mo
%	95.1%	92.2%	90.3%	90.3%	88.7%	86.7%	85.2%
#	222	174	139	119	100	74	51

US Market Release	10Sep1998
CE Approval	15Apr1997
Registered USA Implants	10,367
Estimated Active USA Implants	3,523
Fixation Type	Tines
Pace Sense Polarity	Quadripolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	8
Crimp Weld Bond	
Insulation Breach	2
Other	

US Acute Lead Observations


Cardiac Perforation	
Conductor Fracture	
Extracardiac Stimulation	1
Failure To Capture	3
Failure To Sense	2
Impedance Abnormal	
Insulation Breach	
Lead Dislodgement	7
Oversensing	1
Unspecified	

Product Surveillance Registry Results

Number of Leads Enrolled in Study	568
Cumulative Months of Followup	15,833
Number of Leads Active in Study	3

Qualifying Complications

Conductor Fracture	3
Failure To Capture	2
Failure To Sense	3



Years	1	2	3	4	5	6	at 84 mo
%	99.7%	99.3%	99.3%	97.9%	97.0%	97.0%	94.1%
#	288	218	160	132	104	78	56

ICD and CRT-D Charge Time Performance

Medtronic continues its commitment to providing updated information on charge time performance.

Introduction

Information on charge time performance of Medtronic products is presented in this section of the CRHF Product Performance Report. Medtronic implemented the collection of charge time data on July 1, 1999. The data are collected via our ongoing active clinical study of long-term system performance called the Product Surveillance Registry. The study protocol requests device data be routinely taken and sent to Medtronic at no more than 6-month intervals.


In our analysis performed for this report, only charge times resulting from full energy charges are considered. To ensure consistent reporting across devices, the charge time reported at implant represents the last charge time available from date of implant. When more than one charge time is available in a 6-month interval, a conservative approach has been adopted whereby only the maximum charge time in each 6-month interval is reported. As charge time is directly proportional to the time elapsed since the last capacitor reformation, charges occurring within 15 days of a previous charge are excluded. This precludes the reporting of overly optimistic results.

Data from over 20,000 devices contribute to the charge time data in this report. By tracking and reporting this charge time data, Medtronic is able to ascertain the actual performance of its charging circuitry. The insight gained through this information is applied to Medtronic's ongoing efforts to provide charge times that are short and consistent over the life of the product.

Charge time data for ICD and CRT-D models are presented using boxplots at 6-month intervals. The shaded box on the plots represents the middle half of the data – the Interquartile Range (IQR). The white line in the middle of each box is the median charge time. The top of the box representing the IQR is the third quartile or the 75th percentile (i.e., 75% of all charge times fall below this line), whereas the bottom of the box represents the first quartile or the 25th percentile. Vertical lines are drawn from the quartiles to the farthest value not more than 1.5 times the interquartile range. Any values more extreme than the vertical lines are considered outliers.


7232

Model Number	Brand
7232Cx	Maximo VR




D154AWG, D164AWG

Model Number	Brand
D164AWG	Virtuoso DR




D154VWC, D164VWC

Model Number	Brand
D164VWC	Virtuoso VR




D204DRM, D214DRM, D224DRG, D234DRG

Model Number	Brand
D204DRM	Secura DR
D214DRM	Secura DR
D224DRG	Secura DR
D234DRG	Secura DR




D204TRM, D214TRM, D224TRK, D234TRK

Model Number	Brand
D204TRM	Consulta CRT-D
D214TRM	Consulta CRT-D
D224TRK	Consulta CRT-D
D234TRK	Consulta CRT-D




D204VRM, D214VRM, D224VRC, D234VRC

Model Number	Brand
D204VRM	Secura VR
D214VRM	Secura VR
D224VRC	Secura VR
D234VRC	Secura VR




D264DRG, D284DRG, D384DRx, D394DRx

Model Number	Brand
D264DRM	Maximo II DR
D284DRG	Maximo II DR
D384DRG	Cardia DR
D394DRG	Egida DR




D264TRM, D284TRK, D384TRx, D394TRx

Model Number	Brand
D264TRM	Maximo II CRT-D
D284TRK	Maximo II CRT-D
D384TRG	Cardia CRT-D
D394TRG	Egida CRT-D




D264VRM, D284VRC, D384VRx, D394VRx

Model Number	Brand
D264VRM	Maximo II VR
D284VRC	Maximo II VR
D384VRG	Cardia VR
D394VRG	Egida VR




D274DRG, D294DRG

Model Number	Brand
D274DRG	Virtuoso II DR
D294DRG	Virtuoso II DR




D274TRK, D294TRK

Model Number	Brand
D274TRK	Concerto II CRT-D
D294TRK	Concerto II CRT-D




D274VRC, D294VRC

Model Number	Brand
D274VRC	Virtuoso II VR
D294VRC	Virtuoso II VR




D314DRx

Model Number	Brand
D314DRG	Protecta XT DR
D314DRM	Protecta XT DR




D314TRx

Model Number	Brand
D314TRG	Protecta XT CRT-D
D314TRM	Protecta XT CRT-D




D314VRx

Model Number	Brand
D314VRG	Protecta XT VR
D314VRM	Protecta XT VR




D334DRx, D364DRx

Model Number	Brand
D334DRG	Protecta DR
D334DRM	Protecta DR
D364DRG	Protecta DR
D364DRM	Protecta DR




D334TRx, D364TRx

Model Number	Brand
D334TRG	Protecta CRT-D
D334TRM	Protecta CRT-D
D364TRG	Protecta CRT-D
D364TRM	Protecta CRT-D




D334VRx, D364VRx

Model Number	Brand
D334VRG	Protecta VR
D334VRM	Protecta VR
D364VRG	Protecta VR
D364VRM	Protecta VR




D354DRx

Model Number	Brand
D354DRG	Protecta XT DR
D354DRM	Protecta XT DR




D354TRx

Model Number	Brand
D354TRG	Protecta XT CRT-D
D354TRM	Protecta XT CRT-D




D354VRx

Model Number	Brand
D354VRG	Protecta XT VR
D354VRM	Protecta XT VR




DDxxxxx, DR

Model Number	Brand
DDBB1D1	Evera XT
DDBB1D4	Evera XT
DDBB2D1	Evera XT
DDBB2D4	Evera XT
DDBC3D1	Evera S
DDBC3D4	Evera S
DDMB1D1	Evera MRI XT
DDMB1D4	Evera MRI XT
DDMB2D1	Evera MRI XT
DDMB2D4	Evera MRI XT
DDMC3D1	Evera MRI S
DDMC3D4	Evera MRI
DDMD3D1	Primo
DDMD3D4	Primo
DDME3D1	Mirro
DDME3D4	Mirro




DTxxxxx, CRT-D

Model Number	Brand
DTBA1D1	Viva XT
DTBA1D4	Viva XT
DTBA1Q1	Viva Quad XT
DTBA1QQ	Viva Quad XT
DTBA2D1	Viva XT
DTBA2D4	Viva XT
DTBA2Q1	Viva Quad XT
DTBA2QQ	Viva Quad XT
DTBB1D1	Viva S
DTBB1D4	Viva S
DTBB1Q1	Viva Quad S
DTBB1QQ	Viva Quad S
DTBB2D1	Viva S
DTBB2D4	Viva S
DTBB2QQ	Viva Quad S
DTBC2D1	Brava
DTBC2D4	Brava
DTBC2Q1	Brava Quad
DTBC2QQ	Brava Quad
DTBX1QQ	Viva Quad C
DTBX2QQ	Viva Quad C
DTMA1D1	Claria MRI
DTMA1D4	Claria MRI
DTMA1Q1	Claria MRI
DTMA1QQ	Claria MRI
DTMA2D1	Claria MRI
DTMA2D4	Claria MRI
DTMA2Q1	Claria MRI
DTMA2QQ	Claria MRI
DTMB1D1	Amplia MRI
DTMB1D4	Amplia MRI
DTMB1Q1	Amplia MRI
DTMB1QQ	Amplia MRI
DTMB2D1	Amplia MRI
DTMB2D4	Amplia MRI
DTMB2Q1	Amplia MRI
DTMB2QQ	Amplia MRI
DTMC1D1	Compia MRI
DTMC1QQ	Compia MRI
DTMC2D1	Compia MRI
DTMC2D4	Compia MRI
DTMC2QQ	Compia MRI



DVxxxxx, VR

Model Number	Brand
DVAB1D1	Visia AF
DVAB1D4	Visia AF
DVAB2D1	Visia AF XT
DVAC3D1	Visia AF S
DVBB1D1	Evera XT
DVBB1D4	Evera XT
DVBB2D1	Evera XT
DVBB2D4	Evera XT
DVBC3D1	Evera S
DVBC3D4	Evera S
DVFB1D1	Visia MRI AF
DVFB1D4	Visia MRI AF
DVFB2D1	Visia MRI AF XT
DVFB2D4	Visia MRI AF XT
DVFC3D1	Visia MRI AF S
DVFC3D4	Visia MRI AF S
DVMB1D4	Evera MRI XT
DVMB2D1	Evera MRI XT
DVMB2D4	Evera MRI XT
DVMC3D1	Evera MRI S
DVMC3D4	Evera MRI S
DVMD3D1	Primo
DVMD3D4	Primo
DVME3D1	Mirro
DVME3D4	Mirro



Device Programming Information - Setting VF ATP During Charging Therapy

Cobalt™ XT, Cobalt and Crome™ ICDs and CRT-Ds

Original Date of Communication: September 2020

This communication provides information about the programming of Ventricular Fibrillation Antitachycardia Pacing (VF ATP) During Charging. When enabled, VF ATP During Charging allows the device to simultaneously deliver ATP therapy while charging to deliver a high-voltage VF therapy, if needed.

For Cobalt and Crome ICD and CRT-D devices, clinicians should confirm that the VF ATP parameter has been set to the desired value. Depending on pre-implant programming sequences, the VF ATP parameter may not be automatically enabled and may require manual programming (see Image 1 below). In prior generations of Medtronic devices, the VF ATP parameter was automatically enabled with all VF therapies.

As of 21-Sept-2020, Medtronic has received one (1) complaint (out of 3,237 devices sold worldwide) related to this issue. No serious adverse events have been reported.

These devices will deliver all programmed high-voltage therapies as expected, regardless of the VF ATP parameter setting. Likewise, all device functions will operate as programmed. If the VF ATP is not enabled, there is risk for a high-voltage therapy to be applied for a Fast VT arrhythmia in the VF detection zone, which could have been treated with ATP During Charging.



Clinician Actions

We realize that each patient requires unique clinical considerations. With deference to those considerations, Medtronic recommends physicians follow normal clinical practices, including:

- At implant, as described in labeling, confirm the appropriate selection has been programmed for the VF ATP parameter.
- At routine follow-up, confirm that the VF ATP parameter is programmed to the desired setting for each patient.

CFx Longevity Estimator Software Error - Software Updates Available June 2020

Subset of IPG, ICD, CRT-P, CRT-D, and Micra TPS devices

Original Date of Communication: June 2020

STATUS UPDATE – OCTOBER 1, 2020

CareLink™ Network update (version CLN18) has been released to correct the CFx longevity estimator error in all CareLink transmissions for the Phase 2 devices listed below.

The first phase of CareLink Network updates occurred in June 2020 and corrected the longevity estimation display error for the following device models.

- Azure™/Astra™ pacemakers
- Percepta™/Serena™/Solara™ CRT-Ps
- Visia AF™/Visia AF MRI™ ICDs
- Amplia MRI™/Claria MRI™/Compia MRI™ CRT-Ds

As of September 2020, the second phase of CareLink™ Network updates have been deployed worldwide, and accurate longevity estimates are now being displayed through CareLink for the following device models:

- Viva™/Brava™ CRT-D
- Evera™/Evera MRI™/Primo MRI™/Mirro MRI™ ICDs
- Micra™ VR TPS

Model 2090 and Encore Programmer software updates to correct the longevity estimation display error for the Phase 2 devices listed above are currently under submission with worldwide regulatory bodies. Once these programmer software updates are approved, Medtronic employees will update Model 2090 and Encore programmers in affected accounts. These programmer updates will then conclude all activities associated with the October 2019 advisory. Devices included in this second phase of updates are not currently supported by Medtronic SmartSync™ Device Manager.

Important Note: Until all programmers are updated, a difference in displayed longevity estimates between programmers and the CareLink Network may be observed. If you have questions regarding the accuracy of any longevity estimate, please contact Medtronic Technical Services.

ORIGINAL COMMUNICATION - JUNE 2020

This communication provides notice on the availability of software updates that will correct the issue disclosed in a communication sent in October 2019. The original communication described the potential for Medtronic programmers and remote monitoring systems to display an inaccurate longevity estimate for a well-defined subset of approximately 53,100 implanted cardiac devices worldwide; and that prophylactic device replacement is not recommended, as device functionality and the RRT indicator are not impacted by the inaccurate longevity estimate.

Two phases of software releases will be required to address the issue (refer to Table 1 below). Device families listed under Phase 1 are receiving the software update at this time. Device families listed under Phase 2 will be addressed in future software releases, anticipated to be approved in late calendar year 2020.

Phase 1 – June 2020	Phase 2 – Late 2020
Azure™/Astra™ (SW030) v 8.1	Viva™/Brava™/ Evera
Serena™/ Solara™/ Percepta™ (SW040) v 8.3	Evera™ MRI/ Primo™ MRI/ Mirro™ MRI
Visia AFT™/ Visia AFT™ MRI (SW035) v 8.2	Micra™ VR TPS
Claria™/ Amplia™/ Compia™ (SW034) v 8.4 (US)	

Table 1: Device family updates by phases

As of 5 June 2020, the Medtronic CareLink™ Network has been updated, and longevity estimates displayed through CareLink for devices in Phase 1 will reflect the correct longevity estimate. Azure IPG and Percepta/Serena/Solara CRT-P patients remotely monitored via the MyCareLink Heart™ mobile app will automatically receive an updated longevity estimate on their mobile app with their next scheduled transmission, or within 92 days, whichever comes first.

Actions for devices in Phase 1

The Independent Physician Quality Panel recommends routine follow up in accordance with standard practice for these devices, as RRT function is normal and the battery longevity is unaffected. There is no need to schedule patients to come in outside of their planned, scheduled visits due to this issue. The corrective fix is implemented in programmers, CareLink, and other systems which display device longevity. The patient’s device does not require an update. Follow the steps below as applicable to your clinic or hospital. A local Medtronic Representative can assist in updating Model 2090/Encore programmers and SmartSync Device Managers in your facilities.

- Model 2090 and Encore™ Programmers

These programmers will require new software to be installed to correct the displayed longevity estimator error. The software applications and version are listed in Table 1 above and can be installed via Medtronic Software Distribution Network (SDN) or via secure USB.

- SmartSync™ Device Managers

These tablet-based programmers will require a software update to be installed via the internet - refer to Appendix A (below) for detailed instructions on how to download and install the updated application software.

Completion of programmer updates may be delayed due to COVID 19 pandemic-related facility restrictions. Based on your facility’s needs and accessibility, Medtronic Representative or authorized personnel will work with your facility as requested to complete the updates. Customers with Paceart systems should contact their support team to ensure the latest device update is applied.

Note: Once a programmer is updated, the correct longevity estimate will display at the patient’s next regularly scheduled clinic visit. Until all SmartSync Device Managers and Model 2090 and Encore programmers are updated, a difference in longevity estimates between programmers and CareLink Network-displayed longevity may be observed.

Recommendations for devices in Phase 2

Continue to follow the patient management recommendations from the October 2019 communication (excerpted below) for the subset of patients within the affected population who are not included in the Phase 1 software updates.

Patient Management Recommendations (October 2019)

We realize that each patient requires unique clinical considerations. In consultation with our Independent Physician Quality Panel, Medtronic provides the following guidance:

- Prophylactic device replacement is not recommended, as device functionality and the RRT indicator are not impacted by the inaccurate longevity estimate.

Until the software update becomes available:

- Continue normal patient follow-up in accordance with standard practice.
- Per labeling, continue to use the RRT notification to identify when device replacement should be scheduled. Where available, utilize the low battery voltage RRT audible alert or wireless CareAlert™.
- At any time, if a lower-than-expected remaining longevity estimate occurs, contact Medtronic Technical Services for assistance – additional analysis of stored device information will be required to assess if the decreased longevity estimate is due to this issue.

APPENDIX A – UPDATING SMARTSYNC™ DEVICE MANAGER

Until all SmartSync Device Managers and Model 2090 and Encore programmers are updated, you may observe a difference in longevity estimates between these programmers and CareLink-displayed longevity.

Updating Medtronic SmartSync™ Device Managers:

1) Connect tablet to internet and open the SmartSync App

- The SmartSync App automatically checks for available updates each time it is opened.

2) If your tablet does not contain the most recent software, you will automatically receive a notification that a new version of the SmartSync App is available (3.2.01):

- If pop-up messages appear with the option to “cancel” or to “update”, select “update”.

o **Medtronic Managed Tablets:**If the App closes, find the Medtronic App Catalog, and **select “Install”** to initiate the download.

o **Customer Owned Tablets:**If the App closes, navigate to the AirWatch App Catalog or App Store and **select “Install”** to initiate the download.

- If you do not receive a notification that a new version of the SmartSync App is available, skip to Step 3.

3) Once you confirm the newest version of the SmartSync App is on your tablet, re-open the SmartSync App.

• The app will automatically provide pop-up notifications informing you if there are new versions of device software applications that must be installed (see table below).

o Select CONTINUE for each pop-up window that appears. If you do not receive any pop-up notifications when you open the SmartSync App, then your tablet contains the most recent versions of all available software.

Device Family	SmartSync Application SW Version
Azure™/Astra™ DR and SR	D00U003, Version 3.2.02
Percepta™/Serena™/Solara™	D00U004, Version 3.2.02

SmartSync Device Manager Telemetry Issue – Software Updates Available June 2020

Azure™ pacemakers, and Percepta™, Serena™, Solara™ CRT-pacemakers

Original Date of Communication: June 2020

STATUS UPDATE – OCTOBER 2020

As of 14 Oct 2020, Medtronic has received thirty (30) complaints due to this issue. No adverse events or patient harm have been reported.

As described in the original advisory communication (June 2020), updates are available for the CareLink SmartSync Device Manager to address this issue. The SmartSync Device Manager software version 3.2.01 update can be obtained by connecting the tablet to the internet and requesting all application downloads. The software update will modify the SmartSync Device Manager to prevent this issue from occurring; no patient actions are required. A local Medtronic Representative can assist or advise your staff on the SmartSync update process as needed

ORIGINAL COMMUNICATION – JUNE 2020

This communication provides notice on software updates available for CareLink SmartSync™ Device Managers supporting Medtronic Azure™ pacemakers, and Percepta™, Serena™, Solara™ cardiac resynchronization therapy pacemakers (CRT-P).

This update addresses a rare communication sequence during the first device interrogation with a SmartSync Device Manager that may result in the temporary suspension of some device features (i.e., battery measurements, Capture Management™, Atrial Lead Position Check™, EffectivCRT™ algorithms, and AdaptivCRT™). This rare interaction results in temporary suspension of automatic threshold testing and output adjustments, and suspension of auto-optimization of CRT therapy. The issue is unlikely to result in clinical impact to the patient, and features are restored upon next programmer device interrogation or presence of a magnet.

As of 8 May 2020, Medtronic has received sixteen (16) complaints due to this issue. The predicted rate of occurrence for this issue is 0.03% on first interrogation of an Azure, Percepta, Serena, or Solara device with a SmartSync programmer. No adverse events or patient harm have been reported. Based on consultation with the Independent Physician Quality Panel and considering that the issue is unlikely to result in clinical impact to the patient, routine patient follow-up in accordance with standard practice is recommended.

Updates are available for the CareLink SmartSync Device Manager to address this issue. The SmartSync Device Manager software version 3.2.01 update can be obtained by connecting the tablet to the internet and requesting all application downloads. The software update will modify the SmartSync Device Manager to prevent this issue from occurring; no patient actions are required.

A local Medtronic Representative can assist or advise your staff on the SmartSync update process as needed.

Azure S DR Atrial Lead Position Check (ALPC) Incorrectly Enabled – Software Update Available June 2020

Subset of Azure™ S DR pacemakers

Original Date of Communication: June 2020

STATUS UPDATE – OCTOBER 2020

As of 14 Oct 2020, there have been eight (8) complaints reported due to the ALPC feature being enabled and over 45,000 devices distributed. No serious adverse events or patient harm have been reported.

ORIGINAL COMMUNICATION – JUNE 2020

This communication provides notice on a software update available for a subset of Azure™ S DR pacemakers manufactured prior to February 2020 to address an issue in which the Atrial Lead Position Check (ALPC) was incorrectly enabled in a subset of this device model. ALPC is intended to operate as an optional feature in device models that offer atrial anti-tachy pacing therapies (ATP). Model Azure S DR does not offer atrial ATP. This update will ensure that ALPC is inactivated in all Azure S DR devices. Device therapies and battery performance are not affected by this issue.

As of 11 May 2020, there have been seven (7) complaints reported due to the ALPC feature being enabled and over 45,000 devices distributed. ALPC has the potential to pace at the programmed pacing rate for approximately 5 minutes at high output during its nightly assessment. No serious adverse events or patient harm have been reported.

Currently, updates are available for CareLink SmartSync™ Device Manager for this issue. The SmartSync Device Manager may receive software version 3.2.01 update by connecting the tablet to the internet. As of 4 June 2020, software application SW030 version 8.1 will be available via Medtronic Software Distribution Network (SDN) for Model 2090 and Encore programmers. In mid-June 2020, software application SW030 version 8.1 will be available via secure USB for Model 2090 and Encore programmers.

Completion of programmer updates may be delayed due to COVID 19 pandemic-related facility restrictions. Based on your facility's needs and accessibility, a Medtronic Representative or authorized personnel will assist with installing software on programmers in your account. Once a programmer is updated, the ALPC feature will be automatically inactivated at the patient's next regularly scheduled interrogation if the device is in scope of this issue. There is no need to schedule patients to come in outside of their planned, scheduled visits due to this issue.

Potential for Partial Reset During Programmer Interrogation

Claria MRI, Amplia MRI, Compia MRI and CRT-Ds

Original Date of Communication: March 2020

Model
CareLink™ 2090 Programmer with Software Application SW034 versions 8.3 and 8.4
CareLink™ 29901 Programmer with Software Application SW034 versions 8.3 and 8.4

STATUS UPDATE – OCTOBER 2020

Medtronic has identified two versions of software that are susceptible to the one-time partial reset during programming interrogation – software applications SW034 version 8.3 and version 8.4. As documented in the original communication (March 2020), the risk for a partial reset on first interrogation with a programmer is approximately 2%. As of October 27, 2020, there are 313 complaints received due to this issue and zero (0) adverse events reported as a result of this behavior.

Recommendations remain unchanged from the original posting. Medtronic recommends continued routine management of your patients. We recognize that a one-time loss of stored device information may limit your ability to assess your patient's clinical status – particularly when an audible alert, symptoms or VF shock delivery has been reported. Please work with your Medtronic Representative to identify data management options that may be available to your clinic.

ORIGINAL COMMUNICATION - MARCH 2020

This notice provides information regarding the potential for a one-time loss of diagnostic information due to a partial electrical reset that may occur for patients implanted with a Medtronic ClariaMRI, Amplia MRI or Compia MRI Cardiac Resynchronization Defibrillator (CRT-D). Based on data available as of March 2020, the calculated occurrence rate of this one-time partial reset is approximately 2%. **Device therapy and programmed settings are not affected by a partial electrical reset.**

A patient with a Claria MRI, Amplia MRI, or Compia MRI may experience a partial electrical reset when the patient has their device interrogated with a programmer that has been updated to software application SW034 version 8.3, and it is the **first** interrogation with this new software.

Background Information

Medtronic analysis identified that the 2% risk for a partial electric reset during the interrogation process is due to an uncommon scenario when a software update is installed simultaneously with routine critical memory scans. Should a reset occur, the clinician will be prompted to "Clear" the reset condition on the programmer (guidance to clear a partial reset is documented in the Instructions for Use for the above-named devices). When the "Clear" option is selected, the programmer will automatically interrogate the device again, and will successfully write the software enhancement to the device memory. Importantly, 98% of download attempts will successfully complete without an electrical reset. **Once the software update has been successfully installed into the device, the potential for a future partial reset due to this interaction no longer exists.**

Additional Details

As documented in the Instructions for Use, a partial electrical reset will result in the loss of stored diagnostic information and episodes. The device longevity estimator will show an "initializing" status for the next seven (7) days, and Recommended Replacement Time (RRT) status will continue to function as normal. Device programmed parameters, and all functions including detection and therapies are maintained. All Claria MRI, Amplia MRI and Compia MRI CRT-D devices are updated with the new software when interrogated for the first time by a programmer with software application SW034 version 8.3.

Medtronic recommends continued routine management of your patients. We recognize that a one-time loss of stored device information may limit your ability to assess your patient's clinical status – particularly when an audible alert, symptoms or VF shock delivery has been reported. Please work with your Medtronic Representative to identify data management options that may be available to your clinic.

Performance Note: Potential For Premature Battery Depletion in a Subset of ICD and CRT-D Devices

STATUS UPDATE – OCTOBER 2020

Through 28 October 2020, the rate of premature battery depletions due to this issue is 0.07%. We have received no reports of permanent harm to patients as a result of this issue.

Confirmed premature battery depletions, regardless of cause, are reported in our semi-annual Product Performance report under the confirmed "Malfunctions" section for each device model. Product Performance information can be accessed directly productperformance.medtronic.com.

Note: Medtronic has determined that the original posting of this Performance Note incorrectly included the Primo MRI and Mirro MRI ICD device models. These device models are not in scope of this communication as all devices manufactured under these model names were manufactured with the enhanced battery design.

ORIGINAL COMMUNICATION - NOVEMBER 2019

Battery Enhancements Implemented

Medtronic identified a rare failure mechanism in the battery design of specific implantable cardioverter defibrillator (ICD) and cardiac resynchronization therapy defibrillator (CRT-D) models that could result in rapid battery depletion. The rapid depletion is caused by a latent shorting mechanism resulting from lithium plating between the anode and cathode elements of the battery. As a result of our understanding of this phenomenon, Medtronic implemented battery design enhancements. All products currently in distribution contain the battery enhancement, however approximately 607,800 devices distributed worldwide were manufactured prior to implementing the battery enhancement and were distributed under the following brand names¹:

- Claria MRI™/Amplia MRI™/Compia MRI™ CRT-Ds
- Viva™/Brava™ CRT-Ds
- Visia AF™/Visia AF MRI™ ICDs
- Evera™/Evera MRI™/Primo MRI™/Mirro MRI™ ICDs

Potential for Premature Battery Depletion in a subset of ICD and CRT-D devices prior to battery enhancement

Approximately 0.04% of devices exhibit this behavior. The battery continues to perform within projected estimates. There have been no reports of permanent harm to patients as a result of this issue.

Under rare circumstances, a small percentage of ICD and CRT-D devices manufactured prior to the battery enhancement may develop lithium plating. If lithium bridges between a positive (cathode) and a negative (anode) element in the battery, an internal short will develop and the battery will deplete rapidly. If this occurs, the device may not meet expected longevity or provide at least three months of device operation between the Recommended Replacement Time (RRT) and End of Service (EOS).

All events have occurred during the mid-portion of device life; typically 1-4 years after implant. Note, there have been no reports of this issue occurring after RRT has triggered under normal conditions. Therefore, when a device reaches RRT based on its programmed settings and use conditions, the device is likely performing as expected and time between RRT and EOS should be as labeled.

Continue to Follow Normal Clinical Practice per Instructions for Use –Pay Attention to Unexpected RRT or Unexpected Changes in Longevity

- Medtronic, in consultation with our Independent Physician Quality Panel, does not recommend prophylactic replacement of any ICD or CRT-D devices manufactured prior to the battery enhancement. Physicians can continue normal patient follow-up in accordance with standard practice.
- Where possible, take advantage of the CareLink™ home monitoring system and the low battery voltage wireless CareAlert to assist with remote management of patients.
- As always, remind patients to seek medical attention if they hear a device audible alert (shipped On with high urgency toning for low battery voltage indicator).
- At each follow-up, verify the status of the implanted system as well as the clinical effectiveness of the device. Monitor changes in device longevity and note any unexpected device status indicators such as RRT and/or EOS, the inability to interrogate the device or to transmit data.
- As with all unexpected events, including a rapid unexplained voltage drop, inform a Medtronic representative immediately if any of the above behaviors are observed. Further device analysis may be warranted to determine if immediate replacement is necessary.
- If there is evidence of rapid battery voltage drop, patients may need to have their devices replaced urgently as device failure may lead to intended therapy not being delivered.

Additional Details

Contact Medtronic Technical Services if you have concerns on a specific patient. A serial number look-up to assist with identifying if an ICD or CRT-D was manufactured prior to the battery enhancement is available at: <https://wwwp.medtronic.com/productperformance/>

Confirmed premature battery depletions, regardless of cause, are reported in our semi-annual Product Performance report under the confirmed “Malfunctions” section for each device model. Product Performance information can be accessed directly at: <https://wwwp.medtronic.com/productperformance/>

Q1) Can any ICD or CRT-D battery that uses lithium experience this rare, latent shorting mechanism?

Yes. Industry-wide, every ICD or CRT-D battery that uses lithium has the potential for plating to develop under normal use conditions and create an internal short. Lithium plating leading to an internal short is influenced by a number of factors including the battery design. There are differences in the battery design (e.g. layout and insulation) for each manufacturer. Note that the lithium plating phenomenon described in this Performance Note is different, and more rare, than lithium “cluster” formations that result from high current pulsing (charging) as has been described in literature.^{2,3,4}

Q2) Are all device models equally susceptible to this rare failure mechanism?

Devices with higher use conditions (such as CRT-D devices) are less susceptible to the failure mode. This is because the free electrolyte element of the battery, which contributes to lithium plating, is consumed by the cathode more rapidly under high current conditions. Additionally, devices that reach RRT as expected, based on programmed settings and use conditions, are also not likely to experience lithium plating since the electrolyte is consumed as part of the normal discharge process of the battery.

¹Device models vary by geography; not all models are available in all geographies.

²Aggarwal, A, et. al. Accelerated Implantable Defibrillator Battery Depletion Secondary to Lithium Cluster Formation: A Case Series. PACE 2016;39:375-7.

³Pokorney, SD, et. al. Novel mechanism of premature battery failure due to lithium cluster formation in implantable cardioverter defibrillators. Heart Rhythm 2014;11:2190-5.

⁴Hayashi, Y, et. al. A case of unexpected early battery depletion caused by lithium cluster formation in implantable cardioverter-defibrillator. J Cardiol Cases 2017;15:184-6.

CFx Longevity Estimator Software Error

Subset of IPG, ICD, CRT-P, CRT-D, and Micra TPS devices

Original Date of Advisory: October 2019

Affected Programmers & Remote Monitoring Software Apps	Affected Devices
2090 CareLink™ Programmer 29901 Encore™ Programmer CareLink Network Application Software 2491 CareLink SmartSync™ Device Manager MyCareLink Heart™ Mobile Application	Subset of the following devices: Claria MRI™/Amplia MRI™/Compia MRI™/Viva™/Brava™ CRT-Ds Visia AF™/ Visia AF MRI™/Evera™/ Evera MRI™/Primo MRI™/Mirro MRI™ ICDs Azure™/Astra™ IPGs Percepta™/Serena™/Solara™ CRT-Ps Micra™TPS

STATUS UPDATE – OCTOBER 2020

As of October 5, 2020, there have been 312 total complaints received related to the software displaying a lower-than-expected longevity estimate. Within the 312 complaints reported, no patient harm was reported and four (4) devices were prematurely explanted after observing an inaccurate longevity estimate.


ORIGINAL ADVISORY – OCTOBER 2019

Medtronic identified the potential for Medtronic programmer and remote monitoring software applications to display an inaccurate remaining longevity estimate for a subset of implanted cardiac device models. This issue does not impact device functionality. Furthermore, the Recommended Replacement Time (RRT) remains an accurate indicator for device replacement.

Through September 18, 2019 there have been three (3) reported complaints and there have been no (0) serious adverse events or deaths.

The inaccurate longevity estimation is limited to a well-defined subset of devices manufactured between October 2018 and April 2019, and only occurs in the middle (plateau) phase of the device life, as illustrated in the graph below. Approximately 53,100 devices worldwide, out of 1.23 million distributed or sold from the identified device families, are susceptible to displaying inaccurate longevity.

The cause of the inaccurate longevity estimate is a slightly lower-than-typical discharge voltage during the plateau phase of the battery depletion curve (dashed line), compared to a typical voltage plateau (solid line), as illustrated in the graph below. During this plateau period, the Carbon Monofluoride (CFx) in the battery cathode is powering the device. Note, longevity estimates early after implantation and later in the device life are unaffected, as shown below. The battery remains within operating specifications.



Software updates to programmers and remote monitoring systems are under development to correct for the inaccuracy in longevity estimates. Medtronic is targeting regulatory approval and release of the software updates to begin in mid-2020. Once available, Medtronic will inform you of the availability of the software and work with you to install the software onto clinic and hospital programmers. Software updates to individual patient devices will not be necessary to correct this issue, since longevity estimation resides on the programmers, mobile app and the CareLink Network.

Internal analysis estimates approximately 11% of the 53,100 identified devices are projected to display an inaccurate longevity estimate before mid-2020.

Patient Management Recommendations

We realize that each patient requires unique clinical considerations. In consultation with our Independent Physician Quality Panel, Medtronic provides the following guidance:

- **Prophylactic device replacement is not recommended**, as device functionality, true longevity and the RRT indicator are not impacted by the inaccurate longevity estimate.

Until the software update becomes available:

- Continue normal patient follow-up in accordance with standard practice.
- Per labeling, continue to use the RRT notification to identify when device replacement should be scheduled. Where available, utilize the low battery voltage RRT audible alert or wireless CareAlert™.
- At any time, if a lower-than-expected remaining longevity estimate occurs, contact Medtronic Technical Services for assistance – additional analysis of stored device information will be required to assess if the decreased longevity estimate is due to this issue.

Note: For Azure IPG or Percepta/Serena/Solara CRT-P patients remotely monitored via the MyCareLink Heart mobile app, patients' mobile app longevity estimates will not change until the software update has been released.

We sincerely regret any difficulties this may cause you and your patients. Medtronic remains dedicated to patient safety and will continue to monitor device performance to ensure we meet your needs and those of your patients.

Performance Note: Potential for no output/no telemetry condition in subset of IPG and CRT-P products due to ceramic capacitor leakage pathway

Azure™ and Astra™ pacemakers, and Percepta™, Serena™ and Solara™ CRT-P

STATUS UPDATE – OCTOBER 2020

As of October 14, 2020, there have been a total of 14 confirmed events worldwide associated with this failure mode. One additional death has been reported since the original May 2019 publication of this issue. The confirmed event included a report of loss of pacing therapy*.

Product manufactured after June 1, 2019, is not susceptible to this issue as these products utilize a different low voltage capacitor. Product manufactured prior to June 1, 2019 (i.e. manufactured with the original low voltage capacitor) continues to perform within our reliability projections as established as part of the product development process.

Confirmed premature battery depletions, regardless of cause, are reported in our semi-annual Product Performance report under the confirmed “Malfunctions” section for each device model. Product Performance information can be accessed directly at: <https://wwwp.medtronic.com/productperformance/>

*Cause of death was reported as acute cerebrovascular accident, which occurred several days prior to hospital admission. Manner of death was reported as natural; loss of pacing therapy could not be ruled out as a contributing factor.

ORIGINAL COMMUNICATION: MAY 2019

Medtronic has identified a rare but potentially serious failure mode in a population of Azure™ and Astra™ pacemakers, and Percepta™, Serena™ and Solara™ cardiac resynchronization therapy pacemakers (CRT-P), manufactured with a specific multilayer ceramic capacitor. These devices continue to perform within reliability projections.

While inherently very reliable, a known failure mode of these capacitors is the potential for internal cracking that can be caused by thermal-mechanical stress during manufacturing. Under rare conditions, internal cracking within a capacitor may result in the development of a leakage pathway, causing high current drain and leading to rapid battery depletion. While the issue presents as rapid battery depletion, this is not a battery performance issue.

As of April 26, 2019, three complaints out of ~266,700 devices distributed worldwide since February 2017, have been received that included a no output /no telemetry scenario resulting from rapid battery depletion. Battery depletion due to this issue can range from several days to several weeks. One of these reported events contributed to a patient death. The three confirmed failures occurred within 9 months post implant. The projected rate for this issue is 0.0028%, with the most susceptible period for a leakage pathway to develop in the capacitor being the first 12 months post implant.

Based on the low predicted rate of failure and the recent implementation of process and component enhancements, Medtronic expects few, if any, additional events to occur. Medtronic, in consultation with our Independent Physician Quality Panel, does not recommend device replacement. Physicians should continue normal patient follow-up in accordance with standard practice, and where possible, continue to utilize the low battery voltage wireless CareAlert™ (shipped ON), together with remote monitoring via CareLink™ home monitor or the MyCareLink Heart™ mobile app. Per the instructions for use, at each follow-up, verify the status of the implanted system as well as the clinical effectiveness of the device. Pay attention to any unexpected changes in remaining longevity estimates or the inability to interrogate the device and/or transmit data.

Contact Medtronic Technical Services if you have concerns on a specific patient.
Brady Technical Services |rs.techservices@medtronic.com| 800-505-4636

Dual Chamber IPG Circuit Error

Adapta, Versa, Sensia, Relia, Attesta, Sphera, and Vitatron A, E, G, Q series Original

Date of Advisory: January 2019

STATUS UPDATE – OCTOBER 2020

- In September 2019 Medtronic released several software updates to correct for this issue. These software applications are:
 - o For Adapta/Versa/Sensia IPGs - Software model SW003 v8.2
 - o For Relia IPGs - SW010 v8.2
 - o For Attesta/Sphera IPGs - SW043 v8.2
 - o For Vitatron IPGs – VSF20 v8.2 and FSF21 v8.2
- Once a device is interrogated by a programmer with the updated software, any pacemaker programmed to a non-susceptible pacing mode, specifically to avoid a circuit error, may be reprogrammed to any pacing mode.
- Once a device is updated (update is installed onto devices via interrogation by a programmer with one of the above software applications), if the circuit error were to occur, the pacing cycle will automatically reset; this may be observed as a single dropped beat.
- As of October 9, 2020, 86,000 devices remain active out of an original population of 156,957 devices worldwide.

Initial Affected Population	Number of Confirmed Advisory Related Events	Estimated Remaining Active Population	Current Malfunction Rate (confirmed malfunctions over total population)
156,957 Worldwide	36 Worldwide	86,000 Worldwide	0.02% Worldwide

ORIGINAL COMMUNICATION - JANUARY 2019

Product

A subset of Medtronic dual chamber pacemakers distributed worldwide between 10 March 2017 and 7 January 2019 under the brand names Adapta™, Versa™, Sensia™, Relia™, Attesta™, Sphera™, and Vitatron™ A, E, G, Q series may experience a circuit error that affects device functionality. Please note that not all devices within these brand names are affected by this recall. You may use the "Search for Information by Serial Number" tool on home page of this web site to determine if a specific device is affected.

Advisory

Devices in the affected subset, when programmed to a dual chamber mode with atrial-sensing, may experience a circuit error that affects device functionality. See Table 1 for modes that are susceptible to this circuit error. For this error to occur, a unique combination of events must take place while the device is processing an atrial-sensed event. If this error occurs, the device will be unable to provide pacing until a ventricular-sensed event (VS) is detected. Once a VS is detected, normal pacing functionality is restored immediately. If a VS is not detected, the device will withhold both atrial and ventricular pacing. In addition, until a VS is detected, the device will be unable to initiate a session with a programmer, initiate a session with a CareLink™ remote monitor, or respond to a magnet. Single chamber and dual chamber pacing modes that do not sense atrial activity are not susceptible to this circuit error (see Table 1).

Table 1: Identification of modes susceptible/not susceptible to circuit error

Modes susceptible to circuit error

DDD, DDDR
DDI, DDIR
VDD
ADI, ADIR
VDI, VDIR
ODO
OAO
MVP - when operating in DDD, DDDR, DDI or DDIR mode

Modes NOT susceptible to circuit error

VVI, VVIR
DVI, DVIR
AAI, AAIR
VOO, VOOR
AOO, AOOR
DOO, DOOR
OVO
VVT, AAT

Through 4 January 2019, Medtronic is aware of four (4) reported occurrences in two (2) patients where a pause in pacing therapy was clinically apparent due to this circuit error. These reported events occurred in three (3) devices from a total of 156,957 devices sold worldwide. No deaths have been reported as a result of this issue.

Patient risk is determined by the patient's underlying cardiac rhythm and whether the device is in a susceptible pacing mode as described above. Through our analysis of this issue, Medtronic estimates that on average, a device in a susceptible pacing mode has a 2.8% chance per month of experiencing a pacing pause of 1.5 seconds or longer. Risk is minimized in patients who have an escape rhythm adequate to prevent syncope during a loss of ventricular pacing, since a VS restores full device functionality. No risk of a pause due to this circuit error exists for patients programmed to a non-susceptible pacing mode.

The root cause for this issue is related to a design change to an integrated circuit in a subset of devices that were distributed between 10 March 2017 and 7 January 2019.

Medtronic is developing a software update that can be installed into affected devices to correct this issue. Medtronic estimates submission of this software update to regulatory agencies by the 2nd half of 2019. Upon subsequent regulatory approval, Medtronic will notify customers of its availability. Until that time, Medtronic is providing the patient management recommendations described below and depicted in Appendix A.

Patient Management Recommendations

We realize that each patient requires unique clinical considerations. In consultation with Medtronic's Independent Physician Quality Panel (IPQP), **Medtronic recommends programming to a non-susceptible pacing mode as the primary mitigation for patients implanted with an affected device until the software update has been installed.** Specific patient risk assessment and programming recommendations are outlined below and provided in Appendix A.

- For patients whose device is programmed to a non-susceptible mode (see Table 1), no action is needed at this time. Continue routine clinical monitoring.
- For patients whose device is programmed to a susceptible mode and are continually in persistent atrial fibrillation, reprogramming the device to the non-susceptible VVI or VVIR mode is recommended to eliminate risk due to this issue until the software update has been installed. Continue routine clinical monitoring.
- For patients whose device is programmed to a susceptible mode and either: have no underlying ventricular escape rhythm; or are at risk for a symptomatic pause until a ventricular escape beat occurs, programming to a non-susceptible mode is recommended to eliminate risk due to this issue until the software update has been installed. Continue routine clinical monitoring.
- For patients who do not tolerate programming to a non-susceptible pacing mode and either: have no underlying ventricular escape rhythm; or are at risk for a symptomatic pause until a ventricular escape beat occurs, continue clinical monitoring in a susceptible mode until the software update is available, or consider device replacement.
 - o The estimated per patient mortality risk due to this issue is 0.021% when programmed to a susceptible pacing


mode over the estimated time until the software update becomes available. This risk is comparable to the Medtronic estimated per-patient mortality risk associated with a device replacement (0.027%) *.

o If a patient reports symptoms consistent with a pacing pause, and you would like assistance assessing whether a patient had a pause due to this issue, contact your Medtronic representative.

- Advise patients remaining in a susceptible mode to seek immediate medical attention if they experience new or unexpected symptoms consistent with a pacing pause.
- Other than reprogramming to a non-susceptible pacing mode, no additional programming options have been identified to mitigate this issue.

*Medtronic Data on File. MDT2260884-CRHF CIED Infection Report; MRCS: MDT2260884, Version 2.0, 11/02/2015.

Appendix A: Programming decision flow chart



Potential Loss of Device Functionality Lower Risk Subset

Amplia, Claria, Compia, and Viva CRT-D, and Evera and Visia ICD

Original Date of Advisory: March 2018

STATUS UPDATE - OCTOBER 2020

Within the 752 lower-risk devices, there have been zero confirmed failures (0%) through October 12, 2020. An estimated 498 devices remain active

Initial Affected Population	Number of Confirmed Advisory Related Events	Estimated Remaining Active Population	Current Malfunction Rate (confirmed malfunctions over total population)
752 Worldwide (all in USA, Puerto Rico or US Virgin Islands.)	0	498	0% Worldwide

ORIGINAL COMMUNICATION - MARCH 2018

Product

In January 2018, Medtronic completed notification to physicians about a subset of 48 Medtronic Cardiac Resynchronization Therapy Defibrillators (CRT-Ds) and Implantable Cardioverter Defibrillators (ICDs) underwent a specific sequence of manufacturing processes that could result in an unexpected loss of device functionality, including high-voltage therapy.

Within this Lower-Risk Subset of 752 devices, if the device delivered the maximum number of shocks until battery depletion, we estimate 0.5% of these devices would experience arcing during high voltage charging, with failure occurring within the first two (2) high-voltage charges in 0.18% of the devices. See table below for comparison of device subsets.

Through 8 March 2018, there have been zero (0) complaints related to internal arcing in these 752 devices. While the risk for failure is lower in this group of devices, it is not possible to identify which of these 752 devices may fail or when they may fail. Successful delivery of previous high-voltage therapy does not ensure future performance.

You may use the "Search for Information by Serial Number" tool at <http://wwwp.medtronic.com/productperformance/> to determine if a specific device is affected.

Table – Device Subsets

January 2018 48 Implanted Higher-Risk Devices	March 2018 752 Lower-Risk Devices
One field failure has been observed with no deaths reported	No field failures have been observed
7.7% of these devices are projected to fail during the first two high-voltage charges	0.18% of these devices are projected to fail during the first two high-voltage charges
Medtronic communicated a recommendation to strongly consider prophylactic replacement in these devices.	Patient management recommendations follow below.

Patient Management Recommendations – Lower Risk Subset

We realize that each patient requires unique clinical considerations. In consultation with Medtronic's Independent Physician Quality Panel (IPQP), Medtronic provides the following recommendations to physicians for patients who have been implanted with one of the identified devices:

- Prophylactic device replacement should be considered for patients at higher risk, including patients whose clinical history indicates prior need for high-voltage therapy and/or for pacemaker-dependent patients.
- Physicians should carefully weigh the risks and benefits of device replacement. The estimated per patient risk for mortality due to this issue is 0.02% to 0.04% considering the risk of device failure and the likelihood of a patient requiring high voltage therapy. This is comparable to the estimated per patient mortality risk of complications associated with a device replacement (0.04%)[i],[ii].
- For patients in whom it is determined that replacement is not warranted:
 - Consider programming changes to reduce the potential for high-voltage charges associated with arrhythmia detection and therapies, such as enabling ATP before charging for fast ventricular rhythms or programming a separate fast VT via VF zone with ATP. For assistance with patient-specific programming needs, contact Medtronic Technical Services at 800-723-4636.
 - -Continue three-month in-clinic or remote follow-ups to verify device functionality. Inability to interrogate a device or a failed remote monitoring transmission may be an indication that internal arcing has occurred. Devices that have failed will not send an alert as telemetry and all device functionality is immediately lost if internal arcing occurs.
 - Advise patients to seek medical attention immediately if they experience new or unexpected symptoms suspicious for a ventricular arrhythmia.
- Medtronic Data on File. MDT2260884-CRHF CIED Infection Report; MRCS: MDT2260884, Version 2.0, 11/02/2015.
- Birnie, D et al. Complications associated with defibrillation threshold testing: The Canadian experience. Heart Rhythm, Volume 5, Issue 3, Pages 387-390.

Potential Loss of Device Functionality

Amplia, Claria, Compia, and Viva CRT-D, and Evera and Visia ICD

Original Date of Advisory: January 2018

STATUS UPDATE - OCTOBER 2020

Within the 48 devices, there has been 1 confirmed failure (2.1%) through October 12, 2020. An estimated 3 devices remain active.

Initial Affected Population	Number of Confirmed Advisory Related Events	Estimated Remaining Active Population	Current Malfunction Rate (confirmed malfunctions over total population)
48 Worldwide (all USA)	1	3	2.1% Worldwide

ORIGINAL COMMUNICATION - JANUARY 2018

Product

A subset of 48 Medtronic Cardiac Resynchronization Therapy Defibrillators (CRT-Ds) and Implantable Cardioverter Defibrillators (ICDs) underwent a specific sequence of manufacturing processes that could result in an unexpected loss of device functionality, including high-voltage therapy. You may use the "Search for Information by Serial Number" tool on home page of this web site to determine if a specific device is affected. No other Medtronic devices are included in this advisory.

Advisory

These 48 devices were sent through a manufacturing sequence that introduced the potential for internal arcing during high-voltage charging, leading to the immediate and permanent loss of device functionality. Through 12 January 2018, Medtronic has confirmed one (1) implanted device failure resulting in loss of high-voltage therapy related to this issue, where the patient was rescued with external defibrillation.

Due to the nature of this issue, it is not possible to identify which of these 48 devices may fail or when they may fail. Further, we cannot predict how many high-voltage charges can occur prior to a potential failure. Based on testing of a limited number of available devices that underwent this manufacturing sequence, this failure was observed during high-voltage cycle testing to battery depletion in 23% of these devices, with failure observed within the first two (2) high-voltage charges in 7.7% of the tested devices. Successful delivery of previous high-voltage therapy does not guarantee future performance.

PATIENT MANAGEMENT RECOMMENDATIONS

We realize that each patient requires unique clinical considerations. In consultation with Medtronic's Independent Physician Quality Panel (IPQP), Medtronic provides the following recommendation:

Prophylactic device replacement should be strongly considered for patients who have been implanted with one of the devices in the affected subset.

Potential Rapid Battery Depletion Due To Circuit Component

Viva™ CRT-D and Evera™ ICD

Original Date of Advisory: August 2016

STATUS UPDATE - OCTOBER 2020

Within the 78 devices, there have been 10 confirmed failures (13%) through October 12, 2020. Medtronic modeling predicts an additional three (3) failures may occur in the remaining active population. An estimated 26 devices remain active.

Initial Affected Population	Number of Confirmed Advisory Related Events	Estimated Remaining Active Population	Current Malfunction Rate (confirmed malfunctions over total population)
78 Worldwide	10 Worldwide	26 Worldwide	13% Worldwide

ORIGINAL COMMUNICATION - AUGUST 2016

Product

A specific subset of 78 Viva CRT-D and Evera ICD may experience rapid battery depletion due to a low resistance path developing within a circuit component. You may use the "Search for Information by Serial Number" tool at <http://wwwp.medtronic.com/productperformance> to determine if a specific device is affected.

Advisory

Devices in the affected population may experience rapid battery depletion due to a low resistance path developing within a circuit component. This is not related to a failure within the battery.

Development of a low resistance path in the circuit component in some cases has been reported to cause battery depletion in seven (7) days or less and may present clinically during a patient follow-up visit as:

- One or more electrical resets, which will display as an observation on the programmer.
- No pacing or defibrillation therapy output.
- No telemetry.
- Programmer screen display of "SERIOUS DEVICE MEMORY FAILURE."

Patient audible alerts and CareAlerts™ may not reliably notify the patient or clinician, due to this issue.

Reported complications have included shortness of breath, pocket heating, low heart rate, and early device explant.

Patient Management Recommendations

We realize that each patient requires unique clinical consideration and we support your judgment in caring for your patients. After consultation with Medtronic's Independent Physician Quality Panel, Medtronic offers the following options for managing patients implanted with an affected device:

Advise patients to seek medical attention immediately if they experience symptoms (e.g., fainting or lightheadedness) or if the audible patient alert sounds.

For pacemaker-dependent patients or those at a higher risk of Ventricular Tachycardia (VT) or Ventricular Fibrillation (VF):

- Physicians should consider device replacement.

For patients where the physician does not believe device explant is the best course of action, Medtronic offers these additional options:

- Program the audible alerts for “Low Battery Voltage RRT” to “On-High”. It is possible that alerts may not sound if the battery is depleted. Therefore physicians should also consider one of the following:
 - Provide a handheld magnet to patients to frequently check device status.
 - Requires one or more audible alerts be programmed ON.
 - Device operation may be monitored frequently (e.g., daily) by patients placing the magnet over the device for **1-2 seconds and then removing the magnet**. If the device is functional, a steady tone will sound for approximately 10 seconds. If no tone or an oscillating high/low tone is heard, advise patients to seek care immediately.
 - Prescribe either a CareLink™ transmission be performed by the patient, or a maintenance transmission by the clinic, on a more frequent basis (e.g., weekly or daily) based on the unique patient considerations. The clinic should review these transmissions upon receipt.
 - If the transmission is unsuccessful the patient should be brought into the clinic for immediate follow-up as this may be an indication that the device battery has depleted to a level where it can no longer support telemetry.
 - Review transmissions for any signs of this issue (e.g., one or more electrical resets, or notification that a device alert has occurred).
 - Each transmission will decrease battery longevity by approximately one day.

Potential High Battery Impedance

InSync® III Model 8042

Original Date of Advisory: November 2015

Product

All InSync® III Model 8042 Pacemakers

Advisory

Medtronic has identified an issue related to long-term battery performance. Through 27 October 2015, Medtronic has confirmed 30 devices (0.03%) worldwide have been impacted by this issue, for which the root cause is unexpected high battery impedance.

Unexpected high battery impedance can result in the battery's inability to supply sufficient electrical current, impacting device function. Twelve (12) of the 30 devices had reports of unexpected loss of pacing capture. The other 18 devices experienced some form of erratic behavior, including early elective replacement indication (ERI), significant fluctuations in remaining longevity estimates, and inaccurate lead impedances. Through 27 October 2015, events associated with this issue have occurred in devices with implant durations of 53 months or more. Medtronic has received one report of a patient death, where it is possible, but unconfirmed, that this issue was a contributing factor.

If pacing capture is compromised, some patients may experience a return of heart failure symptoms due to loss of biventricular pacing. In cases involving pacemaker-dependent patients, a loss of pacing capture could result in serious injury or death.

The Physician Letter for this issue is available at <http://www.medtronic.com/insync-iii-crt-p>

Patient Management Recommendations (As of November 2015)

We realize that each patient requires unique clinical consideration. After consultation with Medtronic's Independent Physician Quality Panel (IPQP), Medtronic offers the following recommendations for patients with an InSync III CRT-pacemaker:

- Prophylactic device replacement in non-pacemaker-dependent patients is not recommended.
- For pacemaker-dependent patients, physicians should carefully weigh the risks and benefits of device replacement to mitigate this issue on an individual patient basis
 - The estimated per patient mortality risk of this issue (0.007% to 0.02%) is comparable to the estimated per patient mortality risk of complications associated with an incremental, early device replacement (0.005%).
- Continue routine patient follow up in accordance with standard practice, and advise patients to seek medical attention immediately if they experience new or unexpected symptoms.

STATUS UPDATE - OCTOBER 2020

As of October 15, 2020, approximately 700 devices remain active worldwide, from an original implant population of 96,800. In the United States, 300 active devices remain. Our modeling predicts an estimated failure rate between 0.16% and 0.6% for the remaining active devices. Due to the unpredictable nature of this issue, it is not possible to identify which devices might fail or when they might fail. The issue cannot be mitigated by programming changes or increasing patient follow-up frequency. InSync III CRT-pacemakers are no longer distributed.

Initial Affected Population	Number of Confirmed Advisory Related Events	Estimated Remaining Active Population	Current Malfunction Rate (confirmed malfunctions over total population)
96,800 Worldwide (39,900 United States)	171 Worldwide (95 United States)	700 Worldwide (300 United States)	0.18% Worldwide (0.24% United States)

Potential Conductor Wire Fracture

6930, 6931, 6948, 6949 Sprint Fidelis Defibrillation Leads

Original Date of Advisory: October 2007

Product

All Model 6930, 6931, 6948, and 6949 implantable defibrillation leads.

Advisory

There are two primary locations where chronic conductor fractures have occurred on Sprint Fidelis leads: 1) the distal portion of the lead, affecting the anode (ring electrode) and 2) near the anchoring sleeve tie-down, predominantly affecting the cathode (helix tip electrode), and occasionally the high voltage conductor. These two locations account for approximately 90% of the chronic fractures identified in Returned Product Analysis (RPA). The remaining 10% of chronic fractures occurred in the DF-1 connector leg and the proximal portion of the RV coil. High voltage conductor fractures could result in the inability to deliver defibrillation therapy. Anode or cathode conductor fractures (at either location) may present clinically as increased impedance, oversensing, increased interval counts, multiple inappropriate shocks, and/or loss of pacing output.

Patient Management Recommendations (Updated April 2011)

The Lead Integrity Alert (LIA) provides three days advance notice prior to inappropriate therapy to 76% of patients with lead fractures¹. As a result, we strongly recommend that all Sprint Fidelis patients who have the ability to upgrade to Lead Integrity Alert do so promptly. Also ensure that high voltage lead impedance alerts (maximum of 100 ohms) are programmed. When a lead fracture is suspected or confirmed, immediate patient attention is strongly recommended. Physicians should inform their patients to seek medical attention without delay if they experience unexpected shocks.

- **If a Fidelis lead fracture of any type has occurred, we recommend implanting a new high voltage lead with or without extraction of the Fidelis lead.**
- In patients with normal device function and no manifestation of lead fracture, no action is recommended. The risk of prophylactic intervention appears to be greater than serious injury resulting from lead fracture even for pacemaker dependent patients, except in select individual patient circumstances as determined by the physician.
- In the event of a device change-out or upgrade procedure, with no manifestation of lead fracture, consider the patient age and lead model data above, as well as patient life expectancy, co-morbidities, ease of extraction related to implant time, patient preference, etc., for the following options:
 - Leave a properly performing lead intact.
 - Implant a new ICD lead without extraction of the existing lead.
 - Carefully consider all factors before prophylactic placement of a pace-sense lead. Data shows an increased risk of high voltage conductor fracture if a pace-sense conductor fracture has previously occurred. This data is available at <http://www.medtronic.com/us-en/healthcare-professionals/products/product-performance/sprint-fidelis-11-2015-update.html>
 - Individual patient circumstances may warrant extracting and implanting a new ICD lead. If warranted, Medtronic's Independent Physician Quality Panel recommends the lead extraction procedure be performed by a physician with extensive lead extraction experience.²

STATUS UPDATE - OCTOBER 2020

As of October 15, 2020, of the initial implant population of 205,600 in the United States, approximately 45,000 remain implanted. According to Product Surveillance Registry results, lead survival is estimated to be 66.6% (+6.0/-5.5%) at 156 months. As the implanted population ages and the sample size increases for each time interval, the accuracy of the estimated survival probability will increase as shown by tighter confidence intervals.

Initial Affected Population	Number of Confirmed Advisory Related Events	Estimated Remaining Active Population
279,500 Worldwide (205,600 United States)	7,270 Worldwide (5,188 United States)	62,000 Worldwide (45,000 United States)

Footnotes:

1: Swerdlow C, Gunderson, B, et al. "Downloadable Algorithm to Reduce Inappropriate Shocks Caused by Fractures of Implantable Cardioverter-Defibrillator Leads", Circulation, November 2008, 118: 2122-2129.

2: "Transvenous Lead Extraction: Heart Rhythm Society Expert Consensus on Facilities, Training, Indications, and Patient Management", Heart Rhythm, Vol 6, No 7, July 2009.

Mailer Kits Available for Returning Product

Medtronic urges all physicians to return explanted products and to notify Medtronic when a product is no longer in use, regardless of reason for explant or removal from use. The procedures for returning products vary by geographic location.

Mailer kits with prepaid US postage are available for use within the United States to send CRT, ICD, IPG, and leads to Medtronic's CRHF Returned Product Analysis Lab. These mailers are sized to accommodate the devices and leads from a single patient or clinical event and are designed to meet US postal regulations for mailing biohazard materials.

If the product being returned is located outside the United States, please contact your local Medtronic representative for instructions.

Medtronic also requests the return of devices from non-clinical sources, such as funeral homes, and will assume responsibility for storage and disposal of the product once received.

Mailer kits can be obtained by contacting the Returned Product Lab.

CRHF Returned Product Analysis Laboratory
Phone: 1 (800) 328-2518, ext. 44800
Email: crdm.returnedproduct@medtronic.com



Medtronic
710 Medtronic Parkway
Minneapolis, MN 55432-5604
USA
Tel: (763) 514-4000
Fax: (763) 514-4879

Toll-free: 1 (800) 328-2518
(24-hour technical support for
physicians and medical
professionals)

medtronic.com