

Capsule Endoscopy

CRM Technical Services Standard Letter

© Medtronic 2024

The guidance provided in this letter is for healthcare providers and Medtronic representatives and applies to the following Medtronic device models:

<u>MRI Conditional Device Model Number, Model Name, Description</u>
MC1AVR1 Micra™ AV MR Conditional dual chamber transcatheter pacing system with SureScan™ technology (VDD)
MC1VR01 Micra™ MR Conditional single chamber transcatheter pacing system with SureScan™ technology (VVIR)
MC2AVR1 Micra™ AV2 MR Conditional dual chamber transcatheter pacing system with SureScan™ technology (VDD)
MC2VR01 Micra™ VR2 MR Conditional single chamber transcatheter pacing system with SureScan™ technology (VVIR)

This Standard Letter addresses Capsule Endoscopy:

Capsule endoscopy, also known as *video capsule endoscopy*, uses an ingestible digital camera that captures a video record of the patient's digestive tract. The camera is in a capsule with light-emitting diodes, a battery, and a transmitter. Transmission of the video data occurs in short bursts of radiofrequency energy, approximately 2 per s, for an 8-hour diagnostic period.

Note: Contact Medtronic Technical Services to confirm that your capsule endoscopy system is safe for your patient.

PATIENT MANAGEMENT GUIDANCE

This document is useful to health care professionals who perform medical procedures on patients with Medtronic implanted cardiac device systems and who consult with the patients' cardiologists.

Labeling

Micra™ AV MC1AVR1 MR Conditional Device Manual. Manual Document Number: M042501C001 REV. A, www.medtronic.com/manuals

Micra™ MC1VR01 MR Conditional Clinician Manual. Manual Document Number: M042502C001 REV. A, www.medtronic.com/manuals

Micra™ AV2 MC2AVR1 MR Conditional Device Manual. Manual Document Number: M019277C001 REV. E, www.medtronic.com/manuals

Micra™ VR2 MC2VR01 MR Conditional Device Manual. Manual Document Number: M019292C001 REV. E, www.medtronic.com/manuals

Additional comments

For further information please contact the following:

Technical questions: Medtronic Technical Services can answer additional questions regarding these device operations.

Capsule Endoscopy

CRM Technical Services Standard Letter

© Medtronic 2024

Device labeling: For additional device-specific guidance, consult the labeling associated with the device available on the Medtronic Manual Library website at www.medtronic.com/manuals

How to contact U.S. CRM Technical Services: Email: tshelp@medtronic.com

Pacemakers: (800) 505-4636, ICDs: (800) 723-4636, Programmer Support: (800) 638-1991

This information is verified for devices approved in the U.S. and may differ by country. For product-specific information on device operation and indications for use, reference the appropriate product labeling.