

Capsule Endoscopy

CRM Technical Services Standard Letter

© Medtronic 2024

The guidance provided in this letter is for healthcare providers and Medtronic representatives and applies to the following Medtronic device models:

Non-MRI Conditional Device Model Number, Model Name, Description

C6TR01 VIVA™ CRT-P Digital pacemaker with cardiac resynchronization therapy (OAE-DDDR)

P1501DR ENRHYTHM® Dual Chamber Rate Responsive Pacemaker with RapidRead™ Telemetry (OAE-DDDR)

C2TR01 SYNCRA® CRT-P Digital pacemaker with cardiac resynchronization therapy (OAE-DDDR)

C4TR01 CONSULTA® CRT-P Digital pacemaker with cardiac resynchronization therapy (OAE-DDDR)

MRI Conditional Device Model Number, Model Name, Description

RVDR01 REVO MRI® SURESCAN® Dual Chamber Pacemaker with SureScan® Technology (OAE-DDDR)

This Standard Letter addresses Capsule Endoscopy:

Capsule endoscopy, pH capsule procedures - Capsule endoscopy is a procedure in which a capsule containing a tiny camera is swallowed by the patient to take pictures of the patient's digestive tract. Capsule endoscopy and pH capsule procedures should pose no risk of electromagnetic interference.

Note: Contact Medtronic Technical Services to confirm that your capsule endoscopy system is safe for your patient.

PATIENT MANAGEMENT GUIDANCE

This document is useful to health care professionals who perform medical procedures on patients with Medtronic implanted cardiac device systems and who consult with the patients' cardiologists.

Labeling

VIVA™ CRT-P C6TR01 Clinician Manual. Manual Document Number: M956337A001 REV. C, www.medtronic.com/manuals

ENRHYTHM® P1501DR Reference Manual. Manual Document Number: M950679A001 REV. B, www.medtronic.com/manuals

SYNCRA® CRT-P C2TR01 Clinician Manual. Manual Document Number: M950693A001 REV. C, www.medtronic.com/manuals

CONSULTA® CRT-P C4TR01 Clinician Manual. Manual Document Number: M950678A001 REV. C, www.medtronic.com/manuals

REVO MRI® SURESCAN® RVDR01 Implant Manual. Manual Document Number: M954624A001 REV. A, www.medtronic.com/manuals

Capsule Endoscopy

CRM Technical Services Standard Letter

© Medtronic 2024

Additional comments

For further information please contact the following:

Technical questions: Medtronic Technical Services can answer additional questions regarding these device operations.

Device labeling: For additional device-specific guidance, consult the labeling associated with the device available on the Medtronic Manual Library website at www.medtronic.com/manuals

How to contact U.S. CRM Technical Services: Email: tshelp@medtronic.com

Pacemakers: (800) 505-4636, ICDs: (800) 723-4636, Programmer Support: (800) 638-1991

This information is verified for devices approved in the U.S. and may differ by country. For product-specific information on device operation and indications for use, reference the appropriate product labeling.