MRI Magnetic Resonance Imaging

CRHF Technical Services Standard Letter

Rev. 1.0, September 2016, © Medtronic 2016

The information provided in this letter is for healthcare providers and Medtronic representatives and applies to the following Medtronic device models:

Model number, model name, description (Non-MRI Conditional devices)

P1501DR ENRHYTHM®, dual chamber IPG

C2TR01 SYNCRA™ CRT-P Cardiac Resynchronization Therapy Pacemaker

C4TR01 CONSULTA® CRT-P Cardiac Resynchronization Therapy Pacemaker

C6TR01 VIVA™ CRT-P Cardiac Resynchronization Therapy Pacemaker

Magnetic resonance imaging (MRI) – An MRI is a type of medical imaging that uses magnetic fields to create an internal view of the body. Do not conduct MRI scans on patients who have this device or lead implanted. MRI scans may result in serious injury, induction of tachyarrhythmias, or implanted system malfunction or damage.

Labeling

ENRHYTHM® P1501DR Reference Manual. Manual Document Number: M950679A001 REV. B, www.medtronic.com/manuals SYNCRA® CRT-P C2TR01 Clinician Manual. Manual Document Number: M950693A001 REV. C, www.medtronic.com/manuals CONSULTA® CRT-P C4TR01 Clinician Manual. Manual Document Number: M950678A001 REV. C, www.medtronic.com/manuals

VIVA [™]CRT-P C6TR01 Clinician Manual. Manual Document Number: M956337A001 REV. C, www.medtronic.com/manuals

PATIENT MANAGEMENT GUIDANCE

This document is useful to health care professionals who perform medical procedures on patients with Medtronic implanted cardiac device systems and who consult with the patients' cardiologists.

Additional comments

For further information please contact the following:

- Technical questions: Medtronic Technical Services can answer additional questions regarding these device operations.
 Device labeling: For additional device-specific guidance, consult the labeling associated with the device available on the Medtronic Manual Library website at www.medtronic.com/manuals.
- Patient questions: Patients who have questions can contact Medtronic Heart Rhythm Patient Services at 1-800-551-5544, Option 3, email at pshelp@medtronic.com, or see www.medtronic.com/rhythms for a variety of resources.

How to contact U.S. CRHF Technical Services: Pacemakers: (800) 505-4636, ICDs: (800) 723-4636, Instruments: (800) 638-1991. Email: tshelp@medtronic.com

This information is verified for devices approved in the U.S. and may differ by country. For product-specific information on device operation and indications for use, reference the appropriate product labeling.

